

150x70mm

USA Service Center: BIONIME USA CORPORATION
1450 E. Spruce Street, Bldg. #B,
Ontario, CA 91761, USA
Tel: 888-481-8485
www.gediabetes.com

Manufacturer:
BIONIME CORPORATION
No. 100, Sec. 2, Daqing St., South Dist.,
Taichung City 40242, Taiwan (R.O.C)
Made in Taiwan

101-3GM232-001
EN(US)

Rightest blood glucose Meter GM232C



User's Manual

Preface

Thank you for selecting the **Rightest** Blood Glucose Monitoring System GM232C. This manual provides all the information you need to operate this product for accurate test results. Please read this entire manual before you start any testing.

For people living with diabetes, it is important to regularly monitor blood glucose levels to effectively reduce complications from the disease. The easy-to-use **Rightest** Blood Glucose Monitoring System GM232C provides accurate, reliable test results to help you better manage your diabetes.

The **Rightest** Blood Glucose Monitoring System GM232C is designed for in vitro diagnostic use only (for self-testing by a single user outside the body). Testing requires only a small amount of fresh capillary whole blood from either the fingertip, palm or forearm.

The **Rightest** Blood Glucose Monitoring System GM232C is manufactured and supported by Bionime Corporation. If you have any questions or concerns, please contact the Bionime Customer Service toll free at 888-481-8485. We will make every effort to assist you.

Preface

The **Rightest** Blood Glucose Monitoring System GM232C is a personal blood glucose monitoring system to be used for self-testing only. The glucose meter and lancing device are for single patient use. Do not use on multiple patients. Do not share meter or lancing device with anyone including other family members.

A healthcare professional should be contacted when Customer Service is not available.

Please forward your warranty card to Bionime customer service.
Bionime customer service
1450 E. Spruce Street, Bldg. #B,
Ontario, CA 91761, USA

Caution

- Before using the **Rightest** Blood Glucose Monitoring System GM232C to test your blood glucose, please read all of the instructions and conduct all of the tests including the quality control test (Refer to page 38).
- Please perform the quality control test regularly to make sure the test results are accurate.
- The **Rightest** Blood Glucose Meter can only be used with **Rightest** Blood Glucose Test Strips GM232C. Other brands' test strips should not be used under any circumstances. The use of other brands' strips may give inaccurate results.
- The **Rightest** Blood Glucose Monitoring System GM232C is intended for in vitro diagnostic use only. The **Rightest** Blood Glucose test results using fresh capillary whole blood samples from the fingertip, palm and forearm are calibrated to be the equivalent to that of plasma testing.
- The **Rightest** Blood Glucose Monitoring System GM232C is intended for self-testing. It should not be used to screen for or diagnose diabetes mellitus.
- If the **Rightest** Blood Glucose Meter GM232C and Test Strips are exposed to a substantial change in temperature, please wait 45 minutes before measurement.
- Follow all environmental protection regulations when disposing of batteries.
- The **Rightest** Blood Glucose Meter GM232C will be set as mg/dL by default when sold in United States. If meter display shows "mmol/L" during setting or test, please contact GE Diabetes Customer Support Center. Use of the wrong unit of measurement may cause incorrect treatment.

Caution

- The minimum blood sample size to test using the *Rightest* Blood Glucose Monitoring System GM232C is 0.75µL: (•)

Sample Size Example



Blood sample size above 3.0µL might contaminate the test strip port and the meter.

Blood sample size below 0.75µL may cause an inaccurate result or may prevent a meter reading. An Er4 reading will be displayed if the sample size is too small. In this case, repeat the test with a new test strip.

Caution



Important Safety Notes:

- The *Rightest* Blood Glucose Meter GM232C and Lancing Device are for single patient use. Do not share them with anyone including other family members.
- All parts of the kit are considered biohazardous and can potentially transmit infectious diseases, even after following the cleaning and disinfecting procedures. Please refer to the section "Cleaning and Disinfecting Procedures" on page 48.
- Users should wash their hands thoroughly with soap and water before and after handling the meter, lancing device, or test strips.
- User should turn off meter or meter's Bluetooth connection to keep it in flight mode " ✈ " while traveling in flight. User can still perform blood glucose testing while meter is in flight mode " ✈ " .
- Action upon delivery:
 - a) Unpacking; b) checking delivery for completeness; c) checking for damage during transport.

Table of Contents

About the *Rightest* Blood Glucose Monitoring System GM232C..... 8

The *Rightest* Blood Glucose Monitoring System GM232C..... 10

The *Rightest* Blood Glucose Meter GM232C..... 14

The *Rightest* Blood Glucose Test Strip GM232C..... 14

Before Testing

Meter Activation and Battery Change.....16

Setting Up Your Meter - Setting the Date, Time and Volume.....18

Turning On/Off the Meter.....22

Turning On/Off the Bluetooth.....23

Handling the GE006 Blood Glucose Test Strip.....24

Testing Procedure

Getting Ready for Testing.....26

Performing a Blood Glucose Test.....27

Setting Markers.....32

Alternative Site Testing.....33

View Window Appearance.....35

Understanding Test Results and Messages.....36

Quality Control

About Quality Control Testing.....38

Table of Contents

Performing a Quality Control Test..... 41

Understanding Control Test Results..... 43

Test Memory

Recalling Test Results..... 44

Recalling Average Test Results..... 46

Caring for Your Meter..... 47

Cleaning and Disinfecting Procedures..... 48

Error Messages and Troubleshooting..... 52

Limitations..... 56

Real Time Synchronize..... 57

FEDERAL COMMUNICATIONS COMMISSION (FCC) STATEMENT & NCC 警語..... 60

Specifications..... 61

Warranty..... 63

Customer Service..... 64

Expected Blood Glucose Values Without Diabetes..... 65

Component Manufacturer Information..... 66

Log Book..... 67

Warranty Card..... 83

Emergency Card..... 85

The Rightest Blood Glucose Monitoring System GM232C


Contents

Your **Rightest** Blood Glucose Monitoring System GM232C consists of several items. Please identify each item, learn its name and how it is used.

Below are the items included in your **Rightest** Blood Glucose Monitoring System GM232C:


- 1. Getting Started Guide *
 - 2. User's Manual (includes Log Book, Warranty Card, Emergency Card) *
 - 3. **Rightest** Blood Glucose Test Strip GM232C Package Insert *
 - 4. **Rightest** Control Solution GC550 Package Insert *
 - 5. **Rightest** Blood Glucose Meter GM232C (with 2 AAA batteries installed)
 - 6. **Rightest** Blood Glucose Test strips GM232C (10/25 pcs) *
 - 7. **Rightest** Control Solution GC550 *
 - 8. **Rightest** GD500 Lancing Device *
 - 9. Clear Cap *
 - 10. Disposable Sterile Lancets (10 pcs) *
 - 11. Carrying Case (not shown) *
 - 12. Instruction for the lancing device (not shown) *
- (* Different packages have different bundled items. Some of packages might not include * items.)

The *Rightest* Blood Glucose Monitoring System GM232C




If you want to purchase optional items, please contact Bionime customer service at 888-481-8485.


The *Rightest* Blood Glucose Meter GM232C



The *Rightest* Blood Glucose Meter GM232C



The Rightest Blood Glucose Meter GM232C




The Rightest Blood Glucose Meter GM232C


mem	Indicates a test result stored in memory		Indicate when to apply the blood sample
NO AVG	Indicates a test result not included into the average result of the meter		Appears after you insert test strip into meter
AVG	Indicates the average result	Error	Appears when an error occurs
CS	Indicates a control solution test result	88-88	Current date under time mode or testing date under memory mode
mmol/L mg/dL	Unit of test result		Warns when the operational temperature limit is exceeded during testing
888	Test result	AM	Indicates the time in 12H format
	Warns when the batteries are low or must be replaced	88:88	Displays current time under time mode or testing time under memory mode
Before After	Indicates a meal marker with test result		Indicates that the Bluetooth is turned off
	Indicates that the Bluetooth is turned on	Day	Indicates the time period of test results
EN OK ES Manufacturing use only			

The *Rightest* Blood Glucose Test Strip GM232C

The *Rightest* Blood Glucose Meter GM232C can only be used with *Rightest* Blood Glucose Test Strips GM232C and *Rightest* Control Solution GC550. The use of other test strips or control solutions can lead to incorrect results.



The *Rightest* Blood Glucose Test Strip GM232C



**PRECAUTION**


- Close the *Rightest* Blood Glucose Test Strip GM232C vial immediately after removing a test strip.
- Do not reuse *Rightest* Blood Glucose Test Strips GM232C.
- Do not use expired *Rightest* Blood Glucose Test Strips GM232C.
- When you open a new vial of *Rightest* Blood Glucose Test Strips GM232C, record the date on the vial. Discard the vial of test strips after 3 months from opening.
- Store the *Rightest* Blood Glucose Test Strip GM232C between 39-86°F (or 4-30°C) and in a location (10-90% relative humidity). Do not expose to direct sunlight or heat.
- If the *Rightest* Blood Glucose Meter GM232C and Test Strips are exposed to a substantial change in temperature, please wait 45 minutes before measurement.
- For detailed information, please refer to the *Rightest* Blood Glucose Test Strip GM232C Package Insert.

Meter Activation and Battery Change


Your **Rightest** Blood Glucose Meter GM232C comes with two AAA, batteries installed. Two new batteries will provide power to perform approximately 800 tests under normal use with one data transfer. Press the main button or insert a strip to activate your meter.



1. Turn the **Rightest** Blood Glucose Meter GM232C over. Press and slide the battery cover to open.



2. Install the batteries. Be sure to put the batteries in the correct direction.



3. Slide the battery cover back until it snaps into place.

Meter Activation and Battery Change

4. The **Rightest** Blood Glucose Meter GM232C will enter Self-Testing Mode automatically when replacing the battery (all symbols will appear on the screen).
5. Press any button to exit the self-test and enter Setting Mode.
6. You must set the time and date when replacing the battery. See Chapter "Setting Up Your Meter - Setting the Date, Time, and Volume" on page 18. Test results are still stored in the memory.



CAUTION

- Please follow the local regulation to properly discard a used battery.
- Risk of explosion if battery is replaced by an incorrect type.

Setting Up Your Meter - Setting the Date, Time and Volume

You can enter the Setting Mode two ways.

1. Replace the Battery

After removing the battery, press the main button several times until there is no signal on screen, then follow the battery installation steps to replace battery. The *Rightest* Blood Glucose Meter GM232C will perform a self-test. Press the main button to exit the self-test and enter the Setting Mode.

2. With Battery Inserted

Press the main button to turn on the *Rightest* Blood Glucose Meter GM232C. Hold down the main button for 7 seconds. During this time the screen will go blank until you hear a beep. After the beep, the meter will turn on into the Setting Mode. The display screen will show setting data.

 **NOTE**

- Use the Left and Right buttons to select your setting data and press the main button to confirm each selection. After confirming all of the settings, you will return to the Time Mode.

Setting Up Your Meter - Setting the Date, Time and Volume

1. Year setting


With the year format blinking, press the Left or Right button until you see the current year. Press the Main button to confirm. Once the year is confirmed, the Month Setting will appear.

2. Month setting

With the month blinking, press the Left or Right button until you see the current month. Press the Main button to confirm. Once the month is confirmed, the Day Setting will appear.

3. Day setting


With the day blinking, press the Left or Right button until you see the current day. Press the Main button to confirm. Once the day is confirmed, the Time Setting will appear.



Setting Up Your Meter - Setting the Date, Time and Volume


4. Time format 12/24H selection

With the time format blinking, press the Left or Right button to switch between 12H and 24H. Press the Main button to confirm. Once confirmed, the Hour Setting will appear.



5. Hour setting

With the hour blinking, press the Left or Right button until you see the current hour. Press the Main button to confirm. Once the hour is confirmed, the Minute Setting will appear.




6. Minute setting

With the minutes blinking, press the Left or Right button until you see the current minute. Press the Main button to confirm. Once the minute is confirmed, the Volume Setting will appear.

Setting Up Your Meter - Setting the Date, Time and Volume


7. Volume Setting

With the volume blinking, press the Left or Right button to turn the volume on or off. Press the Main button to confirm and finish the settings.



8. Ending Setting

After volume setting, for confirmation you will hear a beep (if volume is turned on). All the settings are saved and completed and will return to time screen.



NOTE

- If you do not change any settings during Meter Set-up for over 2 minutes, the *Rightest* Blood Glucose Meter GM232C will leave setting mode and power off automatically.

Turning On / Off the Meter

1. How to turn on the *Rightest* Blood Glucose Meter GM232C

Press the Main button or Insert a test strip.

2. Manual Power off

To power off the *Rightest* Blood Glucose Meter GM232C, press and hold the Main button for 4 seconds.

3. Auto Power off

The *Rightest* Blood Glucose Meter GM232C will power off automatically after 2 minutes if no buttons are pressed or no strip is inserted.

Turning On/Off the Bluetooth

1. Auto turn on the Bluetooth

After leaving the Setting mode, the Bluetooth will turn on automatically and the “” symbol blinks on the screen.

2. Turn off the Bluetooth

In Time mode, press and hold the Right or Left button for two seconds. The blinking “” symbol will switch to the “” symbol with a beep (If Volume is turned on).

3. Manual turn on the Bluetooth

In Time mode, press and hold the Right or Left button for two seconds. The “” symbol will switch to the blinking “” symbol with a beep (If Volume is turned on).



NOTE

1. The blinking “” symbol should turn to steady immediately when the Bluetooth connection is established.

2. Blinking back and forth “” symbol on the screen means the meter is transferring data.

3. User should turn off meter or meter's Bluetooth connection to keep it in flight mode “” while traveling in flight. User can still perform blood glucose testing while meter is in flight mode “”.




Handling the *Rightest* Blood Glucose Test Strip GM232C

How to handle the *Rightest* Glucose Test Strip GM232C.

Inserting the *Rightest* Blood Glucose Test Strip GM232C:


1. Hold the *Rightest* Blood Glucose Test Strip GM232C between your thumb and middle finger with the view window "  " facing up.
2. Put your forefinger on the side of the strip as shown.
3. Insert the *Rightest* Blood Glucose Test Strip GM232C into test strip port until it clicks and firmly stops.



Handling the *Rightest* Blood Glucose Test Strip GM232C

Removing the *Rightest* Blood Glucose Test Strip GM232C:


1. Hold the *Rightest* Blood Glucose Test Strip GM232C as shown.
2. Rotate the *Rightest* Blood Glucose Test Strip GM232C counterclockwise and pull up simultaneously.
3. Take the *Rightest* Blood Glucose Test Strip GM232C out of the test strip port.
Please follow your health care professional's instructions and discard used strips properly.



Getting Ready for Testing


Before performing a blood glucose test, prepare the items below :

- **Rightest** Blood Glucose Meter GM232C
- **Rightest** Blood Glucose Test Strips GM232C
(Please check the expiration date on the test strip vial. Do not use expired test strips)
- **Rightest** GD500 Lancing device
- Sterile lancet
- Alcohol swab (optional)



Performing a Blood Glucose Test


1. Hold the adjustable depth cap in one hand while holding the base in the other hand. Lift the cap until a gap appears between the cap and base.
2. Pull the two parts in opposite directions to remove the adjustable depth cap.
3. Insert a new disposable lancet firmly into lancet carrier.
4. Twist off and set aside the protective cover of the disposable lancet.
5. Replace the depth adjustable cap.
6. Choose a depth of penetration by rotating the top portion of the depth adjustable cap until your desired setting is visible in the window. Settings are based on skin type " " for soft or thin skin; " " for average skin; " " for thick or calloused skin.



Performing a Blood Glucose Test


- 7. Hold the base in one hand and pull on the plunger with the other hand. The device will be cocked. Release the plunger and it will automatically move back to its original position near the base.
- 8. Wash your hands with warm soapy water and dry thoroughly.
- 9. Take one **Rightest** Blood Glucose Test Strip GM232C from the vial. Close the vial cap immediately.
- 10. Insert the strip into the test strip port of the **Rightest** Blood Glucose Meter GM232C with the view window facing up.
- 11. Once the strip is inserted, all symbols will appear on your meter display and will be accompanied by a beep (if volume is turned on).

NOTE
- The **Rightest** Blood Glucose Meter GM232C will automatically detect the Code number on the strip. You do not have to check the Code number on the meter display and strip vial.




Performing a Blood Glucose Test

- 12. A blood drop icon will appear on the display window and will be accompanied by a beep (if volume is turned on). Apply the blood sample within 2 minutes.
- 13. Place the lancing device against your fingertip and press the release button.
- 14. Touch and hold the blood drop to the edge of sample port until the view window is filled with blood. If the view window is not completely filled with blood the test will not start. If the blood sample was insufficient, discard the test strip and repeat with a new **Rightest** Blood Glucose Test Strip GM232C.
- 15. Countdown Mode will begin on the display window. After 5 seconds, your test result will appear.



Performing a Blood Glucose Test

16. Pull off the depth adjustable cap. Without touching the used disposable lancet, insert the lancet tip into the protective lancet cover.
17. Hold the release button in one hand (see picture 17, step 1) and pull on the plunger with the other hand (see picture 17, step 2) to safely eject the used disposable lancet.
18. Discard the used disposable lancet into an appropriate puncture-proof or biohazard container.
19. Replace the depth adjustable cap after finishing the test.



Performing a Blood Glucose Test



CAUTION

- Do not apply your blood drop to the sample port on the strip until you see the " " appear. The **Rightest** Blood Glucose Meter GM232C is performing an internal test and will display " " and "**Error**" if you apply blood too soon. If this occurs, please repeat the test with a new test strip.
- Record the opening date of a new test strip vial. Discard the vial of test strips 3 months after opening.
- Always keep the metal contacts of the test strip port clean. If any dust or impurities are present, please clean with a small, soft brush.
- The **Rightest** Blood Glucose Monitoring System GM232C and **Rightest** GD500 Lancing Device are intended for a single user only. Do not share these devices with anyone, including other family members.
- All parts of this kit are considered biohazards and can potentially transmit infectious diseases, even after you have performed the cleaning and disinfecting procedure.
- Users should wash hands thoroughly with soap and water after handling the meter, lancing device, and test strips.
- Please refer to the section "Cleaning and Disinfecting Procedures" on page 48 for cleaning and disinfecting instructions.



Setting Markers

According to actual life status, you can set meal markers with test result.
You have to perform meal marker setting immediately after the meter displays the reading of your test result.

- 1. After test, the test result is displayed on the screen with time, date and flashing "Before".
- 2. Press the Right or Left button to choose among "Before" (before meal), "After" (after meal), " " (no marker) and "NO AVG". Then press the Main button to confirm and save this record.

	Before meal: You execute this blood glucose test before a meal.
	After meal: You execute this blood glucose test after a meal.
	(Blank) no marker: regular test or definition by user.
NO AVG	No average: You will exclude this test result from average calculation. Please refer to "Excluding test results from the average calculation" on page 45 for more information.

 **NOTE**
If no button pressed, the marker will be set blank automatically after 15 seconds.

Alternative Site Testing

Alternative site testing: palm or forearm blood sampling

- To perform alternative site testing, install the clear cap for your **Rightest** GD500 Lancing Device. (For detailed information, refer to the instruction manual for the **Rightest** GD500 Lancing Device.)
- To increase the blood flow, massage the intended puncture area of your palm or forearm for a few seconds.
- Immediately after massaging the intended puncture area, press and hold the lancing device with the clear cap against palm or forearm.
- Press the release button.
- Continue holding the lancing device against your palm or forearm and gradually increase pressure for a few seconds until the blood sample size is sufficient. (DO NOT test on the palm or forearm if you are testing for hypoglycemia (Low blood glucose).



Alternative Site Testing



CAUTION

- Test results may vary if blood samples are taken from different sites or under certain conditions where glucose levels are changing rapidly such as: following a drink, a meal, an insulin dose or exercise. In these cases, only the fingertip should be used.
- DO NOT test on the palm or forearm if you are testing for hypoglycemia (Low blood glucose).
- Fingertip samples can show the rapid change in glucose faster than palm or forearm samples.
- Consult your healthcare professional before sampling from your palm or forearm.
- Since the blood flow in the forearm and palm is slower than in the fingertip, use the clear cap provided with the **Rightest** GD500 Lancing Device when testing sites other than fingertip.
- If you use the normal lancing device for your palm or forearm, the blood sample may not be sufficient for meter operation.
- Periodically compare the test system to a laboratory test system that is known to be well maintained and monitored by a healthcare provider.

View Window Appearance

Make sure your blood sample covers the whole area of the view window to get an accurate test result. An insufficient blood sample will result in an error message ("Er4"). If this occurs, repeat the test with a new test strip.



Insufficient blood sample



Sufficient blood sample



CAUTION

- Check the expiration date printed on the strip vial every time you use a test strip. Do not use expired **Rightest** Blood Glucose Test Strips GM232C.
- Use each **Rightest** Blood Glucose Test Strip GM232C immediately after removing it from the vial.
- Do not reuse **Rightest** Blood Glucose Test Strips GM232C.
- Do not perform testing immediately after moving from one ambient temperature to another (e.g. after coming indoors from the outside). Allow 45 minutes for the temperature of the meter and the test strips to equilibrate.
- Apply the blood sample only on the sample port of the test strip.

Understanding Test Results and Messages

Blood glucose test results are shown on the **Rightest** Blood Glucose Meter GM232C as mg/dL.

If your blood glucose result is unusually high or low, or if you question your test results, repeat the test with a new **Rightest** Blood Glucose Test Strip GM232C.

You can also run a Quality Control Test with the **Rightest** Control Solution GC550 to check your **Rightest** Blood Glucose Meter and **Rightest** Test Strip (Refer to “Performing a Quality Control Test” on page 41).

If the test result still remains unusually high or low, contact your healthcare professional immediately.

If you are experiencing symptoms that are not consistent with your blood glucose test results and you have made sure to follow all instructions in this manual, contact your healthcare professional immediately.

Consult your health care provider for appropriate disposal of used test strips and lancets.

Understanding Test Results and Messages

The **Rightest** Blood Glucose Meter GM232C displays results between 20 and 600 mg/dL.

If the test result is below 20 mg/dL, “Lo” will appear on the screen. Please repeat your test with by a new test strip.

If you still get a “Lo” result, contact your healthcare professional.

If the test result is above 600 mg/dL, “Hi” will appear on the screen. Please repeat your test again with a new test strip.

If you still get a “Hi” result, contact your healthcare professional.




About Quality Control Testing

What is a Quality Control Test ?

To ensure proper monitoring function, it is necessary to regularly perform a quality control test. Use one of the **Rightest** Control Solution GC550 when testing your **Rightest** Blood Glucose Monitoring System GM232C in the Control Solution Mode. If the test result is within the Control Solution Range printed on the strip vial label, the **Rightest** Blood Glucose Monitoring System GM232C passes the Quality Control Test and your **Rightest** Blood Glucose Monitoring System GM232C is working properly.

Control Solution Range:

Example of Control Solution Range printed on your test strip vial label.



NOTE
- Level 2 of **Rightest** Control Solution GC550 is provided in your starter kit. If you want to purchase Level 4 or another Level 2 control solutions, please contact the GE Diabetes Customer Support Center toll free at 1-866-613-7085 .

About Quality Control Testing

When should a Quality Control Test Be Performed ?

- To ensure that your **Rightest** Blood Glucose Meter GM232C and **Rightest** Blood Glucose Test Strip GM232C are working properly.
- To confirm that you are following the correct testing procedures.
- To prepare for your initial blood glucose test.
- To check the **Rightest** Blood Glucose Test Strip GM232C when you open a new vial of strips.
- To check your **Rightest** Blood Glucose Meter GM232C after it has been dropped, damaged or exposed to liquids.
- If you suspect that your test results are inaccurate, or if your test results are not consistent with the way you feel.
- To practice testing.

Required Items for Quality Control Tests


To perform a quality control test, prepare the items below :

- **Rightest** Blood Glucose Meter GM232C
- **Rightest** Blood Glucose Test Strips GM232C
- **Rightest** Control Solution GC550

About Quality Control Testing

**CAUTION**
- Each time you open a new bottle of control solution, write the expiration date on the label. The **Rightest** Control Solution GC550 is good for 3 months after opening the bottle, or until the expiration date printed on the label, whichever comes first.


Example



- Wipe the bottle cap with a clean tissue before tightly closing the bottle of control solution.
- Close the bottle of control solution tightly immediately after using.
- Check the expiration date before use. (Refer to the **Rightest** Control Solution GC550 Package Insert).
- Keep control solution bottles out of reach of children.

Performing a Quality Control Test

1. Take one **Rightest** Blood Glucose Test Strip GM232C from the vial and close the vial cap immediately.
2. Insert the **Rightest** Blood Glucose Test Strip GM232C horizontally (not vertically) with the colored view window facing up into the test strip port.
3. While the blood drop icon is flashing on the display window, press and hold the Main button for at least 3 seconds until the "▢▶CS" symbol appears.
4. You will see the blinking "▢▶" icon and "▶CS" icon on the screen, prompting you to apply **Rightest** Control Solution GC550.
5. Shake the bottle of **Rightest** Control Solution GC550 well before opening the cap. Place the cap on a flat surface.
6. Place a drop of control solution onto the top of the cap.
7. Gently touch the sample port of the strip with the control solution from the top of the cap.
8. The screen will display the count time starting from 5 (you will hear a beep if the volume is turned on).



Performing a Quality Control Test

- 9. Tightly replace the cap on the **Rightest** Control Solution GC550 bottle.
- 10. The control solution result will appear. Compare your Quality Control Test result to the Control Solution Range printed on the **Rightest** Blood Glucose Test Strip GM232C vial label.



**CAUTION**

- Your Control Solution Test results will not be included in the average calculations, however, they can still can be recalled and viewed. The Control Solution Test result will be shown with the "CS" icon on the screen.
- The Control Solution Test should be conducted between 59-104°F (15-40°C).
- Before "CS" and "CS" appear, do not touch the control solution to the sample port on the strip. The **Rightest** Blood Glucose Meter GM232C is performing an internal check. Touching the control solution to the sample port before prompted will result in an error message: "Error" and "CS" and be accompanied by beeps (if volume is turned on).
- Do not drip the control solution on to the sample port of the test strip directly. The reagent on the strip might leak into the bottle of control solution and may cause the degeneration of the control solution. This could contaminate the meter via the test strip port.
- Keep the test strip port clean and dry. Clean immediately if the test strip port is stained or is overly exposed to moisture.
- Do not touch the tip of the control solution bottle. If the tip is touched, clean with water.



Understanding Control Test Results

Your control solution test results should fall within the control solution range printed on the test strip vial label. If the results are within the range, the **Rightest** Blood Glucose Monitoring System GM232C is working correctly.

Example of control solution range printed on your test strip vial label.

Control Solution Range	Level 2	Level 4
	88~120 mg/dL	250~340 mg/dL

- Possible reasons your Control Solution results are out of the range :
- Your **Rightest** Control Solution GC550 is expired or was first opened more than 3 months ago.
 - Your **Rightest** Blood Glucose Test Strip GM232C has expired.
 - You left the cap of the Blood Glucose Test Strip vial or the control solution off for a period of time.
 - You did not perform the test procedure correctly.
 - The **Rightest** Blood Glucose Meter GM232C or **Rightest** Blood Glucose Test Strip GM232C have malfunctioned.


If **Rightest** Control Solution GC550 results are out of range, your **Rightest** Blood Glucose Monitoring System GM232C may not be working properly. Repeat the Quality Control Test. If your control solution results are still out of range, do not use **Rightest** Blood Glucose Meter GM232C to test your blood glucose. Please contact the GE Diabetes Customer Support Center toll free at 888-481-8485.

Recalling Test Results

The **Rightest** Blood Glucose Meter GM232C is able to automatically store a maximum of 500 test results with time and date. If your meter has stored 500 results, the newest test result will replace the oldest one.


To recall your test memory, start the meter without a test strip inserted.

1. Press the Main button to switch from the Time Mode to Memory Mode. You will see the "mem" icon in the upper left corner of the display. The latest test result will appear with the number "1".
2. Use either the Left or Right button on the side to review all previous results with date and time. Results display from the most recent (Sequence no. "1") to the oldest (Sequence no. "500") in the lower right hand corner of the screen.
3. The Quality Control Test result can be recalled from the saved data. When "CS" appears under the data, the test was performed using the **Rightest** Control Solution GC550. The control solution test result will not be used when recalling test average results.



Recalling Test Results

The **Rightest** Blood Glucose meter GM232C provides several average test results. View the 1-day, 7-days, 14-days, 30-days, 60-days and 90-days average test results for better blood glucose monitoring.



1. Press the Main button to switch the screen to Average Mode.
2. On the average screen, use either the Left or Right button to view 1-day, 7-days, 14-days, 30-days, 60-days or 90-days averages.
3. The number shown in the lower right hand corner indicates how many test results have been calculated.
4. Excluding test results from the average calculation:
You may exclude the test result from average calculation right after each test. You have to perform this exclusion immediately after the meter displays the reading of your test result. Press the Right or Left button to choose "**NO AVG**" then press the main button to confirm. The test result is now excluded from averaging.

Recalling Average Test Results

5. Quick Searching:
To view test results in sequential order automatically, enter the Memory Mode. Press the Right or Left button for 2 seconds. (The Right button is for viewing from the most current to the oldest test results. The Left button is for viewing from the oldest to the most current test results). Release the button to stop at any particular test result.

 **NOTE**

- The Average function requires that the correct time and date is set. Test results must exist during the desired time interval. For example: To get a 14-day average on 1/30/14, you must have test results dated between 1/17 and 1/30. If no test results exist during that time frame you will not receive an average.
- To exclude test results from average calculation, you need to remain in Test Mode and perform the exclusion right after you get the test result. Once you move to MEM Mode, Average Mode or Time Mode, you will not be able to perform the exclusion.
- The "Lo", "Hi" results, control solution test results and test results under abnormal temperature < 50°F (10°C), > 104°F (40°C) will be excluded from average calculations.

Caring for Your Meter

- Maintenance**
Keep your meter and test strip free of dust, water or any other liquid. Store the meter in the carrying case when not in use. If your meter is dropped or damaged, perform a quality control test with the control solution before performing a blood glucose test.

- Cleaning Meter**
Clean the outside of the meter with disinfecting wipes listed on next page. Do not get the test strip port wet.

- Cleaning Test Strip Port**
If your test strip port is stained with blood, control solution or any liquid, please use disinfecting wipes listed on next page to clean it immediately. Do not use anything wet to clean it. Perform a quality control test to ensure the **Rightest** Blood Glucose Meter GM232C is working properly.

Cleaning and Disinfecting Procedures

Indirect transmission of Human Immunodeficiency Virus (HIV), Hepatitis B Virus (HBV) and Hepatitis C Virus (HCV) during the delivery of healthcare services has been increasingly reported. Persons using blood glucose monitoring systems have been identified as one risk group due to the shared use of fingerstick (lancing) devices and point of care blood testing devices.

The cleaning procedure is to remove dust, blood and body fluid from the surface and should be performed whenever the meter or lancing device is visibly dirty. Performing the cleaning procedure once per week is recommended. The disinfecting procedure is necessary to kill pathogens such as HIV, HBV and HCV on the device. This procedure should be performed periodically; once per week is recommended. The meter and lancing device should be cleaned before being disinfected.

If the meter is being operated by a second person who is providing testing assistance to the user, the meter and lancing device should be decontaminated prior to use by the second person. NOTE: the cleaning procedure can only remove visible contaminants from surfaces. Only the disinfecting procedure can eliminate non-visible pathogens.

CaviWipes Disinfecting Towelettes has been tested and may be used to clean and disinfect the meter and lancing device. Users may purchase the disinfecting wipes from the manufacturer (Metrex), distributors or major on line retail sites such as amazon.com. You can contact Metrex customer service at 1-800-841-1428 and info@metrex.com.

Cleaning and Disinfecting Procedures



NOTE


- Clean and disinfect the outside of the device only. Do not remove battery cover when cleaning and disinfecting.

To clean the meter:

1. Thoroughly wipe the entire surface of the meter, especially port, face/display, housing and buttons with disinfecting wipes listed above to clean any possible dirt, dust, blood and other body fluids.

To disinfect the meter:

2. Take another disinfecting wipe and wipe the meter thoroughly. (Note: All blood and body fluids should be cleaned from surface before performing the disinfecting procedure)
3. Allow the surface to remain wet for 2 minutes.
4. Allow to air dry.



NOTE

- Your **Rightest** Blood Glucose Meter GM232C has been tested to ensure that there is no change in the performance or external materials of the device after 550 cleaning cycles and 550 disinfecting cycles. The testing simulates 2 cleaning and disinfecting cycles per week over the typical life of the meter (5 years).


Cleaning and Disinfecting Procedures

To clean the lancing device:

1. Thoroughly wipe the entire surface of the lancing device with disinfecting wipes listed above to clean any possible dirt, dust, blood and other body fluids.

To disinfect the lancing device:

2. Take another disinfecting wipe and wipe the Lancing device thoroughly. (Note: All blood and body fluids should be cleaned from surface before performing disinfecting procedure)
3. Allow the surface to remain wet for 2 minutes.



NOTE

- Your **Rightest** GD500 Lancing Device has been tested to ensure that there is no change in the performance of the device after 550 cleaning cycles and 550 disinfecting cycles. The testing simulates 2 cleaning and disinfecting cycles per week over the typical life of the meter (5 years).

Cleaning and Disinfecting Procedures



CAUTION


- Users should wash their hands thoroughly with soap and water after handling the meter, lancing device or test strips.
- Please examine your LCD screen, test strip port, buttons and surface of your meter and lancing device after cleaning and disinfecting cycles. Stop using the meter and/or lancing device if any of the following occur:
 - Thin, sliver streaks appear on the screen,
 - The screen becomes cracked, soft, dissolved, brittle or swollen.
- You are unable to turn on/off your meter, operate the left/ right button, the lancing device release button or depth adjustable cap.
- You are unable to enter meter settings, function modes or recall your testing results.

If you have any questions or concerns, please contact your **Rightest** Blood Glucose Monitoring System GM232C authorized representative or call the Bionime customer service toll free at 888-481-8485.

Error Messages and Troubleshooting

Temperature Error

1. In order to get accurate testing results, please test between 50-104 °F (10-40 °C).
2. When the temperature is below 43 °F (6 °C) or over 111 °F (44 °C), the meter will not function and the " **Error** " symbol will blink.
3. If the **Rightest** Blood Glucose Meter GM232C and Test Strips are exposed to a substantial change in temperature, please wait 45 minutes before measurement.



Battery Error


1. The "  " symbol blinks when the battery power is low. Please change the battery as soon as possible. You can still test.
2. The "  " and " **Error** " symbols blink when the battery is too low. Meter cannot test. Please change the battery immediately.



Error Messages and Troubleshooting

Strip Error

1. When the test strip is inserted incorrectly, the meter will beep 4 times while the "  " symbol flashes. Do not apply a sample on the test strip. Please re-insert the unused test strip correctly (refer to "Handling the **Rightest** Blood Glucose Test Strip GM232C" on Page 24 for more information).
2. If, after performing the required steps, the "  " and flashing " **Error** " symbols appear on the screen, it means the **Rightest** Test strip GM232C has been inserted incorrectly more than twice. Please, re-insert the unused test strip again after reviewing the steps on how to insert a test strip correctly on p. 24 or call Customer Service for support. This error message can also indicate that you may have used the wrong test strips. Please check the test strip vial to ensure you are using the **Rightest** Test strips GM232C.



Sampling Error


Please do not apply the blood to the sample port of the test strip before the meter displays "  ". Discard the test strip if the meter shows " **Error** ".



Error Messages and Troubleshooting

- Er1** - The inserted test strip has been used or damaged. Please use a new test strip.
- Er2** - The meter has malfunctioned. Do a Quality Control Test or reinstall the batteries to check if the meter works properly.
- Er3** - The signal transmission is disrupted. Repeat the test using a new test strip.
- Er4** - The blood sample volume is insufficient. Repeat the test using a new test strip.
- Er5** - An issue calibrating the Meter has occurred. Please follow the steps below:
- (1) Remove the test strip from the meter.
 - (2) Turn off the meter (press and hold the Main button for 3 seconds).
 - (3) Press the main button to turn ON the meter.
- If after turning on the Meter, you do not see Er5, your meter is functioning properly and able to perform a test.

If one of the above error messages still appears, please contact Bionime customer service at 888-481-8485.



Error Messages and Troubleshooting

- Meter Malfunction**
- If the **Rightest** Blood Glucose Meter GM232C will not turn on, please follow the steps below:
1. Open the battery cover and remove the batteries.
 2. Wait for 5 minutes and reinsert the batteries as described as instructed on page 16, (Meter Activation and Changing the Battery).
- The meter should work normally after finishing the above steps. If the meter still does not work, please contact Bionime customer service at 888-481-8485.
-  **CAUTION**


A blood sample should ONLY be applied to the test strip after the test strip has been inserted correctly and the Meter's screen is showing the image of the test strip and a blood drop flashing symbol. If the blood drop flashing symbol is NOT showing on the screen, do not apply a sample to the test strip. Please re-insert the unused test strip correctly. The screen MUST show the image of the test strip accompanied by the blood drop symbol flashing (this symbol take 3 seconds to show after correctly inserting the test strip) in order for a sample to be applied. Please consult your User Manual and/or contact Customer Service for support on how to correctly insert a **Rightest** Test strip.

Limitations

- **Rightest** Blood Glucose Monitoring System GM232C is not intended for serum or plasma testing.
- Inaccurate test results may be obtained at altitudes greater than 10,000 feet (3048 meters) above sea level.
- Severe dehydration and excessive water loss may cause inaccurately low results.
- High concentrations of Ascorbic acid (Vitamin C) ≥ 5 mg/dL, Cholesterol ≥ 600 mg/dL, and Uric acid ≥ 10 mg/dL may result in inaccurate test results.
- Not for neonatal use.
- For over-the-counter use. For single-patient use only. For in vitro diagnostic use only.
- Not for screening or diagnosis of diabetes mellitus.
- Not for use on critically ill patients in shock, dehydrated patients or hypo-osmolar patients.
- Alternative site testing (AST) should only be performed during steady-state times (when glucose is not changing rapidly).
- Alternative Site Testing should not be used to calibrate continuous glucose monitoring systems (CGMS).
- Alternative site testing should not be performed if testing for hypoglycemia (Low blood glucose).
- Results from alternative site testing should not be used in insulin dose calculation.


Real Time Synchronize

The **Rightest** Blood Glucose Meter GM232C can be used with the HeartSquare App, a mobile application. After pairing the **Rightest** Blood Glucose Meter GM232C with the mobile phone by Bluetooth, the HeartSquare App will automatically detect any activities on the **Rightest** Blood Glucose Meter GM232C. Upon detection or activation of the **Rightest** Blood Glucose Meter, the HeartSquare App will display the current status of the **Rightest** Blood Glucose Meter GM232C, as well as a step by step walk through of performing a new blood glucose test.



After inserting the test strip, the App will display the status of "Strip IN" on the screen of your mobile phone.

Real Time Synchronize



A blood drop icon will appear on the display window of Meter. And the indication of applying blood will also appear on the screen of your mobile phone.



NOTE

- For more information of HeartSquare App, please visit www.heartsquare.com.
- The screenshots in this User's Manual were taken from a mobile phone using iOS 7.0+. Appearance may vary slightly depending on the operating system and manufacturer of your mobile phone.
- User should turn off meter or meter's Bluetooth connection to keep it in flight mode "✈" while traveling in flight. User can still perform blood glucose testing while meter is in flight mode "✈".

FEDERAL COMMUNICATIONS COMMISSION (FCC) STATEMENT

15.21
You are cautioned that changes or modifications not expressly approved by the part responsible for compliance could void the user's authority to operate the equipment.

15.105(b)
This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FEDERAL COMMUNICATIONS COMMISSION (FCC) STATEMENT

Specification

This device complies with Part 15 of the FCC Rules.
Operation is subject to the following two conditions:
1) this device may not cause harmful interference and
2) this device must accept any interference received, including interference that may cause undesired operation of the device.

NCC警語
低功率電波輻射性電機管理辦法（930322）
第十二條 經型式認證合格之低功率射頻電機，非經許可，公司、商號或使用者均不得擅自變更頻率、加大功率或變更原設計之特性及功能。
第十四條 低功率射頻電機之使用不得影響飛航安全及干擾合法通信；經發現有干擾現象時，應立即停用，並改善至無干擾時方得繼續使用。
前項合法通信，指依電信法規定作業之無線電通信。低功率射頻電機須忍受合法通信或工業、科學及醫療用電波輻射性電機設備之干擾。

Measurement Technology	Oxidase Electrochemical Sensor
Sample	Capillary whole blood
Minimum Sample Volume	0.75 microliters
Measuring Range	20 - 600 mg/dL
Test Time	5 seconds
Memory Capacity	500 blood glucose test results with date and time
Power Saving	Turns off automatically 2 minutes after last user action. To turn off manually, press the main button "  " for 3 seconds.
Operating Temperature	50 - 104°F (10 - 40°C)
Operating Relative Humidity	10 - 90%

Specification

Hematocrit	20-60%
Power Supply	Two AAA Batteries (Input: 3V )
Battery Life	Approximately 800 tests, each test with one data transfer
Meter Dimension	85mm x 57mm x 22.5 mm
Meter Weight	81.0 ± 5g with batteries
Monitor	LCD
Display Area	39 mm x 39.5 mm
Meter Storage/ Transportation Conditions	14 - 140°F (-10 - 60°C), 10-90% relative humidity
Test Strip Storage/ Transportation Conditions	39 - 86°F (4 - 30°C), 10-90% relative humidity

Warranty

Bionime Corporation warrants that this product will be free from defects in materials and workmanship for five years from the date of purchase.

This warranty does not apply to the performance of a **Rightest** Blood Glucose Meter GM232C that has been altered, misused, tampered with or abused in any way.

This warranty applies only to the original purchaser of the **Rightest** Blood Glucose Monitoring System GM232C. Please complete and return the enclosed warranty card.

Different models have different specifications. This warranty applies only to the **Rightest** Blood Glucose Monitoring System GM232C; other models are not covered with this warranty card.

GE is a trademark of General Electric Company and is used under license by Bionime Corporation, No. 100, Sec. 2, Daqing St., South Dist., Taichung City 40242, Taiwan




NOTE

- During blood glucose measurement, the **Rightest** Blood Glucose Meter GM232C itself may come into contact with blood. All parts of the **Rightest** Blood Glucose Monitoring System GM232C are considered biohazardous and can potentially transmit infectious diseases. Please follow your health care professional's direction to properly dispose of the used **Rightest** Blood Glucose Monitoring System GM232C after removing the batteries.

Customer Service

Please review all of the instructions to make sure you are performing the steps correctly. If you have any questions or concerns, please contact the Bionime customer service toll free at 888-481-8485. We will make every effort to provide assistance and solutions for you.

Description of used symbols:



Dispose of waste

The symbol indicating separate collection for electrical equipment consists of the crossed-out wheeled bin. When the end-user wishes to discard this product, it must be sent to separate collection facilities for recovery and recycling. By separating this product from other household-type waste, the volume of waste sent to incinerators or land-fills will be reduced and natural resources will thus be conserved.

Dispose of batteries



Batteries that contain hazardous substances are identified by a **symbol showing a crossed out waste bin and should not be disposed of as normal household waste**. As a consumer, you are required by law to return used batteries. Old batteries can be handed in as special waste at the public recycling points in your borough, or in shops in which batteries of similar kinds are sold. These symbols are found on batteries containing hazardous substances: Pb = battery contains lead, Cd = battery contains cadmium, Hg = battery contains mercury.

According to the current state of knowledge, this equipment does not contain any substances that are particularly harmful for the environment. It must not be treated as normal waste, but must be taken to a collection point where electric and electronic devices are recycled. You can obtain further information from your municipal authorities, local waste disposal services, or the company from which this product was purchased.

Expected Glucose Values Without Diabetes

Expected glucose values without diabetes ⁽¹⁾

Status	Range (mg/dL)
Fasting	<100
Two Hours after meals	<140

References

1) American Diabetes Association: Standard of Medical Care in Diabetes 2014, Diabetes Care, Vol.37, supplement 1, S14-S80, January 2014.

Component Manufacturer Information

Log Book

Blood Glucose Meter, Test Strip, Control Solution
Manufacturer: Bionime Corporation
No. 100, Sec. 2, Daqing St., South Dist., Taichung City 40242, Taiwan
Product complies with In Vitro Diagnostic Medical Device Directive 98/79/EC. (CE0197)
EC-Rep: Bionime GmbH, Klaus Ellensohn, Tschuetschgasse 8, 6833 Klaus / Vlb., Austria

Lancing Device
Manufacturer: Bionime Corporation
No. 100, Sec. 2, Daqing St., South Dist., Taichung City 40242, Taiwan
Product complies with Medical Device Directive 93/42/EEC
EC-Rep: Bionime GmbH, Klaus Ellensohn, Tschuetschgasse 8, 6833 Klaus / Vlb., Austria

Disposable Sterile Lancets
Manufacturer: SteriLance Medical (SuZhou) Inc.
No.68 LiTangHe RD, XiangCheng, SuZhou, JiangSu 215133, P.R. China
Product complies with Medical Device Directive 93/42/EEC. (CE0197)
EC- Rep: EMERGO EUROPE, Molenstraat 15, 2513 BH The Hague, The Netherlands

Name:

Address:

Home Phone:

Work Phone

Doctor:

Doctor's Phone

Pharmacy:

Pharmacy Phone

Insulin/Pills:

Log book date From: To:

In case of emergency contact:

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								

Log Book

DATE		M	T	W	T	F	S	S
Breakfast	Blood Glucose							
	Insulin/ Medication							
Lunch	Blood Glucose							
	Insulin/ Medication							
Dinner	Blood Glucose							
	Insulin/ Medication							
Bedtime	Blood Glucose							
	Insulin/ Medication							
Other	Blood Glucose							
	Insulin/ Medication							
Comments								



Warranty Card

Distributor name and address

PLACE
STAMP
HERE

Warranty Card

Name	Tel	Distributor name and address
Address		
Serial No	Model	
Date of purchase		

Limited warranty: (Please present this card for product replacement)

1. Bionime warrants to the original purchaser only, that the **Rightest** Blood Glucose Meter GM232C shall be free of any defects in materials or workmanship and, provided it is not modified, altered or misused, will perform in accordance with specifications for a period of five (5) years from the original date of purchase.

2. Bionime guarantees the performance of the **Rightest** Blood Glucose Meter GM232C only if used as directed and provided that the failure to perform has not been caused in whole or in part by the use of test strips that are not **Rightest** Blood Glucose Test Strips GM232C manufactured by Bionime. Use only **Rightest** Blood Glucose Test Strips GM232C in your **Rightest** Blood Glucose Meter GM232C..

3. The sole obligation of Bionime under this warranty shall be to replace any defective **Rightest** Blood Glucose Meter GM232C. No other warranties, express or implied, are made. Bionime shall not be responsible for any incidental or consequential damages.

4. You must contact your distributor for assistance and/or instructions for obtaining a replacement **Rightest** Blood Glucose Meter GM232C.

Emergency Card

EMERGENCY CARD
Rightest Blood Glucose Monitoring System GM232C

- User Name:
- Emergency Contact Phone No.:
- Blood Type:
- Doctor/Hospital:

I am a diabetes patient. If you find me in a coma or stupor, please contact nearest emergency services immediately. Or call:

* Please fill out this card and carry with you at all times.

