

450MHz System Performance Check & EUT Evaluation (Face-held)

Measured Fluid Dielectric Parameters (Brain)

August 18, 2003

Frequency	e'	e''
350.000000 MHz	45.7521	38.7562
360.000000 MHz	45.4509	38.0704
370.000000 MHz	45.1822	37.4401
380.000000 MHz	44.9314	36.8515
390.000000 MHz	44.7538	36.3257
400.000000 MHz	44.4593	35.8005
410.000000 MHz	44.1470	35.3449
420.000000 MHz	43.9072	34.8817
430.000000 MHz	43.6288	34.4170
440.000000 MHz	43.3826	34.0265
450.000000 MHz	43.1192	33.6381
460.000000 MHz	42.9593	33.3153
470.000000 MHz	42.7406	32.9169
480.000000 MHz	42.5800	32.4996
490.000000 MHz	42.3572	32.1957
500.000000 MHz	42.1973	31.8093
510.000000 MHz	41.9803	31.4476
520.000000 MHz	41.7785	31.1305
530.000000 MHz	41.6197	30.8343
540.000000 MHz	41.4213	30.4614
550.000000 MHz	41.2221	30.1802