

Device under Test Properties

| Model, Manufacturer | Dimensions [mm]   | IMEI | DUT Type |
|---------------------|-------------------|------|----------|
| Device,             | 50.0 x 10.0 x 8.0 |      | Dipole   |

Exposure Conditions

| Phantom Section, TSL         | Position, Test Distance [mm] | Band | Group, UID | Frequency [MHz], Channel Number | Conversion Factor | TSL Conductivity [S/m] | TSL Permittivity |
|------------------------------|------------------------------|------|------------|---------------------------------|-------------------|------------------------|------------------|
| Flat, Head Simulating Liquid | ,                            |      | CW, 0--    | 2450.000, 0                     | 7.84              | 1.84                   | 39.8             |

Hardware Setup

| Phantom                                 | TSL, Measured Date                     | Probe, Calibration Date     | DAE, Calibration Date   |
|-----------------------------------------|----------------------------------------|-----------------------------|-------------------------|
| Twin-SAM V8.0 (30deg probe tilt) - 2100 | HBBL-600-10000 Charge:xxxx, 2025-05-19 | EX3DV4 - SN7383, 2025-03-14 | DAE4 Sn1318, 2024-10-09 |

Scan Setup

|                     | Area Scan   | Zoom Scan          |
|---------------------|-------------|--------------------|
| Grid Extents [mm]   | 40.0 x 80.0 | 30.0 x 30.0 x 30.0 |
| Grid Steps [mm]     | 10.0 x 10.0 | 5.0 x 5.0 x 1.5    |
| Sensor Surface [mm] | 3.0         | 1.4                |
| Graded Grid         | N/A         | Yes                |
| Grading Ratio       | N/A         | 1.5                |
| MAIA                | N/A         | N/A                |
| Surface Detection   | VMS + 6p    | VMS + 6p           |
| Scan Method         | Measured    | Measured           |

Measurement Results

|                     | Area Scan         | Zoom Scan         |
|---------------------|-------------------|-------------------|
| Date                | 2025-05-19, 13:17 | 2025-05-19, 13:23 |
| psSAR1g [W/kg]      | 4.92              | 4.93              |
| psSAR10g [W/kg]     | 2.29              | 2.29              |
| Power Drift [dB]    | 0.01              | 0.00              |
| Power Scaling       | Disabled          | Disabled          |
| Scaling Factor [dB] |                   |                   |
| TSL Correction      | No correction     | No correction     |
| M2/M1 [%]           |                   | 80.9              |
| Dist 3dB Peak [mm]  |                   | 8.3               |

Warning(s) / Error(s)

| Details    | Area Scan | Zoom Scan |
|------------|-----------|-----------|
| Warning(s) |           |           |
| Error(s)   |           |           |

