

Client: **CCS**
Shenzhen

Certificate No: 26J02Z000053

CALIBRATION CERTIFICATE

Object DAE4 - SN: 867

Calibration Procedure(s) FF-Z11-002-01
Calibration Procedure for the Data Acquisition Electronics (DAEx)


Calibration date: February 02, 2026

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)℃ and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Process Calibrator 753	1971018	11-Jun-25 (CTTL, No.25J02X004977)	Jun-26

Calibrated by:	Name	Function	Signature
	Yu Zongying	SAR Test Engineer	
Reviewed by:	Lin Jun	SAR Test Engineer	
Approved by:	Qi Dianyuan	SAR Project Leader	

Issued: February 03, 2026

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China

Tel: +86-10-62304633-2117

E-mail: emf@caict.ac.cn

<http://www.caict.ac.cn>

Glossary:

DAE

data acquisition electronics

Connector angle

information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters:

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle:* The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The report provide only calibration results for DAE, it does not contain other performance test results.



In Collaboration with

s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China

Tel: +86-10-62304633-2117

E-mail: emf@caict.ac.cn

<http://www.caict.ac.cn>

DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = 6.1 μ V, full range = -100...+300 mV

Low Range: 1LSB = 61nV, full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	406.266 $\pm$ 0.15% (k=2)	405.715 $\pm$ 0.15% (k=2)	406.758 $\pm$ 0.15% (k=2)
Low Range	4.00537 $\pm$ 0.7% (k=2)	3.99691 $\pm$ 0.7% (k=2)	4.02779 $\pm$ 0.7% (k=2)

Connector Angle

Connector Angle to be used in DASY system	222° $\pm$ 1 °
---	----------------

SEM060-14_D2450V2-SN955(2025-06-18)							
Head Liquid							
Frequence(MHz)	Date of Measurement	Return Loss(dB)	Δ %	Impedance real (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2450	2025/6/18	-26.5	/	52.6	/	4.09	/
	2026/6/15	-25.88	-2.34%	51.8	-0.8	4.28	0.19

- End of Appendix -