

Transmitter ZM-540PA, ZM-541PA

OPERATOR'S MANUAL

Transmitter

ZM-540PA, ZM-541PA

1st Edition: 01 Jul 2009
9th Edition: XX Oct 2015

0614-902754H

Printed:

About This Manual

In order to use this product safely and fully understand all its functions, read this manual before using the product.

Keep this manual near the instrument or in the reach of the operator and refer to it whenever the operation is unclear.

Copyright Notice

The entire contents of this manual are copyrighted by Nihon Kohden. All rights are reserved. No part of this document may be reproduced, stored, or transmitted in any form or by any means (electronic, mechanical, photocopied, recorded, or otherwise) without the prior written permission of Nihon Kohden.

This product stores personal patient information. Manage the information appropriately.

Patient names on the screen shots and recording examples in this manual are fictional and any resemblance to any person living or dead is purely coincidental.

The contents of this manual are subject to change without notice.

If you have any comments or suggestions on this manual, please contact us at:

www.nihonkohden.com

Contents

Unpacking.....	i
GENERAL HANDLING PRECAUTIONS.....	I
WARRANTY POLICY.....	II
EMC RELATED CAUTION	VI
Conventions Used in this Manual and Instrument.....	VIII
Intended Use	1
General	1
Receiving Monitor	3
Panel Description	4
Front Panel	4
Rear Panel.....	6
Top Panel.....	7
Bottom Panel	7
Important Safety Information	9
General	9
Output Signal	15
Battery	15
Transmitter Channel Management	16
For Patients Using Implantable Pacemaker	17
ECG Monitoring	17
SpO ₂ Monitoring.....	18
NIBP Monitoring.....	23
Maintenance	26
Preparation on Transmitter	27
Batteries.....	27
Handling Batteries	27
Battery Lifetime.....	27
Installing and Replacing Batteries	28
Situations Requiring Battery Replacement	29
Battery Level Indication	30
Turning On the Transmitter	31
Check Items Before Use.....	31
Check Items After Power On	32
Check Items After Use	33
Turning Off the Transmitter	33
Changing the Transmitter Channel.....	34

Changing Parameter and System Setup Settings	35
Notes on Parameter Settings.....	35
Changing PARAMETER SETUP Settings	36
Parameter Setup Setting List.....	36
Displaying the PARAMETER SETUP Screen.....	37
Changing Parameter Setup Settings	38
Changing SYSTEM SETUP Settings.....	44
System Setup Setting List.....	44
Displaying the SYSTEM SETUP Screen.....	45
Changing System Setup Settings.....	46
Initializing Settings	50
Attaching Electrodes, SpO ₂ Probe and NIBP Cuff to the Patient	51
Attachment Examples.....	51
Selecting Electrode Leads and Disposable Electrodes	52
Connecting the Electrode Lead to the Transmitter.....	53
Electrode Position.....	53
Attaching Electrodes to the Patient and Connecting the	
Electrode Leads to Disposable Electrodes.....	57
Checking ECG on the Transmitter Screen	58
Attaching the SpO ₂ Probe.....	59
Selecting the SpO ₂ Probe	59
Connecting the SpO ₂ Probe to the Transmitter	62
Attaching the Probe to the Patient.....	63
Attaching the NIBP Cuff.....	65
Selecting the NIBP Cuff.....	65
Connecting the NIBP Cuff to the Transmitter	69
Attaching the NIBP Cuff to the Patient	71
Locking the Keys on the Transmitter	77
Monitoring.....	78
Screen Descriptions.....	78
Check Electrodes Screen	79
Numeric and Waveform Screen.....	80
Waveform Review Screen	81
Numeric Review Screen	82
Display Off	82
Basic Monitoring Operation	83
Using the Function Key.....	83
Suspending Alarms on the Receiving Monitor	84
Pausing Monitoring	85
Resuming Monitoring after Pause	87
Confirming the Patient	87
Turning the Display Off	88

Turning the Display On after It was Turned Off	89
ECG and Respiration Monitoring	89
Turning ECG Measurement On/Off	92
Turning Respiration Measurement On/Off.....	92
Electrode Detachment	92
SpO ₂ Monitoring.....	93
Monitoring SpO ₂ during NIBP Measurement.....	98
NIBP Monitoring.....	99
Selecting the Initial Cuff Inflation Pressure	99
Selecting the Measurement Mode and Interval.....	99
Measuring NIBP.....	101
Monitoring SpO ₂ during NIBP Measurement.....	106
Indication and Message List.....	107
Indication	107
Messages	107
Message Display Priority	111
Troubleshooting	112
Transmitter	112
ECG/Respiration	113
SpO ₂	114
NIBP.....	114
Maintenance	116
1. External Check	116
2. Transmitter Channel	117
3. Transmitting/Receiving Signal	117
4. Display	118
5. Key Operation.....	119
6. ECG Check.....	120
7. Respiration Check	121
8. SpO ₂ Check	121
9. NIBP Check	123
10. NIBP Cuff for Attaching Transmitter to Patient Arm	127
Maintenance Check Sheet.....	128
Lifetime and Disposal	129
Disposing of Used Batteries	129
Battery Lifetime.....	129
Disposal	129
Disposing of Electrodes, SpO ₂ Probes and NIBP Cuffs.....	129
Disposing of Transmitter	129
Cleaning, Disinfection and Sterilization	130
Transmitter and Electrode Leads.....	130






Cleaning	130
Disinfection	130
SpO ₂ Probe	131
NIBP Cuff	131
Periodic Inspection	132
Repair Parts Availability Policy	132
Specifications	133
ZM-540PA	133
Measured Parameters	133
Transmitted Data.....	133
Displayed Data.....	133
ECG	133
ECG Display and Heart Rate Count	134
Respiration Measurement.....	134
SpO ₂ Measurement (ISO 9919: 2005 compliant)	134
Noninvasive Blood Pressure, NIBP (IEC 60601-2-30: 1999 compliant)	136
Transmitter	137
Power Requirements	137
Dimensions and Weight.....	137
Environment.....	137
Safety Standards	138
Electromagnetic Compatibility	138
Electromagnetic Emissions	139
Electromagnetic Immunity	139
Recommended Separation Distances between Portable and Mobile RF Communications Equipment	141
Recovery Time after Defibrillation	142
System Composition for EMC Test.....	142
ZM-541PA	143
Measured Parameters	143
Transmitted Data.....	143
Display	143
Displayed Data.....	143
ECG	143
Respiration Measurement.....	144
SpO ₂ Measurement (ISO 9919: 2005 compliant)	144
Noninvasive Blood Pressure, NIBP (IEC 60601-2-30: 1999 compliant)	146
Transmitter	147
Power Requirements	147
Dimensions and Weight.....	147

Environment.....	147
Safety Standards	148
Electromagnetic Compatibility	148
Electromagnetic Emissions	149
Electromagnetic Immunity	149
Recommended Separation Distances between Portable and Mobile RF Communications Equipment	151
Recovery Time after Defibrillation	152
System Composition for EMC Test.....	152
Standard Accessories.....	153
Options	154
Transmitter.....	154
ECG/RESP	154
SpO ₂	155
NIBP.....	156
Transmission Frequencies	157

Unpacking

Check that all the items are included in the package. If there are any missing items, contact your Nihon Kohden representative.

The name and quantity are described under the illustration.

 ZM-540PA or ZM-541PA transmitter (1)	 NIBP cuff for adult, standard (1)	 Strap (1)
 Operator's manual (1)	 Certification (1)	

For the standard accessories and consumables, only use Nihon Kohden specified parts and accessories to satisfy the safety and performance.

For the model, code number and supply code of the standard accessories, refer to “Standard Accessories” (p. 153) in this operator's manual.

GENERAL HANDLING PRECAUTIONS

This device is intended for use only by qualified medical personnel.

Use only Nihon Kohden approved products with this device. Use of non-approved products or in a non-approved manner may affect the performance specifications of the device. This includes, but is not limited to, batteries, recording paper, pens, extension cables, electrode leads, input boxes and AC power.

Please read these precautions thoroughly before attempting to operate the instrument.

- 1. To safely and effectively use the instrument, its operation must be fully understood.**
- 2. When installing or storing the instrument, take the following precautions.**
 - (1) Avoid moisture or contact with water, extreme atmospheric pressure, excessive humidity and temperatures, poorly ventilated areas, and dust, saline or sulphuric air.
 - (2) Place the instrument on an even, level floor. Avoid vibration and mechanical shock, even during transport.
 - (3) Avoid placing in an area where chemicals are stored or where there is danger of gas leakage.
 - (4) The power line source to be applied to the instrument must correspond in frequency and voltage to product specifications, and have sufficient current capacity.
 - (5) Choose a room where a proper grounding facility is available.
- 3. Before Operation**
 - (1) Check that the instrument is in perfect operating order.
 - (2) Check that the instrument is grounded properly.
 - (3) Check that all cords are connected properly.
 - (4) Pay extra attention when the instrument is in combination with other instruments to avoid misdiagnosis or other problems.
 - (5) All circuitry used for direct patient connection must be doubly checked.
 - (6) Check that battery level is acceptable and battery condition is good when using battery operated models.
- 4. During Operation**
 - (1) Both the instrument and the patient must receive continual, careful attention.
 - (2) Turn power off or remove electrodes and/or transducers when necessary to assure the patient's safety.
 - (3) Avoid direct contact between the instrument housing and the patient.
- 5. To Shutdown After Use**
 - (1) Turn power off with all controls returned to their original positions.
 - (2) Remove the cords gently; do not use force to remove them.
 - (3) Clean the instrument together with all accessories for their next use.
- 6. The instrument must receive expert, professional attention for maintenance and repairs. When the instrument is not functioning properly, it should be clearly marked to avoid operation while it is out of order.**
- 7. The instrument must not be altered or modified in any way.**

8. Maintenance and Inspection

- (1) The instrument and parts must undergo regular maintenance inspection at the interval which is specified after the GENERAL HANDLING PRECAUTIONS section.
- (2) If stored for extended periods without being used, make sure prior to operation that the instrument is in perfect operating condition.
- (3) Technical information such as parts list, descriptions, calibration instructions or other information is available for qualified user technical personnel upon request from your Nihon Kohden representative.

9. When the instrument is used with an electrosurgical instrument, pay careful attention to the application and/or location of electrodes and/or transducers to avoid possible burn to the patient.

10. When the instrument is used with a defibrillator, make sure that the instrument is protected against defibrillator discharge. If not, remove patient cables and/or transducers from the instrument to avoid possible damage.

WARRANTY POLICY

Nihon Kohden Corporation (NKC) shall warrant its products against all defects in materials and workmanship for one year from the date of delivery. However, consumable materials such as recording paper, ink, stylus and battery are excluded from the warranty.

NKC or its authorized agents will repair or replace any products which prove to be defective during the warranty period, provided these products are used as prescribed by the operating instructions given in the operator's and service manuals.

No other party is authorized to make any warranty or assume liability for NKC's products.

NKC will not recognize any other warranty, either implied or in writing. In addition, service, technical modification or any other product change performed by someone other than NKC or its authorized agents without prior consent of NKC may be cause for voiding this warranty.

Defective products or parts must be returned to NKC or its authorized agents, along with an explanation of the failure. Shipping costs must be pre-paid.

This warranty does not apply to products that have been modified, disassembled, reinstalled or repaired without Nihon Kohden approval or which have been subjected to neglect or accident, damage due to accident, fire, lightning, vandalism, water or other casualty, improper installation or application, or on which the original identification marks have been removed.

In the USA and Canada other warranty policies may apply.

CAUTION

United States law restricts this product to sale by or on the order of a physician.

Equipment Authorization Requirement

Compliance with FCC/IC Requirements

This device complies with Part 15 of the FCC (Federal Communications Commission) Rules. Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

This device complies with Part 95 Subpart H of the FCC Rules to be used in wireless medical telemetry service.

Operation of this equipment requires the prior coordination with a frequency coordinator designated by FCC for the Wireless Medical Telemetry Service.

CAUTION

(English)

To comply with the FCC/IC radio frequency (RF) exposure compliance requirements, no change to the antenna or the device is permitted. Any change to the antenna or the device could result in the device, exceeding the RF exposure requirements and void user's authority to operate this device.

(Français)

Pour respecter les normes relatives aux radiofréquences FCC/IC, ce dispositif ne doit pas être situé ou utilisé à proximité d'une autre antenne ou d'un autre émetteur si ce n'est selon les procédures mentionnées pour le multi-transmetteur FCC/IC ou décrites dans la documentation.

NOTE

- Use this device only indoors.
- This device has been tested and complies with FCC radiation exposure limits set forth for an uncontrolled environment. The RF transmission power from the antenna conforms to the general public FCC RF Exposure Guidelines limit of Specific Absorption Rate (SAR) 1.6 W/kg. The maximum SAR value measured from this device was extremely smaller than 1.6 W/kg. This device must not be located together with or operated in conjunction with any other unspecified antenna or transmitter.
- The devices require registration and deployment by an authorized frequency coordinator. The ASHE (American Society for Healthcare Engineering) has been designated by the FCC to manage the WMTS frequencies. This device has frequency bands which may not be used in some areas. For details, contact your Nihon Kohden representative. For details on the guidelines, refer to the ASHE home page.

Compliance with IC Requirements

(English)

This device complies with Industry Canada licence-exempt RSSs. Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device.

Installation of this telemetry device is permitted in hospitals and health care facilities only. This device shall not be operated in mobile vehicles (including ambulances and other vehicles associated with health care facilities).

(Français)

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes : (1) l'appareil ne doit pas produire de brouillage, et (2) l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

L'utilisation de cet appareil de télémétrie est permise seulement dans les hôpitaux et établissements de soins de santé. Cet appareil ne doit pas être mis en marche dans des véhicules (y compris les ambulances et autres véhicules associés aux établissements de santé).

ZM-540PA only

(English)

Installation of this telemetry device is permitted in hospitals and health care facilities only. This device shall not be operated in mobile vehicles (including ambulances and other vehicles associated with health care facilities). The installer/user of this device shall ensure that it is at least 80 km from the Dominion Radio Astrophysical Observatory (DRAO) near Penticton, British Columbia. The coordinates of DRAO are: latitude N 49° 19' 15", longitude W 119° 37' 12". For medical telemetry systems not meeting this 80 km separation (e.g. the Okanagan Valley, British Columbia) the installer/user must coordinate with, and obtain the written concurrence of, the Director of DRAO before the equipment can be installed or operated. The Director of DRAO may be contacted at 250-497-2300 (telephone) or 250-497-2355 (fax). (Alternatively, the Manager, Regulatory Standards, Industry Canada, may be contacted.)

(Français)

L'utilisation de cet appareil de télémétrie est permise seulement dans les hôpitaux et établissements de soins de santé. Cet appareil ne doit pas être mis en marche dans des véhicules (y compris les ambulances et autres véhicules associés aux établissements de santé). La personne qui installe/ utilise cet appareil doit s'assurer qu'il se trouve à au moins 80 km de l'Observatoire fédéral de radioastrophysique (OFR) de Penticton en Colombie-Britannique. Les coordonnées de l'OFR sont : latitude N 49° 19' 15", longitude O 119° 37' 12". La personne qui installe/utilise un système de télémétrie médicale ne pouvant respecter cette distance de 80 km (p. ex. dans la vallée de

l'Okanagan (Colombie-Britannique), doit se concerter avec le directeur de l'OFR et obtenir de sa part une autorisation écrite avant que l'équipement ne puisse être installé ou mis en marche. Le directeur de l'OFR peut être contacté au 250-497-2300 (tél.) ou au 250-497-2355 (télécopieur). (Le Directeur des Normes réglementaires d'Industrie Canada peut également être contacté).

ZM-541PA only

(English)

“This equipment shall not be operated in the areas of Sydney, Nova Scotia, or Gander, Newfoundland and Labrador. Please contact the local Industry Canada Office for further information.”

(Français)

« L'utilisation de cet équipement est interdite à Sydney, Nouvelle-Écosse de même qu'à Gander, Terre-Neuve et Labrador. Veuillez contacter le bureau local d'Industrie Canada pour de plus amples renseignements ».

EMC RELATED CAUTION

This equipment and/or system complies with IEC 60601-1-2 International Standard for electromagnetic compatibility for medical electrical equipment and/or system. However, an electromagnetic environment that exceeds the limits or levels stipulated in IEC 60601-1-2, can cause harmful interference to the equipment and/or system or cause the equipment and/or system to fail to perform its intended function or degrade its intended performance. Therefore, during the operation of the equipment and/or system, if there is any undesired deviation from its intended operational performance, you must avoid, identify and resolve the adverse electromagnetic effect before continuing to use the equipment and/or system.

The following describes some common interference sources and remedial actions:

1. Strong electromagnetic interference from a nearby emitter source such as an authorized radio station or cellular phone:
Install the equipment and/or system at another location. Keep the emitter source such as cellular phone away from the equipment and/or system, or turn off the cellular phone.
2. Radio-frequency interference from other equipment through the AC power supply of the equipment and/or system:
Identify the cause of this interference and if possible remove this interference source. If this is not possible, use a different power supply.
3. Effect of direct or indirect electrostatic discharge:
Make sure all users and patients in contact with the equipment and/or system are free from direct or indirect electrostatic energy before using it. A humid room can help lessen this problem.
4. Electromagnetic interference with any radio wave receiver such as radio or television:
If the equipment and/or system interferes with any radio wave receiver, locate the equipment and/or system as far as possible from the radio wave receiver.
5. Interference of lightning:
When lightning occurs near the location where the equipment and/or system is installed, it may induce an excessive voltage in the equipment and/or system. In such a case, disconnect the AC power cord from the equipment and/or system and operate the equipment and/or system by battery power, or use an uninterruptible power supply.
6. Use with other equipment:
When the equipment and/or system is adjacent to or stacked with other equipment, the equipment and/or system may affect the other equipment. Before use, check that the equipment and/or system operates normally with the other equipment.
7. Use of unspecified accessory, transducer and/or cable:
When an unspecified accessory, transducer and/or cable is connected to this equipment and/or system, it may cause increased electromagnetic emission or decreased electromagnetic immunity. The specified configuration of this equipment and/or system complies with the electromagnetic requirements with the specified configuration. Only use this equipment and/or system with the specified configuration.

Caution - continued

8. Use of unspecified configuration:

When the equipment and/or system is used with the unspecified system configuration different than the configuration of EMC testing, it may cause increased electromagnetic emission or decreased electromagnetic immunity. Only use this equipment and/or system with the specified configuration.

9. Measurement with excessive sensitivity:

The equipment and/or system is designed to measure bioelectrical signals with a specified sensitivity. If the equipment and/or system is used with excessive sensitivity, artifact may appear by electromagnetic interference and this may cause mis-diagnosis. When unexpected artifact appears, inspect the surrounding electromagnetic conditions and remove this artifact source.

If the above suggested remedial actions do not solve the problem, consult your Nihon Kohden representative for additional suggestions.

Conventions Used in this Manual and Instrument


Warnings and Cautions

Level	Description
 WARNING	A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.
 CAUTION	A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Explanations of the Symbols in this Manual and Instrument

The following symbols found in this manual/instrument bear the respective descriptions as given.

On Panel

Symbol	Description	Symbol	Description
	Change screen		Defibrillation proof type BF applied part
	Attention, consult operator's manual		Defibrillation proof type CF applied part
	Moves cursor, scrolls data		Serial number
	Direction for attaching battery cover		Date of manufacture
	Direction for inserting battery		RF transmitter Non-ionizing radiation
	Direct current		CSA mark

On LCD

Symbol	Description	Symbol	Description
	Full battery		Battery very weak Cannot measure NIBP Replace battery
	Battery 1/3 full		Alarm suspended
	Battery weak Replace battery		QRS/pulse sync mark

Intended Use

General

The ZM-540PA and ZM-541PA transmitters transmit ECG, respiration and pulse waveforms, SpO₂ and NIBP data from a patient to a Nihon Kohden monitor for continuous monitoring. The front LCD displays ECG (or pulse wave), numeric values of monitoring parameters, NIBP measuring mode and interval, messages and battery condition.* They also display the compressed waveform and numeric data of the latest 10 minutes.

This transmitter is designed for use by qualified medical personnel in a medical facility such as a hospital or clinic. It is not designed for use outdoors or in a home environment, and is not designed to be operated by the patient themselves.

* Essential performance of this transmitter.

The difference between the ZM-540PA and ZM-541PA is the transmission frequency range.

ZM-540PA: 608.0250 to 613.9750 MHz (channel number 9002 to 9478)

ZM-541PA*: 1395.0250 to 1399.9750 MHz (channel number E002 to E398)

1427.0250 to 1431.9750 MHz (channel number E502 to E898)

* For Canada: 1395.0250 to 1399.9750 MHz (channel number E002 to E398)

1427.0250 to 1429.4750 MHz (channel number E502 to E698)

- NOTE
- The transmitter channel can be changed with a QI-901PK channel writer.
 - Read the operator's manual for the receiving monitor together with this manual before use.

WARNING

Do not diagnose a patient based only on data acquired by the transmitter. Overall judgement must be performed by a physician who understands the features, limitations and characteristics of the transmitter and by reading the biomedical signals acquired by other instruments.

WARNING

Do not use this transmitter for monitoring a patient. Monitor the patient on a receiving monitor. This transmitter displays the waveforms and measured values but does not have an alarm function.

⚠ WARNING

Do not use the same transmitter for more than one patient at the same time. Do not connect different sensors from different patients to the same transmitter.

⚠ CAUTION

- Do not use the same channel for different patients. If the same channel is used for two patients, the two patients' data will be lost due to mutual modulation interference, or another patient's data may appear on the receiving monitor screen.
- Do not use two transmitters with adjacent channels in the same hospital. If transmitters with adjacent channels are used, their radio waves interfere with each other.

⚠ CAUTION

Signal loss and artifact may occur because of the multipath cancellation* when using a transmitter.

* Multipath Cancellation (Standing Wave Interference)

When a radio wave reflects off a surface, there may be some points in the room where the reflected and direct waves are exactly out of phase. At these points in the room, the reflected and direct waves cancel each other out so that the signal strength is decreased. Locations where signal loss occurs are called "null spots". If the transmitter is moving or nearby people or objects are moving, null spots can occur anytime and anywhere.

- NOTE
- To prevent interference between channels, assign a channel administrator in the hospital and only he or she should manage channel assignment.
 - Use Nihon Kohden parts and accessories to assure maximum performance from your instrument.
 - For stable signal reception, it is recommended to use a diversity antenna system on the receiving monitor. Otherwise, spike noise from transient fading of electric field strength (for example, people moving) may interfere with the transmitter signal and may be mistaken as an arrhythmia on the receiving monitor.
 - If the transmitter is used outside the specified environment, its performance cannot be guaranteed.

- For details on antennas and antenna construction, contact your Nihon Kohden representative. You can also refer to the Telemetry System Installation Guide.


Receiving Monitor

Any Nihon Kohden receiving monitor (central monitor with multiple patient receiver) can receive signals from this transmitter as long as the protocol version and channel setting are the same on the receiving monitor and transmitter.

- NOTE
- For details on the receiving monitor and upgrade information, contact your Nihon Kohden representative.
 - The transmitter does not give any patient alarm, only a “no battery” alarm. Patient alarms must be managed on the receiving monitor.

Panel Description

Front Panel



1 LCD

Displays numeric values, ECG or pulse wave, NIBP measuring mode and interval, messages and battery status. For details, refer to the “Screen Descriptions” section.

2 Infrared receiver

Used for upgrading the transmitter software.

3 Screen key

Toggles the screen in the following order.

After power on: Start up → Check electrodes → Numeric and waveform → Waveform review → Numeric review → Display off → Check electrodes ...

After auto display off: Numeric and waveform → Waveform review → Numeric review → Display off → Check electrodes → Numeric and waveform ...


On a SETUP or CHECK screen, this key cancels changing setting or exits the screen.

4 Function key

Depending on the setting on the transmitter, this key suspends alarms, pauses monitoring on the receiving monitor or transmits “Patient confirmed” message.

On a SETUP screen, this key registers the selected setting and moves the cursor to the next setting item.

On a CHECK screen, this key starts or stops maintenance test.



Refer to the WARNING
on the next page.

5 NIBP Start/Stop key

Starts/stops NIBP measurement in
selected mode.

6 NIBP Interval key

Selects NIBP measurement mode.

7 Lead/Scroll keys

On the numeric and waveform screen,
these keys change the ECG lead.

On the waveform review screen, these
keys scroll data.

On a SETUP screen, these keys move the
cursor.


8 Battery case

Contains three 1.5 V dry cell batteries
(AA TYPE).

⚠ WARNING

Close the battery case cover during operation. If
the transmitter is used with the battery case cover
open, anyone who touches the opened battery
case may receive an electrical shock when
defibrillation is performed. Touching the opened
battery case may cause electrostatic discharge
and intermittently interfere with the waveform or
data.


Rear Panel




WARNING

This transmitter is not waterproof. If detergent or liquid spills into the transmitter, stop using it and contact your Nihon Kohden representative. If a wet transmitter is used, the patient or operator may receive an electrical shock or injury.

Top Panel



Bottom Panel



 WARNING

Before defibrillation, all persons must keep clear of the bed and must not touch the patient or any equipment or cord connected to the patient. Failure to follow this warning may cause electrical shock or injury.

 WARNING

When performing defibrillation, discharge as far as possible from electrodes, patches and any gel, cream or medicine on the chest of the patient. If there is a possibility that the defibrillator paddle could touch these materials, remove them from the patient. If the defibrillator paddle directly contacts these materials, the discharged energy may cause skin burn to the patient.

 WARNING

When the transmitter is used with an electrosurgical unit (ESU), firmly attach the entire area of the ESU return plate. Otherwise, the current from the ESU flows into the electrodes of the transmitter, causing electrical burn where the electrodes are attached. For details, refer to the ESU manual.

 CAUTION

Do not shake or swing the transmitter while holding the leads or cables connected to the transmitter. The transmitter may come off and injure someone or damage surrounding instruments.

Important Safety Information

General

WARNING

Never use the transmitter in the presence of any flammable anesthetic gas or high concentration oxygen atmosphere. Failure to follow this warning may cause explosion or fire.

WARNING

When performing MRI test, remove all electrodes and probe from the patient which are connected to this transmitter. Failure to follow this warning may cause skin burn on the patient. For details, refer to the MRI manual.

WARNING

Never use the transmitter in a hyperbaric oxygen chamber. Failure to follow this warning may cause explosion or fire.

WARNING

Do not take this transmitter into the MRI test room. This transmitter is not designed to be used during MRI tests.

WARNING

Before defibrillation, all persons must keep clear of the bed and must not touch the patient or any equipment or cord connected to the patient. Failure to follow this warning may cause electrical shock or injury.

 WARNING

When performing defibrillation, discharge as far as possible from electrodes, patches and any gel, cream or medicine on the chest of the patient. If there is a possibility that the defibrillator paddle could touch these materials, remove them from the patient. If the defibrillator paddle directly contacts these materials, the discharged energy may cause skin burn to the patient.

 WARNING

When the transmitter is used with an electrosurgical unit (ESU), firmly attach the entire area of the ESU return plate. Otherwise, the current from the ESU flows into the electrodes of the transmitter, causing electrical burn where the electrodes are attached. For details, refer to the ESU manual.

 WARNING

Do not use the same transmitter for more than one patient at the same time. Do not connect different sensors from different patients to the same transmitter.

 WARNING

This transmitter is not waterproof. If detergent or liquid spills into the transmitter, stop using it and contact your Nihon Kohden representative. If a wet transmitter is used, the patient or operator may receive an electrical shock or injury.

** WARNING**

Close the battery case cover during operation. If the transmitter is used with the battery case cover open, anyone who touches the opened battery case may receive an electrical shock when defibrillation is performed. Touching the opened battery case may cause electrostatic discharge and intermittently interfere with the waveform or data.

** WARNING**

When the signal is unstable, keep the patient under close observation. When the signal is unstable, the monitoring and alarm are not reliable and the receiving monitor cannot detect a sudden change of the patient's condition. This may cause critical changes in the patient condition to be overlooked. Install an appropriate antenna system to ensure stable signal condition.

** WARNING**

While the "ALARMS SUSPENDED" message is displayed on the transmitter, all alarms on the receiving monitor are suspended so keep the patient under close observation.

** WARNING**

When the patient returns to the bed, turn on the transmitter and check that the monitoring is resumed on the receiving monitor.

** WARNING**

If the transmitter is not turned off and monitoring continues for the selected interval, pause monitoring is canceled and monitoring continues. Check that the monitoring is resumed on the receiving monitor.

 WARNING

Do not allow the conductive part of the connector which is connected to the patient to contact other conductive parts including earth. This causes leakage current and incorrect measurement value and leads to wrong diagnosis.

 WARNING

After admitting a patient on the central monitor and attaching electrodes and sensors on the patient and connecting cables to the transmitter, check that there is no error messages and that the waveforms and numeric data are appropriately displayed on the transmitter and central monitor screen. If there is an error message, or waveform or numeric data is not appropriate, check the electrodes and sensors attachment, patient condition and settings on the transmitter and remove the cause.

 CAUTION

Only use Nihon Kohden specified electrodes, electrode leads, SpO₂ probes, and NIBP cuffs. Otherwise, the maximum performance from the transmitter cannot be guaranteed.

 CAUTION

Do not reuse disposable parts and accessories.

 CAUTION

For handling and precautions on electrodes, electrode leads, SpO₂ probes and NIBP cuffs, refer to the manual.

 CAUTION

The measurement values and displayed waveforms on the transmitter and receiving monitor may be different due to timing delay of the display or difference in detection settings.

 CAUTION

Do not shake or swing the transmitter while holding the leads or cables connected to the transmitter. The transmitter may come off and injure someone or damage surrounding instruments.

 CAUTION

Signal loss and artifact may occur because of the multipath cancellation* when using a transmitter.

* Multipath Cancellation (Standing Wave Interference)

When a radio wave reflects off a surface, there may be some points in the room where the reflected and direct waves are exactly out of phase. At these points in the room, the reflected and direct waves cancel each other out so that the signal strength is decreased. Locations where signal loss occurs are called “null spots”. If the transmitter is moving or nearby people or objects are moving, null spots can occur anytime and anywhere.

 CAUTION

Turn off the power of mobile phones, small wireless devices and other devices which produce strong electromagnetic interference around a patient (except for devices allowed by the hospital administrator). Radio waves from devices such as mobile phones or small wireless devices may be mistaken as pulse waves or respiration waves and the displayed data may be incorrect.

 CAUTION

When monitoring respiration is needed, measure respiration with an instrument. The transmitter calculates SpO₂ of arterial blood based on the principle of pulse oximeter and does not measure respiration.

 CAUTION

- Do not use the same channel for different patients. If the same channel is used for two patients, the two patients' data will be lost due to mutual modulation interference, or another patient's data may appear on the receiving monitor screen.
- Do not use two transmitters with adjacent channels in the same hospital. If transmitters with adjacent channels are used, their radio waves interfere with each other.

 CAUTION

When monitoring SpO₂ only, detection of arrhythmia and asystole is not available and arrhythmia alarms such as ASYSTOLE, VF or VT are not available. If the patient requires ECG monitoring, monitor the ECG.

 CAUTION

When monitoring SpO₂ only (without ECG monitoring), turn on both the upper and lower limit alarms for PR and SpO₂ on the receiving monitor. If the patient's pulse is not detected during asystole or other condition, a "CANNOT DETECT PULSE" or "CHECK PROBE" alarm occurs instead of an SpO₂ limit alarm. Furthermore, if the patient has no pulse, noise from probe movement could be misjudged as a pulse and cause an incorrect PR or SpO₂ value to be displayed.

⚠ CAUTION

Always lock the battery cover while using the transmitter. Otherwise the battery cover may come off and correct measurement might not be performed.

Output Signal

⚠ WARNING

Do not use the output signal from the receiving monitor as the synchronization signal for other equipment such as IABP, MRI, echocardiography or defibrillator. There may be time delay between the monitor and the other equipment caused by waveform transmission delay and spike noise may interfere on the output signal and be mistaken as a trigger.

Battery

⚠ CAUTION

Battery replacement must be performed by the operator. When replacing the batteries of a transmitter that is currently used for a patient, disconnect the electrode leads from the transmitter before replacing batteries or do not touch the patient during replacement.

⚠ CAUTION

Refer to the battery and battery charger manuals for details on handling the batteries.

Transmitter Channel Management

WARNING

The following actions must be taken to properly receive the transmitter signal of the correct patient on the receiving instrument. Otherwise, there may be signal loss or signals may mix causing a serious accident, such as monitoring a different patient.

- Assign a channel administrator in the hospital and only he or she should manage channel assignment.
- The channel administrator must manage the channels in the facility so that there is no signal interference.
- When the transmitter channel is changed, the channel administrator must check that the channel on the receiving monitor is also changed and the signal is properly received.
- The channel administrator must replace the channel number label on the transmitter with the new one after changing the channel.

For Patients Using Implantable Pacemaker

WARNING

The bioelectric impedance measurement sensor of a minute ventilation rate-adaptive implantable pacemaker may be affected by transmitter which is connected to the same patient. If this occurs, the pacemaker may pace at its maximum rate and the transmitter may give incorrect data to the monitor. If this occurs, disconnect the electrode leads from the patient or change the setting on the pacemaker by referring to the pacemaker's manual. For more details, contact your pacemaker representative or Nihon Kohden representative.

ECG Monitoring

WARNING

Turn the pacing pulse detection to ON on the receiving monitor when monitoring a pacemaker patient. Otherwise the pacemaker pulse is not rejected. However, even when the pacing pulse detection is set to ON, the pacemaker pulse might not be rejected. When the pacemaker pulse is not rejected, the pacemaker pulse is detected as QRS and false heart rate may be indicated or critical arrhythmia such as asystole may be overlooked. Keep pacemaker patients under close observation.

* For the pacemaker pulse rejection capability of the ZM-540PA and ZM-541PA transmitters refer to the "Specifications - ECG" section.

WARNING

Even when the pacing pulse detection is set to ON on the receiving monitor, the pacemaker pulse can be overlooked or detected as QRS. You cannot confirm the pacemaker operation only from the detected pacemaker pulse.

 CAUTION

Only use Nihon Kohden specified electrodes and electrode leads. When other type of electrodes or electrode leads are used, the "CHECK ELECTRODES" message may be displayed and ECG monitoring may stop.

 CAUTION

When the "CHECK ELECTRODES" message is displayed on the receiving monitor, ECG is not monitored properly and the ECG alarm does not function. Check the electrode, electrode leads, and if necessary, replace with new ones.

SpO₂ Monitoring

 WARNING

SpO₂ measurement may be incorrect in the following cases.

- When the patient's carboxyhemoglobin or methemoglobin increases abnormally.
- When dye is injected in the blood.
- During CPR.
- When measuring at a site with venous pulse.
- When there is body movement.
- When the pulse wave is small (insufficient peripheral circulation).

 WARNING

When monitoring SpO₂ of a patient who is receiving photodynamic therapy, the light from the finger probe sensor may cause a burn. Photodynamic therapy uses a photosensitizing agent that has a side effect of photosensitivity.

The SpO₂ probes manufactured by Nihon Kohden have two wavelengths with peaks in the range of 650 and 950 nm. The maximum light intensity is less than 5.5 mW.

⚠ WARNING

When not monitoring SpO₂, disconnect the SpO₂ cable from the transmitter. Otherwise, noise from the probe sensor may interfere and incorrect data is displayed on the screen.

⚠ WARNING

- When using the TL-201T finger probe, do not fasten the probe and cable to the finger by wrapping with tape. This may cause burn, congestion or pressure necrosis from poor blood circulation.
- When using probes other than the TL-201T finger probe, to avoid poor circulation, do not wrap the tape too tight. Check the blood circulation condition by observing the skin color and congestion at the skin peripheral to the probe attachment site. Even for short-term monitoring, there may be burn or pressure necrosis from poor blood circulation, especially on neonates or low birth weight infants whose skin is delicate. Accurate measurement cannot be performed on a site with poor peripheral circulation.

 WARNING

Check the circulation condition by observing the skin color at the measurement site and pulse waveform. Change the measurement site every 8 hours for disposable probes and every 4 hours for reusable probes (every 8 hours for TL-631T series probe). The skin temperature may increase at the attached site by 2 or 3°C (4 or 5°F) and cause a burn or pressure necrosis. When using the probe on the following patients, take extreme care and change the measurement site more frequently according to symptoms and degree.

- Patient with a fever
- Patient with insufficient peripheral circulation
- Neonate or low birth weight infant with delicate skin

 CAUTION

The disposable probe is not sterilized. Use the disposable probe only for a single patient. Never reuse the disposable probe for another patient because it causes cross infection.

 CAUTION

If the attachment site is dirty with blood or bodily fluids, clean the attachment site before attaching the probe. If there is nail polish on the attachment site, remove the polish. Otherwise, the amount of transmitted light decreases, and measured value may be incorrect or measurement cannot be performed.

 CAUTION

If the skin gets irritated or redness appears on the skin from the probe, change the attachment site or stop using the probe. Take extreme care for the patients with delicate skin.

 CAUTION

While a patient is on medication which causes vasodilation, the pulse waveform may change and in rare cases the SpO₂ value might not be displayed.

 CAUTION

NIBP and SpO₂ can be measured on the same limb, but the SpO₂ monitoring might not be accurate during NIBP measurement. Be careful when reading the SpO₂ values.*

* Monitoring SpO₂ during NIBP Measurement

When the SpO₂ probe is attached to the same limb as the NIBP cuff, the blood flow decreases during NIBP measurement and pulse wave cannot be detected and SpO₂ cannot be monitored properly. When “INHIBIT SpO₂ DURING NIBP” on the PARAMETER SETUP screen is set to ON (factory default setting), SpO₂ monitoring is paused during NIBP measurement to avoid SpO₂ alarm occurrence. However, when monitoring SpO₂ on the same limb as the NIBP, be careful when reading SpO₂ values.

 CAUTION

Normal external light does not affect monitoring but strong light such as a surgical light or sunlight may affect monitoring. If affected, cover the measuring site with a blanket.

 CAUTION

When a message indicates a faulty probe, stop monitoring and replace the probe with a new one.

 CAUTION

Handle the probe cable according to the following cautions. Failure to follow these cautions may cause cable discontinuity or short circuit of the probe cable which may cause incorrect measurement data or inability to perform measurement. Also in rare cases, the probe temperature may increase and cause skin burn on the patient. If the probe cable is damaged, replace the probe with a new one.

- Do not pull or bend the probe cable.
- Do not let caster feet run over the probe cable.

 CAUTION

Do not use a probe which is deteriorated by aging. Accurate measurement cannot be performed.

 CAUTION

Do not use a damaged or disassembled probe. It causes incorrect measurement and may injure the patient.

 CAUTION

When the probe is attached on an appropriate site with sufficient thickness and the error message confirming the probe attachment repeatedly appears, the probe may be deteriorated. Replace it with a new one.

NIBP Monitoring

WARNING

Be careful when measuring NIBP on a patient with known bleeding disorders or coagulation. After NIBP measurement, there may be dot hemorrhage, or circulatory disorder by thrombus where the cuff is attached.

WARNING

NIBP measurement may be incorrect in the following situations.

- When using an ESU
- Body movement
- Small pulse wave
- Too many arrhythmias
- Shaking from an external source
- Rapid blood pressure change
- During CPR
- Slow pulse
- Low blood pressure
- Small pulse pressure
- Cuff is too tight or too loose
- Cuff does not fit the arm
- Cuff is wrapped over thick clothing
- Cuff is deteriorated
- Arterial sclerosis
- Poor perfusion
- Diabetes
- Age
- Pregnancy
- Pre-eclampsia
- Renal disease
- Shivering
- Trembling

 WARNING

When performing NIBP measurements in STAT mode or 5 minute intervals, periodically remove the cuff from the patient for ventilation. The skin temperature may increase at the cuff attachment site by 2 or 3°C (4 or 5°F). When measuring a patient with a fever or peripheral circulation insufficiency, it may cause a burn.

 WARNING

Do not attach the NIBP cuff on a limb which is being used for intravascular access or therapy, or an arterio-venous (A-V) shunt. It may cause reflux of blood or medicinal solution or block injection of medicinal solution due to poor blood circulation.

 WARNING

Do not attach the NIBP cuff on an arm which is the same side as a mastectomy. It may cause circulatory disorder such as swelling from poor blood circulation.

 WARNING

While measuring NIBP, if the NIBP cuff and other medical equipment are attached to the same limb, the medical equipment might not function temporarily.

 WARNING

While measuring NIBP, check the patient condition and confirm that the NIBP cuff does not affect blood circulation.

 WARNING

When performing long term measurement at intervals less than 5 minutes, perform measurements while observing the state of the patient, blood vessels and limb to ensure adequate circulation. Congestion may occur at the measurement site. When performing periodic measurement for a long time, periodically check the circulation condition.

 CAUTION

Do not wrap the cuff on an arm or thigh which is used for injection. NIBP measurement on an arm or thigh which is used for injection may cause reflux of blood and stop injection.

 CAUTION

Do not wrap the cuff too tight. It may cause poor blood circulation and congestion. If the cuff is wrapped too loosely, the NIBP value may increase.

 CAUTION

Do not attach the cuff to the site where there is injury or inflammation. If the skin gets irritated or redness appears on the skin from the cuff, change the attachment site or stop using the cuff. Take extreme care on the patients with delicate skin.

 CAUTION

When using an extension hose, check that the extension hose is not bent or squeezed. Otherwise, the cuff might not inflate or deflate. If the cuff cannot deflate, it may cause congestion on the patient at the cuff attachment site.

 CAUTION

When performing NIBP measurement repeatedly, have a rest between measurements to recover adequate circulation.

Maintenance

 CAUTION

If detergent or liquid spills into the transmitter, stop cleaning or disinfecting it and contact your Nihon Kohden representative. The transmitter needs to be checked for safety and function before use.

 CAUTION

Before maintenance, cleaning or disinfection, turn the transmitter power off and remove the batteries. Failure to follow this instruction may result in electrical shock and transmitter malfunction.

 CAUTION

Never disassemble or repair the transmitter. If there is any problem with the transmitter, contact your Nihon Kohden representative.

 CAUTION

Dispose of Nihon Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise, it may affect the environment. If there is a possibility that the product may have been contaminated with infection, dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.

Preparation on Transmitter

Batteries

Handling Batteries

WARNING

Do not handle the batteries with wet hands.

WARNING

When the transmitter is not in use, remove the batteries. When the batteries are installed, battery power is consumed even when measurement is not performed. When NiMH batteries are left in the transmitter when it is not used, the batteries may overdischarge and leak liquid which makes the batteries unusable and damages the transmitter.

CAUTION

Refer to the battery and battery charger manuals for details on handling the batteries.

- NOTE**
- Remove the batteries from the transmitter before disposing of it.
 - Use either new alkaline batteries or fully charged rechargeable NiMH batteries. Using unspecified batteries, previously used batteries or batteries that have been stored for long periods may result in short battery life or reduced performance resulting in unstable measurement.

Battery Lifetime

Use three AA type alkaline dry cell batteries. NiMH batteries can also be used but the lifetime of alkaline batteries is longer. The lifetime of NiMH batteries is about 1/2 of alkaline batteries (when fully charged).

With new Nihon Kohden recommended alkaline batteries, the transmitter can continuously measure ECG, respiration, SpO₂ and NIBP for approximately 1 day. The measurement is performed at room temperature, NIBP is measured in auto mode at 60 minute intervals and SpO₂ is measured on an index finger of a male patient with weight 60 kg. Operation time depends on the thickness of the SpO₂ probe attachment site.

Recommended batteries

NiMH secondary: Panasonic eneloop

Alkaline primary: Nihon Kohden Medipower (equivalent to Panasonic LR6 (G))


Installing and Replacing Batteries

CAUTION


Battery replacement must be performed by the operator. When replacing the batteries of a transmitter that is currently used for a patient, disconnect the electrode leads from the transmitter before replacing batteries or do not touch the patient during replacement.

If electrode leads are attached to the patient and the person replacing batteries touches the patient during battery replacement, excess patient leakage current may flow into the patient.


- NOTE
- Replace all batteries at the same time.
 - Do not use different types of batteries together.
 - Insert the batteries with the correct polarity (+ and -).
 - The capacity of rechargeable NiMH batteries is reduced if the batteries are recharged before they are fully discharged. For details, refer to the battery operator's manual.




1 Remove the battery case cover.



2 Insert three new or fully charged batteries into the battery case observing the correct polarity.



NOTE: Insert the (-) end of the battery first and press it against the spring. If you try to force the (+) end of the battery in first as shown, it will deform the spring and damage the battery and transmitter.



3 Close the cover.

The transmitter is automatically turned on when the batteries are installed and the cover is closed.

CAUTION

Always lock the battery cover while using the transmitter. Otherwise the battery cover may come off and correct measurement might not be performed.

Situations Requiring Battery Replacement

Replace the batteries when any of the following occurs:

- The transmitter displays the “BATTERY WEAK” message or  icon.
- The transmitter generates a constant alarm (continuous “peep” sound).
- The transmitter LCD does not display anything when the power is turned on.
- The receiving monitor displays a battery replacement message.

NOTE: The battery replacement message or icon on the transmitter disappears after a short time depending on the type of battery. When the battery replacement indication appears, immediately replace the transmitter battery with a new one.

Battery Level Indication


The following icons on the display indicate the battery level. When the screen is turned off, press the Screen key to check the battery level. When “PROTOCOL” of the SYSTEM SETUP screen is set to 57, the battery level indication is transmitted to the receiving monitor.

Indication	Battery Level	Message on the Receiving Monitor
	Fully charged batteries	There is no message on the monitor.
	Batteries are 1/3 full.	
	Batteries are weak. Replace batteries.	Message requiring battery replacement is displayed.
	Batteries are very weak. Cannot measure NIBP. Replace batteries.	
No indication	Dead batteries	No signal can be transmitted to the monitor. There is no indication on the monitor.

Turning On the Transmitter

When the batteries are installed correctly, the power is turned on. A “peep” sounds for one second, the startup screen appears, then the check electrodes screen appears. (There is no “peep” sound when there is no battery power.)

NOTE: Check that the “COMMUNICATION LOSS” message is not displayed on the central monitor.



After checking that the ECG is stable on the check electrodes screen, press the Screen key to display the numeric and waveform screen.



For details on the screen, refer to the “Screen Descriptions” section.

Check Items Before Use

Before turning on the transmitter power, check the following to confirm that the transmitter can be used in normal and safe condition.

Operating Environment

If the transmitter has been stored in conditions such as low temperatures that are outside the usage environment range, place the transmitter in the usage environment for sufficient time for it to adjust to the room temperature before use. If the transmitter may be used constantly, store it within the usage environment range.

Appearance

- There are no damaged or dirty parts on the outside of the transmitter (LCD, keys, sockets, battery case cover, battery case, lock plate, etc.).
- The transmitter is completely dry.
- The electrodes, electrode lead, SpO₂ probe and NIBP cuff are not broken.

Batteries

- The battery polarity is correct.
- The battery case spring is firmly attached and the battery is not loose.
- The battery case cover is firmly closed.

Channel Setting

- The transmitter channel matches the receiving monitor channel.
- There is no nearby transmitter with the same channel.

Other

The transmitter information is correctly sent to the receiving monitor when the transmitter is turned on or off.

Check Items After Power On

After turning on the power, check the following.

Power On

- The transmitter generates a one second “peep” sound and the startup screen appears.
- The transmitter displays the check electrodes screen.
- The transmitter is not too hot.
- The transmitter does not display the “BATTERY WEAK” message.
- The transmitter does not interfere with the operation of other medical instruments.

Daily Check

- The “signal loss” message is not displayed on the receiving monitor when the transmitter is inside the receiving range of the monitor.
- The battery replacement message is not displayed on the monitor.
- The keys on the transmitter function properly.
- The LCD brightness is appropriate. To adjust brightness, refer to the “Changing SYSTEM SETUP Settings” section.

Check Items After Use

To use the transmitter in safe and optimum condition for next time, check the following.

Before Turning Power Off

- Temporarily changed settings are changed back to the previous settings.
- There was no malfunction on the transmitter.

Storage

- ECG electrode leads, SpO₂ probe and NIBP cuff are cleaned and disinfected.
- If the transmitter got wet, liquid is wiped off and the transmitter is thoroughly dried.
- There are enough consumables, such as disposable electrodes.
- The transmitter power is turned off by removing batteries from the transmitter.
- Dead batteries are disposed of properly.

Turning Off the Transmitter

To turn off the power, remove the batteries. When the power is turned off, the saved waveform and numeric data are deleted.

Changing the Transmitter Channel

The channel of the transmitter can be changed with an optional QI-901PK channel writer.


⚠ WARNING

The following actions must be taken to properly receive the transmitter signal of the correct patient on the receiving instrument. Otherwise, there may be signal loss or signals may mix causing a serious accident, such as monitoring a different patient.

- Assign a channel administrator in the hospital and only he or she should manage channel assignment.
- The channel administrator must manage the channels in the facility so that there is no signal interference.
- When the transmitter channel is changed, the channel administrator must check that the channel on the receiving monitor is also changed and the signal is properly received.
- The channel administrator must replace the channel number label on the transmitter with the new one after changing the channel.

- NOTE
- The software version of the QI-901PK channel writer must be 02-01 or later to change the channel on the transmitter.
 - The channel writer must be used outside the patient environment.

The channel is displayed in the upper left corner of the screen.



Changing Parameter and System Setup Settings

The initial settings on the PARAMETER SETUP and SYSTEM SETUP screens can only be changed before monitoring. Changing these settings during monitoring interrupts monitoring.

NOTE: Changing Parameter and System Setup settings must be done by qualified personnel.

Notes on Parameter Settings

When monitoring NIBP and SpO₂, the following setting must be set as indicated in the table to properly transmit the monitoring data to the receiving monitor. Otherwise, SpO₂ cannot be monitored properly during NIBP measurement.

Some receiving monitors require the software to be upgraded. For details, contact your Nihon Kohden representative.

SpO ₂ Probe Attachment Site	INHIBIT SpO ₂ DURING NIBP SETTING
Probe attached to the same limb as the cuff ¹	ON
Probe attached to the limb without cuff	OFF

¹ When the SpO₂ probe is attached to the same limb as the NIBP cuff and the cuff is inflated, the SpO₂ value becomes unstable and SpO₂ or PR alarm may occur.

Changing PARAMETER SETUP Settings

Parameter Setup Setting List

The factory default settings are underlined.

Setting Item	Description	Settings
ECG ELECTRODES	Select the electrode lead type.	IEC, <u>AHA</u>
LEAD TYPE	Select the type of ECG leads.	<u>AUTO</u> , 6 LEADS
ECG MEASUREMENT	Turn ECG monitoring on or off. When electrodes are attached to the patient and ECG leads are connected, ECG monitoring starts even when this setting is set to OFF. If this setting is set to OFF, the same setting on the receiving monitor must also be set to OFF. NOTE: When "PROTOCOL" of the transmitter is set to 57 and the receiving monitor is able to receive protocol 57, ECG measurement on the receiving monitor is automatically set to OFF when this setting is set to OFF on the transmitter.	<u>ON</u> , OFF
RESP MEASUREMENT	Turn respiration monitoring on or off. When this setting is set to OFF, the same setting on the receiving monitor is automatically set to OFF.	<u>ON</u> , OFF
SpO ₂ RESPONSE	Select the SpO ₂ response mode.	FAST, NORMAL, <u>SLOW</u>
INHIBIT SpO ₂ DURING NIBP	Turn SpO ₂ monitoring on or off during NIBP measurement.	<u>ON</u> , OFF
SELECTABLE INTERVALS (min)	Select the NIBP measurement modes for the mode selection.	<u>MANUAL</u> , STAT, <u>5</u> , <u>10</u> , <u>15</u> , <u>30</u> , <u>60</u> , 120, 240
INITIAL INTERVAL (min)	Select the initial NIBP measurement mode at power on.	<u>MANUAL</u> , STAT, 5, 10, 15, 30, 60, 120, 240
NIBP MODE AFTER STAT (min)	Select the NIBP measurement mode after completing STAT measurement.	MANUAL, <u>5</u> , 10, 15, 30
START/FINISH SOUND	Turn ON or OFF the sound for NIBP measurement start/finish.	START: ON, <u>OFF</u> FINISH: ON, <u>OFF</u>
OLD NIBP DATA/ AFTER (min)	Select whether to hide or dim the NIBP data after measurement and how long to wait after measurement to dim or hide it.	DATA: HIDE, <u>DIM</u> AFTER: 5, 10, <u>30</u>
INITIAL CUFF PRESSURE (mmHg)	Select the NIBP cuff inflation pressure.	120, 150, <u>180</u> , 210, 240

Displaying the PARAMETER SETUP Screen

Cursor




MENU screen

Cursor

Setting item

Setting




PARAMETER SETUP screen - page 1

- 1 Turn off the transmitter by removing one battery.
- 2 While pressing the Function key, turn on the transmitter by inserting the battery. The MENU screen appears.
- 3 Press the ► key to move the cursor to “PARAMETER SETUP”.
- 4 Press the Function key to enter PARAMETER SETUP. The current settings are highlighted.
- 5 Change settings.
 - To move the cursor and select the setting item, press the ◀ or ▶ key then press the Function key.
 - To select and register the setting, press the ◀ or ▶ key then press the Function key.
 - To cancel changing the setting of the selected item, press the Screen key.
- 6 When changing settings on the PARAMETER SETUP screen is complete, press the Screen key to return to the MENU screen.
- 7 Press the ◀ or ▶ key to move the cursor to “EXIT”.
- 8 Press the Function key. The numeric and waveform screen appears.

Changing Parameter Setup Settings

ECG ELECTRODES

Select the electrode lead type.



- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “ECG ELECTRODES”.
- 2 Press the Function key. The cursor moves to the selection item.
- 3 Press the ► key to select “IEC” or “AHA”.



- 4 Press the Function key to register the selected setting. The cursor returns to “ECG ELECTRODES”.

LEAD TYPE

Select the type of ECG leads. In normal use, select “AUTO”. When using DIN type lead with 6 electrodes, select “6 LEADS”.

- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “LEAD TYPE”.
- 2 Press the Function key.
- 3 Press the ► key to select “AUTO” or “6 LEADS”.
- 4 Press the Function key to register the selected setting. The cursor returns to “LEAD TYPE”.

ECG MEASUREMENT

Turn ECG monitoring on or off. When electrodes are attached to the patient and ECG leads are connected, ECG monitoring starts even when this setting is set to OFF.

If this setting is set to OFF, the same setting on the receiving monitor must also be set to OFF.

NOTE: When "PROTOCOL" of the transmitter is set to 57 and the receiving monitor is able to receive protocol 57, ECG measurement on the receiving monitor is automatically set to OFF when this setting is set to OFF on the transmitter.

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to "ECG MEASUREMENT".
- 2** Press the Function key.
- 3** Press the ► key to select "ON" or "OFF".
- 4** Press the Function key to register the selected setting. The cursor returns to "ECG MEASUREMENT".

RESP MEASUREMENT

Turn respiration monitoring on or off. When this setting is set to OFF, the same setting on the receiving monitor is automatically set to OFF.

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to "RESP MEASUREMENT".
- 2** Press the Function key.
- 3** Press the ► key to select "ON" or "OFF".
- 4** Press the Function key to register the selected setting. The cursor returns to "RESP MEASUREMENT".

SpO₂ RESPONSE

Select the SpO₂ response mode from FAST, NORMAL or SLOW. For details on the response time, refer to the “Specifications - SpO₂ Measurement (ISO 9919: 2005 compliant)” section in this manual.

NOTE: When measurement condition is unstable due to strenuous movement of the patient, etc., response may become slower in all modes.



PARAMETER SETUP screen - page 2

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to “SpO₂ RESPONSE”. “SpO₂ RESPONSE” is on the second page of the PARAMETER SETUP screen.
- 2** Press the Function key.
- 3** Press the ► key to select “FAST”, “NORMAL” or “SLOW”.
- 4** Press the Function key to register the selected setting. The cursor returns to “SpO₂ RESPONSE”.

INHIBIT SpO₂ DURING NIBP

Turn SpO₂ monitoring on or off during NIBP measurement.

When the SpO₂ probe is attached to the same limb as the NIBP cuff and this setting is set to OFF, the pulse may become unstable and SpO₂ or PR alarm may occur. It is recommended to set this setting to ON so that SpO₂ is not measured during NIBP measurement.

When the SpO₂ probe is attached to the other limb from the NIBP cuff, this setting can be set to OFF.

NOTE: When this “INHIBIT SpO₂ DURING NIBP” is set to OFF, refer to the “Monitoring SpO₂ during NIBP Measurement” section.

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to “INHIBIT SpO₂ DURING NIBP”. “INHIBIT SpO₂ DURING NIBP” is on the second page of the PARAMETER SETUP screen.
- 2** Press the Function key.
- 3** Press the ► key to select “ON” or “OFF”.
- 4** Press the Function key to register the selected setting. The cursor returns to “INHIBIT SpO₂ DURING NIBP”.

SELECTABLE INTERVALS (min)

When the NIBP INTERVAL key is pressed, the measurement mode changes according to the modes selected in this item. MANUAL mode is already selected for the mode selection.



PARAMETER SETUP screen - page 3

- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “SELECTABLE INTERVALS”. “SELECTABLE INTERVALS” is on the third page of the PARAMETER SETUP screen.
- 2 Press the Function key.
- 3 Press the ► key to select or unselect the mode. The selected modes are highlighted.
- 4 Press the Function key to register the selected setting. The cursor returns to “SELECTABLE INTERVALS”.

INITIAL INTERVAL (min)

Select the initial NIBP measurement mode at power on.

- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “INITIAL INTERVAL”. “INITIAL INTERVAL” is on the third page of the PARAMETER SETUP screen.
- 2 Press the Function key.
- 3 Press the ► key to select the mode. Only the mode or interval selected for “SELECTABLE INTERVALS” are available.
- 4 Press the Function key to register the selected setting. The cursor returns to “INITIAL INTERVAL”.

NIBP MODE AFTER STAT (min)

Select the NIBP measurement mode after completing the STAT measurement.

- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “NIBP MODE AFTER STAT”. “NIBP MODE AFTER STAT” is on the third page of the PARAMETER SETUP screen.
- 2 Press the Function key.
- 3 Press the ► key to select the mode. Only the mode or interval selected for “SELECTABLE INTERVALS” are available.
- 4 Press the Function key to register the selected setting. The cursor returns to “NIBP MODE AFTER STAT”.

START/FINISH SOUND

Turn on or off the sound for NIBP measurement start and finish.



PARAMETER SETUP screen - page 4

- 1 On the PARAMETER SETUP screen, press the ► key to move the cursor to “START/FINISH SOUND”. “START/FINISH SOUND” is on the fourth page of the PARAMETER SETUP screen.
- 2 Press the Function key. The cursor moves to “START”.
- 3 Press the ► key to turn “ON” or “OFF” the sound for NIBP measurement start.
- 4 Press the Function key to register the setting for “START”. The cursor moves to “FINISH”.
- 5 Press the ► key to turn “ON” or “OFF” the sound for NIBP measurement finish.
- 6 Press the Function key to register the selected setting. The cursor returns to “START/FINISH SOUND”.

OLD NIBP DATA/AFTER (min)

Select whether to dim or hide the NIBP data after measurement and how long to wait after NIBP measurement to dim or hide it.

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to “OLD NIBP DATA/AFTER”. “OLD NIBP DATA/AFTER” is on the fourth page of the PARAMETER SETUP screen.
- 2** Press the Function key. The cursor moves to “DATA”.
- 3** Press the ► key to select “HIDE” or “DIM” the NIBP data.
- 4** Press the Function key to register the setting for “DATA”. The cursor moves to “AFTER”.
- 5** Press the ► key to select the interval after NIBP measurement to dim or hide.
- 6** Press the Function key to register the selected setting. The cursor returns to “OLD NIBP DATA/AFTER”.

INITIAL CUFF PRESSURE (mmHg)

Select the NIBP cuff inflation pressure.

- 1** On the PARAMETER SETUP screen, press the ► key to move the cursor to “INITIAL CUFF PRESSURE”. “INITIAL CUFF PRESSURE” is on the fourth page of the PARAMETER SETUP screen.
- 2** Press the Function key.
- 3** Press the ► key to select the inflation pressure.
- 4** Press the Function key to register the selected setting. The cursor returns to “INITIAL CUFF PRESSURE”.


Changing SYSTEM SETUP Settings

System Setup Setting List


The factory default settings are underlined.

Setting Item	Description	Settings
PROTOCOL	Select the transmitting protocol. 57: New protocol. A central monitor with an ORG-9100A, ORG-9110A or ORG-9700A multiple patient receiver whose software version 02-01 or later can receive this protocol. 42: Old protocol. A central monitor with an ORG-9100A, ORG-9110A or ORG-9700A multiple patient receiver can receive this protocol. NOTE: When 57 is set, the receiving monitor must be able to receive protocol 57. Otherwise, signals from the transmitter cannot be received.	<u>57</u> , 42
BRIGHTNESS	Select the screen brightness.	<u>DARK</u> , BRIGHT
PRESSURE UNIT	Select the unit for NIBP.	<u>mmHg</u> , kPa
FUNCTION KEY	Select the function of the Function key. SUSPEND ALARM & PAUSE: Suspends alarm on the receiving monitor for 2 minutes. Pauses monitoring on the transmitter and receiving monitor. SUSPEND ALARM: Suspends alarm on the receiving monitor for 2 minutes. CONFIRM: Displays the "PATIENT CONFIRMED" message on the transmitter screen and transmits the message to the receiving monitor. OFF: No function. NOTE: "SUSPEND ALARM & PAUSE" and "CONFIRM" can only be set when PROTOCOL is set to 57.	<u>SUSPEND ALARM & PAUSE</u> , SUSPEND ALARM, CONFIRM, OFF
AUTO RESUME AFTER PAUSE	Select the interval to resume monitoring after PAUSE.	10 s, <u>30 s</u> , 1 min, 2 min, 3 min
SELECTABLE SCREEN TIME OUT PERIOD (min)	Select the display time-out period.	5, 10, 15, <u>30</u> , 60, 120, 240

Displaying the SYSTEM SETUP Screen



MENU screen



SYSTEM SETUP screen - page 1

- 1 Turn off the transmitter by removing one battery.
 - 2 While pressing the Function key, turn on the transmitter by inserting the battery. The MENU screen appears.
 - 3 Press the ► key to move the cursor to “SYSTEM SETUP”.
 - 4 Press the Function key to enter SYSTEM SETUP. The current settings are highlighted.
 - 5 Change settings.
 - To move the cursor and select the setting item, press the ◀ or ▶ key then press the Function key.
 - To select and register the setting, press the ◀ or ▶ key then press the Function key.
 - To cancel changing the setting of the selected item, press the Screen key.
- The SYSTEM SETUP screen has two pages. To display the second page, press the ► key when the cursor is at “PRESSURE UNIT”.
- 6 When changing settings on the SYSTEM SETUP screen is complete, press the Screen key to return to the MENU screen.
 - 7 Press the ◀ or ▶ key to move the cursor to “EXIT”.
 - 8 Press the Function key. The numeric and waveform screen appears.

Changing System Setup Settings

PROTOCOL

Select the transmitting protocol. For differences between protocols, refer to the table below.


57: New protocol. A central monitor with an ORG-9100A, ORG-9110A or ORG-9700A multiple patient receiver whose software version 02-01 or later can receive this protocol.

42: Old protocol. A central monitor with ORG-9100A, ORG-9110A or ORG-9700A multiple patient receiver can receive this protocol.

NOTE: When 57 is set, the receiving monitor must be able to receive protocol 57. Otherwise, signals from the transmitter cannot be received.

Differences Between Protocols

Function	Protocol 42	Protocol 57
Setting ECG MEASUREMENT to OFF on the transmitter automatically turns off the ECG measurement setting on the receiving monitor	No (ECG measurement must be turned off on the receiving monitor)	Yes
Pause monitoring on the receiving monitor from the transmitter	No	Yes
Transmit "PATIENT CONFIRMED" message	No	Yes
Display battery level of the transmitter on the receiving monitor	No	Yes
Transmit SpO ₂ messages	Some messages (refer to the "Indication and Message List" section)	All messages



- 1 Press the ► key to move the cursor to "PROTOCOL".
- 2 Press the Function key.
- 3 Press the ► key to select "42" or "57".



NOTE: FUNCTION KEY (on the second page of the SYSTEM SETUP screen) can be set to "SUSPEND ALARM & PAUSE" or "CONFIRM" only when PROTOCOL is "57". If PROTOCOL is changed to "42", FUNCTION KEY is automatically changed to "OFF".

- 4 Press the Function key to register the selected setting. The cursor returns to "PROTOCOL".

BRIGHTNESS

Select the screen brightness.

- 1 Press the ► key to move the cursor to "BRIGHTNESS".
- 2 Press the Function key.
- 3 Press the ► key to select "DARK" or "BRIGHT".
- 4 Press the Function key to register the selected setting. The cursor returns to "BRIGHTNESS".

PRESSURE UNIT

Select the unit for NIBP.

- 1 Press the ► key to move the cursor to "PRESSURE UNIT".
- 2 Press the Function key.
- 3 Press the ► key to select "mmHg" or "kPa".
- 4 Press the Function key to register the selected setting. The cursor returns to "PRESSURE UNIT".

FUNCTION KEY

Select the function of the Function key. For details on using these functions, refer to “Basic Monitoring Operation” in the “Monitoring” section.

SUSPEND ALARM & PAUSE: Suspends alarm on the receiving monitor for 2 minutes.
Pauses monitoring on the transmitter and receiving monitor.

SUSPEND ALARM: Suspends alarm on the receiving monitor for 2 minutes.

CONFIRM: Displays the “PATIENT CONFIRMED” message on the transmitter screen and transmits the message to the receiving monitor.

OFF: No function.

NOTE: “SUSPEND ALARM & PAUSE” and “CONFIRM” can only be set when PROTOCOL is set to 57.



SYSTEM SETUP screen - page 2

- 1** On the SYSTEM SETUP screen, press the ► key to move the cursor to “FUNCTION KEY”. “FUNCTION KEY” is on the second page of the SYSTEM SETUP screen.
- 2** Press the Function key.
- 3** Press the ► key to select the function.
- 4** Press the Function key to register the selected setting. The cursor returns to “FUNCTION KEY”.

AUTO RESUME AFTER PAUSE

Select the interval to resume monitoring after PAUSE. When either of the following conditions is met, monitoring resumes on the receiving monitor.

- Heart rate is properly monitored for the selected interval.
- SpO₂ is properly monitored for the selected interval.
- NIBP is properly measured and the SYS, DIA or MAP value is displayed.

- 1** Press the ► key to move the cursor to “AUTO RESUME AFTER PAUSE”. “AUTO RESUME AFTER PAUSE” is on the second page of the SYSTEM SETUP screen.
- 2** Press the Function key.
- 3** Press the ► key to select the interval.
- 4** Press the Function key to register the selected setting. The cursor returns to “AUTO RESUME AFTER PAUSE”.

SELECTABLE SCREEN TIME OUT PERIOD (min)

Select the display time-out period. If no key is pressed for the selected time, the display is automatically turned off. The selected time is shown on the Select Screen Time Out Period screen. Refer to the “Turning the Display Off” section for details.



SYSTEM SETUP screen - page 2

- 1** Press the ► key to move the cursor to “SELECTABLE SCREEN TIME OUT PERIOD (min)”. It is on the third page of the SYSTEM SETUP screen.
- 2** Press the Function key.
- 3** Press the ◀ or ▶ key to select time-out period.
- 4** Press the Function key to register the selected setting. The cursor returns to “SELECTABLE SCREEN TIME OUT PERIOD (min)”.

Initializing Settings

Do the following procedure to initialize all settings, except for channel, to the factory default settings.

Cursor



- 1 Turn off the transmitter by removing a battery.
- 2 While pressing the Function key, turn on the transmitter by inserting the battery. The MENU screen appears.
- 3 Press the ► key to move the cursor to "SYSTEM INITIALIZE".



- 4 Press the Function key to enter the SYSTEM INITIALIZE screen.
- 5 Press the Function key. A confirmation message appears.
To return to the MENU screen without initializing, press the Screen key.



- 6 Press the Function key to initialize settings.
To cancel initializing, press the ► key. The screen returns to the MENU screen.

Attaching Electrodes, SpO₂ Probe and NIBP Cuff to the Patient

The transmitter can be attached to an arm of the patient or placed on the bedside. The required length of the electrode leads and SpO₂ probe cable depends on how the transmitter is to be attached to the patient.


NOTE: Monitoring SpO₂ during NIBP Measurement

When the SpO₂ probe is attached to the same limb as the NIBP cuff, the blood flow decreases during NIBP measurement and pulse wave cannot be detected and SpO₂ cannot be monitored properly. When "INHIBIT SpO₂ DURING NIBP" on the PARAMETER SETUP screen is set to ON (factory default setting), SpO₂ monitoring is paused during NIBP measurement to avoid SpO₂ alarm occurrence. However, when monitoring SpO₂ on the same limb as NIBP, be careful when reading SpO₂ values.


When monitoring SpO₂ is important, attach the probe to the limb to which the NIBP cuff or catheter is not attached.

Attachment Examples

When transmitter is attached on an arm



When transmitter is placed on a bed






NOTE: When placing the transmitter on a bedside, place it on a stable and flat place. If the transmitter falls off, it may be damaged.

Selecting Electrode Leads and Disposable Electrodes

CAUTION

Only use Nihon Kohden specified electrodes and electrode leads. When other type of electrodes or electrode leads are used, the “CHECK ELECTRODES” message may be displayed and ECG monitoring may stop.

The following optional electrodes and electrode leads can be used with the transmitter.


Disposable Electrode	Electrode Lead	Note
Vitrode  The example is Vitrode C.	BR-903PA 	Three electrodes Clip type
	BR-913PA 	Three electrodes Clip type
	BR-906PA 	Six electrodes Clip type
	BR-916PA 	Six electrodes Snap type

Refer to the “Options” section when ordering the disposable electrodes and electrode leads.

Connecting the Electrode Lead to the Transmitter

Connect the electrode lead to the ECG/RESP socket on the transmitter.

NOTE: Hold the sides of the transmitter when connecting or disconnecting the electrode lead. Pressing the display while holding the transmitter may damage the transmitter.



CAUTION

Do not shake or swing the transmitter while holding the leads or cables connected to the transmitter. The transmitter may come off and injure someone or damage surrounding instruments.

CAUTION

Hold the connector of the electrode lead when connecting/disconnecting the electrode lead. If you disconnect the electrode lead by pulling the lead, it damages the electrode lead.

Electrode Position

Follow the physician's instructions for electrode placement when available.

For ECG monitoring, electrodes are attached only on the chest to allow patient movement and obtain continuous stable ECG. The following leads are examples. When also monitoring respiration, refer to the "Electrode Position for Respiration Monitoring" section.


NOTE: The optimum electrode positions for ECG measurement are not always optimum for respiration measurement. Select positions that are suitable for both ECG and respiration measurement or positions which give priority to either ECG or respiration measurement.

Electrode Positions for ECG Monitoring

Six Electrodes

The 6-electrode method with lead II and lead V5 is effective for monitoring myocardial ischemia.

You can improve monitoring accuracy considerably by adding lead V4 to this combination. Va and Vb can be at any position of the standard 12 leads V1 to V6, but V4 and V5 are most appropriate for myocardial ischemic monitoring.




Electrode Position	Symbol		Lead Color	
	AHA	IEC	AHA	IEC
Left infraclavicular fossa	LA	L	Black	Yellow
Right infraclavicular fossa	RA	R	White	Red
Below lowest rib on the left anterior axillary line	LL	F	Red	Green
Right anterior axillary line at the same level as LL/F	RL	RF	Green	Black
Fifth intercostal space on the left midclavicular line. (V4 position of standard 12 leads)	Va	Ca	Brown-blue	White-brown
Left anterior axillary line at the same level as Va. (V5 position of standard 12 leads)	Vb	Cb	Brown-orange	White-black


Lead Position

Standard limb leads


Lead I



Lead II




Lead III




Monopolar limb leads


aV_R lead



aV_L lead




aV_F lead




Monopolar chest leads

V1 to V6 leads




Three Electrodes

- Lead MII, which is similar to standard lead II, used when ECG measurement has priority




Electrode Position	Symbol		Lead Color	
	AHA	IEC	AHA	IEC
Left infraclavicular fossa	LA	L	Black	Yellow
Right infraclavicular fossa	RA	R	White	Red
Below lowest rib on the left anterior axillary line	LL	F	Red	Green

- Lead MI, which is similar to standard lead I
- Change F/LL and L/LA of lead MII



- Lead MIII, which is similar to standard lead III
- Change R/RA and L/LA of lead MII.




If the electrode positions above are not available due to chest surgery, attach the electrodes to the root of the limbs or below the clavicles for stable ECG monitoring.

Electrode Positions for Respiration Monitoring

Place the R/RA and F/LL electrodes so that the lungs are between the electrodes.

Position 1


In this position, respiration measurement is available; however, there is a difference in amplitude between different patients.



R or RA	F or LL
Right infraclavicular fossa	Fifth intercostal space on the left midclavicular line, V4

Position 2


In this position, the waveform amplitude is usually large and the ECG lead is similar to Lead MII. This position can be generally recommended.



R or RA	F or LL
Right infraclavicular fossa	Fifth intercostal space on the left midaxillary line, V6

Position 3


In this position, the respiration waveform is optimum, but the ECG lead is unusual.



R or RA	F or LL
Right midaxillary at the horizontal level of V6	Fifth intercostal space on the left midaxillary line, V6

Position 4

In this position, the respiration measurement is influenced by the impedance variation of the abdomen, so the cardiac pulse wave included in the respiration wave is reduced. Note that the waveform is inverted in phase compared with the chest movement (the waveform goes down during inspiration). It is difficult to measure the ECG at the same time.



R or RA	F or LL
Lowest rib on the right anterior axillary line	Lowest rib on the left anterior axillary line

Attaching Electrodes to the Patient and Connecting the Electrode Leads to Disposable Electrodes

Prepare the Patient Skin

Shave off excessive body hair.

To reduce skin impedance, clean the electrode site with cream or with a cotton pad moistened with alcohol. Thoroughly dry the skin with a clean cotton pad.

- NOTE**
- For a patient with frequent body movement, rub the sites with Skinpure skin preparation gel. However, do not use Skinpure skin preparation gel on sensitive skin.
 - Do not place electrodes on a wound or on an inflamed, wrinkled or uneven skin surface.


Attaching Electrodes to the Patient

⚠ CAUTION


Do not reuse disposable parts and accessories.

- NOTE**
- To maintain good contact between the electrode and skin, check that the paste of the disposable electrode is not dry.
 - When contact between the disposable electrode and skin becomes poor, replace electrodes with new ones immediately. Otherwise, contact impedance between the skin and the electrode increases and accurate ECG cannot be obtained.

Refer to the electrode operator's manual for details.



- 1** Carefully remove the backing paper from the electrode. Avoid touching the adhesive surface.
- 2** Place the electrode on the previously cleaned skin. Pay attention to the electrode lead color and symbol.



- 3** Clip the electrode lead to the electrode.
- 4** Fasten the electrode lead wire with surgical tape with an extra length of wire between the tape and the electrode. This lessens the movement of electrode leads by body movement and helps stable monitoring.

Checking ECG on the Transmitter Screen

After attaching electrodes and connecting ECG leads, check that the electrodes are properly attached to the patient and the ECG waveform is acquired on the check electrodes screen. For details on the screen, refer to the "Screen Descriptions" section.

Attaching the SpO₂ Probe

Selecting the SpO₂ Probe

Select an appropriate probe for the patient.

CAUTION


Only use Nihon Kohden specified electrodes, electrode leads and SpO₂ probes. Otherwise, the maximum performance from the transmitter cannot be guaranteed.



CAUTION

Do not use a damaged or disassembled probe. It causes incorrect measurement and may injure the patient.

Reusable Probes

When using a TL-201T finger probe, choose the appropriate cable length for attachment.


Probe	Cable Length	Patient	Attachment Site
Finger probe TL-201T 	0.6 m	Adult or child 20 kg or more	Finger
	1.6 m		
Finger probe TL-630T1, TL-630T3 	TL-630T1: 0.6 m	Adult or child 50 kg or more	
	TL-630T3: 1.6 m		

Probe	Cable Length	Patient	Attachment Site
Finger probe TL-631T1, TL-631T3 	TL-631T1: 0.6 m	Adult or child 20 kg or more	Finger or toe
	TL-631T3: 1.6 m		
Multi-site probe TL-220T 	1.6 m	Adult or infant 3 kg or more	Finger or toe
		Neonate 3 kg or less	Instep and sole






Disposable Probes

⚠ CAUTION

The disposable probe is not sterilized. Use the disposable probe only for a single patient. Never reuse the disposable probe for another patient because it causes cross infection.

Probe	Patient	Attachment Site
SpO ₂ probe TL-051S, TL-052S 	Adult 50 kg or more	Finger
	Neonate 3 kg or less	Instep and sole


Cable length TL-051S: 0.8 m
TL-052S: 1.6 m

Probe	Patient	Attachment Site
SpO ₂ probe TL-061S, TL-062S  Cable length TL-061S: 0.8 m TL-062S: 1.6 m	Adult or child 15 to 50 kg	Finger
	Infant 3 to 15 kg	Toe
SpO ₂ probe TL-271T, TL-271T3  Cable length TL-271T: 0.8 m TL-271T3: 1.6 m	Adult 30 kg or more	Finger or toe
SpO ₂ probe TL-272T, TL-272T3  Cable length TL-272T: 0.8 m TL-272T3: 1.6 m	Child 10 to 50 kg	Finger or toe
SpO ₂ probe TL-273T, TL-273T3  Cable length TL-273T: 0.8 m TL-273T3: 1.6 m	Neonate 3 kg or less	Instep and sole
	Adult 40 kg or more	Finger or toe
SpO ₂ probe TL-274T, TL-274T3  Cable length TL-274T: 0.8 m TL-274T3: 1.6 m	Infant 3 to 20 kg	Finger or toe

Connecting the SpO₂ Probe to the Transmitter

Connect the probe to the SpO₂ socket on the transmitter.

NOTE: Hold the connector of the probe cable when connecting or disconnecting the probe. Pulling the cable may damage the probe.



When transmitter is attached on an arm



CAUTION

Do not shake or swing the transmitter while holding the leads or cables connected to the transmitter. The transmitter may come off and injure someone or damage surrounding instruments.

CAUTION

Do not use a damaged or disassembled probe. It causes incorrect measurement and may injure the patient.

Attaching the Probe to the Patient

Attach the probe to the patient by referring to the probe's manual. Make sure that the light emitter and photo detector of the probe face each other at the attachment site.

WARNING

- When using the TL-201T finger probe, do not fasten the probe and cable to the finger by wrapping with tape. This may cause burn, congestion or pressure necrosis from poor blood circulation.
- When using probes other than the TL-201T finger probe, to avoid poor circulation, do not wrap the tape too tight. Check the blood circulation condition by observing the skin color and congestion at the skin peripheral to the probe attachment site. Even for short-term monitoring, there may be burn or pressure necrosis from poor blood circulation, especially on neonates or low birth weight infants whose skin is delicate. Accurate measurement cannot be performed on a site with poor peripheral circulation.

WARNING

Check the circulation condition by observing the skin color at the measurement site and pulse waveform. Change the measurement site every 8 hours for disposable probes and every 4 hours for reusable probes (every 8 hours for TL-631T series probe). The skin temperature may increase at the attached site by 2 or 3°C (4 or 5°F) and cause a burn or pressure necrosis. When using the probe on the following patients, take extreme care and change the measurement site more frequently according to symptoms and degree.

- Patient with a fever
- Patient with insufficient peripheral circulation
- Neonate or low birth weight infant with delicate skin

 WARNING

When monitoring SpO₂ of a patient who is receiving photodynamic therapy, the light from the finger probe sensor may cause a burn. Photodynamic therapy uses a photosensitizing agent that has a side effect of photosensitivity.

The SpO₂ probes manufactured by Nihon Kohden have two wavelengths with peaks in the range of 650 and 950 nm. The maximum light intensity is less than 5.5 mW.

 CAUTION

If the attachment site is dirty with blood or bodily fluids, clean the attachment site before attaching the probe. If there is nail polish on the attachment site, remove the polish. Otherwise, the amount of transmitted light decreases, and measured value may be incorrect or measurement cannot be performed.

 CAUTION

When the probe is attached on an appropriate site with sufficient thickness and the error message confirming the probe attachment repeatedly appears, the probe may be deteriorated. Replace it with a new one.

 CAUTION

If the skin gets irritated or redness appears on the skin from the probe, change the attachment site or stop using the probe. Take extreme care for the patients with delicate skin.

 CAUTION

Do not use a probe which is deteriorated by aging. Accurate measurement cannot be performed.

Attaching the NIBP Cuff

Selecting the NIBP Cuff


Select the NIBP cuff appropriate for the patient.

NOTE: NIBP cannot be measured on neonates using this transmitter.

Reusable Cuffs


When attaching the transmitter to the patient arm, a special NIBP cuff is required. An optional YN-990P extension hose (1.5 m) is available to extend the length between the NIBP socket on the transmitter and NIBP cuff (e.g. when not attaching the transmitter to the patient arm and placing the transmitter on a bedside).

Reusable Cuff		Model	Width (cm)	Air Hose Length (cm)
For adult	Standard	YP-503P	13	15
	Large	YP-504P	15	



When not attaching the transmitter to the patient arm, the following cuffs can be used. To use these cuffs, an optional YN-990P extension hose (1.5 m) is required.

Reusable Cuff		Model	Width (cm)	Air Hose Length (cm)
For infant		YP-960T	5	15
For child	Small	YP-961T	7	
	Standard	YP-702T	10	
		YP-962T		
For adult	Standard	YP-703T	13	
		YP-963T		
	Large	YP-964T	15	




Disposable Cuffs

CAUTION

Disposable cuffs are not sterilized. If necessary, sterilize the cuff using glutaraldehyde solution.


Disposable Cuff		Model	Width (cm)	Air Hose Length (cm)
For infant		YP-810P	6	17
For child		YP-811P	8	
For adult	Small	YP-812P	10	
	Standard	YP-813P	14	
	Medium large	YP-814P	15	
	Large	YP-815P	17	
	Extra large	YP-816P	18	



Extension Hose

CAUTION

When using an extension hose, check that the extension hose is not bent or squeezed. Otherwise, the cuff might not inflate or deflate. If the cuff cannot deflate, it may cause congestion on the patient at the cuff attachment site.




YN-990P extension hose, 150 cm

Reference for selecting a cuff

The AHA (American Heart Association) recommends that the cuff width be 40% of the circumference of the upper arm. Refer to the following graph and select the cuff which suits the patient's arm.

- NOTE**
- If a range of arm circumference appropriate for the cuff is prescribed, use a cuff within that range.
 - To obtain accurate measured values, select a wide cuff which can be attached to the upper arm. Measuring with a very narrow cuff may result in measured values higher than the actual values.
 - The YP-503P NIBP cuff is for standard size adult. Do not use this cuff when it does not fit the patient.




Connecting the NIBP Cuff to the Transmitter


When Using YP-503P/YP-504P NIBP Cuff

To attach the YP-503P/YP-504P NIBP cuff to the transmitter, the lock plate is required.

YP-503P/YP-504P NIBP cuff




Lock plate




The lock plate is for attaching the NIBP cuff to the transmitter.


NOTE: Do not roll up or put weight on the cuff when the lock plate is attached to it. The lock plate may break if the cuff is rolled up or weight is put on it when the lock plate attached.




- 1** Remove the lock plate from the transmitter.




- 2** Insert the lock plate into the lock plate pocket on the NIBP cuff.



- 3** Attach the transmitter to the lock plate by inserting the tabs on the lock plate into the slots on the transmitter.



- 4** Cover the transmitter with the front cover of the NIBP cuff.





- 5** Connect the air hose to the NIBP socket on the transmitter. Turn the cuff connector joint until it clicks.

When Using YP-810P series Disposable Cuffs or YP-702T/703T or YP-960T series Reusable Cuffs

To use these NIBP cuffs, an optional YN-990P extension hose (1.5 m) is required.

NOTE: Connect the joints properly. If there is an air leak, NIBP cannot be measured properly.



1 Connect the NIBP cuff to the extension hose.

2 Connect the other end of the extension hose to the NIBP socket on the transmitter. Turn the joint clockwise until it clicks.

To disconnect the cuff from the transmitter, turn the hose joint counterclockwise.

Attaching the NIBP Cuff to the Patient

WARNING

Be careful when measuring NIBP on a patient with known bleeding disorders or coagulation. After NIBP measurement, there may be dot hemorrhage, or circulatory disorder by thrombus where the cuff is attached.

CAUTION

Do not wrap the cuff on an arm or thigh which is used for injection. NIBP measurement on an arm or thigh which is used for injection may cause reflux of blood and stop injection.

 CAUTION

Do not attach the cuff to the site where there is injury or inflammation. If the skin gets irritated or redness appears on the skin from the cuff, change the attachment site or stop using the cuff. Take extreme care on the patients with delicate skin.

 CAUTION

Do not wrap the cuff too tight. It may cause poor blood circulation and congestion. If the cuff is wrapped too loosely, the NIBP value may increase.

 CAUTION

Do not reuse disposable parts and accessories.

 CAUTION

NIBP and SpO₂ can be measured on the same limb, but the SpO₂ monitoring might not be accurate during NIBP measurement. Be careful when reading the SpO₂ values.*


* Monitoring SpO₂ during NIBP Measurement

When the SpO₂ probe is attached to the same limb as the NIBP cuff, the blood flow decreases during NIBP measurement and pulse wave cannot be detected and SpO₂ cannot be monitored properly. When “INHIBIT SpO₂ DURING NIBP” on the PARAMETER SETUP screen is set to ON (factory default setting), SpO₂ monitoring is paused during NIBP measurement to avoid SpO₂ alarm occurrence. However, when monitoring SpO₂ on the same limb as the NIBP, be careful when reading SpO₂ values.

- NOTE**
- Measuring NIBP at a site other than the upper arm gives different values from those measured at the upper arm. When making diagnosis based on the NIBP values, measure NIBP on an upper arm.
 - To accurately detect the pulsatile flow of the artery, the cuff should be wrapped around a bare upper arm.
 - Do not use an abnormal cuff. The cuff deteriorates from use and cleaning. Before use, check the cuff and confirm that there is no flaw, crack or hole in it. Be careful not to damage the inflation bag. If the inflation bag has a hole or a flaw, it may burst during use. Dispose of an abnormal cuff and replace it with a new one.
 - Refer to the NIBP cuff manual for details.

Cuff Position


When placing the transmitter on a bed, make sure that the hose is not bent.




Place the cuffed upper arm (brachium) at the same height as the patient's heart. If the cuff is not at the same level as the heart, the weight of the blood affects the blood pressure reading. The pressure difference per unit height is 0.7 mmHg/cm. The blood pressure reading decreases when the arm is higher than the heart and increases when lower.

The best measuring condition is when the patient is lying on his/her back with arms and legs relaxed. If the cuff position cannot be on the same level as the heart, the displayed blood pressure reading must be mathematically adjusted.


Attaching the Transmitter on an Arm (Using the YP-503P/YP-504P NIBP Cuff)



- 1** Attach the NIBP cuff to the transmitter. Refer to the “Connecting the NIBP Cuff to the Transmitter” section.



- 2** Before putting the cuff over the arm, insert the end of the cuff into the belt and then through the D ring as shown at left.



- 3** Put the patient arm through the cuff. Fold back the cuff at the D ring and fasten it using the velcro tape. Make sure that the cuff is not attached on a joint.


NOTE: The cuff must not wrap around the elbow.

Attaching the Strap to the Transmitter


- NOTE**
- Use the strap to prevent the transmitter from falling.
 - Do not attach the clip to hard objects such as thick cloth or zipper. It will break the clip.

Attach a strap provided with the transmitter to the NIBP cuff and patient clothes.

To open the clip, firmly pull out the tab in direction of the arrow.




To adjust the strap length, push down the tab on the adjuster and slide.




- 1 Adjust the length of the strap.
- 2 Clip one end of the strap to the belt for the strap on the NIBP cuff.
- 3 Clip the other end of the strap to the patient's clothes as shown left.


Belt for the strap on the NIBP cuff



Placing the Transmitter on the Bed (Using the YP-810P series Disposable Cuffs or YP-702T/703T or YP-960T series Reusable Cuffs)



- 1 Put the cuff on the upper arm so that the ▼ mark of "ARTERY ▼" aligns with the artery of the patient.



- 2 Wrap the cuff so that "INDEX ➡" comes within the "← RANGE ➡".


If "Index ➡" is not within the "← RANGE ➡", change the cuff size.

Locking the Keys on the Transmitter

To prevent the patient from pressing the keys on the transmitter during monitoring, you can lock the keys.

To lock the keys:

Press the ◀ and ▶ keys at the same time and hold for more than 3 seconds. The “Key locked” screen appears.



When the screen time-out period is set to “1 min” (factory default), the “Key locked” screen is displayed for 5 seconds, then the display turns off if there is no key operation.

When the screen time-out period is set to a certain number of minutes, the “Key locked” screen is displayed for 5 seconds, then it changes to the numeric and waveform screen. If there is no key operation, the display turns off when the remaining time elapses. Refer to the “Turning the Display Off” section for details.

To unlock the keys:

Press the ◀ and ▶ keys at the same time and hold for more than 3 seconds.

Monitoring

CAUTION

The measurement values and displayed waveforms on the transmitter and receiving monitor may be different due to timing delay of the display or difference in detection settings.


- NOTE
- The waveforms displayed on the check electrodes screen are just to check that the electrode or probe is attached to the patient properly. Do not use the displayed waveforms for other purposes.
 - Do not let the transmitter directly touch the patient's skin. The transmitter temperature rises and this may cause burn to the patient.

Screen Descriptions

When the transmitter is turned on, the startup screen appears, then the check electrodes screen appears to check the electrode attachment.

The screen changes in the following order when the Screen key is pressed.

Check electrodes → numeric and waveform → waveform review → numeric review → display off → check electrodes . . .





Check Electrodes Screen

You can check whether the electrodes are properly attached to the patient and the ECG waveform is acquired.

When 6 leads are used, the I, II, Va and Vb lead waveforms are displayed.

When 3 leads are used, only the lead II waveform is displayed.




When electrodes are detached, the “CHECK ELECTRODES” message and detached electrode position appear on the screen.


NOTE: When ECG measurement is set to OFF on the PARAMETER SETUP screen, the check electrodes screen does not appear.

Numeric and Waveform Screen

Numeric values and waveforms of the monitoring parameters are displayed. You can change the ECG lead with the ◀ and ▶ keys.



When ECG and respiration measurement is turned off




NOTE: The pulse wave amplitude varies according to the ratio of the pulsation component to the entire transmitted IR signal. When the pulsation component ratio is 1%, the pulse wave amplitude is about 5 mm at $\times 1$ sensitivity on the screen.

Waveform Review Screen

ECG full disclosure for up to 10 minutes can be saved and reviewed. When ECG measurement is turned off and SpO₂ is monitored, the pulse waveform is saved.

When ECG lead is changed on the numeric and waveform screen, the ECG full disclosure of the changed lead is saved.

The saved data is deleted when the transmitter is turned off.



To scroll the waveform, press the ◀ or ▶ key. The waveform is scrolled by 30 seconds.

Numeric Review Screen

Numeric data of heart rate (or pulse rate when ECG is turned off), SpO₂ and respiration rate for up to 10 minutes are saved at 1 minute intervals.


NOTE: NIBP measured values are not saved.

The saved data is deleted when the transmitter is turned off.

	chE898			
		HR (bpm)	SpO ₂ (%)	RR (/min)
Older data	-10[min]	80	97	20
	- 9[min]	80	97	20
	- 8[min]	80	97	20
	- 7[min]	80	97	20
Time before the numeric review screen is displayed.	- 6[min]	80	97	20
	- 5[min]	80	97	20
	- 4[min]	80	97	20
	- 3[min]	80	97	20
	- 2[min]	80	97	20
Newer data	- 1[min]	80	97	20


Display Off

The display can be turned off any time. Refer to the “Turning the Display Off” section for details.



Basic Monitoring Operation

Using the Function Key



One of the following functions can be assigned to the Function key on the SYSTEM SETUP screen. Refer to the “Changing SYSTEM SETUP Settings” section.

SUSPEND ALARM: Suspends alarms on the receiving monitor, before they occur, for 2 minutes.

PAUSE: Pauses monitoring on the transmitter and receiving monitor.

CONFIRM: Transmits the signal that the patient is confirmed and displays the “PATIENT CONFIRMED” message on the transmitter.

NOTE: To use the Function key for pause or confirm, “PROTOCOL” on the SYSTEM SETUP screen of the transmitter must be set to 57 and the receiving monitor must be able to receive protocol 57. A central monitor with an ORG-9100A series multiple patient receiver whose software version 02-01 or later can receive this protocol.

Suspending Alarms on the Receiving Monitor

WARNING

While the “ALARMS SUSPENDED” message is displayed on the transmitter, all alarms on the receiving monitor are suspended so keep the patient under close observation.

When the FUNCTION KEY is set to “SUSPEND ALARM” or “SUSPEND ALARM & PAUSE” on the SYSTEM SETUP screen, alarms can be suspended for 2 minutes on the receiving monitor before they occur.


To suspend alarms:

- 1 Press the Function key. The “Suspend alarms” confirmation screen appears.



- 2 Press the ► key to suspend alarms.
To cancel suspending alarms and return to the previous screen, press the Screen key.

When the alarms are suspended, the “ALARMS SUSPENDED” message and alarm suspended icon with the remaining minutes in alarm suspension appear on the transmitter screen.



To cancel suspending alarms during 2 minute alarm suspension:

- 1 Press the Function key while the “ALARMS SUSPENDED” message is displayed. The confirmation screen appears.



- 2 Press the ► key to cancel alarm suspension.
Press the Screen key to not cancel alarm suspension.

Pausing Monitoring

When FUNCTION KEY is set to “SUSPEND ALARM & PAUSE” on the SYSTEM SETUP screen, you can pause monitoring on the receiving monitor from the transmitter when the patient cannot be monitored, such as during X-ray examination.

NOTE: To use the Function key for pause, “PROTOCOL” on the SYSTEM screen of the transmitter must be set to 57 and the receiving monitor must be able to receive protocol 57.

Monitoring on the receiving monitor resumes when one of the following conditions is met. When OFF is selected for “AUTO RESUME AFTER PAUSE” on the SYSTEM SETUP screen, monitoring does not automatically resume.

- Heart rate is properly monitored on the transmitter for the interval selected for “AUTO RESUME AFTER PAUSE”
- SpO₂ is properly monitored on the transmitter for the interval selected for “AUTO RESUME AFTER PAUSE”
- NIBP is properly measured and the SYS, DIA or MAP value is displayed on the transmitter

To pause monitoring:

- 1 Press the Function key. The “Suspend alarms” confirmation screen appears.



- 2 Press the Function key for 3 seconds to display the “Pause monitoring” confirmation screen.



- 3 Press the ► key to pause monitoring.
To cancel pause monitoring, press the Screen key.
- 4 Wait about 5 seconds until the “Turn power off” screen appears.



5 Turn off the transmitter.

If the transmitter is not turned off and monitoring continues for the interval set for “AUTO RESUME AFTER PAUSE” on the SYSTEM SETUP screen, pause monitoring is cancelled and monitoring continues.

Resuming Monitoring after Pause

To resume monitoring after pause, check that the electrodes, electrode leads and probe are attached to the patient then turn on the transmitter.

⚠ WARNING


When the patient returns to the bed, turn on the transmitter and check that the monitoring is resumed on the receiving monitor.

⚠ WARNING

If the transmitter is not turned off and monitoring continues for the selected interval, pause monitoring is canceled and monitoring continues. Check that the monitoring is resumed on the receiving monitor.

Confirming the Patient


When the FUNCTION KEY is set to “CONFIRM” on the SYSTEM SETUP screen, you can transmit signal to the receiving monitor to indicate that the patient is confirmed by a medical staff by pressing the Function key.



Turning the Display Off

The display can be turned off any time. To turn off the display:

- 1 Press the Screen key several times until the following screen appears.



- 2 Select the timing for turning off the display with the ◀ or ▶ key. The selected item is highlighted in blue. A 10 second countdown starts. You can select a different time within the 10 second countdown.

1 min (factory default): Turns the display off 1 minute later.

5, 10, 15, 30, 60, 120 or 240 min: Turns the display off when the selected time elapses. To set the time, refer to “SELECTABLE SCREEN TIME OUT PERIOD (min)” in the “Changing System Setup Settings” section.

CONTINUE: Keep the display turned on.

NOTE: If longer than “1 min” is selected, it reduces the battery lifetime.

- 3 Wait 10 seconds until the countdown ends. When the countdown ends, the numeric and waveform screen appears.

Or, press the Screen key before the countdown ends. The check electrodes screen appears.

When the selected time elapses without any key operation, the display turns off automatically. If a key is pressed, the countdown resets.

- When displaying the check electrodes screen after selecting the time-out period, the screen automatically changes to the numeric and waveform screen 2 minutes later if there is no key operation, then the display turns off when the remaining time elapses.
- When displaying another screen after selecting the time-out period, the screen changes to the numeric and waveform screen 1 minute later if there is no key operation, then the display turns off when the remaining time elapses.
- When “1 min” is selected, the display turns off without changing to the numeric and waveform screen.

When the display is turned off automatically or the power is turned off, the setting returns to “1 min” (factory default).

Turning the Display On after It was Turned Off

Press the Screen key. One of the following screen appears.


- The previous screen: The Screen key is pressed within 5 minutes after the display turned off.
- Numeric and waveform screen: The Screen key is pressed more than 5 minutes after the display turned off.
- “Key locked” screen: The Screen key is pressed after the display turned off and the “Key locked” screen was the last screen before the display turned off.

ECG and Respiration Monitoring

When the electrodes are attached and the ECG leads are connected, the heart rate, ECG waveform, respiration rate and respiration waveform appear on the receiving monitor. Refer to the operator's manual of the receiving monitor for details.

When 6 leads are used on this transmitter, up to 8 leads (I, II, III, aVR, aVL, aVF, Va and Vb) of ECG waveforms can be displayed on the receiving monitor. The heart rate is also measured.

When 3 leads are used, one channel ECG waveform of lead II can be displayed on the receiving monitor. Refer to the operator's manual of the monitor for details.



If the receiving monitor does not detect the pacemaker pulse, change the electrode positions. Attach the electrode lead clip to change the monitoring electrode.

 WARNING

The bioelectric impedance measurement sensor of a minute ventilation rate-adaptive implantable pacemaker may be affected by transmitter which is connected to the same patient. If this occurs, the pacemaker may pace at its maximum rate and the transmitter may give incorrect data to the monitor. If this occurs, disconnect the electrode leads from the patient or change the setting on the pacemaker by referring to the pacemaker's manual. For more details, contact your pacemaker representative or Nihon Kohden representative.

 WARNING

Turn the pacing pulse detection to ON on the receiving monitor when monitoring a pacemaker patient. Otherwise the pacemaker pulse is not rejected. However, even when the pacing pulse detection is set to ON, the pacemaker pulse might not be rejected. When the pacemaker pulse is not rejected, the pacemaker pulse is detected as QRS and false heart rate may be indicated or critical arrhythmia such as asystole may be overlooked. Keep pacemaker patients under close observation.

For the pacemaker pulse rejection capability of the ZM-540PA and ZM-541PA transmitters, refer to the "Specifications - ECG" section.

 WARNING

Even when the pacing pulse detection is set to ON on the receiving monitor, the pacemaker pulse can be overlooked or detected as QRS. You cannot confirm the pacemaker operation only from the detected pacemaker pulse.

⚠ WARNING

When the transmitter is used with an electrosurgical unit (ESU), firmly attach the entire area of the ESU return plate. Otherwise, the current from the ESU flows into the electrodes of the transmitter, causing electrical burn where the electrodes are attached. For details, refer to the ESU manual.

⚠ CAUTION

Turn off the power of mobile phones, small wireless devices and other devices which produce strong electromagnetic interference around a patient (except for devices allowed by the hospital administrator). Radio waves from devices such as mobile phones or small wireless devices may be mistaken as pulse waves or respiration waves and the displayed data may be incorrect.

⚠ CAUTION

When the "NOISE" message is displayed on the receiving monitor, ECG is not monitored and the ECG alarm does not function. Remove the cause by checking the electrodes, electrode leads, patient's body movement, EMG and peripheral instruments grounding. Also make sure that an electric blanket is not used.

- NOTE**
- Noise generated from an electrosurgical unit may interfere on an ECG waveform, but will not damage it.
 - If an electric blanket is used and incorrect heart rate is displayed on the monitor, turn off the pacing spike detection on the monitor.
 - Turn the pacing spike detection to ON on the receiving monitor when monitoring a pacemaker patient. Pacing pulse is detected by the transmitter and transmitted to the monitor. If the pacing spike detection is turned OFF, QRS and pacemaker spike might not be distinguished and pacemaker failure might not be recognized.
 - If defibrillation may be necessary, set the filter to Monitor or Maximum on the receiving monitor. If defibrillation is performed while the setting is OFF, the waveform recovery may be slow.
 - ECG cannot be monitored on a neonate using this transmitter.

Turning ECG Measurement On/Off

ECG measurement can be turned on or off on the PARAMETER SETUP screen. When electrodes are attached to the patient and ECG leads are connected, ECG monitoring starts even when ECG is turned off.

When “PROTOCOL” on the SYSTEM screen is set to 57:

ECG measurement on the receiving monitor is automatically set to OFF. Also, ECG measurement on the transmitter cannot be turned on or off from the receiving monitor.

When “PROTOCOL” on the SYSTEM screen is set to 42:

If ECG measurement is turned off on the transmitter, ECG measurement on the receiving monitor must also be turned off.

Turning Respiration Measurement On/Off

Respiration measurement can be turned on or off on the PARAMETER SETUP screen. If respiration measurement is turned off, respiration measurement on the receiving monitor is also turned off.

Electrode Detachment

In the following conditions, the “CHECK ELECTRODES” message is displayed on the transmitter and receiving monitor.

- Electrode is detached from skin.
- Electrode lead is disconnected from the electrode.
- Polarization voltage between the electrode and skin is excessively high.

In these cases, check the cause and if necessary, replace electrodes with new ones.


CAUTION

When the “CHECK ELECTRODES” message is displayed on the receiving monitor, ECG is not monitored properly and the ECG alarm does not function. Check the electrode, electrode leads, and if necessary, replace with new ones.

SpO₂ Monitoring

When monitoring starts, SpO₂ and the pulse waveform are sent to the monitor and SpO₂ and the pulse level bar graph are displayed on the transmitter screen. When ECG is not measured, the pulse waveform and pulse rate are also displayed.

Check that pulse wave is displayed and that the “CHECK PROBE” message is not displayed. If the probe is detached, SpO₂, pulse rate and pulse wave are not displayed on the display.



NOTE: During SpO₂ monitoring, the transmitter constantly checks the light intensity of the probe and adjusts the light intensity to maintain optimum measurement condition. During adjustment, the pulse wave becomes flat for about 0.5 seconds.



Measuring Principle

SpO₂ is measured by attaching a probe to a place where light can easily penetrate, such as a finger. The probe has two light emitter diodes and one photo detector.

Light of two wavelengths, 660 nm and 940 nm, are emitted from the emitter diodes.

Oxygenated and deoxygenated blood absorb different wavelengths of light as shown in fig. 1 and 2. The light absorption of the skin, tissue, bone and venous blood is constant and can be eliminated. The pulsatile change represents arterial blood only, as shown in fig. 3, and this is measured to determine SpO₂. The amount of unabsorbed light at each wavelength is detected by the photo detector. SpO₂ is calculated from the ratio of absorbed light of the two wavelengths (fig. 2).

SpO₂ can be measured correctly when there is an arterial pulse which causes change in the transmitted light of both wavelengths. When the patient has asystole and no arterial pulse, there is a chance that some external factor such as movement of the SpO₂ probe may cause change in the transmitted light and produce a pulse-like signal which causes an incorrect SpO₂ value to be displayed. This incorrect SpO₂ value depends on the optical properties of the change in transmitted light.




Fig. 1




Fig. 2




Fig. 3

⚠ WARNING

SpO₂ measurement may be incorrect in the following cases.

- When the patient's carboxyhemoglobin or methemoglobin increases abnormally.
- When dye is injected in the blood.
- During CPR.
- When measuring at a site with venous pulse.
- When there is body movement.
- When the pulse wave is small (insufficient peripheral circulation).

⚠ WARNING

Check the circulation condition by observing the skin color at the measurement site and pulse waveform. Change the measurement site every 8 hours for disposable probes and every 4 hours for reusable probes (every 8 hours for TL-631T series probe). The skin temperature may increase at the attached site by 2 or 3°C (4 or 5°F) and cause a burn or pressure necrosis. When using the probe on the following patients, take extreme care and change the measurement site more frequently according to symptoms and degree.

- Patient with a fever
- Patient with insufficient peripheral circulation
- Neonate or low birth weight infant with delicate skin

⚠ WARNING

When monitoring SpO₂ of a patient who is receiving photodynamic therapy, the light from the finger probe sensor may cause a burn. Photodynamic therapy uses a photosensitizing agent that has a side effect of photosensitivity.

The SpO₂ probes manufactured by Nihon Kohden have two wavelengths with peaks in the range of 650 and 950 nm. The maximum light intensity is less than 5.5 mW.

 WARNING

When not monitoring SpO₂, disconnect the SpO₂ cable from the transmitter. Otherwise, noise from the probe sensor may interfere and incorrect data is displayed on the screen.

 CAUTION

Turn off the power of mobile phones, small wireless devices and other devices which produce strong electromagnetic interference around a patient (except for devices allowed by the hospital administrator). Radio waves from devices such as mobile phones or small wireless devices may be mistaken as pulse waves or respiration waves and the displayed data may be incorrect.

 CAUTION

Handle the probe cable according to the following cautions. Failure to follow these cautions may cause cable discontinuity or short circuit of the probe cable which may cause incorrect measurement data or inability to perform measurement. Also in rare cases, the probe temperature may increase and cause skin burn on the patient. If the probe cable is damaged, replace the probe with a new one.

- Do not pull or bend the probe cable.
- Do not let caster feet run over the probe cable.

 CAUTION

Normal external light does not affect monitoring but strong light such as a surgical light or sunlight may affect monitoring. If affected, cover the measuring site with a blanket.

 CAUTION

When the probe is attached on an appropriate site with sufficient thickness and the error message confirming the probe attachment repeatedly appears, the probe may be deteriorated. Replace it with a new one.

 CAUTION

When a message indicates a faulty probe, stop monitoring and replace the probe with a new one.

 CAUTION

While a patient is on medication which causes vasodilation, the pulse waveform may change and in rare cases the SpO₂ value might not be displayed.

 CAUTION

When monitoring SpO₂ only, detection of arrhythmia and asystole is not available and arrhythmia alarms such as ASYSTOLE, VF or VT are not available. If the patient requires ECG monitoring, monitor the ECG.

 CAUTION

When monitoring SpO₂ only (without ECG monitoring), turn on both the upper and lower limit alarms for PR and SpO₂ on the receiving monitor. If the patient's pulse is not detected during asystole or other condition, a "CANNOT DETECT PULSE" or "CHECK PROBE" alarm occurs instead of an SpO₂ limit alarm. Furthermore, if the patient has no pulse, noise from probe movement could be misjudged as a pulse and cause an incorrect PR or SpO₂ value to be displayed.

- NOTE**
- In order to maintain sufficient blood circulation, keep the measurement site warm by covering it with a blanket or something similar. Warming the site is effective, especially for a patient with a small pulse amplitude.
 - When monitoring a patient who has an IABP and SpO₂ cannot be measured, monitor the patient on a wired monitor. If the monitor has a sensitivity mode, set the mode to "MAX".
 - Unlike ECG monitoring, detection of arrhythmia and asystole is not available when monitoring SpO₂ only. If SpO₂ is monitored without ECG, PR and SpO₂ alarms do not occur during asystole because PR and SpO₂ values are not measured when there is no pulse. Furthermore, if the patient has no pulse, noise from probe movement could be misjudged as a pulse and cause an incorrect PR or SpO₂ value to be displayed.

Monitoring SpO₂ during NIBP Measurement

CAUTION

NIBP and SpO₂ can be measured on the same limb, but the SpO₂ monitoring might not be accurate during NIBP measurement. Be careful when reading the SpO₂ values.*

- * When the SpO₂ probe is attached to the same limb as the NIBP cuff, the blood flow decreases during NIBP measurement and pulse wave cannot be detected and SpO₂ cannot be monitored properly. When "INHIBIT SpO₂ DURING NIBP" on the PARAMETER SETUP screen is set to ON (factory default setting), SpO₂ monitoring is paused during NIBP measurement to avoid SpO₂ alarm occurrence. However, when monitoring SpO₂ on the same limb as NIBP, be careful when reading SpO₂ values.

When monitoring SpO₂ is important, attach the probe to the limb to which the NIBP cuff or catheter is not attached.

When SpO₂ monitoring is paused during NIBP measurement, the SpO₂ value just before the start of NIBP measurement and an  mark are displayed on the transmitter for 30 seconds.

When NIBP measurement is not completed after 30 seconds, "— — —" is displayed for the SpO₂ value. The same data also appears on the monitor screen.

- NOTE**
- When continuous SpO₂ monitoring is necessary, attach the probe to the limb to which the NIBP cuff is not attached and set "INHIBIT SpO₂ DURING NIBP" on the PARAMETER SETUP screen to OFF.
 - When the probe is attached to the same limb as the NIBP cuff, set the sync source to a parameter other than SpO₂ on the receiving monitor.
 - When monitoring SpO₂ during STAT NIBP measurement, attach the probe to the limb to which the NIBP cuff is not attached.

NIBP Monitoring

Selecting the Initial Cuff Inflation Pressure

The initial cuff inflation pressure can be changed on the PARAMETER SETUP screen. The default setting is 180 mmHg. To change the setting, refer to the “Changing PARAMETER SETUP Settings” section.


Selecting the Measurement Mode and Interval

Measurement Modes

There are three measurement modes: manual, auto and STAT. The selected mode or interval is displayed on the screen.

The measurement mode and interval can be changed by pressing the NIBP Interval key.

When the key is pressed, the NIBP mode setting screen appears. The measurement modes selected at “SELECTABLE INTERVALS” on the PARAMETER SETUP screen are displayed (key color: white). Select the measurement mode with the ◀ and ▶ keys or NIBP Interval key and press the Function key.



To select the modes to be displayed on the NIBP mode setting screen, refer to the “Changing PARAMETER SETUP Settings” section.

Manual Measurement

In Manual mode, a single NIBP measurement is performed when the NIBP Start/Stop key is pressed.

STAT (Continuous) Measurement

In STAT mode, measurement is continuously repeated for 15 minutes after the NIBP Start/Stop key is pressed.

When the STAT measurement for 15 minutes is completed, the measurement mode automatically changes to the Manual mode or Auto mode of selected interval depending on the “NIBP MODE AFTER STAT” setting on the PARAMETER SETUP screen. The default setting is Manual mode. Refer to the “Changing Parameter Setup Settings” section.

The STAT measurement completes within 15 minutes. When more than 12 minutes elapse from the start of measurement, there will be no more measurement performed and the measurement mode changes to the mode selected for “NIBP MODE AFTER STAT” on the PARAMETER SETUP screen.

Auto Measurement

In Auto mode, measurement is performed automatically at the preset time intervals.

In Auto mode, a single measurement can be performed by pressing the NIBP Start/Stop key between auto measurements.

Measuring NIBP

WARNING

Be careful when measuring NIBP on a patient with known bleeding disorders or coagulation. After NIBP measurement, there may be dot hemorrhage, or circulatory disorder by thrombus where the cuff is attached.

WARNING

NIBP measurement may be incorrect in the following situations.

- When using an ESU
- Body movement
- Small pulse wave
- Too many arrhythmias
- Shaking from an external source
- Rapid blood pressure change
- During CPR
- Slow pulse
- Low blood pressure
- Small pulse pressure
- Cuff is too tight or too loose
- Cuff does not fit the arm
- Cuff is wrapped over thick clothing
- Cuff is deteriorated
- Arterial sclerosis
- Poor perfusion
- Diabetes
- Age
- Pregnancy
- Pre-eclampsia
- Renal disease
- Shivering
- Trembling

 WARNING

When performing NIBP measurements in STAT mode or 5 minute intervals, periodically remove the cuff from the patient for ventilation. The skin temperature may increase at the cuff attachment site by 2 or 3°C (4 or 5°F). When measuring a patient with a fever or peripheral circulation insufficiency, it may cause a burn.

 WARNING

When performing long term measurement at intervals less than 5 minutes, perform measurements while observing the state of the patient, blood vessels and limb to ensure adequate circulation. Congestion may occur at the measurement site. When performing periodic measurement for a long time, periodically check the circulation condition.

 CAUTION

When performing NIBP measurement repeatedly, have a rest between measurements to recover adequate circulation.


- NOTE
- When measuring patients who are conscious, help the patient to relax. Measurement may not be accurate if the patient's arm is tense or if the patient talks.
 - The data for measurement on a leg tends to be higher than measurement on the arm. When making diagnosis based on the NIBP values, measure NIBP on an upper arm.
 - Do not apply pressure to the cuff or air hose. NIBP may not be measured correctly because of noise or NIBP measurement may stop due to the NIBP safety circuit.
 - When the transmitter is attached to the patient arm and the NIBP measurement is performed when moving, tell the patient to relax and keep quiet. Otherwise, measurement may be stopped or remeasurement is repeated due to body movement.
 - If there is an abnormal noise generated during measurement, stop using the transmitter and contact your Nihon Kohden representative.

- Do not measure NIBP of a patient on whom an IABP is being used. Measurement may be incorrect due to the mixing of the patient's own pulse and IABP pulse.
- NIBP cannot be measured on a neonate using this transmitter.


For Accurate NIBP Measurement

- Make sure the patient is comfortably seated. Do not perform measurement immediately after the patient is seated. Keep the patient in the resting state for five minutes before measurement.
- Make sure the patient has legs uncrossed, feet down on the floor, and back and back arm supported during measurement.

- 1 Press the NIBP Interval key to display the NIBP mode setting screen.




- 2 Select the measurement mode by pressing the NIBP Interval key or and keys.



- 3 Press the Function key.

- 4** Press the NIBP Start/Stop key to perform measurement.



The cuff is inflated and the inflation pressure is displayed on the screen.




- | | |
|-----------------|--|
| In manual mode: | Measurement is performed once. |
| In STAT mode: | Measurement is performed repeatedly for 15 minutes. |
| In auto mode: | The first measurement is performed when the NIBP Start/Stop key is pressed. The second measurement is performed when the current time in the transmitter reaches the selected time interval. |

To stop measurement during measurement, press the NIBP Start/Stop key again.

In STAT mode, after completing the STAT measurement, the measurement mode changes to the mode set for "NIBP MODE AFTER STAT" on the PARAMETER SETUP screen.

In auto mode, to stop measurement in auto mode, change the mode to manual. To cancel one measurement, press the NIBP Start/Stop key during measurement.

After the measurement is complete, the measured data is displayed on the screen and is transmitted to the monitor.



When ECG and SpO₂ are not monitored (ECG measurement is turned off and SpO₂ probe is not connected to the transmitter), the pulse rate at the end of NIBP measurement is displayed.

A buzzer can be set to sound at the start and end of NIBP measurement. Refer to the “Changing PARAMETER SETUP Settings” section.

Auto Mode Measurement

When auto mode measurement is selected, “STANDBY” message is displayed on the screen until the NIBP Start key is pressed for the first time.



A time bar appears to indicate the interval between auto mode measurements.



During auto mode measurement, the measurement mode can be changed. During the interval, press the NIBP Interval key to change the mode. When “MANUAL” is displayed for more than one second, the measurement in auto mode is stopped.

Data Display After NIBP Measurement

When the time set at “OLD NIBP DATA” on the PARAMETER SETUP screen elapses after the last measurement, the NIBP data is dimmed or hidden. Whether to dim or hide the old data can also be selected at “OLD NIBP DATA”. Refer to the “Changing PARAMETER SETUP Settings” section.

Monitoring SpO₂ during NIBP Measurement

When the SpO₂ probe is attached to the same limb as the NIBP cuff, the blood flow decreases during NIBP measurement and pulse wave cannot be detected and SpO₂ cannot be monitored properly. When “INHIBIT SpO₂ DURING NIBP” on the PARAMETER SETUP screen is set to ON (factory default setting), SpO₂ monitoring is paused during NIBP measurement to avoid SpO₂ alarm occurrence. However, when monitoring SpO₂ on the same limb as the NIBP, be careful when reading SpO₂ values.

Indication and Message List

Indication

Indication	Cause	Countermeasure
	Fully charged batteries	—
	Batteries are 1/3 full.	
	Batteries are weak.	Replace the batteries.
	Batteries are very weak. Cannot measure NIBP.	
	Alarms on the receiving monitor were suspended by pressing the Function key on the transmitter.	Alarms resume when the suspend interval elapses. To cancel alarm suspension, press the Function key again.

Messages

When “PROTOCOL” on the SYSTEM screen is set to 57, all messages are transmitted to the receiving monitor. When “PROTOCOL” is set to 42, the messages marked with * are not transmitted to the receiving monitor.

Message	Cause	Countermeasure
AIR LEAK	The cuff and extension hose are not properly connected	Connect them properly.
	The cuff hose (or extension hose) is not properly connected to the NIBP socket	
	The cuff or extension hose is damaged	Replace with a new one.
ALARMS SUSPENDED	Alarms on the receiving monitor is suspended by pressing the Function key on the transmitter	Alarms resume when the 2 minute suspend interval elapses. To cancel alarm suspension, press the Function key again.
BATTERY WEAK	Dead batteries	Replace batteries.

Message	Cause	Countermeasure
CANNOT DETECT PULSE* (displayed in blue)	Poor blood circulation for measuring the SpO ₂ value	Check the patient condition, probe attachment or change the attachment site.
	The probe is attached too tightly and is obstructing the blood circulation	Reattach the probe.
	The probe is not attached to the patient properly	Attach the probe to the patient properly.
	“LIGHT INTERFERENCE”, “CHECK PROBE SITE” or “DETECTING PULSE” message is displayed for more than 30 seconds	Refer to the cause and countermeasure for each message in this Messages table and remove the cause.
CANNOT DETECT PULSE (displayed in pink)	The patient's pulse wave is too small to measure NIBP	Measure by palpation or auscultation.
	The cuff is not wrapped on the patient properly	Wrap the cuff on the patient properly.
CHECK ELECTRODES	Electrode lead is disconnected from the electrode	Firmly connect the electrode lead to the electrode.
	Electrode lead is disconnected from the transmitter	Firmly connect the electrode lead to the transmitter.
	Electrode lead discontinuity	Replace the electrode lead with a new one.
	Electrode is not firmly attached to the skin	Replace the electrode with a new one.
	Polarization voltage is abnormally high	
CHECK PROBE	The probe is not attached to the patient properly	Attach the probe to the patient properly.
	The probe is not attached at the appropriate site	Attach the probe to an appropriate site indicated in the probe manual.
	The probe is disconnected from the transmitter	Connect the probe cable to the transmitter.
	The probe is past its expiration date	Replace the probe with a new one.
CHECK PROBE SITE*	The probe is not attached at the appropriate site	Attach the probe to an appropriate site indicated in the probe manual.
	The probe is deteriorated	Replace the probe with a new one.
	The probe is past its expiration date	
CUFF OCCLUSION	Transmitter malfunction	Immediately remove the cuff from the patient and contact your Nihon Kohden representative.

Message	Cause	Countermeasure
DETECTING PULSE	Searching for the correct pulse wave for SpO ₂ monitoring	Wait until the pulse wave is detected.
	The SpO ₂ value cannot be obtained because the waveform is unstable	Attach the probe to the patient properly.
	The probe is not attached to the patient properly	
HIGH CUFF PRESS	Enormous pressure was applied by the pressure of the cuff	Remove the cause.
INFLATION PRESS LOW	Insufficient cuff inflation pressure	Wait for the remeasurement to be performed with increased cuff inflation pressure.
LIGHT INTERFERENCE	The SpO ₂ measurement site is under fluorescent light, surgical light, sunlight, or other strong light	Cover the measurement site with a blanket or cloth.
	Considerable body movement	When the message is displayed frequently, check the patient condition and, if necessary, change the attachment site.
	The probe is not attached to the patient properly	
	SpO ₂ monitoring is paused for NIBP measurement	Wait for NIBP measurement to finish.
MEAS TIME OUT	The NIBP measuring time exceeded the specified time due to arrhythmia, body movement, vibration or, cuff or air hose being squeezed	Remove the cause if the cause is body movement, vibration or squeezing of cuff or hose.
NIBP MODULE ERROR	Module malfunction	Contact your Nihon Kohden representative.
NO NIBP CHANGE BATTERIES	NIBP cannot be measured due to low battery	Replace batteries with new ones.
PATIENT CONFIRMED*	Function key is pressed and the "PATIENT CONFIRMED" signal is transmitted to the receiving monitor (when "PATIENT" is assigned as the function for the Function key on the SYSTEM SETUP screen)	—
PROBE FAILURE*	The probe is past its expiration date	Replace the probe with a new one.
	Probe is damaged or short-circuited	
REMEASURING	NIBP is being remeasured due to arrhythmia, body movement, vibration or, cuff or air hose being squeezed	If the message still appears after remeasurement, remove the cause if the cause is body movement, vibration or squeezing of cuff or hose.

Message	Cause	Countermeasure
SAFETY CIRCUIT ERROR	The NIBP safety circuit error	Immediately remove the cuff from the patient and contact your Nihon Kohden representative.
SAFETY CIRCUIT RUNNING (When this message is displayed, measurement cannot be performed for 40 seconds.)	NIBP measurement stopped by the safety circuit	Check that the hose is not bent or squeezed.
		Wait 40 seconds, then perform remeasurement. If the message still appears, contact your Nihon Kohden representative.
SpO ₂ MODULE ERROR*	Transmitter failure	Contact your Nihon Kohden representative.
SYS OUT OF RANGE	The systolic blood pressure exceeded the 40 to 280 mmHg range and the blood pressure cannot be measured.	Measure by palpation or auscultation.
WEAK PULSE* (displayed in blue)	Poor peripheral circulation	Check the patient condition and change the probe attachment site.
	The probe is attached too tightly and is obstructing the blood circulation	Check the probe attachment condition and if necessary, reattach the probe.
WEAK PULSE (displayed in pink)	The patient's pulse wave is too small to measure NIBP	Measure NIBP by palpation or auscultation.
	The cuff is wrapped too loosely	Wrap the cuff properly.
	The cuff size is not appropriate	Use the appropriate cuff.
ZEROING	NIBP zero balance is being adjusted	Do not touch the cuff during zeroing. Wait for the message to disappear.

Message Display Priority

When more than one message condition occurs on the transmitter, only the message with the highest priority is displayed.

Priority	Message
High ↑ ↓ Low	PATIENT CONFIRMED
	SAFETY CIRCUIT RUNNING
	CUFF OCCLUSION
	PROBE FAILURE
	CHECK ELECTRODES
	NIBP MODULE ERROR
	SYS OUT OF RANGE
	HIGH CUFF PRESS
	AIR LEAK
	MEAS TIME OUT
	CANNOT DETECT PULSE (NIBP)
	SpO ₂ MODULE ERROR
	CHECK PROBE
	CHECK PROBE SITE
	CANNOT DETECT PULSE (SpO ₂)
	LIGHT INTERFERENCE
	REMEASURING
	INFLATION PRESS LOW
	WEAK PULSE (NIBP)
	ZEROING
	NO NIBP CHANGE BATTERIES
	DETECTING PULSE
	WEAK PULSE (SpO ₂)
	ALARMS SUSPENDED
	BATTERY WEAK
	BATTERY WEAK

Troubleshooting

If a problem occurs, use the following tables to find and fix it. If the problem still remains after troubleshooting according to these tables, contact your Nihon Kohden representative.

Transmitter

Problem	Cause	Countermeasure
Nothing is displayed on the LCD after turning the power on.	Batteries are not installed correctly. The battery polarity is wrong.	Install the batteries correctly.
	Batteries are completely discharged.	Replace the batteries with new ones.
LCD is difficult to see (too dark or too light).	LCD brightness is not appropriate.	Change the LCD brightness on the SYSTEM SETUP screen. Refer to the "Changing SYSTEM SETUP Settings" section.
Nothing is displayed on the receiving monitor after turning the transmitter power on.	The channel of the transmitter and monitor does not match.	Set the correct channel on the monitor.
	The software version of the multiple patient receiver or central monitor is old.	Upgrade the multiple patient receiver or central monitor software to receive signal from the transmitter. The software version must be 01-09 or later.
	Protocol on the transmitter and monitor does not match.	Set the same protocol on the transmitter and monitor.
Signal receiving condition is poor.	Another transmitter with the same channel is used nearby.	Turn the transmitter power off. If the monitor still receives a signal, there is a high probability that another transmitter of the same channel is used nearby. Follow the instruction of your channel administrator and use another transmitter with a different channel.
	Signals of another patient are mixing.	Follow the instructions of your channel administrator and use another transmitter of a different channel.
	The transmitter is damaged.	Contact your Nihon Kohden representative.

ECG/Respiration

Problem	Cause	Countermeasure
The heart rate is unstable.	Pacing detection setting on the monitor is not correct.	Turn off the pacing detection setting on the monitor. When monitoring a pacemaker patient, turn on pacing detection.
The "CHECK ELECTRODES" message appears on the receiving monitor.	Electrode lead is disconnected from the electrode.	Firmly connect the electrode lead to the electrode.
	Electrode lead discontinuity	Replace the electrode lead with a new one.
	Electrode is not firmly attached to the skin.	Replace the electrode with a new one.
	Polarization voltage is abnormally high.	Use Nihon Kohden specified electrodes.
ECG baseline is thick (AC hum).	The gel on the electrode is dried out.	Replace the electrode with a new one.
	The gel on the electrode is coming off.	
	An electric blanket is used.	Cover the blanket with a shield cover.
	The hum filter is set to OFF on the monitor.	Set the filter to ON.
The heart rate of a patient who is using an electric blanket is unstable on the receiving monitor.	Pacing pulse detection is turned ON on the receiving monitor.	Turn OFF the pacing pulse detection on the receiving monitor.
No heart rate or ECG is displayed.	"ECG MEASUREMENT" on the PARAMETER SETUP screen is set to OFF.	If ECG monitoring is necessary, set "ECG MEASUREMENT" to ON.
Respiration waveform measurement is unstable.	The gel on the electrode is dried out.	Replace the electrode with a new one.
	The gel on the electrode is coming off.	
No respiration rate is displayed.	"RESP MEASUREMENT" on the PARAMETER SETUP screen is set to OFF.	If respiration monitoring is necessary, set "RESP MEASUREMENT" to ON.

SpO₂

Problem	Cause	Countermeasure
SpO ₂ data is unstable and not reliable.	The probe size is not appropriate for the patient.	Use the appropriate probe for the patient.
	Probe attachment condition is poor. The probe is partly detached from the skin. External light is entering the probe.	Firmly attach the probe according to the procedure in the probe operator's manual.
	Measurement site is dirty. Patient is wearing nail polish.	Remove dirt and nail polish.
	The probe is attached to the same limb that is used for NIBP measurement.	Attach the probe to the opposite limb. Avoid a site where blood circulation condition changes greatly.

NIBP

Problem	Cause	Countermeasure
Cuff inflation pressure is less than 10 mmHg.	The cuff hose is not connected to the NIBP socket properly.	Connect the cuff hose to the socket properly.
	The cuff is not wrapped around the arm or is wrapped too loosely.	Wrap the cuff around the upper arm.
The cuff does not inflate when the NIBP Start/Stop key is pressed.	The cuff hose is not connected to the NIBP socket.	Connect the cuff hose to the socket firmly.
	The cuff hose or extension hose may be folded or squeezed when the cuff pressure display on the screen increases quickly but the actual cuff does not inflate.	Check the cuff hose and air hose.
Abnormal measurement results are displayed.	The cuff size is not correct.	Select the cuff which fits the patient's limb circumference.
	The cuff is not wrapped around the arm correctly.	Wrap the cuff around the upper arm, not too tightly or too loosely.
	NIBP data is not correct because of body movement.	Prevent the patient from moving during measurement.
	Vibration on the cuff.	Check that nothing is touching the cuff during measurement.
		Change the measuring site.

Problem	Cause	Countermeasure
The cuff is suddenly deflated during inflation.	The NIBP Start/Stop key is pressed during inflation.	—
Auto mode measurement does not start even when the time interval has passed.	The NIBP Interval key is pressed and the measurement mode is changed.	Check the measurement mode and interval.
The cuff suddenly inflates.	The measurement mode is set to auto mode.	Check the time interval. If necessary, stop measurement.
Cannot connect cuff to the air hose.	Unspecified cuff is used.	Use a cuff specified by Nihon Kohden.
Cannot measure NIBP.	Vibration on the cuff.	Check that nothing is touching the cuff during measurement.
	The cuff hose or extension hose is bent or squeezed.	Remove the cause.
	The cuff has worn out.	Use a new cuff.
Blood congestion occurs.	Measuring over a long period of time at short intervals.	Increase the measuring interval.
		Do not measure NIBP over a long time.
Thrombus occurs.	Measuring on a patient with known bleeding disorders or coagulation.	Do not perform NIBP measurement on such a patient.
NIBP data on the screen is --- or dark.	The time set for "OLD NIBP DATA" on the PARAMETER SETUP screen elapsed from the last measurement.	When NIBP is measured again, the data is displayed in normal brightness.
Three loud pip sounds indicating NIBP measurement cannot be started.	The cuff is not deflated enough to start another measurement.	Wait 30 seconds and measure again.

Maintenance

To use the transmitter in safe and optimum condition, perform maintenance check at least once every year.

The transmitter contains parts which gradually deteriorate with use. Original performance might not be delivered if any part of the transmitter is deteriorated. Perform regular maintenance checks to assure continued safe operation.

The following units are necessary for some checking items.

- AX-400G vital sign simulator
- AX-300T SpO₂ checker
- Electric or mercury manometer
- 700 mL dummy cuff
- Receiving monitor

CAUTION

Never disassemble or repair the transmitter. If there is any problem with the transmitter, contact your Nihon Kohden representative.

CAUTION

Before maintenance, cleaning or disinfection, turn the transmitter power off and remove the batteries. Failure to follow this instruction may result in electrical shock and transmitter malfunction.

- NOTE
- The measurement accuracy of the above units must be managed to perform accurate maintenance check.
 - For details on the operation of the above units, refer to the manuals provided with these units.

A maintenance check sheet is provided at the end of this section. Make a copy of this check sheet before performing maintenance check.

1. External Check

- There is no damage or dirt on the outside of the transmitter.
- The battery case cover is not damaged and can be closed firmly.

- No keys are damaged.
- NIBP socket is not damaged.
- No electrode leads are damaged.
- There are no blood or chemicals on the transmitter.
- The springs in the battery compartment are not damaged or detached.
- Terminals in the battery case are not corroded.

2. Transmitter Channel

- The channel number label on the transmitter is not torn or removed.
- The channel of the transmitter matches the label.

The transmitter channel is displayed in the upper left corner of the screen. The channel number also appears on the startup screen.



Startup screen



Numeric and waveform screen

3. Transmitting/Receiving Signal

Use the AX-400G vital sign simulator and receiving monitor.

- 1 Connect the vital sign simulator to the transmitter only with the electrode leads.



- 2 Place the transmitter 2 to 3 m from the receiving monitor.
- 3 Set the channel on the receiving monitor to the channel of the transmitter.

- 4 Turn on the transmitter and vital sign simulator.
- 5 Check that the ECG of the transmitter appears on the receiving monitor.
- 6 Turn off the transmitter.
- 7 Check that the ECG disappears from the receiving monitor.

4. Display

Check that there are no dots missing on the screen.


- 1 Turn off the transmitter.
- 2 While pressing the Function key, turn on the transmitter. The MENU screen appears.
- 3 Press the ► key to move the cursor to “MANUAL CHECK” and press the Function key.



- 4 Press the ◀ or ▶ key to move the cursor to “LCD TEST” and press the Function key.



- 5** Each time the ► key is pressed, the screen changes as below. Check that no dots are missing.



- 6** Press the Screen key to return to the MANUAL CHECK screen.

- 7** Press the Screen key again to return to the MENU screen.

5. Key Operation


- 1** Turn off the transmitter.
- 2** While pressing the Function key, turn on the transmitter. The MENU screen appears.
- 3** Press the ► key to move the cursor to "MANUAL CHECK" and press the Function key.



- 4** Press the ◀ or ▶ key to move the cursor to “KEY CHECK” and press the Function key.



- 5** Press each key one at a time and check that the pressed key is highlighted on the screen.




Example:
When the Function key is pressed, the key name is highlighted.

- 6** After checking, press and hold the Screen key to return to the MANUAL CHECK screen.

6. ECG Check

- 1** Connect the vital sign simulator to the transmitter only with the electrode leads.



- 2** Place the transmitter 1 m from the receiving monitor.

- 3** Turn on the transmitter and vital sign simulator.
- 4** Check that the ECG of the transmitter appears on the receiving monitor.

7. Respiration Check

- 1** Connect the vital sign simulator to the transmitter only with the electrode leads.




- 2** Place the transmitter 1 m from the receiving monitor.
- 3** Turn on the transmitter and vital sign simulator.
- 4** Check that the respiration waveform of the transmitter appears on the receiving monitor.

8. SpO₂ Check

SpO₂ can be checked with the AX-300T SpO₂ checker or AX-400G vital sign simulator.

With the AX-300T SpO₂ Checker

- 1** Connect the SpO₂ checker to the transmitter only with the SpO₂ connection cable.



- 2** Place the transmitter 1 m from the receiving monitor.
- 3** Turn on the transmitter and SpO₂ checker.
- 4** Check that the pulse bar graph appears on the transmitter screen.

- 5** Check that SpO₂ and pulse rate on the transmitter is within the following range.


SpO ₂ on the SpO ₂ Checker		Range
SpO ₂ ¹	97%	95 to 99%SpO ₂ (±2 digits)
	80%	78 to 82%SpO ₂ (±2 digits)
	70%	66 to 74%SpO ₂ (±4 digits)
Pulse rate	60 beats/min	57 to 63 beats/min (±3% ±1 beat/min)
	120 beats/min	115 to 125 beats/min (±3% ±1 beat/min)

¹ The SpO₂ check by the SpO₂ checker is affected by the checker's tolerance to the SpO₂ measuring accuracy of the transmitter. (The measurement accuracy is described in the "Specifications" section.) For details, refer to the SpO₂ checker manual.

- 6** Check that the SpO₂ and pulse waveform of the transmitter appear on the receiving monitor.

With the AX-400G Vital Sign Simulator

- 1** Connect the vital sign simulator to the transmitter only with the SpO₂ connection cable.



- 2** Place the transmitter 1 m from the receiving monitor.
- 3** Turn on the transmitter and vital sign simulator.
- 4** Check that the pulse bar graph appears on the transmitter screen.

- 5** Check that SpO₂ and ECG on the transmitter is within the following range.

Vital Sign Simulator		Range
SpO ₂ ¹	97%	95 to 99%SpO ₂ (±2 digits)
	80%	78 to 82%SpO ₂ (±2 digits)
	70%	66 to 74%SpO ₂ (±4 digits)
ECG	NORMAL 80	76 to 84 beats/min (±4 beats/min)

¹ The SpO₂ check by the vital sign simulator is affected by the simulator's tolerance to the SpO₂ measuring accuracy of the transmitter. (The measurement accuracy is described in the "Specifications" section.) For details, refer to the vital sign simulator manual.


- 6** Check that the SpO₂ and ECG of the transmitter appear on the receiving monitor.

9. NIBP Check

Check that the transmitter displays the correct cuff pressure and that there is no air leak. Also check the pressure sensor. The following procedure uses an electric manometer.

- 1** Connect the electric manometer and dummy cuff to the transmitter.

NOTE: Air leaks from the rubber ball during inflation. Use a clip on the air hose of the rubber ball to stop air leaking.




- 2** Turn on the electric manometer.
- 3** While pressing the Function key, turn on the transmitter. The MENU screen appears.

- 4 Press the ► key to move the cursor to “MANUAL CHECK” and press the Function key.



- 5 Press the ◀ or ▶ key to move the cursor to “NIBP HARDWARE TEST” and press the Function key.



- 6 Press the ◀ or ▶ key to move the cursor to “AIR LEAK TEST” and press the Function key.



- 7 Press the ◀ or ▶ key to move the cursor to “AIRLEAK(AUTO)” and press the Function key.




The transmitter inflates the cuff up to about 250 mmHg and measures air leakage from 60 seconds to 120 seconds after inflation.


- 8 Check the following.
- The value for “AIR LEAK (AUTO)” is below 10 mmHg.
 - The difference between the pressure value displayed on the manometer and transmitter is within ± 6 mmHg.

- 9 Press the Screen key to return to the NIBP HARDWARE TEST screen.

- 10 Press the ◀ or ▶ key to move the cursor to “PRESSURE TEST” and press the Function key.



- 11** Remove the clip from the air hose of the rubber ball.



- 12** By squeezing the rubber ball, increase pressure up to 300 mmHg on the manometer.

- 13** Check the following.

- The difference between the pressure value displayed on the manometer and “PRESS. (SAFETY)” value on the transmitter is within ± 15 mmHg.
- The difference between the pressure value displayed on the manometer and “PRESS” value on the transmitter is within ± 6 mmHg.



10. NIBP Cuff for Attaching Transmitter to Patient Arm

The NIBP cuff is a consumable. Check the following and when necessary, replace it with a new one.

Appearance

- There are no dirty parts.
- There are no broken stitches on the cuff.
- The label on the cuff is readable.
- The velcro tape on the cuff is not removed and there are no broken stitches.
- The lock plate is not damaged and functions properly.

Inflation bag

- The inflation bag is not torn or damaged.
- There is no water inside the inflation bag.
- The connector on the inflation bag is not damaged.

Maintenance Check Sheet

Hospital/Organization: _____

Service Personnel: _____

Instrument Name: Transmitter

Instrument Model: ZM-540PA, ZM-541PA

Instrument Serial Number: _____

Hardware Revision Number: _____

Software Revision Number: _____

1. External Check	Pass	Fail
2. Transmitter Channel	Pass	Fail
3. Transmitting/Receiving Signal	Pass	Fail
4. Display	Pass	Fail
5. Key Operation	Pass	Fail
6. ECG Check	Pass	Fail
7. Respiration Check	Pass	Fail
8. SpO ₂ Check	Pass	Fail
9. NIBP Check	Pass	Fail
10. NIBP Cuff for Attaching Transmitter to Patient Arm	Pass	Fail

Overall Judgement

- ☐ OK
- ☐ Can be used but needs maintenance
- ☐ Maintenance required. Cannot be used.

Lifetime and Disposal

CAUTION

Dispose of Nihon Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise, it may affect the environment. If there is a possibility that the product may have been contaminated with infection, dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.

Disposing of Used Batteries

Battery Lifetime

Replace the batteries when the battery replacement indication appears on the transmitter. When using rechargeable batteries, recharge them.

Disposal

NOTE: Remove the batteries before disposing of the transmitter.

Before disposing of the batteries, check with your local solid waste officials for details in your area for proper disposal. It may be illegal to dispose of these batteries in the municipal waste stream.

Disposing of Electrodes, SpO₂ Probes and NIBP Cuffs

Refer to the manual for each item.

Disposing of Transmitter

Remove the batteries from the transmitter and dispose of the transmitter following your local laws for disposal.

Cleaning, Disinfection and Sterilization

Transmitter and Electrode Leads

CAUTION

If detergent or liquid spills into the transmitter, stop cleaning or disinfecting it and contact your Nihon Kohden representative. The transmitter needs to be checked for safety and function before use.

- NOTE
- The transmitter cannot be sterilized. Sterilizing the transmitter may damage it.
 - When cleaning or disinfecting the transmitter, disconnect it from the patient and remove the batteries beforehand. Otherwise it may result in malfunction.
 - Do not let the NIBP socket get wet during cleaning and disinfection.
 - Do not use a hair dryer etc. to dry the transmitter when it is wet. Doing so may cause the transmitter to deform or become damaged. It may also cause loss of the waterproofing function.
 - Do not use volatile liquids such as thinner or benzine because these will cause the materials to deform or crack.
 - Avoid using flammable disinfectants such as ethanol in a closed place. Ventilate the room if you use flammable disinfectants.
 - Wipe the transmitter with a dry cloth before use and dry it completely after cleaning and disinfecting.

Before cleaning or disinfecting the transmitter, disconnect it from the patient and remove the batteries from the transmitter. Be careful not to let any liquid get inside the transmitter.

Cleaning

Wipe the transmitter and electrode leads with a soft cloth moistened with disinfecting alcohol or neutral detergent diluted with water. After cleaning, dry them completely.

Use cotton swab moistened with neutral detergent diluted with water to clean inside the battery compartment.

Disinfection

CAUTION

Do not immerse the electrode lead connector in liquid.

Wipe the outside surface of the transmitter and electrode lead with a non-abrasive cloth moistened with any of the disinfectants listed below. For details on the disinfectants, refer to the instruction provided with the disinfectants. Use the recommended concentration.

<u>Disinfectant</u>	<u>Concentration (%)</u>
Glutaraldehyde solution	2.0
Alkyldiaminoethylglycine hydrochloride	0.5
Benzalkonium chloride	0.2
Benzethonium chloride solution	0.2
Chlorhexidine gluconate solution	0.5

SpO₂ Probe

Refer to the probe manual.

NIBP Cuff

Refer to the cuff manual.

Periodic Inspection

If the periodic inspection is not performed, degradation or loss of function may go unnoticed and lead to misdiagnosis.

Service personnel should perform the periodic inspection at least once every year. Make sure that the transmitter operates properly and replace the consumables.

If you found abnormalities as a result of inspection and the transmitter is suspected to be faulty, attach an “Unusable” or “Repair request” label to the transmitter and contact your Nihon Kohden representative.

Repair Parts Availability Policy

Nihon Kohden Corporation (NKC) shall stock repair parts (parts necessary to maintain the performance of the instrument) for a period of 8 years from the date of delivery. In that period NKC or its authorized agents will repair the instrument. This period may be shorter than 8 years if a board or part necessary for the faulty section is not available.

Specifications

ZM-540PA

Measured Parameters

Waveforms:	ECG, impedance method respiration, pulse
Numeric data:	Heart rate, respiration rate, SpO ₂ , NIBP, pulse rate

Transmitted Data

Waveforms:	ECG, respiration, pulse wave
Numeric data:	SpO ₂ , NIBP, pulse rate
Status information:	Battery replacement, battery level*, alarm suspended, pause monitoring*, patient confirmed*, ECG lead, pacing detection, electrode detachment, electrode impedance*, ECG off*, respiration method (impedance)*, SpO ₂ status, NIBP status, channel ID, time constant (3.2 s), type of transmitter, transmitter code number*, transmitter serial number*
* Transmitted only when the protocol is "57".	

Display

Display size:	2.2 inch TFT color LCD
Viewing area:	44.16 (H) × 33.12 (V) mm
Resolution:	320 (H) × 240 (V) dots

Displayed Data

Numeric and waveform screen:	ECG (one waveform from lead I, II, III, Va or Vb), heart rate, pulse rate, respiration rate, SpO ₂ , NIBP (systolic, diastolic, MAP), message, battery level, QRS/pulse sync mark, pulse bar graph, NIBP measurement mode and status information, ECG lead
Waveform review screen:	ECG or pulse wave of past 10 minutes
Numeric review screen:	Heart rate or pulse rate, respiration rate and SpO ₂ at 1 minute interval for past 10 minutes
Check electrodes screen:	ECG for checking electrode attachment

ECG

ECG Measurement

Channels:	4
Input dynamic range:	±10 mV or more
Electrode offset potential tolerance:	±500 mV or more
Input impedance:	5 MΩ or more
Common mode rejection ratio:	95 dB or more
	IEC 60601-2-27 50.102.10 compliant

Pacing pulse detection:	amplitude ± 2 to 700 mV, duration 0.1 to 2 ms IEC 60601-2-27: 2005 compliant
Defibrillation-proof:	Based upon pacemaker pulse rejection capability ECG input protected against 400 Ws/DC 5 kV IEC 60601-2-27 17.101 compliant
ECG recovery time after defibrillation:	within 10 s
Electrode condition:	Displays CHECK ELECTRODES message
Tall T-wave rejection capability:	Complies with the heights of T-waves from 0 to 1.6 mV IEC 60601-2-27: 2005 50.102.17 compliant
Pacemaker pulse rejection capability, without overshoot:	Complies with the amplitudes of pacemaker pulses ± 2 to ± 700 mV and widths 0.1 to 2 ms (As defined by method B of IEC 60601-2-27: 2005 50.102.13, this corresponds to the pacemaker pulses amplitudes and widths of ± 2.8 mV/2 ms to amplitudes ± 45.6 mV/0.1 ms.)
Pacemaker pulse rejection capability, with overshoot:	Overshoot amplitudes and time constants of ± 0.12 mV/100 ms to ± 2 mV/4 ms (As defined by method B of IEC 60601-2-27: 2005 50.102.13, this corresponds to the pacemaker pulses amplitudes and widths of ± 2.8 mV/2 ms to amplitudes ± 45.6 mV/0.1 ms.)

ECG Display and Heart Rate Count

Frequency characteristic:	filter on: 1 to 18 Hz, filter off: 0.05 to 60 Hz
Heart rate detection method:	Average
QRS detection:	70 to 120 ms: amplitude ≥ 0.5 mV, rate 30 to 200 beats/min 40 to 120 ms: amplitude ≥ 0.5 mV, rate 30 to 250 beats/min
Heart rate counting range:	0, 15 to 300 beats/min
Heart rate counting accuracy*:	± 2 beats/min, (0, 15 to 300 beats/min)
* Essential performance of this transmitter	

Respiration Measurement

Measuring method:	Impedance method
Measuring lead:	Between R and F
Impedance range:	2 k Ω or less
Respiration rate measuring accuracy*:	± 2 counts/min (at 0 to 150 counts/min)
Respiration rate counting range:	0 to 150 counts/min
* Essential performance of this transmitter	

SpO₂ Measurement (ISO 9919: 2005 compliant)

Measuring range:	0 to 100%SpO ₂
Declared range:	70 to 100%SpO ₂
Minimum display range:	1%SpO ₂
Display update cycle:	Every 3 seconds
Measuring accuracy (rms)*:	Accuracy assurance temperature: 18 to 40°C

Total accuracy including probe: $80\% \text{SpO}_2 \leq \% \text{SpO}_2 \leq 100\% \text{SpO}_2: \pm 2\% \text{SpO}_2$
 $70\% \text{SpO}_2 \leq \% \text{SpO}_2 < 80\% \text{SpO}_2: \pm 3\% \text{SpO}_2$
under $70\% \text{SpO}_2$: not specified

Accuracy of the transmitter: $80\% \text{SpO}_2 \leq \% \text{SpO}_2 \leq 100\% \text{SpO}_2: \pm 1\% \text{SpO}_2$
 $50\% \text{SpO}_2 \leq \% \text{SpO}_2 < 80\% \text{SpO}_2: \pm 2\% \text{SpO}_2$
under $50\% \text{SpO}_2$: not specified

Pulse rate measuring range: 30 to 300 bpm

Pulse rate display range: 30 to 300 bpm

Pulse rate accuracy (rms)*: $\pm 3\% \pm 1 \text{ bpm}$


* Essential performance of this transmitter


- NOTE**
- The SpO_2 measuring accuracy was tested using the TL-201T, TL-260T, TL-271T and TL-631T SpO_2 probes. The testing was performed during induced hypoxia on healthy volunteers (Ethnicity: 10 Caucasians, 2 Africans, 1 Asian and 3 Indians), (Skin: 8 light, 4 medium, 4 dark), (Age: 21 to 34), (5 women and 11 men) under the condition of no motion. Arterial blood was sampled and measured by a CO-oximeter. The difference between SpO_2 measured by the SpO_2 probe and functional SaO_2 measured by a CO-oximeter was calculated using the root-mean-square (rms) method. This measurement accuracy figure represents 2/3 of all test measurements.
 - A pulse oximeter tester that generates simulated signals can be used to check the difference from the design specification, but it cannot be used as a replacement for human signals for testings accuracy.

Response time:


Selectable from "Slow", "Normal" and "Fast".

The following graphs show response time for $0.6\%/s$ SpO_2 change.





The following graph shows response time for 10 bpm/s pulse rate change.



Noninvasive Blood Pressure, NIBP (IEC 60601-2-30: 1999 compliant)

Measuring method:	Oscillometric
Measurement mode:	Manual, STAT (≤ 15 min), Periodic
Intended patient type:	Adult, child
Measuring range:	0 to 300 mmHg
Pressure display range:	0 to 300 mmHg
Measuring accuracy*:	± 3 mmHg ($0 \text{ mmHg} \leq \text{NIBP} \leq 300 \text{ mmHg}$) AAMI SP-10: 2002 compliant
Cuff inflation time:	≤ 20 s (700 cc), 0 to 200 mmHg ≤ 15 s (70 cc), 0 to 200 mmHg
Pressure retention:	≤ 5 mmHg (250 cc at 250 mmHg inflation for 10 seconds)
Air leakage:	≤ 3 mmHg/min (700 cc at 300 mmHg inflation)
Maximum temperature at cuff attachment site:	43°C (109°F)
Power discontinuity:	Deflate immediately after power down

Safety:	Maximum pressurization value cuff inflation limiter:	300 to 330 mmHg
	Cuff inflation time limiter:	≤ 180 s
	Interval time limiter:	≤ 30 s

* Essential performance of this transmitter

Transmitter

FCC regulation:	FCC part 95 Subpart H Wireless Medical Telemetry Service (WMTS)
IC regulation:	INDUSTRY CANADA RSS-GEN Issue 3 INDUSTRY CANADA RSS-210 Issue 8 Annex 4 Medical Telemetry, Wireless Microphones, Auditory Assistance, Goods Tracking and Law Enforcement
Field strength limits:	< 200 mV/m (at 3 m)
Undesired emissions:	
	FCC part 95 95.1115 (b) (1), (2):
	≤ 960 MHz: 200 μV/m (at 3 m)
	≥ 960 MHz: 500 μV/m (at 3 m)
	IC RSS-GEN Issue 3, RSS-210 Issue 8 Annex 4:
	30 to 88 MHz: 100 μV/m (at 3 m)
	88 to 216 MHz: 150 μV/m (at 3 m)
	216 to 960 MHz: 200 μV/m (at 3 m)
	≥ 960 MHz: 500 μV/m (at 3 m)
Antenna:	Internal
Transmission channel:	Indicated on the transmitter
Transmission frequency range:	608.0250 to 613.9750 MHz
Channel spacing:	50 kHz or 37.5 kHz (12.5 kHz when interleaved)
Modulation:	FSK (frequency shift keying)
Type of emission:	F1D
Occupied bandwidth:	< 20 kHz
Effective radiated power:	1.0 mW

Power Requirements

Rated voltage:	3.6 V
Operating voltage:	3.2 to 4.8 V
Battery type:	Three AA (R6) type NiMH secondary batteries Three AA (R6) type alkaline dry cell primary batteries
Battery lifetime (with alkaline batteries, at room temperature):	approximately 1 day (measuring ECG, respiration, SpO ₂ of approximately 60 kg weight adult male patient at the index finger, NIBP at 60 minute intervals)

Dimensions and Weight

Dimensions:	114 W × 125 H × 63 D (mm)
Weight:	about 340 g (excluding batteries and other accessories) about 410 g (including batteries, excluding other accessories)

Environment

Operating environment

Temperature:	5 to 40°C (41 to 104°F)
Humidity:	30 to 85% (noncondensing)
Atmospheric pressure:	700 to 1060 hPa

Storage and transport environment

Temperature:	−20 to +65°C (−4 to +149°F)
Humidity:	10 to 95%
Atmospheric pressure:	700 to 1060 hPa

Safety Standards

Safety standard:	CAN/CSA-C22.2 No. 601-1 M90
	CAN/CSA-C22.2 No. 601-1. 1S1-94
	CAN/CSA-C22.2 No. 601-1. 1B-90
	CAN/CSA-C22.2 No. 60601-2-49-04
	CAN/CSA-C22.2 No. 60601-2-27-06
	CAN/CSA-C22.2 No. 60601-2-30-02
	IEC 60601-1:1988
	IEC 60601-1 Amendment 1: 1991
	IEC 60601-1 Amendment 2: 1995
	IEC 60601-2-27: 2005
	IEC 60601-2-30: 1999
	IEC 60601-2-49: 2001
	ISO 9919: 2005
	UL 60601-1: 2003

Type of protection against electrical shock:

INTERNALLY POWERED EQUIPMENT

Degree of protection against electrical shock:

ECG and impedance method respiration:

DEFIBRILLATION-PROOF TYPE CF APPLIED
PART

SpO₂ and NIBP:

DEFIBRILLATION-PROOF TYPE BF APPLIED
PART

Degree of protection against harmful ingress of water:

IPX0 (non-protected)

Degree of safety of application in the presence of a FLAMMABLE ANAESTHETIC
MIXTURE WITH AIR, OR WITH OXYGEN OR NITROUS OXIDE:

Equipment not suitable for use in the presence of
FLAMMABLE ANAESTHETIC MIXTURE WITH
AIR, OR WITH OXYGEN OR NITROUS OXIDE
CONTINUOUS OPERATION

Mode of operation:

Electromagnetic Compatibility

IEC 60601-1-2: 2001

IEC 60601-1-2 Amendment 1: 2004

Electromagnetic Emissions

This Model ZM-540PA is intended for use in the electromagnetic environment specified below. The customer or the user of the ZM-540PA should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The ZM-540PA uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The ZM-540PA is suitable for use in all establishments, including domestic establishments.
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Not applicable	

Electromagnetic Immunity

This Model ZM-540PA is intended for use in the electromagnetic environment specified below. The customer or the user of the ZM-540PA should assure that it is used in such an environment.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/ burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	Not applicable	—
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	Not applicable	—
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5% U_T (>95% dip in U_T) for 0.5 cycles 40% U_T (60% dip in U_T) for 5 cycles 70% U_T (30% dip in U_T) for 25 cycles <5% U_T (>95% dip in U_T) for 5 s	Not applicable	—
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

NOTE: U_T is the AC mains voltage prior to application of the test level.

Avoiding Electromagnetic Interference (Impedance Respiration)

Impedance respiration measurement is very sensitive and affected by electromagnetic interference. Technological limitations do not allow immunity levels higher than 1 V/m for radiated RF electromagnetic fields. Electromagnetic fields with field strengths above 1 V/m may cause measurement error. Do not use electrically radiating equipment near the impedance respiration measurements.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any part of the ZM-540PA, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m 80 MHz to 2.5 GHz (1 V/m 80 MHz to 2.5 GHz for respiration)	Recommended separation distance $d = 1.2\sqrt{P}$ $d = 1.2\sqrt{P}$ 80 MHz to 800 MHz $d = 2.3\sqrt{P}$ 800 MHz to 2.5 GHz ($d = 3.5\sqrt{P}$ 80 MHz to 800 MHz for respiration) $d = 7.0\sqrt{P}$ 800 MHz to 2.5 GHz for respiration) where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey¹, should be less than the compliance level in each frequency range². Interference may occur in the vicinity of equipment marked with the following symbol: 

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

- ¹ Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the ZM-540PA is used exceeds the applicable RF compliance level above, the ZM-540PA should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the ZM-540PA.
- ² Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 1 V/m for respiration and 3 V/m for all other functions.

Recommended Separation Distances between Portable and Mobile RF Communications Equipment

The ZM-540PA is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the ZM-540PA can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the ZM-540PA as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150 kHz to 80 MHz $d = 1.2\sqrt{P}$	80 MHz to 800 MHz $d = 1.2\sqrt{P}$ (For respiration: $d = 3.5\sqrt{P}$)	800 MHz to 2.5 GHz $d = 2.3\sqrt{P}$ (For respiration: $d = 7.0\sqrt{P}$)
0.01	0.12	0.12 (0.35 ¹)	0.23 (0.7 ¹)
0.1	0.38	0.38 (1.1 ¹)	0.73 (2.2 ¹)
1	1.2	1.2 (3.5 ¹)	2.3 (7.0 ¹)
10	3.8	3.8 (11 ¹)	7.3 (22 ¹)
100	12	12 (35 ¹)	23 (70 ¹)

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

(¹ For respiration)

NOTE 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Recovery Time after Defibrillation

The transmitter returns to the normal operating mode within 10 seconds after defibrillation. The stored settings are not affected.

System Composition for EMC Test

The ZM-540PA transmitter is tested to comply with IEC 60601-1-2:2001 and Amendment 1:2004 with the following composition. If any part which is not specified by Nihon Kohden is used, the EMC specifications might not comply.

Units	Cable Length
ZM-540PA transmitter	—
YP-503P NIBP cuff	0.15 m
BR-906P ECG electrode lead	0.8 m
TL-201T finger probe	1.6 m

ZM-541PA

Measured Parameters

Waveforms:	ECG, impedance method respiration, pulse
Numeric data:	Heart rate, respiration rate, SpO ₂ , NIBP, pulse rate

Transmitted Data

Waveforms:	ECG, respiration, pulse wave
Numeric data:	SpO ₂ , NIBP, pulse rate
Status information:	Battery replacement, battery level*, alarm suspended, pause monitoring*, patient confirmed*, ECG lead, pacing detection, electrode detachment, electrode impedance*, ECG off*, respiration method (impedance)*, SpO ₂ status, NIBP status, channel ID, time constant (3.2 s), type of transmitter, transmitter code number*, transmitter serial number*
* Transmitted only when the protocol is "57".	

Display

Display size:	2.2 inch TFT color LCD
Viewing area:	44.16 (H) × 33.12 (V) mm
Resolution:	320 (H) × 240 (V) dots

Displayed Data

Numeric and waveform screen:	ECG (one waveform from lead I, II, III, Va or Vb), heart rate, pulse rate, respiration rate, SpO ₂ , NIBP (systolic, diastolic, MAP), message, battery level, QRS/pulse sync mark, pulse bar graph, NIBP measurement mode and status information, ECG lead
Waveform review screen:	ECG or pulse wave of past 10 minutes
Numeric review screen:	Heart rate or pulse rate, respiration rate and SpO ₂ at 1 minute interval for past 10 minutes
CHECK ELECTRODE screen:	ECG for checking electrode attachment

ECG

ECG Measurement

Channels:	4
Input dynamic range:	±10 mV or more
Electrode offset potential tolerance:	±500 mV or more
Input impedance:	5 MΩ or more
Common mode rejection ratio:	95 dB or more
	IEC 60601-2-27 50.102.10 compliant
Pacing pulse detection:	amplitude ±2 to 700 mV, duration 0.1 to 2 ms
	IEC 60601-2-27: 2005 compliant
	Based upon pacemaker pulse rejection capability
Defibrillation-proof:	ECG input protected against 400 Ws/DC 5 kV
	IEC 60601-2-27 17.101 compliant

ECG recovery time after defibrillation: within 10 s
 Electrode condition: Displays CHECK ELECTRODES message
 Tall T-wave rejection capability: Complies with the heights of T-waves from 0 to 1.6 mV
 IEC 60601-2-27: 2005 50.102.17 compliant

Pacemaker pulse rejection capability, without overshoot:

Complies with the amplitudes of pacemaker pulses ± 2 to ± 700 mV and widths 0.1 to 2 ms (As defined by method B of IEC 60601-2-27: 2005 50.102.13, this corresponds to the pacemaker pulses amplitudes and widths of ± 2.8 mV/2 ms to amplitudes ± 45.6 mV/0.1 ms.)

Pacemaker pulse rejection capability, with overshoot:

Overshoot amplitudes and time constants of ± 0.12 mV/100 ms to ± 2 mV/4 ms (As defined by method B of IEC 60601-2-27: 2005 50.102.13, this corresponds to the pacemaker pulses amplitudes and widths of ± 2.8 mV/2 ms to amplitudes ± 45.6 mV/0.1 ms.)

ECG Display and Heart Rate Count

Frequency characteristic: filter on: 1 to 18 Hz, filter off: 0.05 to 60 Hz
 Heart rate detection method: Average
 QRS detection: 70 to 120 ms: amplitude ≥ 0.5 mV, rate 30 to 200 beats/min
 40 to 120 ms: amplitude ≥ 0.5 mV, rate 30 to 250 beats/min
 Heart rate counting range: 0, 15 to 300 beats/min
 Heart rate counting accuracy*: ± 2 beats/min, (0, 15 to 300 beats/min)

* Essential performance of this transmitter

Respiration Measurement

Measuring method: Impedance method
 Measuring lead: Between R and F
 Impedance range: 2 k Ω or less
 Respiration rate measuring accuracy*: ± 2 counts/min (at 0 to 150 counts/min)
 Respiration rate counting range: 0 to 150 counts/min

* Essential performance of this transmitter

SpO₂ Measurement (ISO 9919: 2005 compliant)

Measuring range: 0 to 100%SpO₂
 Declared range: 70 to 100%SpO₂
 Minimum display range: 1%SpO₂
 Display update cycle: Every 3 seconds
 Measuring accuracy (rms)*: Accuracy assurance temperature: 18 to 40°C
 Total accuracy including probe: 80%SpO₂ $\leq$ %SpO₂ $\leq$ 100%SpO₂: $\pm 2\%$ SpO₂
 70%SpO₂ $\leq$ %SpO₂ $<$ 80%SpO₂: $\pm 3\%$ SpO₂
 under 70%SpO₂: not specified
 Accuracy of the transmitter: 80%SpO₂ $\leq$ %SpO₂ $\leq$ 100%SpO₂: $\pm 1\%$ SpO₂
 50%SpO₂ $\leq$ %SpO₂ $<$ 80%SpO₂: $\pm 2\%$ SpO₂
 under 50%SpO₂: not specified

Pulse rate measuring range: 30 to 300 bpm
 Pulse rate display range: 30 to 300 bpm
 Pulse rate accuracy (rms)*: $\pm 3\% \pm 1$ bpm


* Essential performance of this transmitter


- NOTE**
- The SpO₂ measuring accuracy was tested using the TL-201T, TL-260T, TL-271T and TL-631T SpO₂ probes. The testing was performed during induced hypoxia on healthy volunteers (Ethnicity: 10 Caucasians, 2 Africans, 1 Asian and 3 Indians), (Skin: 8 light, 4 medium, 4 dark), (Age: 21 to 34), (5 women and 11 men) under the condition of no motion. Arterial blood was sampled and measured by a CO-oximeter. The difference between SpO₂ measured by the SpO₂ probe and functional SaO₂ measured by a CO-oximeter was calculated using the root-mean-square (rms) method according to ISO 9919: 2005. This measurement accuracy figure represents 2/3 of all test measurements.
 - A pulse oximeter tester that generates simulated signals can be used to check the difference from the design specification, but it cannot be used as a replacement for human signals for testings accuracy.

Response time:


Selectable from “Slow”, “Normal” and “Fast”.

The following graphs show response time for 0.6%/s SpO₂ change.





The following graph shows the response time example when pulse rate changes 10 bpm/s.



Noninvasive Blood Pressure, NIBP (IEC 60601-2-30: 1999 compliant)

Measuring method:	Oscillometric
Measurement mode:	Manual, STAT (≤ 15 min), Periodic
Intended patient type:	Adult, child
Measuring range:	0 to 300 mmHg
Pressure display range:	0 to 300 mmHg
Measuring accuracy*:	± 3 mmHg ($0 \text{ mmHg} \leq \text{NIBP} \leq 300 \text{ mmHg}$) AAMI SP-10: 2002 compliant
Cuff inflation time:	≤ 20 s (700 cc), 0 to 200 mmHg ≤ 15 s (70 cc), 0 to 200 mmHg
Pressure retention:	≤ 5 mmHg (250 cc at 250 mmHg inflation for 10 seconds)
Air leakage:	≤ 3 mmHg/min (700 cc at 300 mmHg inflation)
Power discontinuity:	Deflate immediately after power down
Safety:	Maximum pressurization value cuff inflation limiter: 300 to 330 mmHg
	Cuff inflation time limiter: ≤ 180 s
	Interval time limiter: ≤ 30 s

* Essential performance of this transmitter

Transmitter

FCC regulation:	FCC part 95 Subpart H Wireless Medical Telemetry Service (WMTS)
IC regulation:	INDUSTRY CANADA RSS-GEN Issue 3 INDUSTRY CANADA RSS-210 Issue 8 Annex 4 Medical Telemetry, Wireless Microphones, Auditory Assistance, Goods Tracking and Law Enforcement
Field strength limits:	< 740 mV/m (at 3 m)
Undesired emissions:	below 960 MHz: < 200 μ V/m (at 3 m) above 960 MHz: < 500 μ V/m (at 3 m)
Antenna:	Internal
Transmission channel:	Indicated on the transmitter
Transmission frequency range*:	1395.0250 to 1399.9750 MHz 1427.0250 to 1431.9750 MHz *For Canada: 1395.0250 to 1399.9750 MHz 1427.0250 to 1429.4750 MHz
Channel spacing:	50 kHz or 37.5 kHz (12.5 kHz when interleaved)
Modulation:	FSK (frequency shift keying)
Type of emission:	F1D
Occupied bandwidth:	< 20 kHz
Effective radiated power:	5.0 mW Can be changed to 1.0 mW if required

Power Requirements

Rated voltage:	3.6 V
Operating voltage:	3.2 to 4.8 V
Battery type:	Three AA (R6) type NiMH secondary batteries Three AA (R6) type alkaline dry cell primary batteries
Battery lifetime (with alkaline batteries, at room temperature):	approximately 1 day (measuring ECG, respiration, SpO ₂ of approximately 60 kg weight adult male patient at the index finger, NIBP at 60 minute intervals)

Dimensions and Weight

Dimensions:	114 W × 125 H × 63 D (mm)
Weight:	about 340 g (excluding batteries and other accessories) about 410 g (including batteries, excluding other accessories)

Environment

Operating environment

Temperature:	5 to 40°C (41 to 104°F)
Humidity:	30 to 85% (noncondensing)
Atmospheric pressure:	700 to 1060 hPa

Storage and transport environment

Temperature:	−20 to +65°C (−4 to +149°F)
Humidity:	10 to 95%
Atmospheric pressure:	700 to 1060 hPa

Safety Standards

Safety standard:	CAN/CSA-C22.2 No. 601-1 M90
	CAN/CSA-C22.2 No. 601-1. 1S1-94
	CAN/CSA-C22.2 No. 601-1. 1B-90
	CAN/CSA-C22.2 No. 60601-2-49-04
	CAN/CSA-C22.2 No. 60601-2-27-06
	CAN/CSA-C22.2 No. 60601-2-30-02
	IEC 60601-1:1988
	IEC 60601-1 Amendment 1: 1991
	IEC 60601-1 Amendment 2: 1995
	IEC 60601-2-27: 2005
	IEC 60601-2-30: 1999
	IEC 60601-2-49: 2001
	ISO 9919: 2005
	UL 60601-1: 2003
Type of protection against electrical shock:	INTERNALLY POWERED EQUIPMENT
Degree of protection against electrical shock:	
ECG and impedance method respiration:	DEFIBRILLATION-PROOF TYPE CF APPLIED PART
SpO ₂ and NIBP:	DEFIBRILLATION-PROOF TYPE BF APPLIED PART
Degree of protection against harmful ingress of water:	IPX0 (non-protected)
Degree of safety of application in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, OR WITH OXYGEN OR NITROUS OXIDE:	Equipment not suitable for use in the presence of FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, OR WITH OXYGEN OR NITROUS OXIDE
Mode of operation:	CONTINUOUS OPERATION

Electromagnetic Compatibility

IEC 60601-1-2: 2001
IEC 60601-1-2 Amendment 1: 2004

Electromagnetic Emissions

This Model ZM-541PA is intended for use in the electromagnetic environment specified below. The customer or the user of the ZM-541PA should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The ZM-541PA uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The ZM-541PA is suitable for use in all establishments, including domestic establishments.
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Not applicable	

Electromagnetic Immunity

This Model ZM-541PA is intended for use in the electromagnetic environment specified below. The customer or the user of the ZM-541PA should assure that it is used in such an environment.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/ burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	Not applicable	—
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	Not applicable	—
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5% U_T (>95% dip in U_T) for 0.5 cycles 40% U_T (60% dip in U_T) for 5 cycles 70% U_T (30% dip in U_T) for 25 cycles <5% U_T (>95% dip in U_T) for 5 s	Not applicable	—
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

NOTE: U_T is the AC mains voltage prior to application of the test level.

Avoiding Electromagnetic Interference (Impedance Respiration)

Impedance respiration measurement is very sensitive and affected by electromagnetic interference. Technological limitations do not allow immunity levels higher than 1 V/m for radiated RF electromagnetic fields. Electromagnetic fields with field strengths above 1 V/m may cause measurement error. Do not use electrically radiating equipment near the impedance respiration measurements.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any part of the ZM-541PA, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance $d = 1.2\sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m 80 MHz to 2.5 GHz (1 V/m 80 MHz to 2.5 GHz for respiration)	$d = 1.2\sqrt{P}$ 80 MHz to 800 MHz $d = 2.3\sqrt{P}$ 800 MHz to 2.5 GHz ($d = 3.5\sqrt{P}$ 80 MHz to 800 MHz for respiration $d = 7.0\sqrt{P}$ 800 MHz to 2.5 GHz for respiration) where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey ¹ , should be less than the compliance level in each frequency range ² . Interference may occur in the vicinity of equipment marked with the following symbol: 

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

- ¹ Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the ZM-541PA is used exceeds the applicable RF compliance level above, the ZM-541PA should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the ZM-541PA.
- ² Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 1 V/m for respiration and 3 V/m for all other functions.

Recommended Separation Distances between Portable and Mobile RF Communications Equipment

The ZM-541PA is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the ZM-541PA can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the ZM-541PA as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150 kHz to 80 MHz $d = 1.2\sqrt{P}$	80 MHz to 800 MHz $d = 1.2\sqrt{P}$ (For respiration: $d = 3.5\sqrt{P}$)	800 MHz to 2.5 GHz $d = 2.3\sqrt{P}$ (For respiration: $d = 7.0\sqrt{P}$)
0.01	0.12	0.12 (0.35 ¹)	0.23 (0.7 ¹)
0.1	0.38	0.38 (1.1 ¹)	0.73 (2.2 ¹)
1	1.2	1.2 (3.5 ¹)	2.3 (7.0 ¹)
10	3.8	3.8 (11 ¹)	7.3 (22 ¹)
100	12	12 (35 ¹)	23 (70 ¹)

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

(¹ For respiration)

NOTE 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Recovery Time after Defibrillation

The transmitter returns to the normal operating mode within 10 seconds after defibrillation. The stored settings are not affected.


System Composition for EMC Test

The ZM-541PA transmitter is tested to comply with IEC 60601-1-2:2001 and Amendment 1:2004 with the following composition. If any part which is not specified by Nihon Kohden is used, the EMC specifications might not comply.


Units	Cable Length
ZM-541PA transmitter	—
YP-503P NIBP cuff	0.15 m
BR-906P ECG electrode lead	0.8 m
TL-201T finger probe	1.6 m

Standard Accessories

1




2




No.	Name	Model	Qty	Supply Code
1	NIBP cuff for adult, standard	YP-503P	1	S937C
2	Strap	—	1	Y236

The following parts are available for replacement.

3



4



Lock plate is a standard accessory of the YP-503P and YP-504P NIBP cuffs.

No.	Name	Model	Qty	Code No.
3	Battery case cover	—	1	6143-903101
4	Lock plate			6113-049585

Options

CAUTION

Only use Nihon Kohden specified electrodes, electrode leads and SpO₂ probes. Otherwise, the maximum performance from the transmitter cannot be guaranteed.

Transmitter

- QI-901PK channel writer

ECG/RESP

Application	Model	Qty	Supply Code
3 electrodes, clip type, lead length 80 cm	BR-903PA	1	K911A
3 electrodes, snap type, lead length 80 cm	BR-913PA	1	K910B
6 electrodes, clip type, lead length 80 cm	BR-906PA	1	K912A
6 electrodes, snap type, lead length 80 cm	BR-916PA	1	K915A

SpO₂

Name	Lead Length	Model	Qty	Supply Code
Finger probe (reusable)	0.6 m	TL-201T	1	P225H
	1.6 m			P225F
Multi-site probe (reusable)	1.6 m	TL-220T		P225G
Finger probe (reusable)	0.6 m	TL-631T1		P311A
	1.6 m	TL-631T3		P311C
SpO ₂ probe (for adult, disposable)	0.8 m	TL-271T	24	P203A
	1.6 m	TL-271T3		P203E
SpO ₂ probe (for child, disposable)	0.8 m	TL-272T		P203B
	1.6 m	TL-272T3		P203F
SpO ₂ probe (for neonate/adult, disposable)	0.8 m	TL-273T		P203C
	1.6 m	TL-273T3		P203G
SpO ₂ probe (for infant, disposable)	0.8 m	TL-274T		P203D
	1.6 m	TL-274T3		P203H
SpO ₂ probe (for neonate/adult, disposable)	0.8 m	TL-051S	5	P228A
	1.6 m	TL-052S		P228B
SpO ₂ probe (for child/infant, disposable)	0.8 m	TL-061S		P229A
	1.6 m	TL-062S		P229B
COTTONY tape	—	—	20	P259
Foam tape for TL-051S/052S/061S/062S			4 × 25	P260C
Attachment tape for TL-220T/631T			3 × 30	P263
Probe fastener		YS-093P2	30	P267

NIBP

Name		Width (cm)	Air Hose Length (cm)	Model	Qty	Supply Code
Cuff for adult, for attaching transmitter to patient arm	Standard	13	15	YP-503P ¹	1	S937C
	Large	15	15	YP-504P ¹	1	S937D
Cuff for infant		5	15	YP-960T	1	S943A
Cuff for child	Small	7	15	YP-961T	1	S943B
	Standard	10	15	YP-702T	1	S949A
			15	YP-962T	1	S943C
Cuff for adult	Standard	13	15	YP-703T	1	S949B
			15	YP-963T	1	S944B
	Large	15	15	YP-964T	1	S944C
Disposable cuff for infant		6	17	YP-810P	20	S945C
Disposable cuff for child		8	17	YP-811P	20	S945D
Disposable cuff for adult	Small	10	17	YP-812P	20	S946E
	Standard	14	17	YP-813P	20	S946F
	Medium large	15	17	YP-814P	20	S946G
	Large	17	17	YP-815P	20	S946H
	Extra large	18	17	YP-816P	20	S946I
Extension hose		—	150	YN-990P	1	S903

¹ The lock plate is provided with these NIBP cuffs.

Transmission Frequencies

Channel: 9002 to 9478

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
9002	608.0250	9037	608.4625	9072	608.9000
9003	608.0375	9038	608.4750	9073	608.9125
9004	608.0500	9039	608.4875	9074	608.9250
9005	608.0625	9040	608.5000	9075	608.9375
9006	608.0750	9041	608.5125	9076	608.9500
9007	608.0875	9042	608.5250	9077	608.9625
9008	608.1000	9043	608.5375	9078	608.9750
9009	608.1125	9044	608.5500	9079	608.9875
9010	608.1250	9045	608.5625	9080	609.0000
9011	608.1375	9046	608.5750	9081	609.0125
9012	608.1500	9047	608.5875	9082	609.0250
9013	608.1625	9048	608.6000	9083	609.0375
9014	608.1750	9049	608.6125	9084	609.0500
9015	608.1875	9050	608.6250	9085	609.0625
9016	608.2000	9051	608.6375	9086	609.0750
9017	608.2125	9052	608.6500	9087	609.0875
9018	608.2250	9053	608.6625	9088	609.1000
9019	608.2375	9054	608.6750	9089	609.1125
9020	608.2500	9055	608.6875	9090	609.1250
9021	608.2625	9056	608.7000	9091	609.1375
9022	608.2750	9057	608.7125	9092	609.1500
9023	608.2875	9058	608.7250	9093	609.1625
9024	608.3000	9059	608.7375	9094	609.1750
9025	608.3125	9060	608.7500	9095	609.1875
9026	608.3250	9061	608.7625	9096	609.2000
9027	608.3375	9062	608.7750	9097	609.2125
9028	608.3500	9063	608.7875	9098	609.2250
9029	608.3625	9064	608.8000	9099	609.2375
9030	608.3750	9065	608.8125	9100	609.2500
9031	608.3875	9066	608.8250	9101	609.2625
9032	608.4000	9067	608.8375	9102	609.2750
9033	608.4125	9068	608.8500	9103	609.2875
9034	608.4250	9069	608.8625	9104	609.3000
9035	608.4375	9070	608.8750	9105	609.3125
9036	608.4500	9071	608.8875	9106	609.3250

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
9107	609.3375	9146	609.8250	9185	610.3125
9108	609.3500	9147	609.8375	9186	610.3250
9109	609.3625	9148	609.8500	9187	610.3375
9110	609.3750	9149	609.8625	9188	610.3500
9111	609.3875	9150	609.8750	9189	610.3625
9112	609.4000	9151	609.8875	9190	610.3750
9113	609.4125	9152	609.9000	9191	610.3875
9114	609.4250	9153	609.9125	9192	610.4000
9115	609.4375	9154	609.9250	9193	610.4125
9116	609.4500	9155	609.9375	9194	610.4250
9117	609.4625	9156	609.9500	9195	610.4375
9118	609.4750	9157	609.9625	9196	610.4500
9119	609.4875	9158	609.9750	9197	610.4625
9120	609.5000	9159	609.9875	9198	610.4750
9121	609.5125	9160	610.0000	9199	610.4875
9122	609.5250	9161	610.0125	9200	610.5000
9123	609.5375	9162	610.0250	9201	610.5125
9124	609.5500	9163	610.0375	9202	610.5250
9125	609.5625	9164	610.0500	9203	610.5375
9126	609.5750	9165	610.0625	9204	610.5500
9127	609.5875	9166	610.0750	9205	610.5625
9128	609.6000	9167	610.0875	9206	610.5750
9129	609.6125	9168	610.1000	9207	610.5875
9130	609.6250	9169	610.1125	9208	610.6000
9131	609.6375	9170	610.1250	9209	610.6125
9132	609.6500	9171	610.1375	9210	610.6250
9133	609.6625	9172	610.1500	9211	610.6375
9134	609.6750	9173	610.1625	9212	610.6500
9135	609.6875	9174	610.1750	9213	610.6625
9136	609.7000	9175	610.1875	9214	610.6750
9137	609.7125	9176	610.2000	9215	610.6875
9138	609.7250	9177	610.2125	9216	610.7000
9139	609.7375	9178	610.2250	9217	610.7125
9140	609.7500	9179	610.2375	9218	610.7250
9141	609.7625	9180	610.2500	9219	610.7375
9142	609.7750	9181	610.2625	9220	610.7500
9143	609.7875	9182	610.2750	9221	610.7625
9144	609.8000	9183	610.2875	9222	610.7750
9145	609.8125	9184	610.3000	9223	610.7875

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
9224	610.8000	9263	611.2875	9302	611.7750
9225	610.8125	9264	611.3000	9303	611.7875
9226	610.8250	9265	611.3125	9304	611.8000
9227	610.8375	9266	611.3250	9305	611.8125
9228	610.8500	9267	611.3375	9306	611.8250
9229	610.8625	9268	611.3500	9307	611.8375
9230	610.8750	9269	611.3625	9308	611.8500
9231	610.8875	9270	611.3750	9309	611.8625
9232	610.9000	9271	611.3875	9310	611.8750
9233	610.9125	9272	611.4000	9311	611.8875
9234	610.9250	9273	611.4125	9312	611.9000
9235	610.9375	9274	611.4250	9313	611.9125
9236	610.9500	9275	611.4375	9314	611.9250
9237	610.9625	9276	611.4500	9315	611.9375
9238	610.9750	9277	611.4625	9316	611.9500
9239	610.9875	9278	611.4750	9317	611.9625
9240	611.0000	9279	611.4875	9318	611.9750
9241	611.0125	9280	611.5000	9319	611.9875
9242	611.0250	9281	611.5125	9320	612.0000
9243	611.0375	9282	611.5250	9321	612.0125
9244	611.0500	9283	611.5375	9322	612.0250
9245	611.0625	9284	611.5500	9323	612.0375
9246	611.0750	9285	611.5625	9324	612.0500
9247	611.0875	9286	611.5750	9325	612.0625
9248	611.1000	9287	611.5875	9326	612.0750
9249	611.1125	9288	611.6000	9327	612.0875
9250	611.1250	9289	611.6125	9328	612.1000
9251	611.1375	9290	611.6250	9329	612.1125
9252	611.1500	9291	611.6375	9330	612.1250
9253	611.1625	9292	611.6500	9331	612.1375
9254	611.1750	9293	611.6625	9332	612.1500
9255	611.1875	9294	611.6750	9333	612.1625
9256	611.2000	9295	611.6875	9334	612.1750
9257	611.2125	9296	611.7000	9335	612.1875
9258	611.2250	9297	611.7125	9336	612.2000
9259	611.2375	9298	611.7250	9337	612.2125
9260	611.2500	9299	611.7375	9338	612.2250
9261	611.2625	9300	611.7500	9339	612.2375
9262	611.2750	9301	611.7625	9340	612.2500

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
9341	612.2625	9380	612.7500	9419	613.2375
9342	612.2750	9381	612.7625	9420	613.2500
9343	612.2875	9382	612.7750	9421	613.2625
9344	612.3000	9383	612.7875	9422	613.2750
9345	612.3125	9384	612.8000	9423	613.2875
9346	612.3250	9385	612.8125	9424	613.3000
9347	612.3375	9386	612.8250	9425	613.3125
9348	612.3500	9387	612.8375	9426	613.3250
9349	612.3625	9388	612.8500	9427	613.3375
9350	612.3750	9389	612.8625	9428	613.3500
9351	612.3875	9390	612.8750	9429	613.3625
9352	612.4000	9391	612.8875	9430	613.3750
9353	612.4125	9392	612.9000	9431	613.3875
9354	612.4250	9393	612.9125	9432	613.4000
9355	612.4375	9394	612.9250	9433	613.4125
9356	612.4500	9395	612.9375	9434	613.4250
9357	612.4625	9396	612.9500	9435	613.4375
9358	612.4750	9397	612.9625	9436	613.4500
9359	612.4875	9398	612.9750	9437	613.4625
9360	612.5000	9399	612.9875	9438	613.4750
9361	612.5125	9400	613.0000	9439	613.4875
9362	612.5250	9401	613.0125	9440	613.5000
9363	612.5375	9402	613.0250	9441	613.5125
9364	612.5500	9403	613.0375	9442	613.5250
9365	612.5625	9404	613.0500	9443	613.5375
9366	612.5750	9405	613.0625	9444	613.5500
9367	612.5875	9406	613.0750	9445	613.5625
9368	612.6000	9407	613.0875	9446	613.5750
9369	612.6125	9408	613.1000	9447	613.5875
9370	612.6250	9409	613.1125	9448	613.6000
9371	612.6375	9410	613.1250	9449	613.6125
9372	612.6500	9411	613.1375	9450	613.6250
9373	612.6625	9412	613.1500	9451	613.6375
9374	612.6750	9413	613.1625	9452	613.6500
9375	612.6875	9414	613.1750	9453	613.6625
9376	612.7000	9415	613.1875	9454	613.6750
9377	612.7125	9416	613.2000	9455	613.6875
9378	612.7250	9417	613.2125	9456	613.7000
9379	612.7375	9418	613.2250	9457	613.7125

Channel (HEX)	Frequency (MHz)
9458	613.7250
9459	613.7375
9460	613.7500
9461	613.7625
9462	613.7750
9463	613.7875
9464	613.8000
9465	613.8125
9466	613.8250
9467	613.8375
9468	613.8500
9469	613.8625
9470	613.8750
9471	613.8875
9472	613.9000
9473	613.9125
9474	613.9250
9475	613.9375
9476	613.9500
9477	613.9625
9478	613.9750

Channel: E002 to E398

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E002	1395.0250	E040	1395.5000	E078	1395.9750
E003	1395.0375	E041	1395.5125	E079	1395.9875
E004	1395.0500	E042	1395.5250	E080	1396.0000
E005	1395.0625	E043	1395.5375	E081	1396.0125
E006	1395.0750	E044	1395.5500	E082	1396.0250
E007	1395.0875	E045	1395.5625	E083	1396.0375
E008	1395.1000	E046	1395.5750	E084	1396.0500
E009	1395.1125	E047	1395.5875	E085	1396.0625
E010	1395.1250	E048	1395.6000	E086	1396.0750
E011	1395.1375	E049	1395.6125	E087	1396.0875
E012	1395.1500	E050	1395.6250	E088	1396.1000
E013	1395.1625	E051	1395.6375	E089	1396.1125
E014	1395.1750	E052	1395.6500	E090	1396.1250
E015	1395.1875	E053	1395.6625	E091	1396.1375
E016	1395.2000	E054	1395.6750	E092	1396.1500
E017	1395.2125	E055	1395.6875	E093	1396.1625
E018	1395.2250	E056	1395.7000	E094	1396.1750
E019	1395.2375	E057	1395.7125	E095	1396.1875
E020	1395.2500	E058	1395.7250	E096	1396.2000
E021	1395.2625	E059	1395.7375	E097	1396.2125
E022	1395.2750	E060	1395.7500	E098	1396.2250
E023	1395.2875	E061	1395.7625	E099	1396.2375
E024	1395.3000	E062	1395.7750	E100	1396.2500
E025	1395.3125	E063	1395.7875	E101	1396.2625
E026	1395.3250	E064	1395.8000	E102	1396.2750
E027	1395.3375	E065	1395.8125	E103	1396.2875
E028	1395.3500	E066	1395.8250	E104	1396.3000
E029	1395.3625	E067	1395.8375	E105	1396.3125
E030	1395.3750	E068	1395.8500	E106	1396.3250
E031	1395.3875	E069	1395.8625	E107	1396.3375
E032	1395.4000	E070	1395.8750	E108	1396.3500
E033	1395.4125	E071	1395.8875	E109	1396.3625
E034	1395.4250	E072	1395.9000	E110	1396.3750
E035	1395.4375	E073	1395.9125	E111	1396.3875
E036	1395.4500	E074	1395.9250	E112	1396.4000
E037	1395.4625	E075	1395.9375	E113	1396.4125
E038	1395.4750	E076	1395.9500	E114	1396.4250
E039	1395.4875	E077	1395.9625	E115	1396.4375

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E116	1396.4500	E155	1396.9375	E194	1397.4250
E117	1396.4625	E156	1396.9500	E195	1397.4375
E118	1396.4750	E157	1396.9625	E196	1397.4500
E119	1396.4875	E158	1396.9750	E197	1397.4625
E120	1396.5000	E159	1396.9875	E198	1397.4750
E121	1396.5125	E160	1397.0000	E199	1397.4875
E122	1396.5250	E161	1397.0125	E200	1397.5000
E123	1396.5375	E162	1397.0250	E201	1397.5125
E124	1396.5500	E163	1397.0375	E202	1397.5250
E125	1396.5625	E164	1397.0500	E203	1397.5375
E126	1396.5750	E165	1397.0625	E204	1397.5500
E127	1396.5875	E166	1397.0750	E205	1397.5625
E128	1396.6000	E167	1397.0875	E206	1397.5750
E129	1396.6125	E168	1397.1000	E207	1397.5875
E130	1396.6250	E169	1397.1125	E208	1397.6000
E131	1396.6375	E170	1397.1250	E209	1397.6125
E132	1396.6500	E171	1397.1375	E210	1397.6250
E133	1396.6625	E172	1397.1500	E211	1397.6375
E134	1396.6750	E173	1397.1625	E212	1397.6500
E135	1396.6875	E174	1397.1750	E213	1397.6625
E136	1396.7000	E175	1397.1875	E214	1397.6750
E137	1396.7125	E176	1397.2000	E215	1397.6875
E138	1396.7250	E177	1397.2125	E216	1397.7000
E139	1396.7375	E178	1397.2250	E217	1397.7125
E140	1396.7500	E179	1397.2375	E218	1397.7250
E141	1396.7625	E180	1397.2500	E219	1397.7375
E142	1396.7750	E181	1397.2625	E220	1397.7500
E143	1396.7875	E182	1397.2750	E221	1397.7625
E144	1396.8000	E183	1397.2875	E222	1397.7750
E145	1396.8125	E184	1397.3000	E223	1397.7875
E146	1396.8250	E185	1397.3125	E224	1397.8000
E147	1396.8375	E186	1397.3250	E225	1397.8125
E148	1396.8500	E187	1397.3375	E226	1397.8250
E149	1396.8625	E188	1397.3500	E227	1397.8375
E150	1396.8750	E189	1397.3625	E228	1397.8500
E151	1396.8875	E190	1397.3750	E229	1397.8625
E152	1396.9000	E191	1397.3875	E230	1397.8750
E153	1396.9125	E192	1397.4000	E231	1397.8875
E154	1396.9250	E193	1397.4125	E232	1397.9000

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E233	1397.9125	E272	1398.4000	E311	1398.8875
E234	1397.9250	E273	1398.4125	E312	1398.9000
E235	1397.9375	E274	1398.4250	E313	1398.9125
E236	1397.9500	E275	1398.4375	E314	1398.9250
E237	1397.9625	E276	1398.4500	E315	1398.9375
E238	1397.9750	E277	1398.4625	E316	1398.9500
E239	1397.9875	E278	1398.4750	E317	1398.9625
E240	1398.0000	E279	1398.4875	E318	1398.9750
E241	1398.0125	E280	1398.5000	E319	1398.9875
E242	1398.0250	E281	1398.5125	E320	1399.0000
E243	1398.0375	E282	1398.5250	E321	1399.0125
E244	1398.0500	E283	1398.5375	E322	1399.0250
E245	1398.0625	E284	1398.5500	E323	1399.0375
E246	1398.0750	E285	1398.5625	E324	1399.0500
E247	1398.0875	E286	1398.5750	E325	1399.0625
E248	1398.1000	E287	1398.5875	E326	1399.0750
E249	1398.1125	E288	1398.6000	E327	1399.0875
E250	1398.1250	E289	1398.6125	E328	1399.1000
E251	1398.1375	E290	1398.6250	E329	1399.1125
E252	1398.1500	E291	1398.6375	E330	1399.1250
E253	1398.1625	E292	1398.6500	E331	1399.1375
E254	1398.1750	E293	1398.6625	E332	1399.1500
E255	1398.1875	E294	1398.6750	E333	1399.1625
E256	1398.2000	E295	1398.6875	E334	1399.1750
E257	1398.2125	E296	1398.7000	E335	1399.1875
E258	1398.2250	E297	1398.7125	E336	1399.2000
E259	1398.2375	E298	1398.7250	E337	1399.2125
E260	1398.2500	E299	1398.7375	E338	1399.2250
E261	1398.2625	E300	1398.7500	E339	1399.2375
E262	1398.2750	E301	1398.7625	E340	1399.2500
E263	1398.2875	E302	1398.7750	E341	1399.2625
E264	1398.3000	E303	1398.7875	E342	1399.2750
E265	1398.3125	E304	1398.8000	E343	1399.2875
E266	1398.3250	E305	1398.8125	E344	1399.3000
E267	1398.3375	E306	1398.8250	E345	1399.3125
E268	1398.3500	E307	1398.8375	E346	1399.3250
E269	1398.3625	E308	1398.8500	E347	1399.3375
E270	1398.3750	E309	1398.8625	E348	1399.3500
E271	1398.3875	E310	1398.8750	E349	1399.3625

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E350	1399.3750	E389	1399.8625
E351	1399.3875	E390	1399.8750
E352	1399.4000	E391	1399.8875
E353	1399.4125	E392	1399.9000
E354	1399.4250	E393	1399.9125
E355	1399.4375	E394	1399.9250
E356	1399.4500	E395	1399.9375
E357	1399.4625	E396	1399.9500
E358	1399.4750	E397	1399.9625
E359	1399.4875	E398	1399.9750
E360	1399.5000		
E361	1399.5125		
E362	1399.5250		
E363	1399.5375		
E364	1399.5500		
E365	1399.5625		
E366	1399.5750		
E367	1399.5875		
E368	1399.6000		
E369	1399.6125		
E370	1399.6250		
E371	1399.6375		
E372	1399.6500		
E373	1399.6625		
E374	1399.6750		
E375	1399.6875		
E376	1399.7000		
E377	1399.7125		
E378	1399.7250		
E379	1399.7375		
E380	1399.7500		
E381	1399.7625		
E382	1399.7750		
E383	1399.7875		
E384	1399.8000		
E385	1399.8125		
E386	1399.8250		
E387	1399.8375		
E388	1399.8500		

Channel: E502 to E898

NOTE: In Canada, transmission frequencies from 1429.5 MHz to 1431.9750 MHz cannot be used due to the IC (Industry Canada) regulations. ZM-541PA can only use channels up to E698 in Canada.

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E502	1427.0250	E537	1427.4625	E572	1427.9000
E503	1427.0375	E538	1427.4750	E573	1427.9125
E504	1427.0500	E539	1427.4875	E574	1427.9250
E505	1427.0625	E540	1427.5000	E575	1427.9375
E506	1427.0750	E541	1427.5125	E576	1427.9500
E507	1427.0875	E542	1427.5250	E577	1427.9625
E508	1427.1000	E543	1427.5375	E578	1427.9750
E509	1427.1125	E544	1427.5500	E579	1427.9875
E510	1427.1250	E545	1427.5625	E580	1428.0000
E511	1427.1375	E546	1427.5750	E581	1428.0125
E512	1427.1500	E547	1427.5875	E582	1428.0250
E513	1427.1625	E548	1427.6000	E583	1428.0375
E514	1427.1750	E549	1427.6125	E584	1428.0500
E515	1427.1875	E550	1427.6250	E585	1428.0625
E516	1427.2000	E551	1427.6375	E586	1428.0750
E517	1427.2125	E552	1427.6500	E587	1428.0875
E518	1427.2250	E553	1427.6625	E588	1428.1000
E519	1427.2375	E554	1427.6750	E589	1428.1125
E520	1427.2500	E555	1427.6875	E590	1428.1250
E521	1427.2625	E556	1427.7000	E591	1428.1375
E522	1427.2750	E557	1427.7125	E592	1428.1500
E523	1427.2875	E558	1427.7250	E593	1428.1625
E524	1427.3000	E559	1427.7375	E594	1428.1750
E525	1427.3125	E560	1427.7500	E595	1428.1875
E526	1427.3250	E561	1427.7625	E596	1428.2000
E527	1427.3375	E562	1427.7750	E597	1428.2125
E528	1427.3500	E563	1427.7875	E598	1428.2250
E529	1427.3625	E564	1427.8000	E599	1428.2375
E530	1427.3750	E565	1427.8125	E600	1428.2500
E531	1427.3875	E566	1427.8250	E601	1428.2625
E532	1427.4000	E567	1427.8375	E602	1428.2750
E533	1427.4125	E568	1427.8500	E603	1428.2875
E534	1427.4250	E569	1427.8625	E604	1428.3000
E535	1427.4375	E570	1427.8750	E605	1428.3125
E536	1427.4500	E571	1427.8875	E606	1428.3250

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E607	1428.3375	E646	1428.8250	E685	1429.3125
E608	1428.3500	E647	1428.8375	E686	1429.3250
E609	1428.3625	E648	1428.8500	E687	1429.3375
E610	1428.3750	E649	1428.8625	E688	1429.3500
E611	1428.3875	E650	1428.8750	E689	1429.3625
E612	1428.4000	E651	1428.8875	E690	1429.3750
E613	1428.4125	E652	1428.9000	E691	1429.3875
E614	1428.4250	E653	1428.9125	E692	1429.4000
E615	1428.4375	E654	1428.9250	E693	1429.4125
E616	1428.4500	E655	1428.9375	E694	1429.4250
E617	1428.4625	E656	1428.9500	E695	1429.4375
E618	1428.4750	E657	1428.9625	E696	1429.4500
E619	1428.4875	E658	1428.9750	E697	1429.4625
E620	1428.5000	E659	1428.9875	E698	1429.4750
E621	1428.5125	E660	1429.0000	E699	1429.4875
E622	1428.5250	E661	1429.0125	E700	1429.5000
E623	1428.5375	E662	1429.0250	E701	1429.5125
E624	1428.5500	E663	1429.0375	E702	1429.5250
E625	1428.5625	E664	1429.0500	E703	1429.5375
E626	1428.5750	E665	1429.0625	E704	1429.5500
E627	1428.5875	E666	1429.0750	E705	1429.5625
E628	1428.6000	E667	1429.0875	E706	1429.5750
E629	1428.6125	E668	1429.1000	E707	1429.5875
E630	1428.6250	E669	1429.1125	E708	1429.6000
E631	1428.6375	E670	1429.1250	E709	1429.6125
E632	1428.6500	E671	1429.1375	E710	1429.6250
E633	1428.6625	E672	1429.1500	E711	1429.6375
E634	1428.6750	E673	1429.1625	E712	1429.6500
E635	1428.6875	E674	1429.1750	E713	1429.6625
E636	1428.7000	E675	1429.1875	E714	1429.6750
E637	1428.7125	E676	1429.2000	E715	1429.6875
E638	1428.7250	E677	1429.2125	E716	1429.7000
E639	1428.7375	E678	1429.2250	E717	1429.7125
E640	1428.7500	E679	1429.2375	E718	1429.7250
E641	1428.7625	E680	1429.2500	E719	1429.7375
E642	1428.7750	E681	1429.2625	E720	1429.7500
E643	1428.7875	E682	1429.2750	E721	1429.7625
E644	1428.8000	E683	1429.2875	E722	1429.7750
E645	1428.8125	E684	1429.3000	E723	1429.7875

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E724	1429.8000	E763	1430.2875	E802	1430.7750
E725	1429.8125	E764	1430.3000	E803	1430.7875
E726	1429.8250	E765	1430.3125	E804	1430.8000
E727	1429.8375	E766	1430.3250	E805	1430.8125
E728	1429.8500	E767	1430.3375	E806	1430.8250
E729	1429.8625	E768	1430.3500	E807	1430.8375
E730	1429.8750	E769	1430.3625	E808	1430.8500
E731	1429.8875	E770	1430.3750	E809	1430.8625
E732	1429.9000	E771	1430.3875	E810	1430.8750
E733	1429.9125	E772	1430.4000	E811	1430.8875
E734	1429.9250	E773	1430.4125	E812	1430.9000
E735	1429.9375	E774	1430.4250	E813	1430.9125
E736	1429.9500	E775	1430.4375	E814	1430.9250
E737	1429.9625	E776	1430.4500	E815	1430.9375
E738	1429.9750	E777	1430.4625	E816	1430.9500
E739	1429.9875	E778	1430.4750	E817	1430.9625
E740	1430.0000	E779	1430.4875	E818	1430.9750
E741	1430.0125	E780	1430.5000	E819	1430.9875
E742	1430.0250	E781	1430.5125	E820	1431.0000
E743	1430.0375	E782	1430.5250	E821	1431.0125
E744	1430.0500	E783	1430.5375	E822	1431.0250
E745	1430.0625	E784	1430.5500	E823	1431.0375
E746	1430.0750	E785	1430.5625	E824	1431.0500
E747	1430.0875	E786	1430.5750	E825	1431.0625
E748	1430.1000	E787	1430.5875	E826	1431.0750
E749	1430.1125	E788	1430.6000	E827	1431.0875
E750	1430.1250	E789	1430.6125	E828	1431.1000
E751	1430.1375	E790	1430.6250	E829	1431.1125
E752	1430.1500	E791	1430.6375	E830	1431.1250
E753	1430.1625	E792	1430.6500	E831	1431.1375
E754	1430.1750	E793	1430.6625	E832	1431.1500
E755	1430.1875	E794	1430.6750	E833	1431.1625
E756	1430.2000	E795	1430.6875	E834	1431.1750
E757	1430.2125	E796	1430.7000	E835	1431.1875
E758	1430.2250	E797	1430.7125	E836	1431.2000
E759	1430.2375	E798	1430.7250	E837	1431.2125
E760	1430.2500	E799	1430.7375	E838	1431.2250
E761	1430.2625	E800	1430.7500	E839	1431.2375
E762	1430.2750	E801	1430.7625	E840	1431.2500

Channel (HEX)	Frequency (MHz)	Channel (HEX)	Frequency (MHz)
E841	1431.2625	E880	1431.7500
E842	1431.2750	E881	1431.7625
E843	1431.2875	E882	1431.7750
E844	1431.3000	E883	1431.7875
E845	1431.3125	E884	1431.8000
E846	1431.3250	E885	1431.8125
E847	1431.3375	E886	1431.8250
E848	1431.3500	E887	1431.8375
E849	1431.3625	E888	1431.8500
E850	1431.3750	E889	1431.8625
E851	1431.3875	E890	1431.8750
E852	1431.4000	E891	1431.8875
E853	1431.4125	E892	1431.9000
E854	1431.4250	E893	1431.9125
E855	1431.4375	E894	1431.9250
E856	1431.4500	E895	1431.9375
E857	1431.4625	E896	1431.9500
E858	1431.4750	E897	1431.9625
E859	1431.4875	E898	1431.9750
E860	1431.5000		
E861	1431.5125		
E862	1431.5250		
E863	1431.5375		
E864	1431.5500		
E865	1431.5625		
E866	1431.5750		
E867	1431.5875		
E868	1431.6000		
E869	1431.6125		
E870	1431.6250		
E871	1431.6375		
E872	1431.6500		
E873	1431.6625		
E874	1431.6750		
E875	1431.6875		
E876	1431.7000		
E877	1431.7125		
E878	1431.7250		
E879	1431.7375		



Manufacturer

NIHON KOHDEN CORPORATION

1-31-4 Nishiochiai, Shinjuku-ku Tokyo 161-8560, Japan
Phone +81 3-5996-8036
Fax +81 3-5996-8100

North and South America

NIHON KOHDEN AMERICA, INC.

15353 Barranca Parkway, Irvine, CA 92618, U.S.A.
Toll-free +1-800-325-0283
Phone +1 949-580-1555
Fax +1 949-580-1550

NIHON KOHDEN LATIN AMERICA S.A.S

Carrera 16 No.93A-36 Of. 802. Bogota D.C., Colombia
Phone +57 1-300-1742
Fax +57 1-300-1825

NIHON KOHDEN DO BRASIL LTDA.

Rua Gomes de Carvalho
1507-Bloco A 9 andar-cjto 91A Vila Olimpia
CEP 04547-005 Sao Paulo-SP. Brasil
Phone +55 11-3841-9943
Fax +55 11-3044-0463

Europe

 European Representative

NIHON KOHDEN EUROPE GmbH

Raiffeisenstrasse 10, D-61191 Rosbach, Germany
Phone +49 6003-827-0
Fax +49 6003-827-599

NIHON KOHDEN FRANCE SARL

8, rue Francois Delage, 94 230 Cachan, France
Phone +33 1-49-08-05-50
Fax +33 1-49-08-93-32

NIHON KOHDEN IBERICA S.L.

C/Ulises 75A, E-28043 Madrid, Spain
Phone +34 91-7-161080
Fax +34 91-3-004676

NIHON KOHDEN ITALIA S.r.l.

Via Fratelli Bronzetti 28, 24124 Bergamo, Italy
Phone +39 035-219543
Fax +39 035-232546

NIHON KOHDEN UK LTD

Trident Court 118, 1 Oakcroft Road
Chessington, Surrey KT9 1BD, UK
Phone +44 20-8391-6800
Fax +44 20-8391-6809

Asia

SHANGHAI KOHDEN

MEDICAL ELECTRONIC INSTRUMENT CORP.

567 Huancheng Bei Road
Shanghai Comprehensive Industrial Development Zone
Shanghai 201401, China
Phone +86 21-5743-6998
Fax +86 21-5743-6939

NIHON KOHDEN SINGAPORE PTE LTD

1 Maritime Square, #10-34 HarbourFront Centre
Singapore 099253
Phone +65 6376-2210
Fax +65 6376-2264

NIHON KOHDEN INDIA PVT. LTD.

308, Tower A, Spazedge, Sector 47, Sohna Road
Gurgaon-122 002 Haryana, India
Toll-free +91 1800-103-8182
Phone +91 124-493-1000
Fax +91 124-493-1029

NIHON KOHDEN MIDDLE EAST FZE

P.O.Box 261516, JAFZA View 19
Room 308 Jebel Ali Free Zone, Dubai, U.A.E
Phone +971 4-884-0080
Fax +971 4-880-0122

NIHON KOHDEN KOREA, INC.

5F Miso Bldg.
890-47 Daechi-dong, Gangnam-gu, Seoul, 135-280 Korea
Phone +82 2-3273-2310
Fax +82 2-3273-2352

Contact information is accurate as of Jan 2015. Visit www.nihonkohden.com for the latest information.

The model and serial number of your instrument are identified on the rear or bottom of the unit.

Write the model and serial number in the spaces provided below. Whenever you call your representative concerning this instrument, mention these two pieces of information for quick and accurate service.

Model _____

Serial Number _____

Your Representative

ZM-540PA, ZM-541PA_0614-902754H