

System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	40.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.87
Conversion Factor	7.5	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7810 2026-05-06	Phantom	Twin-SAM V4.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1434 2026-04-30	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.09	5.01
psSAR10g [W/Kg]	2.26	2.28
Power Drift [±0.2 dB]	N/A	-0.16



System Performance Check Report for D2300V2 - SN1058

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	40.0
Group UID	CW 0-	TSL Conductivity [S/m]	1.68
Conversion Factor	7.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7810 2026-05-06	Phantom	Twin-SAM V4.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1434 2026-04-30	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		84.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.22	2.20
psSAR10g [W/Kg]	1.06	1.08
Power Drift [±0.2 dB]	N/A	-0.11



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	38.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.64
Conversion Factor	7.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7810 2026-05-06	Phantom	Twin-SAM V4.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1434 2026-04-30	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		79.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.21	2.26
psSAR10g [W/Kg]	1.05	1.11
Power Drift [±0.2 dB]	N/A	-0.10



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	37.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.87
Conversion Factor	7.5	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7810 2026-05-06	Phantom	Twin-SAM V4.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1434 2026-04-30	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.52	2.55
psSAR10g [W/Kg]	1.16	1.17
Power Drift [±0.2 dB]	N/A	-0.03



System Performance Check Report for D2300V2 - SN1160

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.7
Group UID	CW 0-	TSL Conductivity [S/m]	1.64
Conversion Factor	7.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7810 2026-05-06	Phantom	Twin-SAM V4.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1434 2026-04-30	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.39	2.37
psSAR10g [W/Kg]	1.15	1.17
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D2300V2 - SN1058

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.67
Conversion Factor	7.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7482 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1787 2026-04-10	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	4.66	4.69
psSAR10g [W/Kg]	2.23	2.27
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	40.0
Group UID	CW 0-	TSL Conductivity [S/m]	1.65
Conversion Factor	7.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7482 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1787 2026-04-10	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.25	2.29
psSAR10g [W/Kg]	1.06	1.12
Power Drift [±0.2 dB]	N/A	-0.09



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.4
Group UID	CW 0-	TSL Conductivity [S/m]	1.90
Conversion Factor	7.0	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7482 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1787 2026-04-10	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.44	2.50
psSAR10g [W/Kg]	1.08	1.15
Power Drift [±0.2 dB]	N/A	-0.10



System Performance Check Report for D2300V2 - SN1160

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	38.4
Group UID	CW 0-	TSL Conductivity [S/m]	1.62
Conversion Factor	7.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7482 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1787 2026-04-10	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.18	2.24
psSAR10g [W/Kg]	1.04	1.10
Power Drift [±0.2 dB]	N/A	-0.09



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.88
Conversion Factor	7.0	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7482 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1787 2026-04-10	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.49	2.60
psSAR10g [W/Kg]	1.10	1.20
Power Drift [±0.2 dB]	N/A	-0.11



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	42.4
Group UID	CW 0-	TSL Conductivity [S/m]	1.63
Conversion Factor	7.73	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7569 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1547 2026-04-14	TSL Type	DAK 3.5 Head 22.44 deg.C 2026-Jun-12 02_36_16
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.25	2.29
psSAR10g [W/Kg]	1.06	1.12
Power Drift [±0.2 dB]	N/A	-0.07



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	42.0
Group UID	CW 0-	TSL Conductivity [S/m]	1.85
Conversion Factor	7.46	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7569 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1547 2026-04-14	TSL Type	DAK 3.5 Head 22.44 deg.C 2026-Jun-12 02_36_16
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		79.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.44	2.50
psSAR10g [W/Kg]	1.09	1.14
Power Drift [±0.2 dB]	N/A	-0.07



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	41.4
Group UID	CW 0-	TSL Conductivity [S/m]	1.86
Conversion Factor	7.46	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7569 2026-04-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1547 2026-04-14	TSL Type	DAK 3.5 Head 25 deg.C 2026-Jun-16 05_18_22
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.29	5.49
psSAR10g [W/Kg]	2.34	2.51
Power Drift [±0.2 dB]	N/A	0.04



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	42.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.85
Conversion Factor	7.05	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3885 2025-10-16	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1892 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.48	2.58
psSAR10g [W/Kg]	1.15	1.24
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	42.2
Group UID	CW 0-	TSL Conductivity [S/m]	1.85
Conversion Factor	7.05	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3885 2025-10-16	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1892 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.51	2.53
psSAR10g [W/Kg]	1.15	1.21
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.3
Group UID	CW 0-	TSL Conductivity [S/m]	1.73
Conversion Factor	7.31	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3885 2025-10-16	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1892 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.72	2.64
psSAR10g [W/Kg]	1.31	1.29
Power Drift [±0.2 dB]	-0.11	-0.10



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	36.0
Group UID	CW 0-	TSL Conductivity [S/m]	2.72
Conversion Factor	5.76	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.06	6.20
psSAR10g [W/Kg]	2.37	2.45
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for D3900V2 - SN1115

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	40.0
Group UID	CW 0-	TSL Conductivity [S/m]	3.05
Conversion Factor	5.63	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.44	3.53
psSAR10g [W/Kg]	1.22	1.26
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D3500V2 - SN1182

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	35.3
Group UID	CW 0-	TSL Conductivity [S/m]	2.71
Conversion Factor	5.76	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		77.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	7.07	6.73
psSAR10g [W/Kg]	2.73	2.63
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	40.8
Group UID	CW 0-	TSL Conductivity [S/m]	2.90
Conversion Factor	5.69	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.14	6.37
psSAR10g [W/Kg]	2.25	2.41
Power Drift [±0.2 dB]	N/A	-0.10



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	40.4
Group UID	CW 0-	TSL Conductivity [S/m]	3.08
Conversion Factor	5.63	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.19	6.51
psSAR10g [W/Kg]	2.18	2.35
Power Drift [±0.2 dB]	N/A	-0.06



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	40.0
Group UID	CW 0-	TSL Conductivity [S/m]	2.94
Conversion Factor	5.69	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3686 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1357 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		77.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.80	6.01
psSAR10g [W/Kg]	2.18	2.33
Power Drift [±0.2 dB]	N/A	-0.03



System Performance Check Report for D3500V2 - SN1182

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	40.6
Group UID	CW 0-	TSL Conductivity [S/m]	2.74
Conversion Factor	5.79	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.40	6.62
psSAR10g [W/Kg]	2.48	2.58
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for CLA13 - SN1008

Exposure Conditions

Frequency [MHz] Channel Number	13.000 0	TSL Permittivity	56.7
Group UID	CW 0-	TSL Conductivity [S/m]	0.693
Conversion Factor	13.93	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	ELI V4.0 (20deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	DAK 12 20 deg.C 2026-Jun-13 23_05_58
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		14.9

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	0.540	0.505
psSAR10g [W/Kg]	0.434	0.309
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3500V2 - SN1060

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	38.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.71
Conversion Factor	5.79	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.77	6.74
psSAR10g [W/Kg]	2.58	2.60
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.76
Conversion Factor	6.54	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.70	2.69
psSAR10g [W/Kg]	1.29	1.30
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	41.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.74
Conversion Factor	5.79	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.37	3.55
psSAR10g [W/Kg]	1.39	1.37
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	38.8
Group UID	CW 0-	TSL Conductivity [S/m]	1.96
Conversion Factor	6.28	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.89	2.92
psSAR10g [W/Kg]	1.31	1.33
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D2300V2 - SN1160

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.68
Conversion Factor	6.54	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.5

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.74	2.65
psSAR10g [W/Kg]	1.31	1.29
Power Drift [±0.2 dB]	N/A	-0.06



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	38.6
Group UID	CW 0-	TSL Conductivity [S/m]	2.94
Conversion Factor	5.76	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.42	3.48
psSAR10g [W/Kg]	1.29	1.33
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	39.3
Group UID	CW 0-	TSL Conductivity [S/m]	3.13
Conversion Factor	5.71	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.12	3.20
psSAR10g [W/Kg]	1.14	1.19
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	38.3
Group UID	CW 0-	TSL Conductivity [S/m]	2.91
Conversion Factor	5.76	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3749 2026-01-15	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1784 2026-04-13	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.12	3.20
psSAR10g [W/Kg]	1.20	1.24
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for D3700V2 - SN1110

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	38.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.92
Conversion Factor	6.35	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.11	6.42
psSAR10g [W/Kg]	2.28	2.41
Power Drift [±0.2 dB]	N/A	-0.14



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	38.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.96
Conversion Factor	6.35	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.15	6.32
psSAR10g [W/Kg]	2.28	2.37
Power Drift [±0.2 dB]	N/A	-0.09



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	38.1
Group UID	CW 0-	TSL Conductivity [S/m]	3.15
Conversion Factor	6.25	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.42	6.65
psSAR10g [W/Kg]	2.27	2.37
Power Drift [±0.2 dB]	N/A	-0.07



System Performance Check Report for D3500V2 - SN1182

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.72
Conversion Factor	6.35	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		79.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.60	5.75
psSAR10g [W/Kg]	2.15	2.23
Power Drift [±0.2 dB]	N/A	-0.08



System Performance Check Report for D3900V2 - SN1115

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	39.2
Group UID	CW 0-	TSL Conductivity [S/m]	3.11
Conversion Factor	6.25	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.79	5.98
psSAR10g [W/Kg]	2.06	2.14
Power Drift [±0.2 dB]	N/A	-0.06



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	40.3
Group UID	CW 0-	TSL Conductivity [S/m]	2.71
Conversion Factor	6.35	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7448 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1380 2026-02-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		78.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.38	6.27
psSAR10g [W/Kg]	2.43	2.44
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	37.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.73
Conversion Factor	6.63	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		5.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.89	6.18
psSAR10g [W/Kg]	2.28	2.28
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	37.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.92
Conversion Factor	6.65	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.22	6.30
psSAR10g [W/Kg]	2.33	2.33
Power Drift [±0.2 dB]	N/A	-0.04



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	36.6
Group UID	CW 0-	TSL Conductivity [S/m]	3.24
Conversion Factor	6.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.09	3.21
psSAR10g [W/Kg]	1.09	1.14
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	36.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.02
Conversion Factor	7.25	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.55	2.51
psSAR10g [W/Kg]	1.15	1.12
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D3500V2 - SN1060

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	37.0
Group UID	CW 0-	TSL Conductivity [S/m]	2.93
Conversion Factor	6.63	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.50	3.42
psSAR10g [W/Kg]	1.35	1.31
Power Drift [±0.2 dB]	N/A	-0.13



System Performance Check Report for D3900V2 - SN1115

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	36.2
Group UID	CW 0-	TSL Conductivity [S/m]	3.33
Conversion Factor	6.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		72.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.28	3.40
psSAR10g [W/Kg]	1.19	1.21
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	38.7
Group UID	CW 0-	TSL Conductivity [S/m]	1.97
Conversion Factor	7.25	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	DAKS 3.5 Head 20.95 deg.C 2026-Jul-16 08_54_43
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		78.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.59	2.61
psSAR10g [W/Kg]	1.18	1.18
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	35.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.95
Conversion Factor	6.65	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3990 2026-02-17	Phantom	Twin-SAM V5.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1545 2026-02-06	TSL Type	DAK 3.5 Head 20.72 deg.C 2026-Jul-20 09_47_18
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.14	3.19
psSAR10g [W/Kg]	1.20	1.21
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	40.6
Group UID	CW 0-	TSL Conductivity [S/m]	2.94
Conversion Factor	6.31	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7897 2025-10-14	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1377 2025-09-09	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.95	3.17
psSAR10g [W/Kg]	1.18	1.21
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D3500V2 - SN1060

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.2
Group UID	CW 0-	TSL Conductivity [S/m]	2.83
Conversion Factor	6.38	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7897 2025-10-14	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1377 2025-09-09	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.27	3.36
psSAR10g [W/Kg]	1.30	1.30
Power Drift [±0.2 dB]	N/A	-0.07



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	38.7
Group UID	CW 0-	TSL Conductivity [S/m]	3.18
Conversion Factor	6.28	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7897 2025-10-14	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1377 2025-09-09	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.16	3.31
psSAR10g [W/Kg]	1.15	1.17
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.8
Group UID	CW 0-	TSL Conductivity [S/m]	1.96
Conversion Factor	6.74	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7897 2025-10-14	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1377 2025-09-09	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.49	2.51
psSAR10g [W/Kg]	1.15	1.14
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	38.1
Group UID	CW 0-	TSL Conductivity [S/m]	2.73
Conversion Factor	6.38	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7897 2025-10-14	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1377 2025-09-09	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.88	3.11
psSAR10g [W/Kg]	1.18	1.19
Power Drift [±0.2 dB]	N/A	0.02



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.9
Group UID	CW 0-	TSL Conductivity [S/m]	1.87
Conversion Factor	6.82	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7979 2026-03-03	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1943 2026-02-23	TSL Type	DAK 3.5 24.42 deg.C 2026-Jun-12 17_21_13
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.42	2.52
psSAR10g [W/Kg]	1.08	1.16
Power Drift [±0.2 dB]	-0.06	-0.14



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.94
Conversion Factor	6.82	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7979 2026-03-03	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1943 2026-02-23	TSL Type	DAK 3.5 Head 22.18 deg.C 2026-Jun-22 12_25_27
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.5

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.89	2.92
psSAR10g [W/Kg]	1.32	1.36
Power Drift [±0.2 dB]	-0.01	-0.01



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	36.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.65
Conversion Factor	7.1	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7979 2026-03-03	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1943 2026-02-23	TSL Type	DAK 3.5 22.2 deg.C 2026-Jun-30 00_53_50
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		10.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.17	5.22
psSAR10g [W/Kg]	2.53	2.57
Power Drift [±0.2 dB]	-0.00	0.01



System Performance Check Report for D2300V2 - SN1058

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	39.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.65
Conversion Factor	7.1	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7979 2026-03-03	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1943 2026-02-23	TSL Type	DAKS 3.5 Head 22.62 deg.C 2026-Jul-16 10_23_45
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		10.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.47	2.46
psSAR10g [W/Kg]	1.19	1.21
Power Drift [±0.2 dB]	0.00	-0.02



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	40.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.99
Conversion Factor	7.04	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7981 2026-03-17	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1944 2026-03-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.59	5.67
psSAR10g [W/Kg]	2.56	2.63
Power Drift [±0.2 dB]	0.00	0.00



System Performance Check Report for D2300V2 - SN1058

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.8
Group UID	CW 0-	TSL Conductivity [S/m]	1.65
Conversion Factor	7.24	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7981 2026-03-17	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1944 2026-03-05	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		83.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.9

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	4.88	4.97
psSAR10g [W/Kg]	2.32	2.46
Power Drift [±0.2 dB]	-0.04	-0.12



System Performance Check Report for D2300V2 - SN1002

Exposure Conditions

Frequency [MHz] Channel Number	2300.000 0	TSL Permittivity	41.3
Group UID	CW 0-	TSL Conductivity [S/m]	1.62
Conversion Factor	7.24	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7981 2026-03-17	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1944 2026-03-05	TSL Type	DAK 3.5 Head 20.75 deg.C 2026-Jun-29 08_45_22
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.23	2.31
psSAR10g [W/Kg]	1.06	1.13
Power Drift [±0.2 dB]	-0.05	-0.11



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	39.8
Group UID	CW 0-	TSL Conductivity [S/m]	1.87
Conversion Factor	7.04	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7981 2026-03-17	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1944 2026-03-05	TSL Type	DAK 3.5 Head 21.11 deg.C 2026-Jul-02 07_45_42
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.44	2.55
psSAR10g [W/Kg]	1.08	1.17
Power Drift [±0.2 dB]	-0.04	-0.12



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	40.3
Group UID	CW 0-	TSL Conductivity [S/m]	2.95
Conversion Factor	6.34	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 Head 20.4 deg.C 2026-Jun-16 13_46_19
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.01	3.14
psSAR10g [W/Kg]	1.11	1.18
Power Drift [±0.2 dB]	-0.03	-0.10



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	39.9
Group UID	CW 0-	TSL Conductivity [S/m]	3.14
Conversion Factor	6.31	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 Head 20.4 deg.C 2026-Jun-16 13_46_19
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		72.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.05	6.40
psSAR10g [W/Kg]	2.15	2.29
Power Drift [±0.2 dB]	-0.02	-0.08



System Performance Check Report for D3500V2 - SN1060

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	38.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.75
Conversion Factor	6.38	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 Head 20.36 deg.C 2026-Jun-19 02_26_40
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.81	2.92
psSAR10g [W/Kg]	1.07	1.14
Power Drift [±0.2 dB]	-0.04	-0.06



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	37.0
Group UID	CW 0-	TSL Conductivity [S/m]	1.90
Conversion Factor	6.9	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 20.24 deg.C 2026-Jul-06 19_51_03
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.56	5.76
psSAR10g [W/Kg]	2.50	2.63
Power Drift [±0.2 dB]	-0.06	-0.12



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	36.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.92
Conversion Factor	6.34	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 20.24 deg.C 2026-Jul-06 20_26_17
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.79	6.15
psSAR10g [W/Kg]	2.19	2.34
Power Drift [±0.2 dB]	-0.06	-0.10



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	42.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.92
Conversion Factor	6.9	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 22.25 deg.C 2026-Jul-17 08_37_02
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.65	2.70
psSAR10g [W/Kg]	1.18	1.24
Power Drift [±0.2 dB]	-0.08	-0.08



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	41.1
Group UID	CW 0-	TSL Conductivity [S/m]	2.71
Conversion Factor	6.38	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7983 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1947 2026-03-23	TSL Type	DAK 3.5 22.25 deg.C 2026-Jul-17 08_37_02
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.93	3.04
psSAR10g [W/Kg]	1.12	1.19
Power Drift [±0.2 dB]	-0.05	-0.07



System Performance Check Report for D3900V2 - SN1115

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	39.3
Group UID	CW 0-	TSL Conductivity [S/m]	3.10
Conversion Factor	6.44	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 21.75 deg.C 2026-Jun-12 11_41_19
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		71.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.85	3.02
psSAR10g [W/Kg]	1.05	1.12
Power Drift [±0.2 dB]	0.00	-0.01



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	37.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.91
Conversion Factor	7.12	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 22.82 deg.C 2026-Jun-22 18_53_38
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		79.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.96	2.95
psSAR10g [W/Kg]	1.34	1.35
Power Drift [±0.2 dB]	-0.12	-0.03



System Performance Check Report for D3700V2 - SN1110

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	37.8
Group UID	CW 0-	TSL Conductivity [S/m]	2.97
Conversion Factor	6.5	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 22.82 deg.C 2026-Jun-22 18_53_38
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.55	3.64
psSAR10g [W/Kg]	1.32	1.38
Power Drift [±0.2 dB]	0.00	0.00



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	37.8
Group UID	CW 0-	TSL Conductivity [S/m]	3.15
Conversion Factor	6.44	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 22.26 deg.C 2026-Jun-25 19_04_06
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.34	6.58
psSAR10g [W/Kg]	2.25	2.37
Power Drift [±0.2 dB]	0.01	0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	40.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.91
Conversion Factor	6.5	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 22.23 deg.C 2026-Jul-03 05_31_57
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.61	3.67
psSAR10g [W/Kg]	1.35	1.40
Power Drift [±0.2 dB]	-0.01	0.00



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.73
Conversion Factor	6.57	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 24 deg.C 2026-Jul-13 19_05_51
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.1

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.05	6.12
psSAR10g [W/Kg]	2.31	2.39
Power Drift [±0.2 dB]	0.00	0.01



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	41.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.93
Conversion Factor	7.12	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 21.56 deg.C 2026-Jul-17 14_34_49
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.03	3.00
psSAR10g [W/Kg]	1.36	1.37
Power Drift [±0.2 dB]	-0.03	-0.01



System Performance Check Report for D3500V2 - SN1182

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	40.0
Group UID	CW 0-	TSL Conductivity [S/m]	2.75
Conversion Factor	6.57	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 21.56 deg.C 2026-Jul-17 14_34_49
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.07	3.14
psSAR10g [W/Kg]	1.20	1.25
Power Drift [±0.2 dB]	0.00	-0.01



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	39.6
Group UID	CW 0-	TSL Conductivity [S/m]	2.95
Conversion Factor	6.5	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7977 2026-03-02	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1942 2026-02-23	TSL Type	DAK 3.5 Head 21.56 deg.C 2026-Jul-17 14_34_49
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.00	3.07
psSAR10g [W/Kg]	1.16	1.20
Power Drift [±0.2 dB]	0.00	-0.01



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	40.6
Group UID	CW 0-	TSL Conductivity [S/m]	2.73
Conversion Factor	6.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 Head 20 deg.C 2026-Jun-15 17_40_50
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.93	6.19
psSAR10g [W/Kg]	2.32	2.42
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D3500V2 - SN1060

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	38.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.84
Conversion Factor	6.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 Head 20.28 deg.C 2026-Jun-29 20_18_22
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.32	3.43
psSAR10g [W/Kg]	1.29	1.34
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	40.7
Group UID	CW 0-	TSL Conductivity [S/m]	1.95
Conversion Factor	6.73	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 19.43 deg.C 2026-Jul-06 11_33_43
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		79.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.47	2.52
psSAR10g [W/Kg]	1.12	1.15
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	38.4
Group UID	CW 0-	TSL Conductivity [S/m]	3.18
Conversion Factor	6.12	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 19.43 deg.C 2026-Jul-06 11_33_43
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.59	3.76
psSAR10g [W/Kg]	1.27	1.35
Power Drift [±0.2 dB]	N/A	-0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	39.5
Group UID	CW 0-	TSL Conductivity [S/m]	2.99
Conversion Factor	6.17	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 19.58 deg.C 2026-Jul-09 09_43_50
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.59	3.71
psSAR10g [W/Kg]	1.34	1.40
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	40.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.90
Conversion Factor	6.73	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7982 2026-03-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1945 2026-03-05	TSL Type	DAK 3.5 Head 20.43 deg.C 2026-Jul-20 10_55_13
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.92	2.95
psSAR10g [W/Kg]	1.32	1.36
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	36.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.74
Conversion Factor	6.84	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7710 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1674 2026-04-27	TSL Type	DAK 3.5 Head 19.82 deg.C 2026-Jun-16 20_00_58
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		77.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.34	6.34
psSAR10g [W/Kg]	2.43	2.49
Power Drift [±0.2 dB]	N/A	-0.02



System Performance Check Report for D3900V2 - SN1115

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	39.0
Group UID	CW 0-	TSL Conductivity [S/m]	3.17
Conversion Factor	6.78	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7710 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1674 2026-04-27	TSL Type	DAK 3.5 Head 23.57 deg.C 2026-Jun-18 21_18_39
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.76	6.99
psSAR10g [W/Kg]	2.39	2.52
Power Drift [±0.2 dB]	N/A	0.03



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	37.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.79
Conversion Factor	6.84	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7710 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1674 2026-04-27	TSL Type	DAK 3.5 Head 20.4 deg.C 2026-Jun-22 16_17_10
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		77.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.03	6.05
psSAR10g [W/Kg]	2.33	2.38
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	38.2
Group UID	CW 0-	TSL Conductivity [S/m]	2.94
Conversion Factor	6.92	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7710 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1674 2026-04-27	TSL Type	DAK 3.5 22 deg.C 2026-Jun-26 00_49_08
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.93	6.35
psSAR10g [W/Kg]	2.20	2.38
Power Drift [±0.2 dB]	N/A	-0.08



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	37.9
Group UID	CW 0-	TSL Conductivity [S/m]	3.17
Conversion Factor	6.78	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7710 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1674 2026-04-27	TSL Type	DAK 3.5 22 deg.C 2026-Jun-26 00_49_08
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		73.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	6.17	6.60
psSAR10g [W/Kg]	2.16	2.35
Power Drift [±0.2 dB]	N/A	-0.07



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	41.0
Group UID	CW 0-	TSL Conductivity [S/m]	2.02
Conversion Factor	6.91	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7947 2025-09-10	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1798 2026-04-27	TSL Type	DAK 3.5 Head 23.93 deg.C 2026-Jun-16 19_30_26
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		80.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.94	5.97
psSAR10g [W/Kg]	2.66	2.72
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.81
Conversion Factor	6.22	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7947 2025-09-10	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1798 2026-04-27	TSL Type	DAK 3.5 Head 19.73 deg.C 2026-Jun-22 09_13_30
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		76.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	7.12	7.21
psSAR10g [W/Kg]	2.70	2.78
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	39.0
Group UID	CW 0-	TSL Conductivity [S/m]	3.01
Conversion Factor	6.17	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7947 2025-09-10	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1798 2026-04-27	TSL Type	DAK 3.5 Head 19.73 deg.C 2026-Jun-22 09_13_30
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		74.5
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	7.27	7.45
psSAR10g [W/Kg]	2.70	2.79
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D3700V2 - SN1131

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	39.4
Group UID	CW 0-	TSL Conductivity [S/m]	2.90
Conversion Factor	6.17	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7947 2025-09-10	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1798 2026-04-27	TSL Type	DAK 3.5 Head 20.63 deg.C 2026-Jul-02 08_30_31
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		75.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.3

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.46	3.50
psSAR10g [W/Kg]	1.28	1.33
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D2600V2 - SN1006

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	42.1
Group UID	CW 0-	TSL Conductivity [S/m]	1.93
Conversion Factor	6.91	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7947 2025-09-10	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1798 2026-04-27	TSL Type	DAK 3.5 20.26 deg.C 2026-Jul-20 08_24_50
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.1
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.90	2.55
psSAR10g [W/Kg]	1.29	1.16
Power Drift [±0.2 dB]	N/A	-0.09



System Performance Check Report for D3500V2 - SN1182

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.9
Group UID	CW 0-	TSL Conductivity [S/m]	2.75
Conversion Factor	6.2	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7463 2026-04-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1917 2025-09-10	TSL Type	DAK 3.5 Head 19.6 deg.C 2026-Jun-16 11_02_00
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		77.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.72	5.81
psSAR10g [W/Kg]	2.21	2.26
Power Drift [±0.2 dB]	N/A	-0.08



System Performance Check Report for D3900V2 - SN1102

Exposure Conditions

Frequency [MHz] Channel Number	3900.000 0	TSL Permittivity	38.9
Group UID	CW 0-	TSL Conductivity [S/m]	3.12
Conversion Factor	5.97	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7463 2026-04-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1917 2025-09-10	TSL Type	DAKS 3.5 Head 19.97 deg.C 2026-Jul-16 07_40_02
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		83.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.74	3.19
psSAR10g [W/Kg]	1.03	1.22
Power Drift [±0.2 dB]	N/A	0.05



System Performance Check Report for D3500V2 - SN1011

Exposure Conditions

Frequency [MHz] Channel Number	3500.000 0	TSL Permittivity	39.7
Group UID	CW 0-	TSL Conductivity [S/m]	2.76
Conversion Factor	6.2	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7463 2026-04-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1917 2025-09-10	TSL Type	DAK 3.5 19.02 deg.C 2026-Jul-20 12_00_39
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		85.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.38	5.98
psSAR10g [W/Kg]	2.09	2.46
Power Drift [±0.2 dB]	N/A	0.04



System Performance Check Report for D3700V2 - SN1039

Exposure Conditions

Frequency [MHz] Channel Number	3700.000 0	TSL Permittivity	39.3
Group UID	CW 0-	TSL Conductivity [S/m]	2.96
Conversion Factor	6.08	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7463 2026-04-24	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1917 2025-09-10	TSL Type	DAK 3.5 19.02 deg.C 2026-Jul-20 12_00_39
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 29.1
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		85.2
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		8.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	5.43	6.22
psSAR10g [W/Kg]	2.07	2.50
Power Drift [±0.2 dB]	N/A	0.03



System Performance Check Report for D5GHzV2 - SN1138

Exposure Conditions

Frequency [MHz] Channel Number	5250.000 0	TSL Permittivity	37.3
Group UID	CW 0-	TSL Conductivity [S/m]	4.48
Conversion Factor	5.89	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7915 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1883 2025-10-15	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		65.8
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.72	3.97
psSAR10g [W/Kg]	1.08	1.18
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D1750V2 - SN1220

Exposure Conditions

Frequency [MHz] Channel Number	1750.000 0	TSL Permittivity	43.7
Group UID	CW 0-	TSL Conductivity [S/m]	1.24
Conversion Factor	8.25	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7915 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1883 2025-10-15	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		85.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		10.8

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	1.70	1.71
psSAR10g [W/Kg]	0.914	0.948
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D1900V2 - SN5d260

Exposure Conditions

Frequency [MHz] Channel Number	1900.000 0	TSL Permittivity	43.4
Group UID	CW 0-	TSL Conductivity [S/m]	1.32
Conversion Factor	7.82	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7915 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1883 2025-10-15	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		86.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.7

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	1.92	1.95
psSAR10g [W/Kg]	0.999	1.05
Power Drift [±0.2 dB]	N/A	0.05



System Performance Check Report for D5GHzV2 - SN1168

Exposure Conditions

Frequency [MHz] Channel Number	5250.000 0	TSL Permittivity	39.2
Group UID	CW 0-	TSL Conductivity [S/m]	4.64
Conversion Factor	5.89	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7915 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1883 2025-10-15	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		65.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.6

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.44	3.71
psSAR10g [W/Kg]	1.01	1.12
Power Drift [±0.2 dB]	N/A	-0.00



System Performance Check Report for D1750V2 - SN1220

Exposure Conditions

Frequency [MHz] Channel Number	1750.000 0	TSL Permittivity	43.7
Group UID	CW 0-	TSL Conductivity [S/m]	1.25
Conversion Factor	8.12	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7914 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1893 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		85.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		10.8

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	1.78	1.80
psSAR10g [W/Kg]	0.954	0.992
Power Drift [±0.2 dB]	N/A	-0.05



System Performance Check Report for D1900V2 - SN5d260

Exposure Conditions

Frequency [MHz] Channel Number	1900.000 0	TSL Permittivity	43.5
Group UID	CW 0-	TSL Conductivity [S/m]	1.33
Conversion Factor	7.68	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7914 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1893 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		84.3
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		10.8

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.03	2.03
psSAR10g [W/Kg]	1.07	1.09
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D5GHzV2 - SN1138

Exposure Conditions

Frequency [MHz] Channel Number	5250.000 0	TSL Permittivity	35.9
Group UID	CW 0-	TSL Conductivity [S/m]	4.55
Conversion Factor	5.78	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7914 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1893 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		62.7
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.2

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.72	4.17
psSAR10g [W/Kg]	1.06	1.20
Power Drift [±0.2 dB]	N/A	-0.15



System Performance Check Report for D5GHzV2 - SN1168

Exposure Conditions

Frequency [MHz] Channel Number	5250.000 0	TSL Permittivity	38.5
Group UID	CW 0-	TSL Conductivity [S/m]	4.59
Conversion Factor	5.78	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN7914 2025-11-07	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4ip Sn1893 2025-11-12	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		65.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.4

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	4.27	4.36
psSAR10g [W/Kg]	1.22	1.27
Power Drift [±0.2 dB]	N/A	-0.04



System Performance Check Report for D2600V2 - SN1036

Exposure Conditions

Frequency [MHz] Channel Number	2600.000 0	TSL Permittivity	41.6
Group UID	CW 0-	TSL Conductivity [S/m]	1.85
Conversion Factor	7.52	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3989 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1472 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		81.9
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.84	2.83
psSAR10g [W/Kg]	1.27	1.32
Power Drift [±0.2 dB]	N/A	0.01



System Performance Check Report for D5GHzV2 - SN1003

Exposure Conditions

Frequency [MHz] Channel Number	5600.000 0	TSL Permittivity	34.7
Group UID	CW 0-	TSL Conductivity [S/m]	4.74
Conversion Factor	5.65	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3989 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1472 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		63.6
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.2

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.87	4.21
psSAR10g [W/Kg]	1.12	1.20
Power Drift [±0.2 dB]	N/A	0.04



System Performance Check Report for D5GHzV2 - SN1168

Exposure Conditions

Frequency [MHz] Channel Number	5600.000 0	TSL Permittivity	36.7
Group UID	CW 0-	TSL Conductivity [S/m]	4.90
Conversion Factor	5.65	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3989 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1472 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		63.0
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.2

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	4.19	4.54
psSAR10g [W/Kg]	1.18	1.29
Power Drift [±0.2 dB]	N/A	-0.12



System Performance Check Report for D2450V2 - SN706

Exposure Conditions

Frequency [MHz] Channel Number	2450.000 0	TSL Permittivity	40.8
Group UID	CW 0-	TSL Conductivity [S/m]	1.70
Conversion Factor	7.62	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup


Probe Calibration Date	EX3DV4 - SN3989 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1472 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		82.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		9.0

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	2.36	2.46
psSAR10g [W/Kg]	1.09	1.18
Power Drift [±0.2 dB]	N/A	0.00



System Performance Check Report for D5GHzV2 - SN1138

Exposure Conditions

Frequency [MHz] Channel Number	5600.000 0	TSL Permittivity	35.6
Group UID	CW 0-	TSL Conductivity [S/m]	4.75
Conversion Factor	5.65	Phantom Section TSL	Flat Head Simulating Liquid

Hardware Setup

Probe Calibration Date	EX3DV4 - SN3989 2026-01-15	Phantom	Twin-SAM V8.0 (30deg probe tilt)
DAE Calibration Date	DAE4 Sn1472 2026-01-08	TSL Type	HBBL-600-10000
Software Version	17.0.4.7034		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	24.0 x 24.0 x 22.9
Grid Steps [mm]	10.0 x 10.0	4.0 x 4.0 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
M2/M1 [> 30%]		61.4
Dist 3dB Peak [> Grid Steps Δx, Δy mm]		7.2

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	3.48	3.99
psSAR10g [W/Kg]	0.996	1.11
Power Drift [±0.2 dB]	N/A	0.18

