

APPENDIX A: VERIFICATION PLOTS

ELEMENT

DUT: Dipole 835.000 MHz; Type: D835V2 - SN4d040

Communication System: UID: 0, CW; Frequency: 835.000 MHz

Medium: 835 Head; Medium parameters used:

$f = 835.000$ MHz; $\text{cond} = 0.920$ S/m; $\text{perm} = 40.8$; $\text{density} = 1000$ kg/m³

Phantom Section: Flat; Space: 15 mm

Test Date: 08/12/2025; Ambient Temp: 23.5°C; Tissue Temp: 22.1°C

Probe: EX3DV4 - SN7360; ConvF:(9.16,9.02,10.26); Calibrated: 2025-03-07

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn534; Calibrated: 2025-03-11

Phantom: Twin-SAM V8.0; Serial: 2034

Measurement SW: DASY Module SAR V16.4.0.5005

835.0 MHz System Verification at 23.0 dBm (200 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 2.91 W/kg

SAR(1 g) = 1.95 W/kg

Deviation (1 g) = 1.67%



ELEMENT

DUT: Dipole 835.000 MHz; Type: D835V2 - SN4d040

Communication System: UID: 0, CW; Frequency: 835.000 MHz

Medium: 835 Head; Medium parameters used:

$f = 835.000$ MHz; $\text{cond} = 0.930$ S/m; $\text{perm} = 42.0$; $\text{density} = 1000$ kg/m³

Phantom Section: Flat; Space: 15 mm

Test Date: 08/21/2025; Ambient Temp: 23.0°C; Tissue Temp: 21.7°C

Probe: EX3DV4 - SN7360; ConvF:(9.16,9.02,10.26); Calibrated: 2025-03-07

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn534; Calibrated: 2025-03-11

Phantom: Twin-SAM V8.0; Serial: 2034

Measurement SW: DASY Module SAR V16.4.0.5005

835.0 MHz System Verification at 23.0 dBm (200 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 3.00 W/kg

SAR(1 g) = 2.02 W/kg

Deviation (1 g) = 5.32%



ELEMENT

DUT: Dipole 1750.000 MHz; Type: D1750V2 - SN1104

Communication System: UID: 0, CW; Frequency: 1750.000 MHz
Medium: 1750 Head; Medium parameters used:
 $f = 1750.000$ MHz; $\text{cond} = 1.36$ S/m; $\text{perm} = 39.2$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/12/2025; Ambient Temp: 23.5°C; Tissue Temp: 22.1°C

Probe: EX3DV4 - SN7360; ConvF:(7.74,7.62,8.67); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

1750.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 6.54 W/kg

SAR(1 g) = 3.73 W/kg

Deviation (1 g) = 4.78%



ELEMENT

DUT: Dipole 1750.000 MHz; Type: D1750V2 - SN1104

Communication System: UID: 0, CW; Frequency: 1750.000 MHz
Medium: 1900 Head; Medium parameters used:
 $f = 1750.000$ MHz; $\text{cond} = 1.36$ S/m; $\text{perm} = 39.3$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/21/2025; Ambient Temp: 24.00°C; Tissue Temp: 23.3°C

Probe: EX3DV4 - SN7552; ConvF:(7.37,8.17,7.95); Calibrated: 2025-05-14
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1676; Calibrated: 2025-05-13
Phantom: Twin-SAM V8.0; Serial: 1944
Measurement SW: DASY Module SAR V16.4.0.5005

1750.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 6.60 W/kg

SAR(1 g) = 3.68 W/kg

Deviation (1 g) = 3.37%



ELEMENT

DUT: Dipole 1900.000 MHz; Type: D1900V2 - SN5d030

Communication System: UID: 0, CW; Frequency: 1900.000 MHz
Medium: 1900 Head; Medium parameters used:
 $f = 1900.000$ MHz; $\text{cond} = 1.45$ S/m; $\text{perm} = 39.5$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/18/2025; Ambient Temp: 23.5°C; Tissue Temp: 23.2°C

Probe: EX3DV4 - SN7552; ConvF:(7.12,7.89,7.68); Calibrated: 2025-05-14
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1676; Calibrated: 2025-05-13
Phantom: Twin-SAM V8.0; Serial: 1944
Measurement SW: DASY Module SAR V16.4.0.5005

1900.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 7.69 W/kg

SAR(1 g) = 4.17 W/kg

Deviation (1 g) = 5.30%



ELEMENT

DUT: Dipole 1900.000 MHz; Type: D1900V2 - SN5d131

Communication System: UID: 0, CW; Frequency: 1900.000 MHz
Medium: 835 Head; Medium parameters used:
 $f = 1900.000$ MHz; $\text{cond} = 1.45$ S/m; $\text{perm} = 40.0$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/21/2025; Ambient Temp: 23.0°C; Tissue Temp: 21.7°C

Probe: EX3DV4 - SN7360; ConvF:(7.34,7.23,8.22); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

1900.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 90.0): Measurement grid: $dx=10.0$ mm, $dy=15.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=6.0$ mm, $dy=6.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 7.40 W/kg

SAR(1 g) = 4.17 W/kg

Deviation (1 g) = 4.77%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1042

Communication System: UID: 0, CW; Frequency: 2600.000 MHz

Medium: 750 Head; Medium parameters used:

$f = 2600.000$ MHz; $\text{cond} = 1.94$ S/m; $\text{perm} = 39.8$; $\text{density} = 1000$ kg/m³

Phantom Section: Flat; Space: 10 mm

Test Date: 08/27/2025; Ambient Temp: 21.5°C; Tissue Temp: 21.0°C

Probe: EX3DV4 - SN7357; ConvF:(7.25,6.45,7.66); Calibrated: 2025-04-11

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn1582; Calibrated: 2025-04-10

Phantom: Twin-SAM V8.0; Serial: 1736

Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 10.9 W/kg

SAR(1 g) = 5.30 W/kg

Deviation (1 g) = -3.46%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1068

Communication System: UID: 0, CW; Frequency: 2600.000 MHz
Medium: 2450 Head; Medium parameters used:
 $f = 2600.000$ MHz; $\text{cond} = 2.03$ S/m; $\text{perm} = 37.8$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/27/2025; Ambient Temp: 22.5°C; Tissue Temp: 22.7°C

Probe: EX3DV4 - SN3746; ConvF:(6.68,6.4,6.93); Calibrated: 2024-10-15
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1237; Calibrated: 2024-10-10
Phantom: Twin-SAM V4.0; Serial: 1598
Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 13.7 W/kg

SAR(1 g) = 6.02 W/kg

Deviation (1 g) = 6.55%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1068

Communication System: UID: 0, CW; Frequency: 2600.000 MHz
Medium: 2450 Head; Medium parameters used:
 $f = 2600.000$ MHz; $\text{cond} = 1.94$ S/m; $\text{perm} = 37.3$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/28/2025; Ambient Temp: 20.8°C; Tissue Temp: 21.1°C

Probe: EX3DV4 - SN7427; ConvF:(6.52,7.03,7.04); Calibrated: 2025-02-10
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1408; Calibrated: 2024-10-10
Phantom: Twin-SAM V8.0; Serial: 1935
Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 11.1 W/kg

SAR(1 g) = 5.48 W/kg

Deviation (1 g) = -3.01%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1068

Communication System: UID: 0, CW; Frequency: 2600.000 MHz
Medium: 2450 Head; Medium parameters used:
 $f = 2600.000$ MHz; $\text{cond} = 2.03$ S/m; $\text{perm} = 38.3$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/29/2025; Ambient Temp: 22.5°C; Tissue Temp: 23.6°C

Probe: EX3DV4 - SN7532; ConvF:(6.78,6.7,7.52); Calibrated: 2025-04-16
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn501; Calibrated: 2025-04-09
Phantom: Twin-SAM V8.0; Serial: 2058
Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 13.2 W/kg

SAR(1 g) = 5.66 W/kg

Deviation (1 g) = 0.18%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1042

Communication System: UID: 0, CW; Frequency: 2600.000 MHz
Medium: 2450 Head; Medium parameters used:
 $f = 2600.000$ MHz; $\text{cond} = 1.89$ S/m; $\text{perm} = 39.8$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/29/2025; Ambient Temp: 20.0°C; Tissue Temp: 19.1°C

Probe: EX3DV4 - SN7360; ConvF:(6.89,6.79,7.72); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 11.6 W/kg

SAR(1 g) = 5.65 W/kg

Deviation (1 g) = 2.91%



ELEMENT

DUT: Dipole 2600.000 MHz; Type: D2600V2 - SN1068

Communication System: UID: 0, CW; Frequency: 2600.000 MHz
Medium: 2450 Head; Medium parameters used:
 $f = 2600.000$ MHz; $\text{cond} = 1.90$ S/m; $\text{perm} = 40.8$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

09/02/2025; Ambient Temp: 22.5°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN7552; ConvF:(6.68,7.41,7.21); Calibrated: 2025-05-14
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1676; Calibrated: 2025-05-13
Phantom: Twin-SAM V8.0; Serial: 1944
Measurement SW: DASY Module SAR V16.4.0.5005

2600.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 11.1 W/kg

SAR(1 g) = 5.55 W/kg

Deviation (1 g) = -1.77%



ELEMENT

DUT: Dipole 3500.000 MHz; Type: D3500V2 - SN1126

Communication System: UID: 0, CW; Frequency: 3500.000 MHz
Medium: 3600 Head; Medium parameters used:
 $f = 3500.000$ MHz; $\text{cond} = 2.80$ S/m; $\text{perm} = 37.7$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/15/2025; Ambient Temp: 23.0°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN7360; ConvF:(6.75,6.65,7.56); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

3500.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 17.3 W/kg

SAR(1 g) = 6.76 W/kg

Deviation (1 g) = 2.11%



ELEMENT

DUT: Dipole 3500.000 MHz; Type: D3500V2 - SN1126

Communication System: UID: 0, CW; Frequency: 3500.000 MHz
Medium: 3600 Head; Medium parameters used:
 $f = 3500.000$ MHz; $\text{cond} = 2.79$ S/m; $\text{perm} = 37.6$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/19/2025; Ambient Temp: 22.1°C; Tissue Temp: 20.5°C

Probe: EX3DV4 - SN7360; ConvF:(6.75,6.65,7.56); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

3500.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 17.4 W/kg

SAR(1 g) = 6.89 W/kg

Deviation (1 g) = 4.08%



ELEMENT

DUT: Dipole 3500.000 MHz; Type: D3500V2 - SN1126

Communication System: UID: 0, CW; Frequency: 3500.000 MHz
Medium: 3600 Head; Medium parameters used:
f = 3500.000 MHz; cond = 2.79 S/m; perm = 38.1; density = 1000 kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/25/2025; Ambient Temp: 19.5°C; Tissue Temp: 19.0°C

Probe: EX3DV4 - SN7682; ConvF:(6.74,6.74,6.95); Calibrated 2025-05-14
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1683; Calibrated: 2025-05-13
Phantom: Twin-SAM V8.0; Serial: 1357
Measurement SW: DASY Module SAR V16.4.0.5005

3500.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: dx=10.0 mm, dy=10.0 mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: dx=5.0 mm, dy=5.0 mm, dz=1.4 mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 17.5 W/kg

SAR(1 g) = 6.95 W/kg

Deviation (1 g) = 4.98%



ELEMENT

DUT: Dipole 3500.000 MHz; Type: D3500V2 - SN1126

Communication System: UID: 0, CW; Frequency: 3500.000 MHz
Medium: 3600 Head; Medium parameters used:
 $f = 3500.000$ MHz; $\text{cond} = 2.80$ S/m; $\text{perm} = 39.2$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/28/2025; Ambient Temp: 20.8°C; Tissue Temp: 20.0°C

Probe: EX3DV4 - SN7682; ConvF:(6.74,6.74,6.95); Calibrated: 2025-05-14
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn1683; Calibrated: 2025-05-13
Phantom: Twin-SAM V8.0; Serial: 1357
Measurement SW: DASY Module SAR V16.4.0.5005

3500.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 17.4 W/kg

SAR(1 g) = 6.79 W/kg

Deviation (1 g) = 2.57%



ELEMENT

DUT: Dipole 3700.000 MHz; Type: D3700V2 - SN1002

Communication System: UID: 0, CW; Frequency: 3700.000 MHz
Medium: 3600 Head; Medium parameters used:
 $f = 3700.000$ MHz; $\text{cond} = 2.99$ S/m; $\text{perm} = 37.3$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/15/2025; Ambient Temp: 23.0°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN7360; ConvF:(6.62,6.52,7.42); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

3700.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 18.3 W/kg

SAR(1 g) = 6.87 W/kg

Deviation (1 g) = 1.18%



ELEMENT

DUT: Dipole 3700.000 MHz; Type: D3700V2 - SN1002

Communication System: UID: 0, CW; Frequency: 3700.000 MHz
Medium: 3600 Head; Medium parameters used:
 $f = 3700.000$ MHz; $\text{cond} = 2.98$ S/m; $\text{perm} = 37.3$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 08/19/2025; Ambient Temp: 22.1°C; Tissue Temp: 20.5°C

Probe: EX3DV4 - SN7360; ConvF:(6.62,6.52,7.42); Calibrated: 2025-03-07
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn534; Calibrated: 2025-03-11
Phantom: Twin-SAM V8.0; Serial: 2034
Measurement SW: DASY Module SAR V16.4.0.5005

3700.0 MHz System Verification at 20.0 dBm (100 mW)

Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 29.1): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 17.8 W/kg

SAR(1 g) = 6.81 W/kg

Deviation (1 g) = 0.29%

