APPENDIX B: TISSUE SPECIFICATIONS

Measurement Procedure for Tissue verification:

1) The network analyzer and probe system were configured and calibrated.
2) The probe was immersed in the tissue. The tissue was placed in a nonmetallic container. Trapped air bubbles beneath the flange were minimized by placing the probe at a slight angle.
3) The complex admittance with respect to the probe aperture was measured
4) The complex relative permittivity ε' can be calculated from the below equation (Pournaropoulos and Misra):

$$Y = \frac{j2\omega\varepsilon_r\varepsilon_0}{[\ln(b/a)]^2}\int_a^b\int_a^b\int_0^\pi \cos\phi' \frac{\exp\left[-j\omega r(\mu_0\varepsilon_r'\varepsilon_0)^{1/2}\right]}{r} d\phi' d\rho' d\rho$$

where Y is the admittance of the probe in contact with the sample, the primed and unprimed coordinates refer to source and observation points, respectively, $r^2 = \rho^2 + \rho'^2 - 2\rho\rho'\cos\phi'$, ω is the angular frequency, and $j = \sqrt{-1}$.

3 Composition / Information on ingredients

3.2 Mixtures
Description: Aqueous solution with surfactants and inhibitors
Declarable, or hazardous components:

CAS: 107-21-1 EINECS: 203-473-3 Reg.nr.: 01-2119456816-28-0000	**Ethanediol** STOT RE 2, H373; Acute Tox. 4, H302	>1.0-4.9%
CAS: 68608-26-4 EINECS: 271-781-5 Reg.nr.: 01-2119527859-22-0000	**Sodium petroleum sulfonate** Eye Irrit. 2, H319	< 2.9%
CAS: 107-41-5 EINECS: 203-489-0 Reg.nr.: 01-2119539582-35-0000	**Hexylene Glycol / 2-Methyl-pentane-2,4-diol** Skin Irrit. 2, H315; Eye Irrit. 2, H319	< 2.9%
CAS: 68920-66-1 NLP: 500-236-9 Reg.nr.: 01-2119489407-26-0000	**Alkoxylated alcohol, > C_{16}** Aquatic Chronic 2, H411; Skin Irrit. 2, H315; Eye Irrit. 2, H319	< 2.0%

Additional information:
For the wording of the listed risk phrases refer to section 16.
Not mentioned CAS-, EINECS- or registration numbers are to be regarded as Proprietary/Confidential.
The specific chemical identity and/or exact percentage concentration of proprietary components is withheld as a trade secret.

Figure B-1

Note: Liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

Unless otherwise specified, no part of this report may be reproduced or utilized in any part, form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from Element Materials Technology. If you have any questions about this or have an enquiry about obtaining additional rights to this report or assembly of contents thereof, please contact ct.info@element.com.

Schmid & Partner Engineering AG

s p e a g

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 44 245 9700, Fax +41 44 245 9779
www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Item Name	**Head Tissue Simulating Liquid (HBBL600-10000V6)**
Product No.	SL AAH U16 CA (Batch: 250317-1)
Manufacturer	SPEAG

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

Test Condition

Ambient Condition	22°C ; 30% humidity
TSL Temperature	22°C
Test Date	20-Mar-25
Operator	CL

Additional Information

TSL Density
TSL Heat-capacity

Results

	Measured			Target		Diff.to Target [%]	
f [MHz]	e'	e"	sigma	eps	sigma	Δ-eps	Δ-sigma
600	45.3	24.6	0.82	42.7	0.88	6.0	-7.0
750	44.7	20.9	0.87	41.9	0.89	6.6	-2.6
800	44.6	20.0	0.89	41.7	0.90	7.0	-0.8
825	44.5	19.8	0.90	41.6	0.91	7.0	-0.7
835	44.5	19.4	0.91	41.5	0.91	7.1	-0.2
850	44.4	19.2	0.91	41.5	0.92	7.0	-0.7
900	44.3	18.5	0.93	41.5	0.97	6.7	-4.1
1400	43.4	14.7	1.14	40.6	1.18	6.9	-3.3
1450	43.3	14.5	1.17	40.5	1.20	6.9	-2.5
1500	43.1	14.0	1.25	40.3	1.28	6.9	-2.7
1625	43.1	13.9	1.26	40.3	1.30	7.0	-3.0
1640	43.1	13.9	1.27	40.3	1.31	7.1	-2.8
1650	43.0	13.9	1.27	40.2	1.31	6.9	-3.3
1700	42.9	13.6	1.30	40.2	1.34	6.8	-3.1
1750	42.8	13.7	1.33	40.1	1.37	6.8	-3.0
1800	42.8	13.6	1.36	40.0	1.40	7.0	-2.9
1810	42.8	13.6	1.37	40.0	1.40	7.0	-2.1
1825	42.8	13.6	1.38	40.0	1.40	7.0	-1.4
1850	42.7	13.5	1.39	40.0	1.40	6.8	-0.7
1900	42.6	13.4	1.42	40.0	1.40	6.5	1.4
1950	42.5	13.4	1.45	40.0	1.40	6.3	3.6
2000	42.5	13.4	1.49	40.0	1.40	6.3	6.4
2050	42.4	13.3	1.52	39.9	1.44	6.2	5.2
2100	42.3	13.3	1.55	39.8	1.49	6.2	4.1
2150	42.3	13.3	1.59	39.7	1.53	6.5	3.7
2200	42.2	13.3	1.62	39.6	1.58	6.4	2.7
2250	42.1	13.3	1.66	39.6	1.62	6.4	2.3
2300	42.0	13.3	1.70	39.5	1.67	6.4	2.0
2350	42.0	13.3	1.74	39.4	1.71	6.7	1.7
2400	41.9	13.3	1.77	39.3	1.76	6.6	0.8
2450	41.8	13.3	1.81	39.2	1.80	6.6	0.6
2500	41.8	13.3	1.85	39.1	1.85	6.8	-0.2
2550	41.7	13.3	1.89	39.1	1.91	6.7	-1.0
2600	41.6	13.4	1.93	39.0	1.96	6.6	-1.7
3500	40.0	14.0	2.72	37.9	2.91	5.5	-6.5
3700	39.7	14.1	2.91	37.7	3.12	5.4	-6.6
5200	37.2	15.8	4.56	36.0	4.66	3.5	-1.6
5250	37.1	15.9	4.63	35.9	4.71	3.3	-1.6
5300	37.0	15.9	4.69	35.9	4.76	3.2	-1.5
5500	36.6	16.1	4.93	35.6	4.96	2.8	-0.7
5600	36.4	16.2	5.05	35.5	5.07	2.3	-0.1
5700	36.2	16.4	5.19	35.4	5.17	2.1	0.4
5800	36.0	16.4	5.30	35.3	5.27	2.0	0.6
6000	35.8	16.5	5.51	35.1	5.48	2.2	0.6
6500	34.8	17.1	6.20	34.5	6.07	1.1	2.1
7000	34.0	17.5	6.81	33.9	6.65	0.3	2.5
7500	33.1	17.8	7.43	33.3	7.24	-0.6	2.6
8000	32.2	18.1	8.04	32.7	7.84	-1.5	2.6
8500	31.4	18.3	8.67	32.1	8.45	-2.4	2.6
9000	30.5	18.6	9.31	31.5	9.06	-3.3	2.6
9500	29.6	18.8	9.94	31.0	9.71	-4.3	2.3
10000	28.8	19.0	10.56	30.4	10.36	-5.2	1.9





Figure B-2
600 – 10000 MHz Head Tissue Equivalent Matter

Unless otherwise specified, no part of this report may be reproduced or utilized in any part, form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from Element Materials Technology. If you have any questions about this or have an enquiry about obtaining additional rights to this report or assembly of contents thereof, please contact ct.info@element.com.