

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880 MHz; ($\sigma = 1.39\text{mho/m}$; $\epsilon_r = 39$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.58, 4.58, 4.58); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.543mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 4.56 V/m; Power Drift = -0.133 dB

Peak SAR (extrapolated) = 0.862 W/kg

SAR(1 g) = 0.496mW/g; SAR(10 g) = 0.274mW/g

Maximum value of SAR (measured) = 0.548mW/g



0 dB = 0.548mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880 MHz; ($\sigma = 1.39\text{mho/m}$; $\epsilon_r = 39$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.58, 4.58, 4.58); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Tilt, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.140mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.63 V/m; Power Drift = -0.095 dB

Peak SAR (extrapolated) = 0.208 W/kg

SAR(1 g) = 0.123mW/g; SAR(10 g) = 0.073mW/g

Maximum value of SAR (measured) = 0.126mW/g



0 dB = 0.126mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:2.075

Medium: Body 836.6 MHz; ($\sigma = 0.982$ mho/m; $\epsilon_r = 53.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/26/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(5.5, 5.5, 5.5); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Front, Fixed Ant., Standard Battery

Area Scan (71x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.253mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.47 V/m; Power Drift = 0.073 dB

Peak SAR (extrapolated) = 0.328 W/kg

SAR(1 g) = 0.239mW/g; SAR(10 g) = 0.172mW/g

Maximum value of SAR (measured) = 0.253mW/g



0 dB = 0.253mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:2.075

Medium: Body 836.6 MHz; ($\sigma = 0.982$ mho/m; $\epsilon_r = 53.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/26/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(5.5, 5.5, 5.5); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Fixed Ant., Standard Battery

Area Scan (71x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.397mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.03 V/m; Power Drift = -0.023 dB

Peak SAR (extrapolated) = 0.532 W/kg

SAR(1 g) = 0.377mW/g; SAR(10 g) = 0.267mW/g

Maximum value of SAR (measured) = 0.400mW/g



0 dB = 0.400mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Body 836.6 MHz; ($\sigma = 0.982$ mho/m; $\epsilon_r = 53.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/26/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(5.5, 5.5, 5.5); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Fixed Ant., Standard Battery

Area Scan (71x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.331mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.41 V/m; Power Drift = 0.001 dB

Peak SAR (extrapolated) = 0.449 W/kg

SAR(1 g) = 0.317mW/g; SAR(10 g) = 0.223mW/g

Maximum value of SAR (measured) = 0.337mW/g



0 dB = 0.377mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:4.15

Medium: Body 836.6 MHz; ($\sigma = 0.982$ mho/m; $\epsilon_r = 53.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/26/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(5.5, 5.5, 5.5); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Fixed Ant., Standard Battery

Area Scan (71x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.480mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.77 V/m; Power Drift = -0.051 dB

Peak SAR (extrapolated) = 0.646 W/kg

SAR(1 g) = 0.460mW/g; SAR(10 g) = 0.326mW/g

Maximum value of SAR (measured) = 0.486mW/g



0 dB = 0.486mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:2.77

Medium: Body 836.6 MHz; ($\sigma = 0.982$ mho/m; $\epsilon_r = 53.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/26/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(5.5, 5.5, 5.5); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Fixed Ant., Standard Battery

Area Scan (71x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.466 mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.65 V/m; Power Drift = -0.093 dB

Peak SAR (extrapolated) = 0.618 W/kg

SAR(1 g) = 0.441mW/g; SAR(10 g) = 0.312mW/g

Maximum value of SAR (measured) = 0.468mW/g



0 dB = 0.468mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:2.075

Medium: Body 1880 MHz; ($\sigma = 1.51$ mho/m; $\epsilon_r = 53$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.33, 4.33, 4.33); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Fixed Ant., Standard Battery

Area Scan (81x121x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.228mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.34 V/m; Power Drift = 0.022 dB

Peak SAR (extrapolated) = 0.351 W/kg

SAR(1 g) = 0.210mW/g; SAR(10 g) = 0.127mW/g

Maximum value of SAR (measured) = 0.226mW/g



0 dB = 0.226mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:2.075

Medium: Body 1880 MHz; ($\sigma = 1.51$ mho/m; $\epsilon_r = 53$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.33, 4.33, 4.33); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Rear, Fixed Ant., Standard Battery

Area Scan (81x121x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.216mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.61 V/m; Power Drift = -0.071 dB

Peak SAR (extrapolated) = 0.327 W/kg

SAR(1 g) = 0.200mW/g; SAR(10 g) = 0.123mW/g

Maximum value of SAR (measured) = 0.216mW/g



0 dB = 0.216mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Body 1880 MHz; ($\sigma = 1.51$ mho/m; $\epsilon_r = 53$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.33, 4.33, 4.33); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Fixed Ant., Standard Battery

Area Scan (81x121x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.118mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 3.46 V/m; Power Drift = 0.081 dB

Peak SAR (extrapolated) = 0.186 W/kg

SAR(1 g) = 0.111mW/g; SAR(10 g) = 0.068mW/g

Maximum value of SAR (measured) = 0.118mW/g



0 dB = 0.118mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: Body 1880 MHz; ($\sigma = 1.51 \text{ mho/m}$; $\epsilon_r = 53$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.33, 4.33, 4.33); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Fixed Ant., Standard Battery

Area Scan (81x121x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.174mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 4.21 V/m; Power Drift = 0.035 dB

Peak SAR (extrapolated) = 0.269 W/kg

SAR(1 g) = 0.162mW/g; SAR(10 g) = 0.098mW/g

Maximum value of SAR (measured) = 0.172mW/g



0 dB = 0.172mW/g

LG Electronics Inc.

DUT: GM600; Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:2.77

Medium: Body 1880 MHz; ($\sigma = 1.51$ mho/m; $\epsilon_r = 53$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/25/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ES3DV3 - SN3066; ConvF(4.33, 4.33, 4.33); Calibrated: 2009-02-16

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn645; Calibrated: 2009-11-19

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Fixed Ant., Standard Battery

Area Scan (81x121x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.262mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.23 V/m; Power Drift = 0.056 dB

Peak SAR (extrapolated) = 0.413 W/kg

SAR(1 g) = 0.244mW/g; SAR(10 g) = 0.148mW/g

Maximum value of SAR (measured) = 0.260mW/g



0 dB = 0.260mW/g

APPENDIX C: Calibration Certificates

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **ES3-3066_Feb09**

CALIBRATION CERTIFICATE

Object **ES3DV3 - SN:3066**

Calibration procedure(s) **QA CAL-01.v6 and QA CAL-23.v3
 Calibration procedure for dosimetric E-field probes**

Calibration date: **February 16, 2009**

Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^\circ\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter E4419B	GB41293874	1-Apr-08 (No. 217-00788)	Apr-09
Power sensor E4412A	MY41495277	1-Apr-08 (No. 217-00788)	Apr-09
Power sensor E4412A	MY41498087	1-Apr-08 (No. 217-00788)	Apr-09
Reference 3 dB Attenuator	SN: S5054 (3c)	1-Jul-08 (No. 217-00865)	Jul-09
Reference 20 dB Attenuator	SN: S5086 (20b)	31-Mar-08 (No. 217-00787)	Apr-09
Reference 30 dB Attenuator	SN: S5129 (30b)	1-Jul-08 (No. 217-00866)	Jul-09
Reference Probe ES3DV2	SN: 3013	2-Jan-09 (No. ES3-3013_Jan09)	Jan-10
DAE4	SN: 660	9-Sep-08 (No. DAE4-660_Sep08)	Sep-09

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
RF generator HP 8648C	US3642U01700	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585	18-Oct-01 (in house check Oct-08)	In house check: Oct-09

	Name	Function	Signature
Calibrated by:	Katja Pokovic	Technical Manager	
Approved by:	Niels Kuster	Quality Manager	

Issued: February 16, 2009

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: ES3-3066_Feb09

Page 1 of 9

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
Polarization ϕ	ϕ rotation around probe axis
Polarization ϑ	ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}:** Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E^2 -field uncertainty inside TSL (see below *ConvF*).
- NORM(f)_{x,y,z} = NORM_{x,y,z} * frequency_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCP_{x,y,z}:** DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters:** Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy):** in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset:** The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

ES3DV3 SN:3066

February 16, 2009

Probe ES3DV3

SN:3066

Manufactured:	December 14, 2004
Last calibrated:	March 21, 2008
Recalibrated:	February 16, 2009

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ES3-3066_Feb09

Page 3 of 9

ES3DV3 SN:3066

February 16, 2009

DASY - Parameters of Probe: ES3DV3 SN:3066Sensitivity in Free Space^A

NormX	1.19 ± 10.1%	$\mu V/(V/m)^2$
NormY	1.34 ± 10.1%	$\mu V/(V/m)^2$
NormZ	1.19 ± 10.1%	$\mu V/(V/m)^2$

Diode Compression^B

DCP X	95 mV
DCP Y	92 mV
DCP Z	93 mV

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL 900 MHz Typical SAR gradient: 5 % per mm

Sensor Center to Phantom Surface Distance	3.0 mm	4.0 mm
SAR _{be} [%] Without Correction Algorithm	9.2	5.4
SAR _{be} [%] With Correction Algorithm	0.8	0.4

TSL 1810 MHz Typical SAR gradient: 10 % per mm

Sensor Center to Phantom Surface Distance	3.0 mm	4.0 mm
SAR _{be} [%] Without Correction Algorithm	10.9	6.7
SAR _{be} [%] With Correction Algorithm	0.9	0.5

Sensor Offset

Probe Tip to Sensor Center 2.0 mm


The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).^B Numerical linearization parameter: uncertainty not required.

ES3DV3 SN:3066

February 16, 2009

Frequency Response of E-Field
(TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ ($k=2$)

ES3DV3 SN:3066

February 16, 2009

Receiving Pattern (ϕ), $\vartheta = 0^\circ$




Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)

ES3DV3 SN:3066

February 16, 2009

Dynamic Range f(SAR_{head})
(Waveguide R22, f = 1800 MHz)





Uncertainty of Linearity Assessment: ± 0.6% (k=2)

ES3DV3 SN:3066

February 16, 2009

Conversion Factor Assessment



f [MHz]	Validity [MHz] ^c	TSL	Permittivity	Conductivity	Alpha	Depth	ConvF Uncertainty
900	± 50 / ± 100	Head	41.5 ± 5%	0.97 ± 5%	0.69	1.16	5.57 ± 11.0% (k=2)
1810	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.51	1.53	4.72 ± 11.0% (k=2)
1950	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.54	1.40	4.54 ± 11.0% (k=2)
2450	± 50 / ± 100	Head	39.2 ± 5%	1.80 ± 5%	0.42	1.71	4.20 ± 11.0% (k=2)
900	± 50 / ± 100	Body	55.0 ± 5%	1.05 ± 5%	0.94	1.10	5.53 ± 11.0% (k=2)
1810	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.42	1.67	4.53 ± 11.0% (k=2)
1950	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.31	1.98	4.49 ± 11.0% (k=2)
2450	± 50 / ± 100	Body	52.7 ± 5%	1.95 ± 5%	0.66	1.23	3.87 ± 11.0% (k=2)

^c The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


Certificate No: ES3-3066_Feb09

Page 8 of 9

ES3DV3 SN:3066

February 16, 2009

Deviation from Isotropy in HSL
Error (ϕ , θ), $f = 900 \text{ MHz}$



Uncertainty of Spherical Isotropy Assessment: $\pm 2.6\%$ ($k=2$)

Schmid & Partner Engineering AG

s p e a g

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 44 245 9700, Fax +41 44 245 9779
info@speag.com, <http://www.speag.com>

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ES3DV3

Serial Number:

3066

Place of Assessment:

Zurich

Date of Assessment:


April 27, 2009

Probe Calibration Date:

February 16, 2009

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1810 MHz.

Assessed by:



ES3DV3-SN:3066

Page 1 of 2

April 27, 2009

Schmid & Partner Engineering AG

s p e a g

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 44 245 9700, Fax +41 44 245 9779
info@speag.com, <http://www.speag.com>

Dosimetric E-Field Probe ES3DV3 - SN:3066

Conversion factor ($\pm$ standard deviation)

835 $\pm$ 50 MHz	ConvF	5.72 $\pm$ 7%	$\epsilon_r = 41.5 \pm 5\%$ $\sigma = 0.90 \pm 5\%$ mho/m (head tissue)
1900 $\pm$ 50 MHz	ConvF	4.58 $\pm$ 7%	$\epsilon_r = 40.0 \pm 5\%$ $\sigma = 1.40 \pm 5\%$ mho/m (head tissue)
835 $\pm$ 50 MHz	ConvF	5.50 $\pm$ 7%	$\epsilon_r = 55.2 \pm 5\%$ $\sigma = 0.97 \pm 5\%$ mho/m (body tissue)
1900 $\pm$ 50 MHz	ConvF	4.33 $\pm$ 7%	$\epsilon_r = 53.3 \pm 5\%$ $\sigma = 1.52 \pm 5\%$ mho/m (body tissue)

Important Note:

For numerically assessed probe conversion factors, parameters Alpha and Delta in the DASY software must have the following entries: Alpha = 0 and Delta = 1.

Please see also Section 4.7 of the DASY4 Manual.

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D835V2-471_Jan09**

CALIBRATION CERTIFICATE

Object: **D835V2 - SN: 471**

Calibration procedure(s): **QA CAL-05.v7**
Calibration procedure for dipole validation kits

Calibration date: **January 19, 2009**


Condition of the calibrated item: **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	08-Oct-08 (No. 217-00898)	Oct-09
Power sensor HP 8481A	US37292783	08-Oct-08 (No. 217-00898)	Oct-09
Reference 20 dB Attenuator	SN: 5086 (20g)	01-Jul-08 (No. 217-00864)	Jul-09
Type-N mismatch combination	SN: 5047.2 / 06327	01-Jul-08 (No. 217-00867)	Jul-09
Reference Probe ES3DV2	SN: 3025	28-Apr-08 (No. ES3-3025_Apr08)	Apr-09
DAE4	SN: 601	14-Mar-08 (No. DAE4-601_Mar08)	Mar-09
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-07)	In house check: Oct-09
RF generator R&S SMT-06	100005	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (in house check Oct-08)	In house check: Oct-09

Calibrated by:	Name Jeton Kastrati	Function Laboratory Technician	Signature 
Approved by:	Name Katja Pokovic	Function Technical Manager	Signature 

Issued: January 20, 2009

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D835V2-471_Jan09

Page 1 of 6

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM x,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V5.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	41.3 $\pm$ 6 %	0.91 mho/m $\pm$ 6 %
Head TSL temperature during test	(21.5 $\pm$ 0.2) °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.44 mW / g
SAR normalized	normalized to 1W	9.76 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	9.66 mW / g $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.61 mW / g
SAR normalized	normalized to 1W	6.44 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	6.39 mW / g $\pm$ 16.5 % (k=2)

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.7 Ω - 1.7 j Ω
Return Loss	- 32.5 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.386 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	November 15, 2002

DASY5 Validation Report for Head TSL

Date/Time: 19.01.2009 11:07:17

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:471

Communication System: CW-835; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz

Medium parameters used: $f = 835$ MHz; $\sigma = 0.91$ mho/m; $\epsilon_r = 41.1$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC)

DASY5 Configuration:

- Probe: ES3DV2 - SN3025; ConvF(5.97, 5.97, 5.97); Calibrated: 28.04.2008
- Sensor-Surface: 3.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 14.03.2008
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- Measurement SW: DASY5, V5.0 Build 120; SEMCAD X Version 13.4 Build 45


Pin=250mW; dip=15mm; dist=3.4mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 56.2 V/m; Power Drift = 0.026 dB

Peak SAR (extrapolated) = 3.61 W/kg

SAR(1 g) = 2.44 mW/g; SAR(10 g) = 1.61 mW/g

Maximum value of SAR (measured) = 2.75 mW/g




0 dB = 2.75mW/g

Certificate No: D835V2-471_Jan09

Page 5 of 6

Impedance Measurement Plot for Head TSL



Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D1900V2-5d017_Jul09**

CALIBRATION CERTIFICATE

Object **D1900V2 - SN: 5d017**

Calibration procedure(s) **QA CAL-05.v7**
Calibration procedure for dipole validation kits

Calibration date: **July 20, 2009**

Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	08-Oct-08 (No. 217-00898)	Oct-09
Power sensor HP 8481A	US37292783	08-Oct-08 (No. 217-00898)	Oct-09
Reference 20 dB Attenuator	SN: 5086 (20g)	31-Mar-09 (No. 217-01025)	Mar-10
Type-N mismatch combination	SN: 5047.2 / 06327	31-Mar-09 (No. 217-01029)	Mar-10
Reference Probe ES3DV2	SN: 3025	30-Apr-09 (No. ES3-3025_Apr09)	Apr-10
DAE4	SN: 601	07-Mar-09 (No. DAE4-601_Mar09)	Mar-10
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-07)	In house check: Oct-09
RF generator R&S SMT-06	100005	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (in house check Oct-08)	In house check: Oct-09

Calibrated by: **Claudio Leubler** **Laboratory Technician**

Approved by: **Katja Pokovic** **Technical Manager**

Signature

Issued: July 22, 2009

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D1900V2-5d017_Jul09

Page 1 of 6