

APPENDIX A: Validation Test Data




Figure 1 835 MHz Dipole Validation Test Setup



Figure 2 1900 MHz Dipole Validation Test Setup

LG Electronics Inc.

DUT: Dipole 835MHz;Type: D835V2;Serial: 471

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1
Medium: Head 835 MHz;($\sigma = 0.907$ mho/m; $\epsilon_r = 42.8$; $\rho = 1000$ kg/m³)
Phantom section: Flat Section


Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27
Sensor-Surface: 4mm (Mechanical Surface Detection)
Electronics: DAE4 Sn534; Calibrated: 2010-02-04
Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117
Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

835 MHz Dipole Validation

Area Scan (71x71x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 2.55mW/g

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm
Reference Value = 56.2 V/m; Power Drift = -0.055 dB
Peak SAR (extrapolated) = 3.36 W/kg
SAR(1 g) = 2.32mW/g; SAR(10 g) = 1.53mW/g
Maximum value of SAR (measured) = 2.50mW/g



LG Electronics Inc.

DUT: Dipole 1900MHz; Type: D1900V2; Serial: 5d017

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1
Medium: Head 1900 MHz; ($\sigma = 1.44\text{mho/m}$; $\epsilon_r = 39$; $\rho = 1000\text{ kg/m}^3$)
Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

1900 MHz Dipole Validation

Area Scan (71x71x1): Measurement grid: $dx=15\text{mm}$, $dy=15\text{mm}$

Maximum value of SAR (interpolated) = 12.3mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: $dx=8\text{mm}$, $dy=8\text{mm}$, $dz=5\text{mm}$

Reference Value = 93.5 V/m ; Power Drift = -0.086 dB

Peak SAR (extrapolated) = 16.8 W/kg

SAR(1 g) = 10.1mW/g ; SAR(10 g) = 5.45mW/g

Maximum value of SAR (measured) = 11.5mW/g



0 dB = 11.5mW/g

LG Electronics Inc.

DUT: Dipole 835MHz;Type: D835V2;Serial: 471

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1
Medium: Head 835 MHz;($\sigma = 0.907$ mho/m; $\epsilon_r = 42.8$; $\rho = 1000$ kg/m³)
Phantom section: Flat Section


Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27
Sensor-Surface: 4mm (Mechanical Surface Detection)
Electronics: DAE4 Sn534; Calibrated: 2010-02-04
Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117
Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

835 MHz Dipole Validation

Area Scan (71x71x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 2.55mW/g

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm
Reference Value = 56.2 V/m; Power Drift = -0.055 dB
Peak SAR (extrapolated) = 3.36 W/kg
SAR(1 g) = 2.32mW/g; SAR(10 g) = 1.53mW/g
Maximum value of SAR (measured) = 2.50mW/g



0 dB = 2.50mW/g

LG Electronics Inc.

DUT: Dipole 1900MHz; Type: D1900V2; Serial: 5d017

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1
Medium: Head 1900 MHz; ($\sigma = 1.44\text{mho/m}$; $\epsilon_r = 39$; $\rho = 1000\text{ kg/m}^3$)
Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

1900 MHz Dipole Validation

Area Scan (71x71x1): Measurement grid: $dx=15\text{mm}$, $dy=15\text{mm}$

Maximum value of SAR (interpolated) = 12.3mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: $dx=8\text{mm}$, $dy=8\text{mm}$, $dz=5\text{mm}$

Reference Value = 93.5 V/m ; Power Drift = -0.086 dB

Peak SAR (extrapolated) = 16.8 W/kg

SAR(1 g) = 10.1mW/g ; SAR(10 g) = 5.45mW/g

Maximum value of SAR (measured) = 11.5mW/g



0 dB = 11.5mW/g

APPENDIX B: SAR Test Data

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Head 836.6MHz; ($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Right Section

Test Date: 03/08/2010;; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Touch, Ch.190, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.384mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.40 V/m; Power Drift = -0.032 dB

Peak SAR (extrapolated) = 0.456 W/kg

SAR(1 g) = 0.366mW/g; SAR(10 g) = 0.269mW/g

Maximum value of SAR (measured) = 0.385mW/g



0 dB = 0.385mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Head 836.6MHz;($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Right Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Tilt, Ch.190, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.233mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 13.0 V/m; Power Drift = 0.233 dB

Peak SAR (extrapolated) = 0.276 W/kg

SAR(1 g) = 0.219mW/g; SAR(10 g) = 0.164mW/g

Maximum value of SAR (measured) = 0.235mW/g



0 dB = 0.235mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Head 836.6MHz; ($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.190, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.452mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.43 V/m; Power Drift = -0.052 dB

Peak SAR (extrapolated) = 0.605 W/kg

SAR(1 g) = 0.421mW/g; SAR(10 g) = 0.291mW/g

Maximum value of SAR (measured) = 0.450mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Head 836.6MHz;($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Tilt, Ch.190,Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.265mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 13.1 V/m; Power Drift = -0.018 dB

Peak SAR (extrapolated) = 0.346 W/kg

SAR(1 g) = 0.257mW/g; SAR(10 g) = 0.182mW/g

Maximum value of SAR (measured) = 0.287mW/g



0 dB = 0.287mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880MHz; ($\sigma = 1.42\text{mho/m}$; $\epsilon_r = 39.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Right Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Touch, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.302mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.02 V/m; Power Drift = -0.223 dB

Peak SAR (extrapolated) = 0.399 W/kg

SAR(1 g) = 0.266mW/g; SAR(10 g) = 0.172mW/g

Maximum value of SAR (measured) = 0.286mW/g



0 dB = 0.286mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880MHz; ($\sigma = 1.42\text{mho/m}$; $\epsilon_r = 39.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Right Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Tilt, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.132mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 9.27 V/m; Power Drift = -0.150 dB

Peak SAR (extrapolated) = 0.163 W/kg

SAR(1 g) = 0.112mW/g; SAR(10 g) = 0.069mW/g

Maximum value of SAR (measured) = 0.121mW/g



0 dB = 0.121mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880MHz;($\sigma = 1.42\text{mho/m}$; $\epsilon_r = 39.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.389mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.41 V/m; Power Drift = 0.153 dB

Peak SAR (extrapolated) = 0.530 W/kg

SAR(1 g) = 0.326mW/g; SAR(10 g) = 0.183mW/g

Maximum value of SAR (measured) = 0.357mW/g



0 dB = 0.357mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880MHz;($\sigma = 1.42\text{mho/m}$; $\epsilon_r = 39.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Tilt, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.139mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.48 V/m; Power Drift = -0.081 dB

Peak SAR (extrapolated) = 0.169 W/kg

SAR(1 g) = 0.119mW/g; SAR(10 g) = 0.076mW/g

Maximum value of SAR (measured) = 0.127mW/g



0 dB = 0.127mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Head 836.4 MHz; ($\sigma = 0.907 \text{ mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000 \text{ kg/m}^3$)
Phantom section: Right Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Touch, Ch.4182, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm.

Maximum value of SAR (interpolated) = 0.339mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.92 V/m; Power Drift = -0.007 dB

Peak SAR (extrapolated) = 0.410 W/kg

SAR(1 g) = 0.323mW/g; SAR(10 g) = 0.239mW/g

Maximum value of SAR (measured) = 0.343mW/g



0 dB = 0.343mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1

Medium: Head 836.4 MHz; ($\sigma = 0.907 \text{ mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Right Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Right Head Tilt, Ch.4182, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm.

Maximum value of SAR (interpolated) = 0.226mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 13.2 V/m; Power Drift = 0.040 dB

Peak SAR (extrapolated) = 0.270 W/kg

SAR(1 g) = 0.214mW/g; SAR(10 g) = 0.158mW/g

Maximum value of SAR (measured) = 0.226mW/g



0 dB = 0.226mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Head 836.4 MHz; ($\sigma = 0.907 \text{ mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000 \text{ kg/m}^3$)
Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.4182, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm.

Maximum value of SAR (interpolated) = 0.413mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.47 V/m; Power Drift = -0.255 dB

Peak SAR (extrapolated) = 0.556 W/kg

SAR(1 g) = 0.385mW/g; SAR(10 g) = 0.268mW/g

Maximum value of SAR (measured) = 0.412mW/g



0 dB = 0.412mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Head 836.4 MHz; ($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)
Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Tilt, Ch.4182, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm.

Maximum value of SAR (interpolated) = 0.233mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 12.1 V/m; Power Drift = 0.146 dB

Peak SAR (extrapolated) = 0.277 W/kg

SAR(1 g) = 0.223mW/g; SAR(10 g) = 0.166mW/g

Maximum value of SAR (measured) = 0.234mW/g



0 dB = 0.234mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:4.15

Medium: Body 836.6 MHz; ($\sigma = 0.963 \text{ mho/m}$; $\epsilon_r = 55.1$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Front, Slide Down, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.366mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.11 V/m; Power Drift = -0.055 dB

Peak SAR (extrapolated) = 0.425 W/kg

SAR(1 g) = 0.342mW/g; SAR(10 g) = 0.252mW/g

Maximum value of SAR (measured) = 0.363mW/g



0 dB = 0.363mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:4.15

Medium: Body 836.6 MHz;($\sigma = 0.963\text{mho/m}$; $\epsilon_r = 55.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Front, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.605mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 10.3 V/m; Power Drift = 0.168 dB

Peak SAR (extrapolated) = 0.743 W/kg

SAR(1 g) = 0.563mW/g; SAR(10 g) = 0.402mW/g

Maximum value of SAR (measured) = 0.600mW/g



0 dB = 0.600mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:4.15

Medium: Body 836.6 MHz;($\sigma = 0.963\text{mho/m}$; $\epsilon_r = 55.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.769mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 14.4 V/m; Power Drift = -0.058 dB

Peak SAR (extrapolated) = 0.966 W/kg

SAR(1 g) = 0.722mW/g; SAR(10 g) = 0.517mW/g

Maximum value of SAR (measured) = 0.765mW/g



0 dB = 0.765mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Body 836.6 MHz;($\sigma = 0.963\text{mho/m}$; $\epsilon_r = 55.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.396mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 10.2 V/m; Power Drift = 0.042 dB

Peak SAR (extrapolated) = 0.502 W/kg

SAR(1 g) = 0.377mW/g; SAR(10 g) = 0.269mW/g

Maximum value of SAR (measured) = 0.403mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: Body 1880 MHz;($\sigma = 1.53\text{mho/m}$; $\epsilon_r = 52.7$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.48, 4.48, 4.48); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Slide Down, Fixed Ant., Standard Battery

Area Scan (71x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.097mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.51 V/m; Power Drift = -0.155 dB

Peak SAR (extrapolated) = 0.126 W/kg

SAR(1 g) = 0.090mW/g; SAR(10 g) = 0.059mW/g

Maximum value of SAR (measured) = 0.097mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: Body 1880 MHz;($\sigma = 1.53\text{mho/m}$; $\epsilon_r = 52.7$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.48, 4.48, 4.48); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Front, Slide Up, Fixed Ant., Standard Battery

Area Scan (81x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.180mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.42 V/m; Power Drift = -0.069 dB

Peak SAR (extrapolated) = 0.230 W/kg

SAR(1 g) = 0.165mW/g; SAR(10 g) = 0.110mW/g

Maximum value of SAR (measured) = 0.176mW/g



LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: Body 1880 MHz; ($\sigma = 1.53\text{mho/m}$; $\epsilon_r = 52.7$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.48, 4.48, 4.48); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (81x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.250mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.32 V/m; Power Drift = 0.091 dB

Peak SAR (extrapolated) = 0.298 W/kg

SAR(1 g) = 0.224mW/g; SAR(10 g) = 0.150mW/g

Maximum value of SAR (measured) = 0.247mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Body 1880 MHz;($\sigma = 1.53\text{mho/m}$; $\epsilon_r = 52.7$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.48, 4.48, 4.48); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (81x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.134mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.73 V/m; Power Drift = 0.389 dB

Peak SAR (extrapolated) = 0.174 W/kg

SAR(1 g) = 0.126mW/g; SAR(10 g) = 0.083mW/g

Maximum value of SAR (measured) = 0.136mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Body 836.4 MHz;($\sigma = 0.964$ mho/m; $\epsilon_r = 55.1$; $\rho = 1000$ kg/m³)
Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04


Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.4182, Front, Slide Down , Fixed Ant., Standard Battery

Area Scan (71x111x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 0.188mW/g

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm
Reference Value = 7.23 V/m; Power Drift = -0.075 dB
Peak SAR (extrapolated) = 0.225 W/kg
SAR(1 g) = 0.179mW/g; SAR(10 g) = 0.132mW/g
Maximum value of SAR (measured) = 0.190mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1

Medium: Body 836.4 MHz;($\sigma = 0.964$ mho/m; $\epsilon_r = 55.1$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.4182, Front, Slide Up , Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.284mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.32 V/m; Power Drift = 0.351 dB

Peak SAR (extrapolated) = 0.420 W/kg

SAR(1 g) = 0.319mW/g; SAR(10 g) = 0.226mW/g

Maximum value of SAR (measured) = 0.338mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Body 836.4 MHz;($\sigma = 0.964$ mho/m; $\epsilon_r = 55.1$; $\rho = 1000$ kg/m³)
Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.4182, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.470mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 11.1 V/m; Power Drift = 0.124 dB

Peak SAR (extrapolated) = 0.583 W/kg

SAR(1 g) = 0.437mW/g; SAR(10 g) = 0.310mW/g

Maximum value of SAR (measured) = 0.464mW/g



LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.6 MHz; Duty Cycle: 1:8.3

Medium: Head 836.6MHz;($\sigma = 0.907\text{mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.190, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.452mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.43 V/m; Power Drift = -0.052 dB

Peak SAR (extrapolated) = 0.605 W/kg

SAR(1 g) = 0.421mW/g; SAR(10 g) = 0.291mW/g

Maximum value of SAR (measured) = 0.450mW/g



0 dB = 0.450mW/g

LG Electronics Inc.

DUT: GW382f;Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: Head 1880MHz;($\sigma = 1.42\text{mho/m}$; $\epsilon_r = 39.1$; $\rho = 1000\text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.97, 4.97, 4.97); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.661, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.389mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 5.41 V/m; Power Drift = 0.153 dB

Peak SAR (extrapolated) = 0.530 W/kg

SAR(1 g) = 0.326mW/g; SAR(10 g) = 0.183mW/g

Maximum value of SAR (measured) = 0.357mW/g



0 dB = 0.357mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1

Medium: Head 836.4 MHz; ($\sigma = 0.907 \text{ mho/m}$; $\epsilon_r = 42.8$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Left Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.22, 6.22, 6.22); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1117; Type: SAM; Serial: TP-1117

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Left Head Touch, Ch.4182, Fixed Ant., Standard Battery

Area Scan (61x101x1): Measurement grid: dx=15mm, dy=15mm.

Maximum value of SAR (interpolated) = 0.413mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.47 V/m; Power Drift = -0.255 dB

Peak SAR (extrapolated) = 0.556 W/kg

SAR(1 g) = 0.385mW/g; SAR(10 g) = 0.268mW/g

Maximum value of SAR (measured) = 0.412mW/g



0 dB = 0.412mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.6 MHz; Duty Cycle: 1:4.15

Medium: Body 836.6 MHz; ($\sigma = 0.963 \text{ mho/m}$; $\epsilon_r = 55.1$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.190, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.769mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 14.4 V/m; Power Drift = -0.058 dB

Peak SAR (extrapolated) = 0.966 W/kg

SAR(1 g) = 0.722mW/g; SAR(10 g) = 0.517mW/g

Maximum value of SAR (measured) = 0.765mW/g



0 dB = 0.765mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: Body 1880 MHz; ($\sigma = 1.53 \text{ mho/m}$; $\epsilon_r = 52.7$; $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03/09/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(4.48, 4.48, 4.48); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.661, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (81x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.250mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.32 V/m; Power Drift = 0.091 dB

Peak SAR (extrapolated) = 0.298 W/kg

SAR(1 g) = 0.224mW/g; SAR(10 g) = 0.150mW/g

Maximum value of SAR (measured) = 0.247mW/g



0 dB = 0.247mW/g

LG Electronics Inc.

DUT: GW382f; Type: Cellular / PCS GSM / WCDMA / EDGE Phone with Bluetooth; Serial:#1

Communication System: WCDMA; Frequency: 836.4 MHz; Duty Cycle: 1:1
Medium: Body 836.4 MHz; ($\sigma = 0.964$ mho/m; $\epsilon_r = 55.1$; $\rho = 1000$ kg/m³)
Phantom section: Flat Section

Test Date: 03/08/2010; Ambient Temp: 22.0°C; Tissue Temp: 21.8°C

Probe: ET3DV6 - SN1729; ConvF(6.05, 6.05, 6.05); Calibrated: 2010-01-27

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2010-02-04

Phantom: SAM Twin Phantom TP-1151; Type: SAM; Serial: TP-1151

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 184

Flat Touch, Ch.4182, Rear, Slide Up, Fixed Ant., Standard Battery

Area Scan (91x111x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.470mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 11.1 V/m; Power Drift = 0.124 dB

Peak SAR (extrapolated) = 0.583 W/kg

SAR(1 g) = 0.437mW/g; SAR(10 g) = 0.310mW/g

Maximum value of SAR (measured) = 0.464mW/g



0 dB = 0.464mW/g

APPENDIX D: Calibration Certificates

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **ET3-1729_Jan10**

CALIBRATION CERTIFICATE

Object **ET3DV6 - SN:1729**

Calibration procedure(s) **QA CAL-01.v6, QA CAL-23.v3 and QA CAL-25.v2**
Calibration procedure for dosimetric E-field probes

Calibration date: **January 27, 2010**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter E4419B	GB41293874	1-Apr-09 (No. 217-01030)	Apr-10
Power sensor E4412A	MY41495277	1-Apr-09 (No. 217-01030)	Apr-10
Power sensor E4412A	MY41498087	1-Apr-09 (No. 217-01030)	Apr-10
Reference 3 dB Attenuator	SN: S5054 (3c)	31-Mar-09 (No. 217-01026)	Mar-10
Reference 20 dB Attenuator	SN: S5086 (20b)	31-Mar-09 (No. 217-01028)	Mar-10
Reference 30 dB Attenuator	SN: S5129 (30b)	31-Mar-09 (No. 217-01027)	Mar-10
Reference Probe ES3DV2	SN: 3013	30-Dec-09 (No. ES3-3013_Dec09)	Dec-10
DAE4	SN: 660	29-Sep-09 (No. DAE4-660_Sep09)	Sep-10
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
RF generator HP 8648C	US3642U01700	4-Aug-99 (in house check Oct-09)	In house check: Oct-11
Network Analyzer HP 8753E	US37390585	18-Oct-01 (in house check Oct-09)	In house check: Oct10

Calibrated by: **Jeton Kastrati** **Laboratory Technician**

Approved by: **Katja Pokovic** **Technical Manager**

Issued: January 27, 2010

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: ET3-1729_Jan10

Page 1 of 11

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C	modulation dependent linearization parameters
Polarization ϕ	ϕ rotation around probe axis
Polarization ϑ	ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}**: Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E^2 -field uncertainty inside TSL (see below *ConvF*).
- NORM(f)_{x,y,z}** = NORM_{x,y,z} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCP_{x,y,z}**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- A_{x,y,z}; B_{x,y,z}; C_{x,y,z}; VR_{x,y,z}**: A, B, C are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: ET3-1729_Jan10

Page 2 of 11

ET3DV6 SN:1729

January 27, 2010

Probe ET3DV6

SN:1729

Manufactured:	October 1, 2002
Last calibrated:	January 20, 2009
Recalibrated:	January 27, 2010

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ET3-1729_Jan10

Page 3 of 11

ET3DV6 SN:1729

January 27, 2010

DASY - Parameters of Probe: ET3DV6 SN:1729**Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm ($\mu\text{V}/(\text{V}/\text{m})^2$) ^A	1.64	1.61	1.82	$\pm 10.1\%$
DCP (mV) ^B	94.1	93.4	94.2	

Modulation Calibration Parameters

UID	Communication System Name	PAR		A dB	B dBuV	C	VR mV	Unc ^E (k=2)
10000	CW	0.00	X	0.00	0.00	1.00	300.0	$\pm 1.5\%$
			Y	0.00	0.00	1.00	300.0	
			Z	0.00	0.00	1.00	300.0	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).

^B Numerical linearization parameter: uncertainty not required.

^E Uncertainty is determined using the maximum deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

ET3DV6 SN:1729

January 27, 2010

DASY - Parameters of Probe: ET3DV6 SN:1729

Calibration Parameter Determined in Head Tissue Simulating Media

f [MHz]	Validity [MHz] ^c	Permittivity	Conductivity	ConvF X	ConvF Y	ConvF Z	Alpha	Depth Unc (k=2)
835	± 50 / ± 100	41.9 ± 5%	0.89 ± 5%	6.22	6.22	6.22	0.48	2.20 ± 11.0%
1750	± 50 / ± 100	40.1 ± 5%	1.37 ± 5%	5.27	5.27	5.27	0.51	2.80 ± 11.0%
1900	± 50 / ± 100	40.0 ± 5%	1.40 ± 5%	4.97	4.97	4.97	0.64	2.38 ± 11.0%
2450	± 50 / ± 100	39.2 ± 5%	1.80 ± 5%	4.42	4.42	4.42	0.99	1.68 ± 11.0%

^c The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

ET3DV6 SN:1729

January 27, 2010

DASY - Parameters of Probe: ET3DV6 SN:1729

Calibration Parameter Determined in Body Tissue Simulating Media

f [MHz]	Validity [MHz] ^C	Permittivity	Conductivity	ConvF X	ConvF Y	ConvF Z	Alpha	Depth Unc (k=2)
750	± 50 / ± 100	55.5 ± 5%	0.96 ± 5%	6.06	6.06	6.06	0.58	1.95 ± 11.0%
835	± 50 / ± 100	55.2 ± 5%	0.97 ± 5%	6.05	6.05	6.05	0.51	2.17 ± 11.0%
1750	± 50 / ± 100	53.4 ± 5%	1.49 ± 5%	4.71	4.71	4.71	0.66	3.04 ± 11.0%
1900	± 50 / ± 100	53.3 ± 5%	1.52 ± 5%	4.48	4.48	4.48	0.94	2.39 ± 11.0%
2450	± 50 / ± 100	52.7 ± 5%	1.95 ± 5%	4.07	4.07	4.07	0.99	1.38 ± 11.0%

^C The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


Certificate No: ET3-1729_Jan10

Page 6 of 11

ET3DV6 SN:1729

January 27, 2010

Frequency Response of E-Field
(TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ ($k=2$)

ET3DV6 SN:1729

January 27, 2010

Receiving Pattern (ϕ), $\vartheta = 0^\circ$




Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)

ET3DV6 SN:1729

January 27, 2010

Dynamic Range f(SAR_{head})
(Waveguide R22, f = 1800 MHz)




Uncertainty of Linearity Assessment: ± 0.6% (k=2)

ET3DV6 SN:1729


January 27, 2010

Conversion Factor Assessment



Deviation from Isotropy in HSL

Error (ϕ, θ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

ET3DV6 SN:1729

January 27, 2010

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	Not applicable
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	10 mm
Tip Diameter	6.8 mm
Probe Tip to Sensor X Calibration Point	2.7 mm
Probe Tip to Sensor Y Calibration Point	2.7 mm
Probe Tip to Sensor Z Calibration Point	2.7 mm
Recommended Measurement Distance from Surface	4 mm

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D835V2-471_Jan09**

CALIBRATION CERTIFICATE

Object: **D835V2 - SN: 471**

Calibration procedure(s): **QA CAL-05.v7
Calibration procedure for dipole validation kits**

Calibration date: **January 19, 2009**


Condition of the calibrated item: **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	08-Oct-08 (No. 217-00898)	Oct-09
Power sensor HP 8481A	US37292783	08-Oct-08 (No. 217-00898)	Oct-09
Reference 20 dB Attenuator	SN: 5086 (20g)	01-Jul-08 (No. 217-00864)	Jul-09
Type-N mismatch combination	SN: 5047.2 / 06327	01-Jul-08 (No. 217-00867)	Jul-09
Reference Probe ES3DV2	SN: 3025	28-Apr-08 (No. ES3-3025_Apr08)	Apr-09
DAE4	SN: 601	14-Mar-08 (No. DAE4-601_Mar08)	Mar-09
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-07)	In house check: Oct-09
RF generator R&S SMT-06	100005	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (in house check Oct-08)	In house check: Oct-09

Calibrated by:	Name Jeton Kastrati	Function Laboratory Technician	Signature 
Approved by:	Name Katja Pokovic	Function Technical Manager	Signature 

Issued: January 20, 2009

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D835V2-471_Jan09

Page 1 of 6

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM x,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V5.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	41.3 $\pm$ 6 %	0.91 mho/m $\pm$ 6 %
Head TSL temperature during test	(21.5 $\pm$ 0.2) °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.44 mW / g
SAR normalized	normalized to 1W	9.76 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	9.66 mW / g $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.61 mW / g
SAR normalized	normalized to 1W	6.44 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	6.39 mW / g $\pm$ 16.5 % (k=2)

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.7 Ω - 1.7 j Ω
Return Loss	- 32.5 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.386 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	November 15, 2002

DASY5 Validation Report for Head TSL

Date/Time: 19.01.2009 11:07:17

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:471

Communication System: CW-835; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz

Medium parameters used: $f = 835$ MHz; $\sigma = 0.91$ mho/m; $\epsilon_r = 41.1$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC)

DASY5 Configuration:

- Probe: ES3DV2 - SN3025; ConvF(5.97, 5.97, 5.97); Calibrated: 28.04.2008
- Sensor-Surface: 3.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 14.03.2008
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- Measurement SW: DASY5, V5.0 Build 120; SEMCAD X Version 13.4 Build 45


Pin=250mW; dip=15mm; dist=3.4mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 56.2 V/m; Power Drift = 0.026 dB

Peak SAR (extrapolated) = 3.61 W/kg

SAR(1 g) = 2.44 mW/g; SAR(10 g) = 1.61 mW/g


Maximum value of SAR (measured) = 2.75 mW/g



Certificate No: D835V2-471_Jan09

Page 5 of 6

Impedance Measurement Plot for Head TSL



Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D1900V2-5d017_Jul09**

CALIBRATION CERTIFICATE

Object **D1900V2 - SN: 5d017**

Calibration procedure(s) **QA CAL-05.v7**
Calibration procedure for dipole validation kits

Calibration date: **July 20, 2009**

Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	08-Oct-08 (No. 217-00898)	Oct-09
Power sensor HP 8481A	US37292783	08-Oct-08 (No. 217-00898)	Oct-09
Reference 20 dB Attenuator	SN: 5086 (20g)	31-Mar-09 (No. 217-01025)	Mar-10
Type-N mismatch combination	SN: 5047.2 / 06327	31-Mar-09 (No. 217-01029)	Mar-10
Reference Probe ES3DV2	SN: 3025	30-Apr-09 (No. ES3-3025_Apr09)	Apr-10
DAE4	SN: 601	07-Mar-09 (No. DAE4-601_Mar09)	Mar-10
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-07)	In house check: Oct-09
RF generator R&S SMT-06	100005	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (in house check Oct-08)	In house check: Oct-09

Calibrated by: **Claudio Leubler** **Laboratory Technician**

Approved by: **Katja Pokovic** **Technical Manager**

Signature

Issued: July 22, 2009

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D1900V2-5d017_Jul09

Page 1 of 6

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V5.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	40.9 $\pm$ 6 %	1.43 mho/m $\pm$ 6 %
Head TSL temperature during test	(22.0 $\pm$ 0.2) °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	10.2 mW / g
SAR normalized	normalized to 1W	40.8 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	40.5 mW / g $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.33 mW / g
SAR normalized	normalized to 1W	21.3 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	21.3 mW / g $\pm$ 16.5 % (k=2)

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Appendix**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	50.6 Ω + 2.7 j Ω
Return Loss	- 31.2 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.196 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	June 04, 2002

DASY5 Validation Report for Head TSL

Date/Time: 20.07.2009 13:51:56

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:5d017

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: HSL U11 BB

Medium parameters used: $f = 1900 \text{ MHz}$; $\sigma = 1.43 \text{ mho/m}$; $\epsilon_r = 40.9$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC)

DASY5 Configuration:

- Probe: ES3DV2 - SN3025; ConvF(4.88, 4.88, 4.88); Calibrated: 30.04.2009
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 07.03.2009
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- Measurement SW: DASY5, V5.0 Build 120; SEMCAD X Version 13.4 Build 45


Pin = 250 mW; dip = 10 mm, scan at 3.0 mm/Zoom Scan (dist=3.0 mm, probe 0deg)**(7x7x7)/Cube 0:** Measurement grid: $dx=5\text{mm}$, $dy=5\text{mm}$, $dz=5\text{mm}$

Reference Value = 96.2 V/m; Power Drift = 0.072 dB

Peak SAR (extrapolated) = 18.6 W/kg

SAR(1 g) = 10.2 mW/g; SAR(10 g) = 5.33 mW/g

Maximum value of SAR (measured) = 12.5 mW/g



0 dB = 12.5mW/g

Impedance Measurement Plot for Head TSL

