

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:8.3

Medium: Head 1880.00 MHz;(σ = 1.39 mho/m; ε_r = 40.5; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(5.16, 5.16, 5.16); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Touch, Ch.661, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.096mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 3.13V/m; Power Drift = -0.077dB

Peak SAR (extrapolated) = 0.124W/kg

SAR(1 g) = 0.092mW/g; SAR(10 g) = 0.060mW/g

Maximum value of SAR (measured) = 0.100mW/g



LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:8.3

Medium: Head 1880.00 MHz;(σ = 1.39 mho/m; ε_r = 40.5; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(5.16, 5.16, 5.16); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Tilt, Ch.661, Slider Down, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.081mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 7.29V/m; Power Drift = 0.132dB

Peak SAR (extrapolated) = 0.098W/kg

SAR(1 g) = 0.068mW/g; SAR(10 g) = 0.041mW/g

Maximum value of SAR (measured) = 0.075mW/g



0 dB = 0.075mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;($\sigma = 0.962$ mho/m; $\epsilon_r = 55.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Front, Slider Down, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.160mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 6.00V/m; Power Drift = 0.194dB

Peak SAR (extrapolated) = 0.187W/kg

SAR(1 g) = 0.150mW/g; SAR(10 g) = 0.113mW/g

Maximum value of SAR (measured) = 0.158mW/g



LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;(σ = 0.962 mho/m; ε_r = 55.5; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Front, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.513mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.8 V/m; Power Drift = 0.064dB

Peak SAR (extrapolated) = 0.613W/kg

SAR(1 g) = 0.497mW/g; SAR(10 g) = 0.372mW/g

Maximum value of SAR (measured) = 0.525mW/g



0 dB = 0.525mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;(σ = 0.962 mho/m; ε_r = 55.5; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Rear, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.606mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.7V/m; Power Drift = -0.113dB

Peak SAR (extrapolated) = 0.709W/kg

SAR(1 g) = 0.568mW/g; SAR(10 g) = 0.421mW/g

Maximum value of SAR (measured) = 0.605mW/g



0 dB = 0.605mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;(σ = 0.962 mho/m; ε_r = 55.5; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Front, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.601mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.4V/m; Power Drift = -0.314dB

Peak SAR (extrapolated) = 0.693W/kg

SAR(1 g) = 0.557mW/g; SAR(10 g) = 0.416mW/g

Maximum value of SAR (measured) = 0.584mW/g



LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;(σ = 1.5 mho/m; ε_r = 53.8; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Front, Slider Down, BT Off Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.049mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 4.10V/m; Power Drift = -0.028dB

Peak SAR (extrapolated) = 0.074W/kg

SAR(1 g) = 0.047mW/g; SAR(10 g) = 0.029mW/g

Maximum value of SAR (measured) = 0.050mW/g



0 dB = 0.050mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;(σ = 1.5 mho/m; ε_r = 53.8; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Front, Slider Up, BT Off Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.096mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 4.16V/m; Power Drift = 0.057dB

Peak SAR (extrapolated) = 0.139W/kg

SAR(1 g) = 0.087mW/g; SAR(10 g) = 0.054mW/g

Maximum value of SAR (measured) = 0.093mW/g



LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;(σ = 1.5 mho/m; ε_r = 53.8; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Rear, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.229mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 5.11V/m; Power Drift = -0.148dB

Peak SAR (extrapolated) = 0.342W/kg

SAR(1 g) = 0.213mW/g; SAR(10 g) = 0.132mW/g

Maximum value of SAR (measured) = 0.230mW/g



LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;($\sigma = 1.5$ mho/m; $\epsilon_r = 53.8$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Rear, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.232mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 5.15V/m; Power Drift = -0.105dB

Peak SAR (extrapolated) = 0.347W/kg

SAR(1 g) = 0.216mW/g; SAR(10 g) = 0.134mW/g

Maximum value of SAR (measured) = 0.233mW/g



0 dB = 0.233mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.60 MHz;Duty Cycle: 1:8.3

Medium: Head 836.60 MHz;(σ = 0.904 mho/m; ε_r = 41.9; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.29, 6.29, 6.29); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Touch, Ch.190, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.397mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 10.1V/m; Power Drift = -0.037dB

Peak SAR (extrapolated) = 0.484W/kg

SAR(1 g) = 0.376mW/g; SAR(10 g) = 0.278mW/g

Maximum value of SAR (measured) = 0.404mW/g



0 dB = 0.404mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:8.3

Medium: Head 1880.00 MHz;(σ = 1.39 mho/m; ε_r = 40.5; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(5.16, 5.16, 5.16); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Touch, Ch.661, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.143mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 4.13V/m; Power Drift = 0.190dB

Peak SAR (extrapolated) = 0.207W/kg

SAR(1 g) = 0.130mW/g; SAR(10 g) = 0.070mW/g

Maximum value of SAR (measured) = 0.149mW/g



0 dB = 0.149mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;(σ = 0.962 mho/m; ε_r = 55.5; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Front, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.601mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.4V/m; Power Drift = 0.314dB

Peak SAR (extrapolated) = 0.693W/kg

SAR(1 g) = 0.557mW/g; SAR(10 g) = 0.416mW/g

Maximum value of SAR (measured) = 0.525mW/g



0 dB = 0.525mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;(σ = 1.5 mho/m; ε_r = 53.8; ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Rear, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.228mW/g



Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 5.16V/m; Power Drift = 0.207dB

Peak SAR (extrapolated) = 0.346W/kg

SAR(1 g) = 0.212mW/g; SAR(10 g) = 0.131mW/g

Maximum value of SAR (measured) = 0.228mW/g



0 dB = 0.228mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GSM 850; Frequency: 836.60 MHz;Duty Cycle: 1:8.3

Medium: Head 836.60 MHz;(σ = 0.904 mho/m; ε_r = 41.9; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.29, 6.29, 6.29); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Touch, Ch.190, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.397mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 10.1V/m; Power Drift = -0.037dB

Peak SAR (extrapolated) = 0.484W/kg

SAR(1 g) = 0.376mW/g; SAR(10 g) = 0.278mW/g

Maximum value of SAR (measured) = 0.404mW/g



0 dB = 0.404mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: PCS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:8.3

Medium: Head 1880.00 MHz;(σ = 1.39 mho/m; ε_r = 40.5; ρ = 1000 kg/m³)

Phantom section: Left Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(5.16, 5.16, 5.16); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Left Head Touch, Ch.661, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.143mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 4.13V/m; Power Drift = 0.190dB

Peak SAR (extrapolated) = 0.207W/kg

SAR(1 g) = 0.130mW/g; SAR(10 g) = 0.070mW/g

Maximum value of SAR (measured) = 0.149mW/g



0 dB = 0.149mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 850; Frequency: 836.60 MHz;Duty Cycle: 1:4.15

Medium: Head 836.60 MHz;($\sigma = 0.962$ mho/m; $\epsilon_r = 55.5$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/28/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(6.2, 6.2, 6.2); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 835; Type: SAM 4.0; Serial:TP-1066

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.190, Front, Slider Up, BT Off, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.601mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.4V/m; Power Drift = 0.314dB

Peak SAR (extrapolated) = 0.693W/kg

SAR(1 g) = 0.557mW/g; SAR(10 g) = 0.416mW/g

Maximum value of SAR (measured) = 0.525mW/g



0 dB = 0.525mW/g

LG Electronics Inc.

DUT: KM710d ;Type: Cellular / PCS GSM / EDGE Phone with Bluetooth; Serial:#1

Communication System: GPRS 1900; Frequency: 1880.00 MHz;Duty Cycle: 1:4.15

Medium: Head 1880.00 MHz;($\sigma = 1.5$ mho/m; $\epsilon_r = 53.8$; $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Test Date: 01/29/2008;Ambient Temp: 22. 9°C;Tissue Temp: 21.8°C

Probe: ES3DV6 - SN1729; ConvF(4.66, 4.66, 4.66); Calibrated: 2007-02-21

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn534; Calibrated: 2007-09-24

Phantom: SAM 1800; Type: SAM 4.0; Serial:TP-1244

Measurement SW: DASY4, V4.7 Build 55; Post processing SW: SEMCAD, V1.8 Build 171

Flat Touch, Ch.661, Rear, Slider Up, BT On, Fixed Ant., Standard Battery

Area Scan (61x61x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.228mW/g


Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 5.16V/m; Power Drift = 0.207dB

Peak SAR (extrapolated) = 0.346W/kg

SAR(1 g) = 0.212mW/g; SAR(10 g) = 0.131mW/g

Maximum value of SAR (measured) = 0.228mW/g



0 dB = 0.228mW/g

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **ET3-1729_Feb07**

CALIBRATION CERTIFICATE

Object **ET3DV6 - SN-1729**

Calibration procedure(s) **QA-CAL-01.v5
Calibration procedure for dosimetric E-field probes**

Calibration date: **February 21, 2007**



Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter E4419B	GB41293874	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Power sensor E4412A	MY41495277	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Power sensor E4412A	MY41498087	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Reference 3 dB Attenuator	SN: S5054 (3c)	10-Aug-06 (METAS, No. 217-00592)	Aug-07
Reference 20 dB Attenuator	SN: S5086 (20b)	4-Apr-06 (METAS, No. 251-00558)	Apr-07
Reference 30 dB Attenuator	SN: S5129 (30b)	10-Aug-06 (METAS, No. 217-00593)	Aug-07
Reference Probe ES3DV2	SN: 3013	4-Jan-07 (SPEAG, No. ES3-3013_Jan07)	Jan-08
DAE4	SN: 654	21-Jun-06 (SPEAG, No. DAE4-654_Jun06)	Jun-07
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
RF generator HP 8648C	US3642U01700	4-Aug-99 (SPEAG, in house check Nov-05)	In house check: Nov-07
Network Analyzer HP 8753E	US37390585	18-Oct-01 (SPEAG, in house check Oct-06)	In house check: Oct-07

	Name	Function	Signature
Calibrated by:	Katja Pokovic	Technical Manager	
Approved by:	Niels Kuster	Quality Manager	

Issued: February 22, 2007

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: ET3-1729_Feb07

Page 1 of 9

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
Polarization ϕ	ϕ rotation around probe axis
Polarization ϑ	ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}**: Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E²-field uncertainty inside TSL (see below *ConvF*).
- NORM(f)_{x,y,z}** = NORM_{x,y,z} * *frequency_response* (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCP_{x,y,z}**: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * *ConvF* whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: ET3-1729_Feb07

Page 2 of 9

ET3DV6 SN:1729

February 21, 2007

Probe ET3DV6

SN:1729

Manufactured:	October 1, 2002
Last calibrated:	January 24, 2006
Modified:	February 19, 2007
Recalibrated:	February 21, 2007

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ET3-1729_Feb07

Page 3 of 9

ET3DV6 SN:1729

February 21, 2007

DASY - Parameters of Probe: ET3DV6 SN:1729

Sensitivity in Free Space^A

Diode Compression^B

NormX	1.64 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP X	96 mV
NormY	1.61 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP Y	96 mV
NormZ	1.83 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP Z	95 mV

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL 900 MHz Typical SAR gradient: 5 % per mm

Sensor Center to Phantom Surface Distance		3.7 mm	4.7 mm
SAR _{be} [%]	Without Correction Algorithm	8.5	4.2
SAR _{be} [%]	With Correction Algorithm	0.0	0.1

TSL 1810 MHz Typical SAR gradient: 10 % per mm

Sensor Center to Phantom Surface Distance		3.7 mm	4.7 mm
SAR _{be} [%]	Without Correction Algorithm	13.3	8.9
SAR _{be} [%]	With Correction Algorithm	0.7	0.1

Sensor Offset

Probe Tip to Sensor Center 2.7 mm

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).


^B Numerical linearization parameter: uncertainty not required.

ET3DV6 SN:1729

February 21, 2007

Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide: R22)






Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ (k=2)

ET3DV6 SN:1729

February 21, 2007

Receiving Pattern (ϕ), $\theta = 0^\circ$



Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)


Certificate No: ET3-1729_Feb07

Page 6 of 9

ET3DV6 SN:1729

February 21, 2007

Dynamic Range f(SAR_{head})
(Waveguide R22, f = 1800 MHz)



Uncertainty of Linearity Assessment: $\pm 0.6\%$ ($k=2$)



Certificate No: ET3-1729_Feb07

Page 7 of 9

ET3DV6 SN:1729

February 21, 2007

Conversion Factor Assessment



f [MHz]	Validity [MHz] ^c	TSL	Permittivity	Conductivity	Alpha	Depth	ConvF Uncertainty
900	± 50 / ± 100	Head	41.5 ± 5%	0.97 ± 5%	0.36	2.70	6.29 ± 11.0% (k=2)
1810	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.52	2.89	5.16 ± 11.0% (k=2)
1950	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.51	2.89	4.90 ± 11.0% (k=2)
900	± 50 / ± 100	Body	55.0 ± 5%	1.05 ± 5%	0.39	2.46	6.20 ± 11.0% (k=2)
1810	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.61	2.66	4.66 ± 11.0% (k=2)
1950	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.79	2.16	4.42 ± 11.0% (k=2)

^c The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


Certificate No: ET3-1729_Feb07

Page 8 of 9

ET3DV6 SN:1729

February 21, 2007

Deviation from Isotropy in HSL
Error (ϕ , θ), $f = 900$ MHz



Uncertainty of Spherical Isotropy Assessment: $\pm 2.6\%$ ($k=2$)

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D835V2-471-Jan07**

CALIBRATION CERTIFICATE

Object **D835V2 - SN: 471**

Calibration procedure(s) **QA CAL-05.v6
Calibration procedure for dipole validation kits**

Calibration date: **January 22, 2007**

Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	03-Oct-06 (METAS, No. 217-00608)	Oct-07
Power sensor HP 8481A	US37292783	03-Oct-06 (METAS, No. 217-00608)	Oct-07
Reference 20 dB Attenuator	SN: 5086 (20g)	10-Aug-06 (METAS, No 217-00591)	Aug-07
Reference 10 dB Attenuator	SN: 5047.2 (10r)	10-Aug-06 (METAS, No 217-00591)	Aug-07
Reference Probe ET3DV6 (HF)	SN 1507	19-Oct-06 (SPEAG, No. ET3-1507_Oct06)	Oct-07
DAE4	SN 907	20-Jul-06 (SPEAG, No. DAE4-907_Jul06)	Jul-07
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (SPEAG, in house check Oct-05)	In house check: Oct-07
RF generator Agilent E4421B	MY41000675	11-May-05 (SPEAG, in house check Nov-05)	In house check: Nov-07
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (SPEAG, in house check Oct-06)	In house check: Oct-07

Calibrated by: **Marcel Fehr** Function: **Laboratory Technician** Signature: 

Approved by: **Katja Bokovic** Technical Manager Signature: 

Issued: January 22, 2007

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D835V2-471_Jan07

Page 1 of 6

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D835V2-471_Jan07

Page 2 of 6

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY4	V4.7
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	40.2 $\pm$ 6 %	0.88 mho/m $\pm$ 6 %
Head TSL temperature during test	(22.4 $\pm$ 0.2) °C	—	—

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.29 mW / g
SAR normalized	normalized to 1W	9.16 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	9.12 mW / g $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.50 mW / g
SAR normalized	normalized to 1W	6.00 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	5.96 mW / g $\pm$ 16.5 % (k=2)

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	52.8 Ω - 2.6 j Ω
Return Loss	- 28.5 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.385 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

- ☐ No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	November 15, 2002

DASY4 Validation Report for Head TSL

Date/Time: 22.01.2007 11:42:41

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 471

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz;

Medium parameters used: $f = 835 \text{ MHz}$; $\sigma = 0.88 \text{ mho/m}$; $\epsilon_r = 40.4$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507 (HF); ConvF(6.09, 6.09, 6.09); Calibrated: 19.10.2006
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn907; Calibrated: 20.07.2006
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; ;
- Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172


Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0:Measurement grid: $dx=5\text{mm}$, $dy=5\text{mm}$, $dz=5\text{mm}$

Reference Value = 55.5 V/m; Power Drift = 0.007 dB


Peak SAR (extrapolated) = 3.37 W/kg

SAR(1 g) = 2.29 mW/g; SAR(10 g) = 1.5 mW/g

Maximum value of SAR (measured) = 2.47 mW/g



Impedance Measurement Plot for Head TSL



**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **LG (Dymstec)**

Certificate No: **D1900V2-5d017_Jan07**

CALIBRATION CERTIFICATE

Object **D1900V2- SN: 5d017**

Calibration procedure(s) **QA CAL-05.v6
Calibration procedure for dipole validation kits**

Calibration date: **January 23, 2007**


Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM-442A	GB37480704	03-Oct-06 (METAS, No. 217-00608)	Oct-07
Power sensor HP 8481A	US37292783	03-Oct-06 (METAS, No. 217-00608)	Oct-07
Reference 20 dB Attenuator	SN: 5086 (20g)	10-Aug-06 (METAS, No 217-00591)	Aug-07
Reference 10 dB Attenuator	SN: 5047.2 (10r)	10-Aug-06 (METAS, No 217-00591)	Aug-07
Reference Probe ET3DV6	SN: 1507	19-Oct-06 (SPEAG, No. ET3-1507_Oct06)	Oct-07
Reference Probe ES3DV3	SN: 3025	19-Oct-06 (SPEAG, No. ES3-3025_Oct06)	Oct-07
DAE4	SN 907	20-Jul-06 (SPEAG, No. DAE4-907_Jul06)	Jul-07
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (SPEAG, in house check Oct-05)	In house check: Oct-07
RF generator Agilent E4421B	MY41000675	11-May-05 (SPEAG, in house check Nov-05)	In house check: Nov-07
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (SPEAG, in house check Oct-06)	In house check: Oct-07

	Name	Function	Signature
Calibrated by:	Mike Melli	Laboratory Technician	
Approved by:	Katja Pokovic	Technical Manager	

Issued: January 24, 2007

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D1900V2-5d017_Jan07

Page 1 of 6

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM x,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D1900V2-5d017_Jan07

Page 2 of 6

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY4	V4.7
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	38.8 $\pm$ 6 %	1.43 mho/m $\pm$ 6 %
Head TSL temperature during test	(21.7 $\pm$ 0.2) °C	—	—

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	9.66 mW / g
SAR normalized	normalized to 1W	38.6 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	37.6 mW / g $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.05 mW / g
SAR normalized	normalized to 1W	20.2mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	19.9 mW / g $\pm$ 16.5 % (k=2)

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Appendix**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	$53.2 \Omega + 3.0 j\Omega$
Return Loss	- 27.4 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.196 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	June 4, 2002

DASY4 Validation Report for Head TSL

Date/Time: 23.01.2007 14:37:49

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:5d017

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: HSL U10 BB;

Medium parameters used: $f = 1900$ MHz; $\sigma = 1.43$ mho/m; $\epsilon_r = 38.8$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507 (HF); ConvF(4.97, 4.97, 4.97); Calibrated: 19.10.2006
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn907; Calibrated: 20.07.2006
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA
- Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 93.0 V/m; Power Drift = 0.028 dB

Peak SAR (extrapolated) = 16.9 W/kg

SAR(1 g) = 9.66 mW/g; SAR(10 g) = 5.05 mW/g

Maximum value of SAR (measured) = 11.0 mW/g



Impedance Measurement Plot for Head TSL

