

WMTS TRANSCEIVERS

Note: The information in this section is only applicable for Wireless Medical Telemetry Service (WMTS).

About Transceivers

Wireless Medical Telemetry Services (WMTS) Remote Transceivers provide the link between the patient and the Central Station through the newly approved 608 - 614 MHz Medical Telemetry frequency band. The ambulatory and bedside device transceivers communicate data to the Central Station through the Access Point transceiver. In addition, the transceivers are capable of receiving control commands for self-use or connection transfer.

The PatientNet ambulatory transceiver is the DT-4500. This transceiver is worn by the patient and usually carried in a gown pocket or pouch, and used with a 3, 4, or 5-wire leadset connected to the electrodes on the patient. The DT-4500 is IPX7 compliant, so it can be submerged in 1m of water for up to 30 minutes.

The DT-7000 and DT-7001 are the PatientNet bedside-device transceivers and are physically connected to bedside monitors (other manufacturers' bedside monitors and NPB 7200 series ventilator).

The DR-10000 Access Point transceiver collects data from the ambulatory and bedside transceivers, sends that data to the Central Station, and transmits control data to the transceiver devices.

WARNING: Remove transceivers from patients before MRI and CAT scan procedures, and store the transceivers outside the room where such equipment is located. Close proximity to MRI or CAT scan equipment may result in damage to transceivers.

Programming Transceivers

Before a transceiver can be used with the PatientNet System, it must first be programmed with a Network Number and Monitor I.D. number to match the corresponding Central Station. Consult your facility's system administrator to perform these functions.

WARNING: When programming the DT-4500 through the External Serial Device (I/O) connector, it must be disconnected from the patient. The accessory connector shall be kept covered when not in use with the supplied protective cover. Failure to follow these instructions could lead to excessive voltages and currents being applied to the patient, resulting in cardiac arrest.

The corresponding Central Station must also be programmed to this Network Number and Monitor Identification number.

If you have any questions about the programming status, contact your system administrator.

Displaying Transceiver Status

You can view transceiver status on the Central Station, but not on the PatientNet Viewers (also known as the IRVS and RVS).

1. Press **System** on the Main screen.
2. Press **OpenNet** button on the Passcode screen to display the OpenNet Status screen (fig. 90).

OpenNet Transceiver Status														
Ch	Module	ID	Device ID	Batt	LL	RA	LA	Ua	Ub	Uc	Freq Resp	Link	Device Type	Version TX RF
65	0	3	9.5	0	0	0	0	-	-	-	0.5Hz	OK	4 Wire	1.2 55.123
66	0	3	9.5	0	0	0	0	-	-	-	0.5Hz	OK	3 Wire	1.2 55.123
67	0	3	9.5	0	0	0	0	-	-	-	0.5Hz	OK	Pass 2	1.2 55.123
68	0	3	9.5	0	0	0	0	-	-	-	0.5Hz	OK	3 Wire	1.2 55.123
69	0	3	9.5	0	0	0	0	-	-	-	0.5Hz	OK	Pass 2	1.2 55.123
70	0	3	9.5	0	0	0	0	0	-	-	0.5Hz	OK	5 Wire	1.2 55.123
71	0	3	9.5	0	0	0	0	0	-	-	0.5Hz	OK	5 Wire	1.2 55.123
72	0	3	9.5	0	0	0	0	0	-	-	0.5Hz	OK	5 Wire	1.2 55.123

Assign TX
Program AT
Program IT
Exit

Fig. 90. OpenNet Status screen

Your System Administrator enables or disables the **Assign TX**, **Program AT**, and **Program IT** buttons. Check with your System Administrator for more information.

The columns in the OpenNet Status screen are described below.

Ch	channel number
Module ID	Transceiver ID that the channel is currently set to
Device ID	Device ID of the transceiver from which the channel receives data
Batt	battery voltage of the ambulatory transceiver
LL	left leg electrode impedance value
RA	right arm electrode impedance value
LA	left arm electrode impedance
Va	chest electrode impedance value
Vb, Vc	miscellaneous electrode impedance value
Freq Resp	frequency response programmed into each transceiver
Link	link between the transceiver and the Central Station: OK: the Central Station is receiving data from the transceiver OFF: the Central Station is <i>not</i> receiving data
Device Type	transceiver or bedside monitor associated with the channel
Version TX	the download firmware version currently in the transceiver
Version RX	the RF Module firmware version currently in the transceiver
Ch	channel number

Impedance Values

The DT-4500 Ambulatory Transceivers electrode impedance values, which are displayed on the OpenNet Transceiver Status screen, indicate the *quality* of the signal connection and are not the *actual* impedance values that are measured by the system. The electrode impedance values range from 100 to 200 (optimal). The typical values range between 180 and 200.

If the electrode's impedance value is greater than the defined Quality Threshold value, then its LED is illuminated. The DT-4500 stores the Quality Threshold value and uses this value to determine whether or not the electrode LED should be illuminated when the Attendant Present buttons are pressed. See Figure 91 on page 190 for details on the DT-4500 Buttons and LED indicators.

Note: A Lead Off alarm will occur when an electrode's impedance value drops to 150 (150 is the default Loose Lead Threshold value).

User Warnings, Cautions, and Notes

Before operating the WMTS transceivers, read and follow all warnings and cautions presented in this section.

Warnings

1. Do not use the output of the DT-4500 as a synchronization source for cardiac defibrillation. Delays in presentation of the R-Wave may be as much as 40 milliseconds.
2. Do not monitor pacer patients with a 3 wire leadset when reliable pacer detection is required. Pacer pulse detection can be erratic when only a single vector is monitored. Always use a 5 wire leadset when reliable pacer detection is required.
3. The DT-4500 is a type BF patient applied device. It is not suitable for direct cardiac application, for use in the operating room, or during cardiac surgery. Use in these environments could cause hazardous voltages and currents being applied to the patient's heart, resulting in cardiac arrest.
4. Only authorized type BF devices can be plugged into the accessory connector of the DT-4500 when it is applied to the patient. The accessory connector must be kept covered when not in use with the supplied accessory connector cover. Failure to follow these instructions could lead to hazardous voltages and currents being applied to the patient resulting in cardiac arrest.
5. Total submersion of the patient worn transceiver and/or patient leadset/antenna may severely limit its transmission range causing loss of patient monitoring. When subjecting the patient and transceiver to submersion, he/she should be carefully monitored to ensure that there is no signal loss.
6. Use only VitalCom Power Supply Part Number 395005 with the DT-7000/DT-7001.

7. **The DT-7000/DT-7001 is not designed as a patient contact device. Per FCC rules, the DT-7000/DT-7001 must reside more than 20 cm (7.9 inches) from the patient.**

Cautions

1. **Any changes or modifications to the device that are not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment.**
2. **Electromagnetic interference or power overload, due to electrosurgical or diathermy instruments, may damage the device.**

Notes

1. This equipment has been tested and found to comply with the limits for a CLASS B digital device, pursuant to Part 15 of the FCC Rules and CISPR 11. These limits are designed to provide reasonable protection against harmful interference. This equipment generates, uses, and radiates radio frequency energy, and, if not installed and used in accordance with the instructions contained in this manual, may cause harmful interference to radio and television communications. However, there is no guarantee that interference will not occur in a particular installation.

If this equipment does cause harmful interference, then the user is encouraged to try to correct the interference by one or more of the following measures:

- Move the DT-4500, the bedside device with a DT-7000 or DT-7001, or the device being interfered with, to increase the separation between the two.

Note: Do not attempt to move fixed antennas as this can negatively impact the PatientNet System's operation.

- Connect the equipment into an outlet on a different circuit.
 - Contact your technical service representative for assistance.
2. To ensure that the use of this product does not contribute to interference, it is necessary to use shielded I/O cables. Connecting this device to peripheral devices that do not comply with the CLASS B requirement or using an unshielded peripheral data cable could result in harmful interference to radio or television reception.
 3. The DT-4500, DT-7000, and DT-7001 should be disposed of at the end of their useful life per applicable regulations.

Ambulatory Transceiver (DT-4500)

The DT-4500 transceiver is a battery-operated ambulatory transceiver worn by the patient and used with a 3, 4, or 5-wire leadset that is connected to electrodes on the patient. The transceiver is available to patients who are not confined to a bed, but still require constant monitoring of their ECG waveforms.

Operating Instructions**Push Button Function and Use**

See Figure 91 on page 190 for an image of the DT-4500 controls and LED indicators.

External Serial Devices (I/O) Connector

The External Serial Device (I/O) connector allows an external serial device or programming cable to connect and maintain a logical communication link between the DT-4500 and the Central Station. See page 183 for details on programming the DT-4500 through the I/O connector.

ECG Leadset Connector

The ECG leadset connector allows the ECG leadset to attach to the DT-4500 and maintain a logical communication link between the DT-4500 and the Central Station. See page 195 for details on attaching the ECG leadset to the DT-4500 through the ECG leadset connector.

Remote Record

When depressed, the Remote Record function button will initiate a strip chart recording at the Central Station.

Nurse Call

When depressed, the Nurse Call function button will initiate a Nurse Call Alarm at the Central Station.

Attendant Present / Procedure Alarm Silence (PAS) Unlock Button

The **Attendant Present/PAS Unlock Buttons** consists of two buttons that are located on either side of the transceiver (See Figure 91 on page 190). The Attendant Present push buttons have three functions. Each function is initiated based on how long the buttons are pressed.

- 1. Lead Quality**

Pressing both Attendant Present buttons simultaneously will illuminate the LEDs for each lead that has a minimum level of quality.

- 2. Initiating an Attendant Present Alarm**

Once the transceiver is in the Power-On Mode, pressing the Attendant Present buttons will activate the **Attendant Present** function and initiate an Attendant Present Alarm at the Central Station.

- 3. Unlocking the PAS button**

The PAS button must be unlocked or enabled prior to initiating the Procedure Alarm Silence button. In the “locked” position, the PAS button is disabled.

To “unlock” the PAS button, press, and hold (for about two seconds), the Attendant Present buttons until the Procedure Alarm Silence Status Indicator LED begins flashing. Once the LED indicator starts flashing, the PAS button is in the “unlocked mode” and functional.

Note: The PAS button must be pressed while the LED is still flashing. If it is pressed after the LED has stopped flashing, then the PAS button will automatically be “re-locked”.

Procedure Alarm Silence (PAS) Button

Depressing the PAS button, while the PAS Status Indicator LED is flashing, informs the clinicians at the Central Station area that the attending nurse will be performing a procedure to the patient that may cause inadvertent false alarms at the Central Station (i.e. changing lead wires, electrodes, etc.)

Once the PAS button is pressed, the following events occur at the Central Station.

1. A timer is displayed in the fourth patient block configurable field that displays the length of Procedure Alarm Silence time remaining on the transceiver.

CAUTION: All non-level one alarms are ignored while the PAS alarm is active.

2. "PA SILENCE" is denoted in Full Disclosure for the duration of the PAS period.

Once the PAS button is pressed, the DT-4500 enters the PAS Mode with the following indications:

1. The active time is set for 120 seconds and begins counting down.
2. The active time is transmitted to the Central Station.
3. The PAS Status LED indicates the time remaining through its flash speed. The LED flash speed increases as the PAS time remaining decreases from 120 seconds to 0 seconds.
4. The attendant can reset the PAS active time to 120 seconds by pressing both Attendant Present buttons again.

The Procedure Alarm Silence alarm remains active until one of the following conditions occur:

1. The transceiver no longer sends the procedure alarm silence indicator to the Central Station.
2. A level one alarm is detected and triggered at the Central Station
3. The patient tile alarm text area is clicked on. All alarms are set to ON once this area is clicked.
4. The attendant presses the PAS button while PAS is active. This will automatically cancel the 120 second PAS at the Central Station, and will re-enable the audible alarm tone.

Note: The PAS feature can be disabled by the System Administrator.




Fig. 91. DT-4500 Controls and LED Indicators

Note: *The External Serial Device (I/O) Cable must be removed and the Connector must be covered whenever the DT-4500 is connected to the patient.

LED Indicators Function

Upon Power-On, all LED indicators are illuminated for a brief period. After the specified time period, only those LEDs displaying positive (or negative) transceiver functions, as described in each section below, remain illuminated.

Procedure Alarm Silence Status Indicator

The Procedure Alarm Silence Status Indicator is illuminated when the PAS function is active. The LED flashes while the Procedure Alarm Silence button is unlocked or the PAS active time is running low. The PAS button can only be pressed and activated during this unlocked phase. Refer to the section on “Procedure Alarm Silence (PAS) Button” on page 189 for more information.

External Serial Devices (I/O)

The External Serial Device (I/O) LED is illuminated when an external serial device is connected, detected and maintaining a logical communication link.

Note: When illuminated, be sure that the device is not connected to the patient.

Low Battery (BATT)

The Low Battery (BATT) LED is illuminated while the battery voltage remains good; however, the LED flashes when the battery voltage falls below a predetermined value. When the battery power falls below a predetermined value, then the transceiver will automatically power itself off.

RF link (RF)

The RF link indicator is illuminated while there is RF communication between the DT-4500 transceiver and the Central Station. The LED flashes if there is communication between the DT-4500 transceiver and the Access point, but not the Central Station.

Electrode Status Indicators (RA, LA, RL, LL, V₁)

Each ECG electrode wire is named, color coded (Table 21), and represented by an LED indicator. Each LED is illuminated with a solid light when the electrode is fully active, and is off when no electrode signal is present.

Table 21 Electrode Colors

Electrode Name	Wire Color
RA	White
LA	Black
RL	Green
LL	Red
V1	Brown

The Quality Threshold value has a default setting. If the electrode's impedance value is greater than the default, then its LED is illuminated. The DT-4500 stores the Quality Threshold value and uses this value to determine whether or not the electrode LED should be illuminated when the Attendant Present buttons are pressed. See "Impedance Values" on page 185 for details.

Note: The V6 indicator will be available in future releases; therefore, the V6 LED will not illuminate when the Attendant Present buttons are pressed.

Cleaning

This section provides cleaning and maintenance instructions for DT-4500 transceivers.

Read and follow all precautions when cleaning transceivers.

WARNING: No claims are made concerning the sterility of the DT-4500 Ambulatory Transceivers.

CAUTION: Do not sterilize any part of the transceivers. Gas sterilization, autoclaving, liquid immersion, and other sterilization methods can cause serious damage to the devices that may not be obvious to the user.

Note: DO NOT use abrasive cleaners.

Cleaning the Chassis

The following applies to cleaning the DT-4500.

- The DT-4500 can be cleaned with the patient cable attached, however please ensure that the cleaning agents used to clean the DT-4500 are compatible with the cleaning agents listed for the ECG cable on page 194, or else ensure that the ECG cable does not come into contact with the cleaning agents for the DT-4500.
- To clean around the ECG connector, remove the ECG cable from the unit.

CAUTION: Prior to cleaning the battery compartment and transceiver chassis, remove the battery from the unit.

1. Remove the battery from the transceiver and inspect the battery compartment after each use. Close the battery door.

CAUTION: Prior to rinsing the DT-4500, make sure that the battery compartment door is properly closed and sealed.

2. Transceivers can be cleaned with a gauze pad or cloth moistened with one of the following agents:
 - Soap and Water
 - Quaternary Ammonium
 - Glutaraldehyde 2%
 - Dilute Chlorine Bleach (Sodium hypochlorite), 10% solution, freshly made in past 24 hours
 - Isopropyl Alcohol 70%
 - Ethyl Alcohol
3. Use a cloth moistened with distilled water to rinse away the cleaning solution.
4. Dry thoroughly with a lint-free cloth.

Note: Once a month, or whenever the DT-4500 is submersed or subjected to a stream of liquid, remove the accessory connector cover and remove any moisture that may have collected inside.

Cleaning the Battery Compartment

CAUTION: When cleaning the battery compartment, use only soap, water, or alcohol. Do not use any other cleaning agents inside the battery compartment as they may damage the battery compartment.

CAUTION: Make sure to rinse all cleaned surfaces with distilled water to remove any cleaning agent residue. Dry off the battery contact leads. Ensure that the battery compartment is dry before inserting the battery into the unit.

Under normal operation, the battery compartment should not require frequent cleaning. If the battery compartment does require cleaning, then use the following instructions.

1. Remove the battery from the battery compartment.
2. Clean the transceiver with a gauze pad or cloth moistened with one of the following agents:
 - Water
 - Soap
3. Use a cloth moistened with distilled water to rinse away the cleaning solution.
4. Dry thoroughly with a lint-free cloth. Allow the battery compartment to air dry completely prior to closing the compartment door.

Cleaning the ECG Leadsets

The transceiver ECG Leadsets are manufactured by Affinity Medical.

Contact your technical support representative for additional leadsets.

Warnings

1. **Do not use leadsets which exhibit signs of wear or damage such as cracking or degradation of the connectors or cable insulation.**
2. **Do not sterilize using steam or gamma radiation. Damage to the leadsets will result.**

Cautions

1. **To increase the life of the leadsets, do not pull on the leadsets to disconnect. Pull gently by grasping the connectors.**
2. **Do not immerse the leadsets in water or other liquid to clean. Immersion may cause damage to the leadsets.**
3. **Repeated exposure to EtO sterilization will shorten the effective life of the leadset. The leadsets should be sterilized only when indicated by specific patient or hospital requirements.**

Cleaning

1. Wipe the leadset with a solution of soap and water.
2. Use a cloth moistened with distilled water to rinse away the cleaning solution.
3. Dry thoroughly with a lint-free cloth.

Disinfecting

Use hospital-approved disinfecting procedures such as those recommended by AAMI or AORN.

1. Wipe the leadset with a fresh 10% solution of chlorine bleach and water or a 2% glutaraldehyde solution such as Cidex.
2. Use a cloth moistened with distilled water to rinse away the cleaning solution.
3. Dry thoroughly with a lint-free cloth.

Sterilization

Leadsets may be sterilized by EtO, when indicated. Use the hospital-approved procedure for EtO sterilization, such as those recommended by AAMI. The Leadsets are designed to remain effective after up to 10 exposures to EtO sterilization Cycles.

Use and Maintenance**Transceiver Storage**

Store the transceiver with the leadset attached and hanging freely. If that is not possible, then wrap the leadset loosely around the transceiver. Wrapping the leadset tightly around the transceiver can damage the wires.

Note: The DT-4500 Ambulatory Transceiver contains no user-serviceable parts. Thus, maintenance service is not needed.

Attaching and Removing a Leadset from the Transceiver

To attach, carefully grasp the leadset connector cover, holding it with the small knob facing upward, and push the leadset into the ECG lead wire connector. Make sure that the leadset is completely inserted into the connector and is flush with the DT-4500 chassis.

To remove the leadset, grasp hold of the sides of the leadset connector cover and pull straight out. If the leadset is difficult to remove, then you can slightly move the leadset cover side-to-side until it is released.

Internal Antenna

The DT-4500 transmits in the 608-614 MHz frequency range. The omnidirectional antenna is a part of the leadset system, with each lead wire paired with an antenna wire. Transceiver output power and system operation requirements are defined by the FCC. **Therefore, it is essential that the leadset provided not be modified or altered in any way.**

Installing and Removing a Battery

See “Specifications” on page 247 for acceptable batteries.

Note: Battery service life can be substantially improved by using nine-volt lithium batteries.

To install a 9V transceiver battery, first open the transceiver battery compartment by placing your thumb and forefinger on the compartment latch and flipping it open. Inspect the battery compartment and insure that there is no foreign object present that could block the battery contact or short the battery terminals. Next, place a 9V battery inside the compartment with the prongs touching the compartment contacts. The orientation of the battery prongs against the contacts does not matter, so long as the prongs and contacts are touching. Finally, close the battery compartment door by pressing it until the latch clicks into place and the compartment is secure.

To remove a battery, simply follow the installation steps listed above and discard the used battery per applicable regulations.

Warnings

1. ECG lead wires must be dressed and secured to the patient to prevent the possibility of them encircling the patient’s neck and causing strangulation.
2. When installing or replacing the battery, visually inspect the battery compartment and ensure that there are no foreign objects inside. A conductive object making contact with the battery contacts could cause the battery and battery compartment to overheat, resulting in burns to the patient and to the attendant removing the battery.
3. A foreign object blocking battery contact with the DT-4500 could prevent its operation resulting in failure to monitor the patient.
4. Always perform a battery check procedure after installing or replacing the battery.
5. Lithium Batteries may explode if mistreated. DO NOT recharge, disassemble, or dispose of batteries in fire.

Cautions

1. Transceivers should be carried securely in pouches or in a pocket of a patient’s gown. If the weight of the transceiver pulls on the wires, then the wires can be damaged or worn
2. Make sure that the wires are not twisted around each other; since, this can interfere with transmission and produce noise.
3. Make sure that the lead wires are not inadvertently pinched in the bed rails. This may cut the insulation or break the leadset.

Instrument Transceiver (DT-7000, DT-7001)

WARNING: When using a bedside device with the instrument transceiver (DT-7000/7001), the bedside device is the primary monitor and alarm source. Disabling alarms on the bedside device is not safe clinical practice.

The DT-7000 and DT-7001¹ send data and alarm information from bedside monitors and NPB 7200 series ventilators to the Central Station.

The transceivers support the bedside monitors shown in the PatientNet Customer Release Notes. See your system administrator for details.

Power to the DT-7000 is provided through one of the following:

- the AC power adapter, which provides continuous power
- the bedside monitor, which provides continuous power

Power to the DT-7001 is provided through one of the following:

- the AC power adapter, which provides continuous power
- the bedside monitor, which provides continuous power
- the internal battery, which is replaceable by qualified service technicians

Note: See your hospital's Service Department for battery replacement.

Operating Instructions

The DT-7000 and DT-7001 appearance and functionality are equivalent; however, only the DT-7001 is capable of using an internal battery as a power source.

Push Button Function and Use

See Figure 92 on page 200 for an image of the DT-7000/DT-7001 controls and LED indicators.

External Serial Devices (I/O) Ports

The External Serial Device (I/O) ports allow external serial devices or programming cables to connect and maintain logical communication links between the DT-7000/DT-7001 and the Central Station.

Note: *External Serial Device (I/O) Port 1 is currently functional. I/O Ports 2, 3 and 4 will be functional in future product releases.

1. The DT-7001 will be available in future releases. Please contact your sales representative for the device availability.

Remote Record

When depressed, the Remote Record function button will initiate a strip chart recording at the Central Station.

Nurse Call

When depressed, the Nurse Call function button will initiate a Nurse Call Alarm at the Central Station.

Power Button

Pressing the Power button either places the transceiver in or out of Standby Mode.

Attendant Present / Procedure Alarm Silence (PAS) Unlock Button

The Attendant Present push button has two functions. Each function is initiated based on how long the button is pressed.

Note: The LEDs are constantly illuminated when powered by an external source.

1. Initiating an Attendant Present Alarm

Once the transceiver is out of the Standby Mode, pressing the Attendant Present button will activate the **Attendant Present** function and initiate an Attendant Present Alarm at the Central Station.

2. Unlocking the PAS button

The PAS function must be enabled at the Central Station prior to initiating the Procedure Alarm Silence alarm at the DT-7000/DT-7001.

In the “locked” position, the PAS button is disabled.

To “unlock” the PAS button, press, and hold (for about two seconds), the Attendant Present button until the Procedure Alarm Silence Status Indicator LED begins flashing. Once the LED indicator starts flashing, the PAS button is in the “unlocked mode” and functional.

Note: The PAS button must be pressed while the LED is still flashing. If it is pressed after the LED has stopped flashing, then the PAS button will automatically be “re-locked”.

Procedure Alarm Silence (PAS) Button

Depressing the PAS button, while the PAS Status Indicator LED is flashing, informs the clinicians at the Central Station area that the attending nurse will be performing a procedure to the patient that may cause inadvertent false alarms at the Central Station (i.e. changing lead wires, electrodes, etc.)

Once the PAS button is pressed, the following events occur at the Central Station.

1. A timer is displayed in the fourth patient block configurable field that displays the length of Procedural Alarm Silence time remaining on the transceiver.

CAUTION: All non-level one alarms are ignored while the PAS alarm is active.

2. "PA SILENCE" is denoted in Full Disclosure for the duration of the PAS period.

Once the PAS button is pressed, the DT-7000/7001 enters the PAS Mode with the following indications:

1. The active time is set for 120 seconds and begins counting down.
2. The active time is transmitted to the Central Station.
3. The PAS Status LED indicates the time remaining through its flash speed. The LED flash speed increases as the PAS time remaining decreases from 120 seconds to 0 seconds.
4. The attendant can reset the PAS active time to 120 seconds by pressing both Attendant Present buttons again.

The Procedure Alarm Silence remains active until one of the following conditions occur:

1. The transceiver no longer sends the procedure alarm silence indicator to the Central Station.
2. A level one alarm is detected and triggered at the Central Station
3. The patient tile alarm text area is clicked on. All alarms are set to ON once this area is clicked.
4. The attendant presses the PAS button while PAS is active. This will automatically cancel the 120 second PAS at the Central Station, and will re-enable the audible alarm tone.




Fig. 92. DT-7000/DT-7001 Controls, I/O Ports, and LED Indicators

Note: *Only Port 1 of the External Serial Device (I/O) is functional. I/O Ports 2, 3 and 4 will be functional in future product releases.

LED Indicators Function

Once the transceivers exit Standby Mode, either by pressing the Attendant Present or Power buttons, all LED indicators are illuminated for a brief period. After the specified time period, only those LEDs displaying positive (or negative) transceiver functions, as described in each section below, remain illuminated.

Procedure Alarm Silence Status Indicator

The Procedure Alarm Silence Status Indicator is illuminated when the PAS function is active. The LED flashes while the Procedure Alarm Silence button is unlocked or the PAS active time is running low. The PAS button can only be pressed and activated during this unlocked phase. Refer to the section on “Procedure Alarm Silence (PAS) Button” on page 189 for more information.

External Serial Devices (I/O)

The External Serial Device (I/O) LEDs are labeled 1-4 and are each illuminated when there is an external serial device connected, detected, and maintaining a logical communication link to the corresponding I/O data port (fig. 92).

Low Battery (BATT)

Note: The Low Battery LED is only functional on the DT-7001.

The Low Battery (BATT) LED is illuminated while the battery voltage remains good; however, the LED flashes when the battery voltage falls below a predetermined value. When the battery power falls below a predetermined value, then the transceiver will automatically power itself off.

RF Link (RF)

The RF link indicator is illuminated while there is RF communication between the DT-7000 and DT-7001 transceivers and the Central Station. The LED flashes if there is communication between the transceivers and the Access point, but not the Central Station.

Power Adapter

The Power Adapter LED is illuminated when the transceiver is powered from an external power source that is connected to the Power Connector (fig. 92), and not one of the I/O ports.

Cleaning

This section provides cleaning and maintenance instructions for DT-7000 and DT-7001 transceivers.

Read and follow all precautions when cleaning transceivers.

WARNING: No claims are made concerning the sterility of the DT-7000 and DT-7001 Instrument Transceivers.

CAUTION: Do not sterilize any part of the transceivers. Gas sterilization, autoclaving, liquid immersion, and other sterilization methods can cause serious damage to the devices that may not be obvious to the user.

Note: DO NOT use abrasive cleaners.

Cleaning the Chassis

1. Transceivers can be cleaned with a gauze pad or cloth moistened with one of the following agents:
 - Soap and Water
 - Quaternary Ammonium
 - Glutaraldehyde 2%
 - Dilute Chlorine Bleach (sodium hypochlorite), 10% solution, freshly made in past 24 hours
 - Isopropyl Alcohol 70%
 - Ethyl Alcohol
2. Use a cloth moistened with distilled water to rinse away the cleaning solution.
3. Dry thoroughly with a lint-free cloth.

Use and Maintenance

Connecting to the Bedside Monitor

See Figure 92 on page 200 for an image of the DT-7000/DT-7001 controls and LED indicators.

Note: Before the transceiver is connected to the bedside monitor, the system administrator must program it with the bedside device specific software module.

1. Attach the transceiver to the bedside monitor by sliding the Device Hook over the Mounting Disk until the transceiver snaps into place. The Mounting Disk is provided with the transceiver and is attached to the bedside device through adhesive or hardware tools.
2. Connect the AC power adapter into the power port located on the bottom of the transceiver.
3. Plug the AC power adapter into the wall electrical outlet. If the AC power adapter is not used, the transceiver will operate either from its internal battery (DT-7001 models) or, on some bedsides, from connection to the bedside device.
4. Attach the host end of the I/O cable to I/O port 1 (Ports 2, 3 and 4 will be functional in future releases).
5. Attach the other end of the I/O cable to the bedside monitor.

This page is intentionally left blank.