

EXHIBIT 6

TEST EQUIPMENT USED

Part 15.231, ANSI C63.4 , RSS 210

This is a list of all test equipment used.

Test Equipment list for Honeywell OATS & Conducted Line:

Equipment	Mfg	Model	Cal Date	Cal Due
Spectrum Analyzer	Rohde & Schwarz	FSEA20	10/15/12	10/15/13
Antenna ('Biconilog')	ETS Lindgren	3149	08/25/12	08/25/13
NOTE: THIS ANTENNA WAS CALIBRATED DOWN TO 26 MHz. PER OUR REQUEST.				
Surge Suppressor	Agilent (HP)	HP11947A	05/09/12	05/09/13
LISN	Com-Power	LI-115	10/19/12	10/19/13

PLEASE SEE PAGE 2-5 FOR TEST EQUIPMENT TRACEABILITY

If you need any additional information from Honeywell please contact:

Greg Barbato RF Engineer

(Acting for Ken Eskildsen)

Phone (Direct): (516) 577-5863

Email: [greg.barbato@honeywell.com](mailto:greg.barbato@honeywell.com)

# Certificate of Calibration

Issue Date: 10/15/2012



General Calibration, Inc.  
2 Mars Court, Boonton, New Jersey 07005  
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: MR-22292  
Work Order #: MR590  
Customer #: 001464

## Performed By:

GENERAL CALIBRATION, INC.  
2 MARS COURT

BOONTON, NJ 07005

## Location of Calibration:

HONEYWELL SECURITY (001464)  
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

## Equipment Information:

Job No.: 018674  
Manufacturer: R&S  
Description: SPECTRUM ANALYZER  
Department: ALARMNET  
Temp./RH: 22.0 C / 51.0 %  
Cal. Interval: 12 MONTHS  
Cal Date: 10/15/2012

## Purchase Order: 5294448

Asset Tag No.: 10503  
Model Number: FSEA20  
Serial Number: DE31589  
Inspected By: MR1  
Job Title: METROLOGIST  
Calibration Result: PASSED  
Cal. Due Date: 10/15/2013

## Procedure #GCP: RS FSEA20

## Calibration Notes:

Condition: Found In Tolerance and Left In Tolerance

## Standards Used To Calibrate Equipment

Company	I.D.	Description	Serial Number	Cal. Due Date
GENERAL CALIBRATION	1031	SYNTHESIZED SWEEPER, 26GHZ	3050A02181	04/12/2013
GENERAL CALIBRATION	1124	FUNCTION GENERATOR	2847A11791	05/31/2013
GENERAL CALIBRATION	200	MEASURING RECEIVER	3438A05277	08/08/2013
GENERAL CALIBRATION	332	FREQUENCY COUNTER	CCN200100126	05/21/2013
GENERAL CALIBRATION	403	ATTENUATOR	219-05771	08/15/2013
GENERAL CALIBRATION	432	CALIBRATION KIT TYPE N 50-OHMS	2541A00233	10/29/2012
GENERAL CALIBRATION	434	POWER SPLITTER	11124	10/11/2013
GENERAL CALIBRATION	511	DIGITAL MULTIMETER	2201A04803	04/04/2013
GENERAL CALIBRATION	518	POWER SENSOR	US37292728	06/22/2013
GENERAL CALIBRATION	588	ATTENUATOR	2522A40468	07/18/2013

This is to certify that General Calibration, Inc. is A2LA accredited and that its calibration system is in compliance with ISO/IEC 17025-2005, ANSI NCSL Z540-1, ANSI NCSL Z540-3, and ISO 9001:2008. The test limits stated in the report correspond to the Manufacturer calibration and published specifications of the equipment, at the points tested. Calibration of standards; reference standards and intermediate standards in this calibration have been checked and calibrated against the above working standard(s) which are traceable to the National Institute of Standards and Technology. Any accredited calibration under our A2LA scope of accreditation is denoted with A2LA accredited symbol on the calibration certificate. If the certificate includes any unaccredited items calibrated, they are clearly marked as unaccredited calibration.

Best Uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2.

The results documented in this certificate relate only to the item(s) calibrated or tested.

Approved By

General Calibration, Inc. - Q. A. Manager





# ETS-LINDGREN™

An ESCO Technologies Company

1301 Arrow Point Drive  
Cedar Park, Texas 78613  
(512) 531-6498

Cert I.D.: 93421

## Certificate of Calibration Conformance

Page 1 of 5

The instrument identified below has been individually calibrated in compliance with the following standard(s):

SAE, ARP-958 - 2003, Electromagnetic Interference Measurement Antennas; Standard Calibration Method, Society of Automotive Engineers, Aerospace Recommended Practice. Fixed height, three antenna rotation, 1 meter separation. 3 meter separation performed per Annex C. Vertical calibration performed per above listed methodology.

Environment: Laboratory MTE is maintained in a temperature controlled environment with ambient conditions from 18 to 28 C, relative humidity less than 90%. The instrument under test has been calibrated on an open air test site (OATS) with environment temperature conditions ranging from 0 to 40 C which has no known influences on measurement quality.

<b>Manufacturer:</b>	ETS-Lindgren	<b>Operating Range:</b>	80 MHz - 6 GHz
<b>Model Number:</b>	3149.	<b>Instrument Type:</b>	Biconilog (Type 5)
<b>Serial Number/ ID:</b>	00045682	<b>Date Code:</b>	
<b>Tracking Number:</b>	S 000026017	<b>Alternate ID:</b>	11242
<b>Date Completed:</b>	25-Sep-12	<b>Customer:</b>	HONEYWELL (NY)
<b>Test Type:</b>	3 meter, Horizontal and Vertical		
<b>Calibration Uncertainty:</b>	01m	26 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.2 dB	
k=2, (95% Confidence Level)	03m	26 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.3 dB	
	10m	26 - 1000 MHz, +/-1.0 dB; 1000 - 2000 MHz, +/-1.4 dB; 2000 - 6000 MHz, +/-2.3 dB	

### Test Remarks:

Calibration Traceability: All Measuring and Test Equipment (M/TE) identified below are traceable to the SI units through the National Institute for Standards and Technology (NIST) or other recognized National Metrology Institute. Calibration Laboratory and Quality System controls are compliant with ISO/IEC 17025-2005 and ANSI/NC SL Z540-1-1994.

### Standards and Equipment Used:

#### Make / Model / Name / S/N / Recall Date

Agilent	N5230C	PNA-L Net/Wrk Analyzer	MY49002145	13-Jul-13
---------	--------	------------------------	------------	-----------

#### Condition of Instrument Upon Receipt:

In Tolerance to Internal Quality Standards

#### On Release:

In Tolerance to Internal Quality Standards

James Hansell

Calibration Completed By  
James Hansell, Calibration Technician

Terry D. O'Neill

Attested and Issued on 25-Sep-12  
Terry D. O'Neill, Calibration Manager



# Certificate of Calibration

Issue Date: 5/9/2012



General Calibration, Inc.  
2 Mars Court, Boonton, New Jersey 07005  
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: MR-21481  
Work Order #: MR554  
Customer #: 001464

## Performed By:

GENERAL CALIBRATION, INC.

## Location of Calibration:

HONEYWELL SECURITY (001464)

2 MARS COURT  
BOONTON, NJ 07005

2 CORPORATE CENTER DRIVE  
MELVILLE, NY 11747

## Equipment Information:

Job No.: 096188  
Manufacturer: HP  
Description: TRANSIENT LIMITER  
Department:  
Temp/RH: 22 C / 45 %  
Cal. Interval: 12 MONTHS  
Cal Date: 05/09/2012

## Purchase Order: 5294448

Asset Tag No.: 10131  
Model Number: 11947A  
Serial Number: 3107A02782  
Inspected By: MRI  
Job Title: METROLOGIST  
Calibration Result: PASSED  
Cal. Due Date: 05/09/2013

Procedure #GCP: HP 11947A

## Calibration Notes:

Condition: Found In Tolerance and Left In Tolerance

## Standards Used To Calibrate Equipment

Company	I.D.	Description	Serial Number	Cal. Due Date
GENERAL CALIBRATION	593	SPECTRUM ANALYZER	3017A05102	08/24/2012
GENERAL CALIBRATION	907	CALIBRATION KIT, 26.5 GHZ, 3.5 MM	3101A05556	02/26/2013
GENERAL CALIBRATION	970	FUNCTION GENERATOR	2847A07354	10/04/2012

This is to certify that General Calibration, Inc. is A2LA accredited and that its calibration system is in compliance with ISO/IEC 17025-2005, ANSI NCSL Z540-1, ANSI NCSL Z540-3, and ISO 9001:2008. The test limits stated in the report correspond to the Manufacturer calibration and published specifications of the equipment, at the points tested. Calibration of standards, reference standards and intermediate standards in this calibration have been checked and calibrated against the above working standard(s) which are traceable to the National Institute of Standards and Technology. Any accredited calibration under our A2LA scope of accreditation is denoted with A2LA accredited symbol on the calibration certificate. If the certificate includes any unaccredited items calibrated, they are clearly marked as unaccredited calibration.

Best Uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ .

The results documented in this certificate relate only to the item(s) calibrated or tested.

Approved By Robert A. Hoff  
General Calibration, Inc. - Q. A. Manager

# Certificate of Calibration

Issue Date: 10/19/2012



General Calibration, Inc.  
2 Mars Court, Boonton, New Jersey 07005  
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: MR-22320  
Work Order #: MR590  
Customer #: 001464

## Performed By:

GENERAL CALIBRATION, INC.  
2 MARS COURT

BOONTON, NJ 07005

## Location of Calibration:

HONEYWELL SECURITY (001464)  
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

## Equipment Information:

Job No.: 076905  
Manufacturer: COM-POWER  
Description: LISN  
Department: DAVID KALMUS  
Temp./RH: 22.0 C / 45.0 %  
Cal. Interval: 12 MONTHS  
Cal Date: 10/19/2012

## Purchase Order: 5294448

Asset Tag No.: 11262  
Model Number: LI-115  
Serial Number: 241050  
Inspected By: MR1  
Job Title: METROLOGIST  
Calibration Result: PASSED  
Cal. Due Date: 10/19/2013

## Procedure #GCP: COM-POWER LI-115

### Calibration Notes:

Condition: Found In Tolerance and Left In Tolerance

### Standards Used To Calibrate Equipment

Company	I.D.	Description	Serial Number	Cal. Due Date
GENERAL CALIBRATION	511	DIGITAL MULTIMETER	2201A04803	04/04/2013
GENERAL CALIBRATION	700	DIGITAL MULTIMETER	77820175	02/10/2013

This is to certify that General Calibration, Inc. is A2LA accredited and that its calibration system is in compliance with ISO/IEC 17025-2005, ANSI NCSL Z540-1, ANSI NCSL Z540-3, and ISO 9001:2008. The test limits stated in the report correspond to the Manufacturer calibration and published specifications of the equipment, at the points tested. Calibration of standards; reference standards and intermediate standards in this calibration have been checked and calibrated against the above working standard(s) which are traceable to the National Institute of Standards and Technology. Any accredited calibration under our A2LA scope of accreditation is denoted with A2LA accredited symbol on the calibration certificate. If the certificate includes any unaccredited items calibrated, they are clearly marked as unaccredited calibration.

Best Uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ .

The results documented in this certificate relate only to the item(s) calibrated or tested.

Approved By

*Robert D. McLean*  
General Calibration, Inc. - Q. A. Manager