

ULTRAVIEW DIGITAL TELEMETRY

Contents

General Telemetry Overview	2	Setting Up SpO ₂ Monitoring.....	16
Assigning a Telemetry Channel.....	7	Ensuring Accurate Monitoring.....	18
Tuning a Receiver for a Bedside.....	7	Setting or Adjusting Alarm Limits.....	18
Entering Patient Information.....	7	Setting SpO ₂ Data Averaging Period and	
Acknowledging Signal Loss.....	7	Sampling Interval.....	19
Setting Battery Status Alarms.....	8	Viewing Pulse Rate.....	20
Controlling Patient-Initiated Recordings..	8	SpO ₂ Error Messages.....	20
Digital Telemetry Problem Solving.....	8	SpO ₂ Troubleshooting.....	22
Setting Up ECG Monitoring	9	NIBP Overview	24
Display Detail.....	11	Setting Up NIBP Monitoring.....	24
Monitoring Paced ECG Patients.....	12	Setting Up the ABP Monitor.....	25
Restoring Default Settings.....	12	Setting or Adjusting Alarm Limits.....	28
Changing the Display Resolution.....	13	Displaying Heart Rate Data.....	29
Selecting Options for Lead Display.....	13	Displaying New or Previous Readings..	29
ECG Troubleshooting.....	14	90217 Event Codes.....	30
SpO₂ Overview (90343 only)	16	NIBP Troubleshooting.....	32

General Telemetry Overview

The digital telemetry receiver module (Model 90478) when used in conjunction with Spacelabs Medical telemetry transmitters, PCMS™ or Ultraview™ monitor, and modular receiver housing (Model 90479), provides continuous monitoring of electrocardiographic signals in order to detect abnormal cardiac rhythms, including life-threatening arrhythmias such as asystole, ventricular fibrillation, and ventricular tachycardia. In addition, when used with the digital telemetry multi-parameter transmitter (model 90343), monitoring of electrocardiographic signals is augmented by the availability of continuous or episodic SpO₂ measurements and episodic NIBP measurements.



NOTE

- The transmitters and receivers referred to in this chapter comply with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.
- Changes or modifications not expressly approved by Spacelabs Medical will void the user's authority to operate the equipment.

Transmitters

The transmitter is a small battery-powered device carried by the patient to monitor ECG activity and SpO₂/NIBP (90343 only) data, and transmit this information to the telemetry receiver module.

- The 90343 and 90347 transmit four leads of ECG and use up to five lead wires. However, only two leads may be displayed simultaneously.
- The 90343 is also capable of transmitting numerical NIBP and SpO₂ data. This data is displayed simultaneously with that of the ECG waveform data.

ULTRAVIEW DIGITAL TELEMETRY

Up to five standard disposable silver/silver chloride chest electrodes are connected to the patient: the ECG lead wires are attached to these electrodes and are connected to the transmitter. A patient-operated RECORD button initiates an ECG strip at the system printer if this feature is enabled at the central or bedside.



CAUTION

- This device has a limited bandwidth from .05 to 30 Hz, which may adversely affect the recording of high frequency components in the ECG signal, especially when the morphology of the ECG changes rapidly.
- This device has a limited dynamic range of $\pm 5\text{mV}$, which may render the device vulnerable to saturation by ECG signals with amplitudes higher than 5mV.
- To clean the transmitter use only the following solutions per the manufacturer's labeling: Isopropyl Alcohol (70%), Hydrogen Peroxide, Cidex, Betadine, and Clorox. Use of cleaning solutions other than those listed will VOID the warranty of the digital telemetry transmitter cases.



NOTE

Clean the transmitter after each use. The transmitter does not require any preventative maintenance other than cleaning.

Transmitter Batteries

A 9-Volt alkaline battery (P/N 146-0033-00) is recommended for standard use in the digital telemetry transmitter. A 9-Volt lithium battery (P/N 146-0054-00) may also be used for applications requiring more extended battery service life.

Always observe the battery position and polarity as illustrated at the bottom of the battery compartment. After battery installation, close and latch the compartment cover. The transmitter begins transmitting as soon as the battery is in place.



NOTE

- Whenever the transmitter is not in use, the battery should be removed. Insert a battery only when the transmitter is actually being used with a patient.
- The LOW BATTERY message appears and an alarm tone sounds (if LO BAT is set to ON) when the transmitter battery voltage falls below 7.0 volts. When this message appears, the transmitter has up to 1 hour (depending on transmitter type and selected options) of operating time left, depending on the type of battery used. The low battery indicator will flash on the enhanced digital telemetry transmitters when the transmitter battery voltage falls below 7.0 Volts.

Digital Telemetry Receiver Module

The telemetry receiver module plugs into a bedside, central, or transport monitor, or a digital telemetry module housing. The receiver module receives patient vital sign data from the transmitter. This data is reconstructed by the receiver module, displayed on the monitor and analyzed as described in the *ECG*, *Arrhythmia*, and *ST Analysis* chapters, and in the various SpO_2 and NIBP sections of this chapter.



- *Telemetry systems may be more susceptible to interference than hardwired systems, which may impact patient safety.*
- WARNING** • *Operation of hand-held, wireless telephone equipment (for example, cordless telephones, cellular telephones, etc.) near telemetry systems may cause interference and should be discouraged.*
- *Do not install a telemetry receiver module into a bedside which is currently equipped with any other ECG module, hardwired or telemetry, (or SpO_2 module or NIBP module if the 90343 is operating with that specific receiver module). Doing so may result in inaccurate patient data displays at remote monitors.*

Digital Telemetry Module Housing

The telemetry module housing can hold up to eight separate telemetry receiver modules. Except for the ON/OFF switches, there are no operator controls on the module housing. For normal operation with AC mains power applied, the AC mains indicator light on the front panel of the housing must be illuminated. Operation of the system without AC mains power is limited to ten minutes of battery backup time.

ULTRAVIEW DIGITAL TELEMETRY




Figure Telemetry-1: Ultraview Digital Telemetry System

To set up the central for ECG (if bed name not remembered):

- 1 Touch key label that matches transmitter's frequency
- 2 Select bed/room number for transmitter channel

To set up the central for ECG (PC1 only):

- 1 Touch MONITOR SETUP
- 2 Touch SCREEN FORMAT
- 3 Select zone desired
- 4 Select bed/room number
- 5 Select ECG ON

To tune a receiver module at bedside:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Touch TM SETUP
- 4 Touch SET TM CHANNEL
- 5 Select the digit to change. Use the ↑ ↓ keys to select the value for that digit.
- 6 Repeat for all digits as necessary
- 7 Touch STORE

To admit a patient:

- 1 Touch MONITOR SETUP
- 2 Touch ADMIT/DISCH
- 3 Select bed/room number for channel
- 4 Touch ADMIT
- 5 Select YES
- 6 Select ID, NAME, HEIGHT, WEIGHT, or BSA
- 7 Enter data using pop-up keypad or keyboard
- 8 Touch ENTER
- 9 Repeat steps 6 - 8 until all data has been entered

Assigning a Telemetry Channel

Telemetry transmitters are preassigned with channel frequencies. This channel number is identified on the back of the case and cannot be changed. To receive this telemetry channel, one of the receivers in the telemetry receiver housing must be tuned to its assigned frequency.



NOTE

- Tuning telemetry receiver modules to transmitter channels at the central must be done by a qualified service person.
- Your central can be configured to remember beds that are assigned to individual telemetry channels using the Module Configuration Manager feature. These beds are permanently assigned until you deassign or reassign them.

Tuning a Receiver for a Bedside

The central must be tuned by a qualified service person, but the bedside may be tuned using the ECG TM Setup menu. You may use this menu to tune the receiver module to the pre-assigned channel frequencies on the telemetry transmitter.



NOTE

The module default is set for North America — UHF band operation. If operating in another country you must select the appropriate frequency band using the Module Configuration Manager feature.

Entering Patient Information

The ADMIT/DISCHARGE menu enables you to enter a patient identification (ID) number, name, height, weight, and body surface area (BSA).



NOTE

Admitting a new patient purges data from the previous patient on that telemetry channel.

Acknowledging Signal Loss

When a telemetry signal is lost because the transmitter is out of range or the battery is removed, the receiver initiates a squelch condition. This condition is indicated when a triangular waveform replaces the normal ECG waveform. The notation SQUELCH is included in the edge print for any strip chart recording. The ECG trace automatically begins again if the lost signal returns.

After eight seconds of loss, the IS SIGNAL LOSS PERMANENT? message appears. Selecting NO suspends alarm tones. Selecting YES displays the message DISCHARGE THE PATIENT? Selecting YES again, provides you with the message PURGES DATA-ARE YOU SURE? Selecting YES a third time, discharges the patient from the system and purges all data for that patient. Selecting NO at any point in this sequence returns you to the previous option.

Setting Battery Status Alarms

To control low battery alarms:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Touch TM SETUP
- 4 Select LO BAT ON or OFF

The telemetry battery alarm tone alerts you to a low battery condition in the transmitter. A LOW BATTERY message also appears in the ECG zone involved. You may select to disable the low battery alarm tone, if your bedside or central is configured to do so.

The factory default setting for low battery alarm is ON.

Controlling Patient-Initiated Recordings

To control transmitter's Patient Record function:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Touch TM SETUP
- 4 Select PT RECORD YES or NO

If the Patient Record function is activated (PT RECORD is YES) in the ECG TM SETUP menu, the patient may initiate a recording by pressing the RECORD button on the front of the transmitter.

Digital Telemetry Problem Solving

INTERMITTENT SIGNAL LOSS

This message indicates that the patient may be out of antenna range, or that the battery is depleted.

- Return the patient into antenna range.
- Check that the battery is functioning properly.

LOW BATTERY

The battery is weak. After this message appears, the battery has from a few hours to 24 hours of charge left (depending on the type of battery used). Install new battery.

SIGNAL INTERFERENCE

This message indicates, via the displayed triangle squelch waveform, that a stronger, interfering, signal has been detected.

PERMANENT SIGNAL LOSS

This message indicates that no RF signal is being detected.

Setting Up ECG Monitoring

To initiate ECG monitoring:

- 1 Select a transmitter
- 2 Note its channel number
- 3 Attach lead wires to transmitter
- 4 Attach lead wires to electrodes
- 5 Install a transmitter battery
- 6 Apply electrodes to patient
- 7 Close the transmitter case

Each lead wire must be plugged into the transmitter, connected to an electrode, and then attached to the patient. Match the lead wire color to the color-coded connectors on the top of the transmitter case. Refer to the *ECG* chapter in this manual for details regarding electrode application. Telemetry patients are commonly ambulatory and require optimal skin preparation and lead application to minimize motion artifact. After the electrodes and lead wires have been attached, it is important to tape a loop of lead wire close to the electrode to minimize stress or pulling on the electrode itself - a process called stress-looping.

ECG monitoring begins when the telemetry receiver module detects a signal sent by a telemetry transmitter. The telemetry transmitter sends a signal as soon as its battery is installed.

ECG telemetry reception requires the following minimum conditions:

- The telemetry receiver module must be connected to a PCMS or Ultraview monitor, either directly or through a module housing, with the power ON and a Spacelabs Medical diversity antenna connected.
- ECG electrodes must be properly attached to the patient, and lead wires to the transmitter.
- The transmitter battery must be functional.
- The telemetry receiver module must be tuned to the telemetry transmitter's frequency (channel number).



NOTE

- All system connections must be made by Spacelabs Medical personnel only.
- Leakage currents are not affected by the high level output. The patient is electrically isolated from the patient monitor by the RF link.



WARNING

- Operation of television receivers or other CRT displays near the transmitter (within 2 to 3 feet), or operation of some pacemaker programmers may suppress the ECG waveform, preventing QRS detection and rate counting. An erroneous Asystole alarm may result.
- Signals resulting from devices such as Automatic Implantable Cardiac Defibrillators (AICD) may momentarily blank the ECG trace rather than display an out-of-range signal. In such cases, it may not be apparent that the AICD has fired and the condition of the patient should be checked. In all instances of AICD firing, the bedside or central will redisplay the ECG waveform within 5 seconds.

ECG monitoring in telemetry is identical to hardwired ECG monitoring. Refer to the *ECG*, *Arrhythmia*, and *ST Analysis* chapters of the *PCMS Operations Manual* (P/N 070-1001-xx) for detailed descriptions of configurations, displays, and controls. A brief overview of ECG monitoring follows.

Electrodes

Use silver/silver-chloride electrodes or their equivalent. Always connect all electrodes required for a particular lead. Missing electrodes may result in the loss of ECG tracing. Refer to the *ECG* chapter for information on placing the electrodes.




Use only Spacelabs Medical recommended electrodes. Some electrodes may be subject to large offset potentials due to polarization. Recovery time after application of defibrillator pulses may be especially compromised. Squeeze bulb electrodes commonly used for diagnostic ECG recording may be particularly vulnerable to this effect.



- Visually inspect each lead wire for obvious damage and replace as needed.
- Only use patient cables and lead wires specified by Spacelabs Medical. Other cables and lead wires may degrade performance and may damage the monitor during defibrillation or high frequency electrosurgery. Non-Spacelabs Medical cables and lead wires may also change the required input impedance and D.C. offset voltage, affecting monitor performance.
- Do not use stainless steel electrodes.
- Do not allow conductive parts of electrodes and connectors, including the reference electrode, to contact other conductive parts, including the ground.
- Poor cable dress or improper electrode preparation may cause line isolation monitor transients to resemble actual cardiac waveforms and thus inhibit heart rate alarms. Refer to the ECG chapter in this manual for details on proper electrode preparation and application.

Display Detail

Signal detection is indicated on your monitor when an ECG signal appears next to the ECG parameter key in the zone assigned to receive the transmitted telemetry channel. The transmitter's channel number is always identified above the waveform, to the left of the ECG key.



- ❶ ECG trace for first lead
 - ❷ telemetry channel number
 - ❸ ECG key for first lead
 - ❹ QRS indicator (flashes once per detected beat)
 - ❺ ECG lead designator
 - ❻ display resolution (monitor or extended)
 - ❼ paced operation indication (pacemaker detection is enabled)
 - ❽ abnormalities per minute alarm limit *
 - ❾ ST segment level for first lead **
 - ❿ abnormalities in a row alarm limit *
 - ⓫ ECG rate alarm limits; split screen centrals display a bell symbol when alarms are enabled; bedsidess display the rate alarm limits (120/40)
 - ⓬ abnormalities per minute counter *
 - ⓭ current heart rate
 - ⓮ ECG lead designator for second lead
 - ⓯ ST segment level for second lead**
 - ⓰ ECG key for second lead
 - ⓱ NIBP measurements: systolic/diastolic(mean) @ hours:minutes; SpO2 measurement (90343 only)
- * Only appears with the MultiView I or II option in the adult mode with Arrhythmia detection enabled.
- ** Only appears in adult mode with the ST segment analysis option.

Monitoring Paced ECG Patients

To monitor paced patients:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Select PACED YES

When monitoring pacemaker patients, use the paced feature to automatically enhance pacemaker spikes for display and eliminate them from the heart rate counter. The last setting you select is retained as the default.

If the interval between the pacemaker pulse and the QRS complex is greater than 150 milliseconds, the beat is considered to have originated in the atria and is not classified as a paced beat.

To prevent pacemaker pulses from being counted as actual beats, specialized circuitry removes the pacemaker pulses from the ECG signal and replaces them with pacemaker flags.



NOTE

The optimal leads for monitoring paced patients may vary. In telemetry monitoring, pacemaker spikes are detected on lead II. If pacemaker spikes are not detected, change electrode position.



WARNING

- *ECG detection circuitry may continue to count the pacemaker rate during occurrences of cardiac arrest or some arrhythmias. Do not rely entirely upon ECG rate alarms. Keep pacemaker patients under close surveillance.*
- *The system may insert pacemaker flags into the ECG signal in response to signals that are not pacemaker pulses. Therefore, if you use a Spacelabs Medical monitor to observe pacemaker performance, you must take into account all possible sources of pacemaker flags.*
- *Use the pacemaker manufacturer's performance analyzer as the primary means of evaluating pacemaker operation.*

Restoring Default Settings

To restore default settings:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Touch RESTORE SETTINGS
- 4 Select YES

With the Module Configuration Manager feature, you can restore all default settings. User-configurable options are listed in the *Module Configuration Manager* chapter of this manual.



NOTE

RESTORE SETTINGS changes the user-configurable options for all parameters in the module.

Changing the Display Resolution

To change the display resolution:

- 1 Touch ECG
- 2 Touch SETUP
- 3 Select MONITOR or EXTENDED

The MONITOR/EXTENDED key determines the display resolution of the two ECG traces, whether or not both traces are currently displayed on the monitor.

Key	Display Resolution
Monitor	(0.5 - 30 Hz)
Extended	(0.05 - 30 Hz)



NOTE

Changing the display resolution does not change the waveform bandwidth used to analyze the ECG signals for arrhythmia and ST segment level.

The factory default setting for display resolution is monitor mode.

To select ECG leads:

- 1 Touch ECG
- 2 Touch LEAD SELECT
- 3 Touch 1ST or 2ND LEAD
- 4 Select the desired lead

Selecting Options for Lead Display

One operational mode is available with the 90343 and 90347 multi-lead transmitters. When all electrodes are connected to the patient, leads I, II, III, AVR, AVL, AVF, and Vx, where x = 1 to 6, are available. When no chest lead is applied, leads I, II, III, AVR, AVL, and AVF are available using the remaining connected electrodes.

Connected Electrodes (X)					90343/90347 Valid Lead Vectors
RL	C	LL	LA	RA	
X	X	X	X	X	V1-6, I, II, III, AVR, AVL, AVF
X	X		X	X	I
X	X	X		X	II
X	X	X	X		III
X		X	X	X	I, II, III, AVR, AVL, AVF
X			X	X	I
X		X		X	II
X		X	X		III
	X	X	X	X	I, II, III, AVR, AVL, AVF
	X		X	X	I
	X	X		X	II
	X	X	X		III
		X	X	X	II



NOTE

All combinations of leads not shown above result in no valid lead vectors. In general, for at least one valid vector, either RL or C and two limb leads must be connected.

ECG Troubleshooting

Refer to the ECG Problem Solving section in the ECG chapter of this manual for additional conditions and solutions.

ASYSTOLE

An ASYSTOLE message means it has been 5 seconds or more since a QRS complex has been detected. Check the patient. If the patient is all right, try the following:

- Check that the lead wires are inserted into the proper receptacle.
- Using the continuity tester, check that there is no damage to the lead wires.
- If the amplitude is poor, check the appropriate lead with a 12-lead ECG.
- Check that the transmitter is not too near (within 3 feet) a television receiver or other CRT displays.

ECG VOLTAGE TOO LOW

The ECG amplitude may have dropped below the R-wave detector threshold level. Reposition the electrodes to obtain a QRS amplitude of at least 0.20 mV (adult) and 0.15 mV (neonate).

NOISY SIGNAL

The patient may be moving excessively. Secure the lead wires to the patient.

- Check the electrodes for good skin adhesion.
- Check lead wires at the transmitter for good contact.

ECG Troubleshooting Guide

Clinical Situation	Possible Cause	Solution
Noisy signal	■ ECG frequency response set to extended mode ■ Electrodes dry or poor skin adhesion	■ Select monitor mode ■ Repeat skin preparation and apply new, moist electrodes
Baseline wanders	■ Patient moving excessively ■ Respiration artifact ■ Electrodes dry or poor skin adhesion	■ Secure lead wires by stress-looping to the patient ■ Re-position electrodes ■ Repeat skin preparation and apply new, moist electrodes
Low amplitude ECG	■ Skin improperly prepared ■ Lead selected not providing QRS complex with greatest amplitude ■ Electrodes could be positioned over bone or muscle mass	■ Abrade skin and reapply electrodes ■ Select another lead for monitoring ■ Re-position electrodes
ECG won't learn	■ ECG signal too noisy ■ ECG voltage not within threshold. ECG VOLTAGE TOO LOW message may be displayed	■ Check lead wires and electrodes, then relearn patient rhythm ■ Select a different lead or adjust electrode location
Excessive alarms	■ Electrodes dry or poor skin adhesion ■ Alarm limits set too close to patient's normal heart rate ■ Excessive patient movement or muscle tremor	■ Repeat skin preparation and apply new, moist electrodes ■ Readjust alarm limit ■ Reposition electrodes and secure electrodes with tape if necessary

Refer to the *Problem Solving* sections in this chapter and in the *ECG* chapter for further monitoring tips.

SpO₂ Overview (90343 only)

Pulse oximetry enables you to noninvasively monitor a patient's hemoglobin oxygen saturation. This may be accomplished in either a continuous or episodic manner. The oximetry sensor contains two light emitting diodes (LEDs) that transmit specific wavelengths (660 and 940 nanometers) of light which are received by a photodetector.

Oxygen saturated blood absorbs light differently as compared to unsaturated blood. Thus the amount of light absorbed by the blood can be used to calculate the ratio of oxygenated hemoglobin to total hemoglobin in arterial blood. The monitor displays this ratio as percent SpO₂. Normal values range from 95 to 100%.



- ***A pulse oximeter should NOT be used as an apnea monitor.***
- ***A pulse oximeter should be considered an early warning device. If a trend towards patient deoxygenation is indicated, blood samples should be analyzed by a laboratory co-oximeter to completely understand the patient's condition.***



If your module is equipped with the Module Configuration Manager feature, you can define your own default settings for such characteristics as alarm limits and display configuration. See the Module Configuration Manager chapter in this manual for further details.

Setting Up SpO₂ Monitoring

To set up SpO₂ monitoring:

- 1 Connect the adapter cable to the transmitter
- 2 Attach the sensor to the patient and connect the sensor cable to the adapter cable
- 3 Initiate ECG monitoring
- 4 Select ECG
- 5 Select CHANNEL FORMAT
- 6 Set SpO₂ ON

The model 90343 digital telemetry multi-parameter transmitter uses Spacelabs Medical sensors as well as those from other manufacturers. Refer to the *Sensors* section at the end of this chapter for information concerning specific sensors and their operation.



- **Use only patient sensors specified by Spacelabs Medical. If you use sensors other than those specified, it may degrade performance and could damage the transmitter during defibrillation.**
- **Check the sensor site frequently. Do not allow the sensor to remain on one site for a prolonged time period, especially when monitoring neonates. Refer to sensor manufacturer's instructions.**
- **Never attach an SpO₂ sensor on a limb being monitored with a blood pressure cuff or a limb with restricted blood flow.**
- **A poorly applied sensor may give incorrect saturation values.**
- **Choose a site with sufficient perfusion to ensure accurate oximetry values.**

All sensors require an adapter cable between the sensor and the transmitter. Do not discard the adapter cable when you have finished using a disposable oximetry sensor. Disconnect the sensor cable from the adapter cable before discarding the sensor.

To connect the adapter cable to the transmitter, align the cable with the notch on the front of the transmitter connector and push the cable straight down into the transmitter. When you remove the cable, press the latch release (use a blunt device) on the bottom of the cable and pull the cable straight out. Never *twist* the cable.

Figure Telemetry-2: SpO₂ Adapter Cable to Transmitter

To enable SpO₂ monitoring in the model 90343 digital telemetry multi-parameter transmitter, enable the SpO₂ setting for DIP switch 5 beneath the battery compartment. (The factory default of this switch is ON.)




Figure Telemetry-3: Display Zone — Full Screen




Figure Telemetry-4: Display Zone — Split Screen

- ① Current SpO₂ value (percent)
- ② The bell indicates that alarms are enabled (on split screen central only)

Ensuring Accurate Monitoring

While each sensor requires site specific application procedures, the following general points will aid in ensuring oximetry monitoring success.

- Choose a site that provides proper alignment of the LEDs and receiving photodetector.
- Reduce light interference when monitoring a neonate under bright light by using a diaper or other light block over the sensor.
- Select a site that has unrestricted blood flow and can remain as immobile as possible to reduce or eliminate movement artifact.
- Do not restrict blood flow when securing a sensor with tape.
- Do not select a site near potential electrical interference (electrical cords, for example).

Setting or Adjusting Alarm Limits

To set or adjust alarms:

- 1 Touch ECG
- 2 Touch ALARM LIMITS
- 3 Select SpO₂ ALARMS ON
- 4 Select HI=, LO=, ALM DELAY, and MSG ALARM DELAY
- 5 Use arrow keys to adjust

Pulse oximetry alarm limits and delays are based either on factory default limits or user-defined limits. The factory default settings for alarm limits are 100% for high and 85% for low. For alarm delays, the factory default settings are 15 seconds for alarm limit delay and 20 seconds for message alarm delay. Read the *Alarms* chapter in this manual for details concerning PCMS alarm operation.

Refer to the *Module Configuration Manager* chapter in this manual for SpO₂ parameter tables that list available user settings and factory defaults for this parameter.

ALM DELAY Key

This key sets the number of seconds delay the system will wait before it reports that an alarm limit has been violated. When this feature is off, the key label will read "ALM DELAY OFF". When it is on, the label will read "ALM DELAY xx" where "xx" is the value, in seconds, of the delay.

The caregiver can set the delay time by taking the following steps:

1. Touch ALM DELAY xx (or ALM DELAY OFF).
2. Touch the up and down arrow keys until the value is set as desired. Possible settings are OFF, 5, 10, 15, 20, 25, and 30 seconds.



NOTE

- If the caregiver presses the down arrow key after the lowest value has been reached, the following message will appear on the prompt line:

Minimum alarm delay time has been reached.

- If the caregiver presses the up arrow key after the highest value has been reached, the following message will appear on the prompt line:

Maximum alarm delay time has been reached.

MSG ALM DELAY Key

This key sets the number of seconds delay the system will wait before it issues an alarm tone following any of the following messages:

- SpO₂ MONITOR FAILURE
- SpO₂ FAULTY SENSOR

- SpO₂ SENSOR OFF PATIENT
- SpO₂ INSUFFICIENT SIGNAL
- AMBIENT LIGHT INTERFERENCE
- SpO₂ INSUFFICIENT SIGNAL
- SpO₂ NOISY SIGNAL

When this feature is off, the key label will read "MSG ALM DELAY OFF". When it is on, the label will read "MSG ALM DELAY xx" where "xx" is the value, in seconds, of the delay.

The caregiver can set the message delay time by taking the following steps:

1. Touch MSG ALM DELAY xx (or MSG ALM DELAY OFF).
2. Touch the up and down arrow keys until the value is set as desired. Possible settings are OFF, 10, 20, 30, 40, 50, and 60 seconds.



NOTE

- If the caregiver presses the down arrow key after the lowest value has been reached, the following message will appear on the prompt line:

Minimum message alarm delay time has been reached.

- If the caregiver presses the up arrow key after the highest value has been reached, the following message will appear on the prompt line:

Maximum message alarm delay time has been reached.

Setting SpO₂ Data Averaging Period and Sampling Interval

The SpO₂ data averaging selection is used to smooth the oximetry saturation value for averaging the patient input values over 4, 8, or 16 seconds. This selection is made by setting the DIP switches 1 and 2 beneath the battery compartment in the Model 90343 digital telemetry multi-parameter transmitter. The default value is 8 seconds. Refer to *Figure Telemetry-5: DIP Switch Setting in Battery Compartment*.

DIP Switch 1	DIP Switch 2	Effect
OFF	OFF	4 seconds averaging enabled
OFF	ON	8 seconds averaging enabled (default)
ON	OFF	16 seconds averaging enabled

Figure Telemetry-5: DIP Switch Setting in Battery Compartment

The sampling interval selection permits the caregiver to determine how often an SpO₂ measurement will be taken. Less frequent SpO₂ readings can extend the usable life of the battery. (Refer to the Ultraview Digital Telemetry Products Data Sheet — P/N 061-0801-xx — for more information on battery service life.) This selection is made by setting DIP switches 3 and 4 beneath the battery compartment in the model 90343 digital telemetry multi-parameter transmitter. The default setting is 1 minute intervals.

DIP Switch 3	DIP Switch 4	Effect
OFF	OFF	Continuous sampling
OFF	ON	1 minute sampling interval (default)
ON	OFF	2 minute sampling interval
ON	ON	5 minute sampling interval

Viewing Pulse Rate

To display heart rate from SpO₂ sensor:

- 1 Touch ECG
- 2 Touch CHANNEL FORMAT
- 3 Read heart rate in the area to the left of the SpO₂ ON/OFF key

In normal operations the heart rate for display will be obtained directly from the acquired ECG leads or an alternate rate source. SpO₂ can be used as the alternate source, if it is set for continuous measurement. When it is set for episodic measurement, SpO₂ cannot be used as an alternate rate source.

SpO₂ Error Messages

Error messages indicate a problem or condition which may affect accurate monitoring values. Do not ignore these messages. Correct any fault before continuing.

When the following messages are displayed, the saturation value is immediately changed to ??? and an alarm is triggered, if your module has been configured with an alarm for that message. (Refer to the *Module Configuration Manager* chapter).

SpO2 SENSOR DISCONNECTED

- The transmitter does not detect an adaptor cable connected or a sensor connected to an adaptor cable. If the message persists and the adaptor cable is secure, replace the adapter cable.
- The alarm will stop after approximately 10 seconds.
- On remote view, there may be no audible alarm on the remote mainframe before the local alarm stops.

SpO2 MONITOR FAILURE

- The LED and/or photodiode have failed. Replace the sensor and/or SpO2 adapter cable.

When the following messages appear, the monitor displays the saturation value alternately with the message ??? every two seconds. An alarm will begin after the message alarm delay time has elapsed (Refer to the Module Configuration Manager chapter.)

SpO2 AMBIENT LIGHT INTERFERENCE

- The sensor is receiving external light interference from a bright light source near the sensor. Shield the sensor from the external light source. If the condition persists for more than 30 seconds, ??? will replace the data display.
- The sensor photodiode and LEDs are misaligned on flexible sensors thereby allowing light to enter. Realign the sensor photodiode with LEDs.
- If a message appears with finger clip, replace sensor.

SpO2 INSUFFICIENT SIGNAL

- Insufficient signal for proper operation.
- Poor sensor application or site. Correctly re-apply or reposition to a better perfused site, massage site, or apply new sensor.

SpO2 NOISY SIGNAL


- The sensor signal is disturbed by motion or other interference. Eliminate sensor movement. The message disappears when a value is obtained.
- The sensor is placed adjacent to power cords or other electrically noisy devices. Move the noisy device or move the sensor to another site.

SpO2 SENSOR OFF PATIENT

- The transmitter is unable to detect a valid sensor input signal. Check the patient for proper sensor placement.
- Tissue between the LED and photodiode is too transmissive. If sensor placement seems correct and the message persists, try a sensor site with a thicker tissue bed.

**NOTE**

Adaptor cables and sensors are ordered separately through the Spacelabs Medical Supplies Products..



Nelcor adaptor cable (P/N 012-0587-00)

Nelcor
sensor


90343 digital telemetry multi-parameter transmitter

Figure Telemetry-6: Nelcor and Sensors and Adaptor Cables

SpO₂ Troubleshooting Guide

Clinical Situation	Possible Cause	Solution
No SpO ₂ label is displayed	■ SpO₂ is not enabled at the 90343	■ Be sure transmitter DIP switch 5 setting is ON ■ Be sure DIP switch 8 is OFF
SpO ₂ value displays ???	■ Sensor not connected to patient ■ Adapter cable not connected to module properly ■ Sensor not connected to adapter cable ■ Excessive patient motion ■ Transmitter is in the initialization phase (the first 15 seconds after sensor application)	■ Reattach sensor ■ Correctly connect the adapter cable ■ Correctly connect the sensor ■ Urge patient to remain still while reading is in progress ■ Wait until initialization is complete
Low signal strength	■ Sensor placement not optimum ■ Sensor placed below blood pressure cuff	■ Move sensor to a site which has better perfusion ■ Align LED with sensor photodetector ■ Move sensor to an alternate limb
Intermittent or complete failure to operate	■ Transmitter error ■ Depleted battery	■ Call qualified service person ■ Replace battery
Factors which cause significant variances in sensor accuracy	■ Presence of dysfunctional hemoglobins (COHb, MetHb) ■ Presence of intravascular dyes (indocyanine green, methylene blue) depending on their concentration in the blood stream ■ High ambient light level ■ Electrosurgical interference ■ Patient is significantly anemic (Hb less than 5gm/dl) or patient has received large amounts of IV solutions	■ Follow hospital procedure for determining oxygenation in these patients ■ Follow hospital procedure for determining oxygenation in these patients ■ Reduce light levels near patient ■ Follow hospital procedure for determining oxygenation in these patients during the procedure ■ Follow hospital procedure for determining oxygenation in these patients

NIBP Overview



90217 ABP Monitor

The model 90343 digital telemetry multi-parameter transmitter can send the non-invasive blood pressure (NIBP) patient data acquired by the Model 90217 ambulatory blood pressure (ABP) monitor, to the Model 90478 digital telemetry receiver. The Model 90478 can display the patient's episodic NIBP data and trigger alarms based on thresholds set at the Central Station monitor by the clinician.

The Model 90217 ABP monitor is a small, lightweight, battery-powered unit designed to take blood pressure and heart rate measurements for 24 or 48 hours or for longer periods of time. Please refer to the *90217 Operations Manual* (070-0137-xx) for more detailed information on this specific product; its initialization via a Local Report Generator (Model 90239A) or direct PC interface (Model 90121); Patient Preparation; Data Transfer and Reports; and Event Codes.

Noninvasive blood pressure (NIBP) uses oscillometric monitoring to measure systolic (S), diastolic (D), and mean (M) arterial blood pressures. The PCMS monitor displays these readings and the time the measurement was acquired. You can also display heart rate. The ABP monitor can display up to ten readings at one time and store up to 120 readings.



The Model 90217 ABP Monitor is intended for use with adult patients only.

NOTE

Setting Up NIBP Monitoring

To set up NIBP monitoring:

- 1 Attach appropriate cuff and 90217 ABP monitor to patient
- 2 Attach adapter cable between 90217 and 90343
- 3 Initialize 90217 with 90239A local report generator
- 4 Configure 90343 DIP switch 6 to enable NIBP operations
- 5 Initiate ECG monitoring
- 6 Select ECG
- 7 Select CHANNEL FORMAT
- 8 Set NIBP ON

Proper cuff selection and application is a critical issue in ensuring the accuracy of NIBP readings. To ensure proper cuff selection, first measure the circumference of the limb at its midpoint. Match the limb measurement to the range of appropriate circumferences (in centimeters) specified on each cuff. If the cuff bladder is too wide for the patient, the reading will be falsely lowered; if too narrow, the reading will be falsely elevated. Undersizing the cuff results in the greatest chance of error, so a variety of cuff sizes should be available to accommodate your full patient population.

The cuff should be snugly applied. When the cuff is properly applied to an adult, you should be able to insert one finger between the cuff and the arm. If you can insert two fingers, the cuff is too loose which may result in falsely elevated readings. Make sure the hose is not kinked when the cuff is applied.

During blood pressure measurement, the inflated cuff reduces blood flow to the limb to which it is applied. Do not apply a cuff to a limb that has restricted blood flow. Check the patient periodically.



NOTE

- Do not apply a blood pressure cuff to a limb being monitored with a pulse oximetry sensor because SpO_2 will be affected during NIBP readings. Avoid applying a cuff to a limb that has an intravenous line in place. Do not apply a cuff to a limb that has restricted blood flow.
- Never use extensions or adapters with the neonatal hose. Use only single hose cuffs to ensure proper operation. Spacelabs Medical's hoses are non-conductive with respect to defibrillator discharge effects.

Patient Factors Affecting Readings

Excess patient movement, speech, or muscle contractions as a result of severe pain or shivering can interfere with automated NIBP readings. Ensure that the patient is quiet and not moving during NIBP readings just as you would manual readings. Avoid applying external pressure to the cuff during readings. Institute measures to minimize shivering and alleviate pain.

Some arrhythmias may cause beat-to-beat pressure fluctuations that can make obtaining NIBP readings more difficult. Increased variability of readings can result from these pressure variations. If it becomes difficult to obtain readings in the presence of arrhythmia, pressure should be temporarily verified using another method. Pressure also varies cyclically with normal respiration. With deep respirations or in certain patients this effect may be enhanced, increasing reading variability.

For patients in shock, indirect methods of measuring pressure (auscultatory, oscillometric, Doppler) may not be reliable because of peripheral vascular changes. These changes include peripheral vasoconstriction and diminished peripheral circulation resulting from shunting of blood to central organs. In some cases peripheral pulses or Korotkoff sounds may be diminished or disappear in spite of adequate blood pressure. In such cases, measuring a cuff pressure may be impossible or give misleading results. Direct blood pressure measurements (Invasive) should be considered in patients with signs of shock or any patient who rapidly becomes unstable for unknown reasons.

Setting Up the ABP Monitor

After the monitor has been initialized, prepare the patient for monitoring as follows:

1. Turn on the monitor (wait for the monitor to perform self-tests). When the LCD displays the current time, the monitor is ready for operation.
2. Strap the monitor to the patient on the hip opposite the side on which the cuff is worn. Secure the monitor using the patient's own belt or the ABP pouch strapped over the opposite shoulder. When using the shoulder strap, use the belt supplied with the monitor or the patient's belt to provide additional security.
3. To select the proper cuff, first measure the circumference of the limb at the point where the cuff is to be applied. Match the limb measurement to the range of appropriate circumferences (in centimeters) specified on each cuff (refer to the table below).

Cuff Size	Limb Circumference
Pediatric	13 to 20 cm
Small Adult	17 to 26 cm
Average Adult	24 to 32 cm
Large Adult	32 to 42 cm
Extra-large Adult	38 to 50 cm

4. Position the cuff so that the center of the inflatable bladder is directly over the brachial artery. The center of the bladder location is marked on the outside of the cuff. Once the proper position is determined, the cuff must be tightened to

ensure that it is equally snug at the top and bottom edges and that it is not kinked. This is especially important on larger arms. Insert a finger between the cuff and the limb to ensure it is not too tight. It may be necessary to wrap the cuff with its tail at an angle to achieve uniform tightness. If the cuff is not equally snug at the top and bottom edges, the number of readings available will be limited and the monitor may indicate that the cuff is improperly applied.



- Use only Spacelabs Medical cuffs with this monitor. Using other manufacturer's cuffs may result in inaccurate readings even if the manufacturer's recommended size is observed.
- If the cuff is too small, pressure readings may be falsely high; a cuff that is too large produces a falsely low reading. The bladder can be positioned in the cuff for either the left or right arm.



CAUTION

- Avoid compression or restriction of pressure in the NIBP patient connector tubes. Check that operation of the equipment does not result in prolonged impairment of circulation.
 - Do not apply cuff to areas of breached or injured skin.
 - Cuff hose connections use luer fittings. Be careful not to connect the ABP monitor into an intravenous fluid line when working close to them.
 - This product contains natural latex rubber components to which some people may be allergic. These components include the bladder and the first four inches of tubing extending from the cuff.
5. Once cuff is applied, the arm should be relaxed at the patient's side. To avoid reading errors due to hydrostatic pressure differences, the level of the cuff on the arm should be near the level of the heart.
 6. Lead the hose up the arm with the cuff and place it across the back of the patient. Drape the hose so it does not cause the patient discomfort and is not pinched shut by too tight a radius. *Figure Telemetry-7: Common Cuff Hose Positions* shows the most common positions for the cuff hose.




Figure Telemetry-7: Common Cuff Hose Positions

7. Connect the hose to the monitor.
8. To verify proper monitor operation, take one or more blood pressure readings. Push the START/STOP key to begin a measurement.
9. Interconnect the adapter cable between the communications port on the 90217 and the NIBP port on the 90343 as shown in *Figure Telemetry-8: Transmitter and ABP Monitor Connections*.




Figure Telemetry-8: Transmitter and ABP Monitor Connections

Figure Telemetry-9: Display Zone — Full Screen and Figure Telemetry-10: Display Zone — Split Screen illustrate typical NIBP displays. You can view NIBP readings from any PCMS bedside or central monitor on a network. NIBP displays on a split screen central monitor appear in a format slightly different from that of bedside or full screen central monitors.




Figure Telemetry-9: Display Zone — Full Screen




Figure Telemetry-10: Display Zone — Split Screen

- ❶ Last systolic reading
- ❷ Last diastolic reading
- ❸ Last mean reading
- ❹ Hour of day of last reading
- ❺ Minutes after the hour of last reading

Setting or Adjusting Alarm Limits

To set or adjust alarms:

- 1 Touch ECG
- 2 Touch ALARM LIMITS
- 3 Select NIBP/ALM ON
- 4 Select SYS, DIA, or MEAN
- 5 Select HI= or LO=
- 6 Use arrow keys to adjust

You can define pressure alarm limits for systolic, diastolic and mean values. The default setting for alarms is OFF. Read the *Alarms* chapter in this manual for PCMS system alarm functions. The alarm limits defaults are listed in *Table 1: Alarm Limits*.

Table 1: Alarm Limits

	Hi		Low	
systolic	150 mmHg	(20.0 kPa)	100 mmHg	(13.5 kPa)
diastolic	100 mmHg	(13.5 kPa)	60 mmHg	(8.0 kPa)
mean	110 mmHg	(14.5 kPa)	90 mmHg	(12.0 kPa)

Refer to the *Module Configuration Manager* chapter in this manual for NIBP parameter tables that list available user settings and factory defaults for this parameter.

**To display heart rate from the
ABP monitor:**

- 1 Touch ECG
- 2 Touch CHANNEL FORMAT
- 3 Read heart rate in the area to
the left of the NIBP ON/OFF key

Displaying Heart Rate Data

In normal operation, the heart rate for display will be obtained directly from the acquired ECG leads or an alternate source (SpO₂). The ABP heart rate is not available as an alternate heart rate source due to the episodic nature of the NIBP measurement.

Displaying New or Previous Readings

90217 Event Codes

The ambulatory blood pressure monitor will display an event code whenever an event prevents the unit from successfully completing a blood pressure measurement. The two numerical digits of the event code indicate the reason the measurement was aborted. The table below lists event codes that are displayed on the monitor, as well as event codes that appear on the Event Code Report.

Monitor	Report	Condition
EC00		
	EC10	Excess movement artifact. Frequent EC10 messages may indicate an air leak.
	EC20	A) A very large number of movement artifacts B) Heart rate arrhythmia
	EC30	A) Movement artifact at mean arterial pressure B) Heart rate arrhythmia
	EC40	A) Movement artifact at asystole B) Heart rate arrhythmia
	EC50	A) Movement artifact at diastole B) Heart rate arrhythmia
	EC60	A) Movement artifact B) Heart rate arrhythmia
	EC70	Systole was found to be above the highest cuff pressure. However, this result appears to be an error caused by motion artifact. Therefore, the cuff will not be inflated to a higher pressure on the next measurement attempt.
	EC80	A) Movement artifact B) Heart rate arrhythmia
	EC90	A) Movement artifact B) Heart rate arrhythmia
EC01		
	EC11	Did not pump above the mean arterial level
	EC21	Did not pump above systolic pressure
	EC91	Systole appears higher than the selected maximum cuff pressure limit
EC02		
	EC12	Did not reach initial cuff pressure. The cuff may have been improperly applied or there may be an air leak.
	EC22	Overpressure
	EC32	Overpressure
	EC42	No cuff attached
	EC52	Kinked hose
	EC62	Cuff applied too loosely
	EC72	Kinked hose

Monitor	Report	Condition
	EC82	Kinked hose.
EC03		
	EC03	Patient canceled reading by pressing STOP key. No retry attempt is made following an EC03 code.
	EC13	The Office Check Mode has been reinstated. No retry attempt is made following an EC13 code.
EC04		
	EC04	Blood pressure measurement not completed in the maximum time allowed. Occasional EC04 messages may result from excessive patient movement. Frequent EC04 messages would indicate an improperly applied cuff or a monitor malfunction which requires service.
	ECn4	(where n = 1 to 9) Indicates that one or more of the blood pressure results have been corrupted and subsequently recovered. Frequent occurrence of this message would indicate a malfunction which requires service.
	EC05	The individual blood pressure result has been corrupted and cannot be recovered.
EC15		Equipment malfunction. Return to Spacelabs Medical for service.
EC25		Unit failed to initialize. Please initialize.
EC35		The monitor needs to be reinitialized.
EC05		
EC05 & EC45	EC45	Invalid bleed size. The monitor automatically has changed the bleed size to 8 mmHg.
EC05 & EC55	EC55	An unexpected loss of power possibly caused by a) removal of the batteries during a blood pressure measurement, b) hardware overpressure, or c) a hardware time-out. Frequent EC55 messages would indicate a malfunction which requires service.
EC05 & EC65	EC65	Extremely large artifact.
EC05 & EC75	EC75	Equipment malfunction. Return to Spacelabs Medical for service.
EC05 & EC85	EC85	Equipment malfunction. Return to Spacelabs Medical for service.
EC05 & EC95	EC95	Cuff pressure baseline out of bounds. The monitor should correct the baseline automatically within 10 minutes; or it can be set by initialization of the monitor. If initialization does not correct the condition the monitor must be returned to Spacelabs Medical for calibration.
EC07		
	EC78	Clogged luer filter
EC08		
	EC18	Too few data entries to accurately determine blood pressure. The message may indicate that the cuff is not being worn by the patient (taken off but left connected to the monitor). The message may also indicate that motion artifacts cause the majority of the incomplete data.
	EC28	Diastole above 200 mmHg

Monitor	Report	Condition
	EC38	Pulse pressure less than 16 mmHg
	EC48	A) Movement artifact at mean arterial pressure B) Heart rate arrhythmia
	EC58	A) Movement artifact at diastole B) Heart rate arrhythmia
	EC68	Division by zero
EC09		
	EC19	Contradictory instructions sent to hardware (e.g., "pump on and valve open")
	EC29	Diastolic pressure value cannot be obtained from the data available.
	EC39	Algorithm could not process input data quickly enough resulting in an input queue overflow.
	EC49	This monitor must be initialized.
	EC59	Heart rate value cannot be obtained from the data available.
	EC69	Heart rate value cannot be obtained from the data available.
	EC79	Bleed steps were too small. This may be caused by a partially obstructed air hose. All blood pressure attempts following this message are inhibited. Attempts can be enabled by turning the power switch off then on.
	EC99	Unexpected or contradictory data (such as a negative cuff pressure).
LLL		
	EC16	Low battery detected prior to start of measurement.
	EC26	Low battery detected after measurement started. Usually caused by the pump drawing enough current to lower the battery voltage.
Lbb		The report does not print an event code for this condition, which is a low backup battery. Contact Spacelabs Medical for replacement of the battery.

NIBP Troubleshooting Guide

Clinical Situation	Possible Cause	Solution
No NIBP displays	■ Adapter cable not inserted correctly ■ NIBP not enabled on 90343 or 90478 ■ ABP monitor not properly initialized	■ Remove and re-insert adapter cable ■ Enable NIBP function by setting transmitter DIP switch 6 ON and setting DIP switch 8 OFF ■ Reinitialize ABP Monitor
No NIBP readings can be obtained	■ Incorrect or inoperative cuff in use ■ Cuff tubing is attached to adult outlet, but monitor is configured in the neonatal mode (or vice versa) ■ Tubing is kinked ■ Some arrhythmias (e.g. atrial fibrillation and frequent ventricular ectopy) may cause a single or repeated failure to obtain a reading (may be due to true beat-to-beat variations in pressure) ■ Excessive patient motion or muscle contractions associated with shivering or severe pain ■ Blood pressure outside of measurement range	■ Replace with cuff known to be operative ■ Connect tubing to correct outlet. Correlate monitor mode, cuff and patient type. ■ Locate kink and straighten tubing ■ Document arrhythmia if present, verify pressure with another method, then follow hospital procedure for care of this type of patient ■ Ensure that patient is quiet with minimal movement during NIBP readings. Minimize patient's shivering. ■ Verify extremely high or low pressure with another method
Intermittent or complete failure to operate	■ ABP monitor error	■ Remove ABP monitor from service; record event code; and call qualified service person
Apparent incorrect value	■ Wrong size cuff for patient ■ Cuff is damaged ■ Excessive patient motion, shivering or severe pain ■ False high readings may be the result of venous congestion caused by frequent readings ■ Cuff too loose or positioned incorrectly	■ Measure patient's limbs at the midpoint. Match limb measurement to range specified on cuff (undersizing the cuff results in the greatest degree of error). ■ Replace with good cuff ■ Ensure patient is quiet with minimal movement during NIBP readings. Minimize patient's shivering. ■ Reduce frequency of readings ■ Tighten cuff or reposition appropriately
Variable readings occur	■ Some arrhythmias may cause beat-to-beat pressure and NIBP readings ■ Larger than normal influence of respiratory phases on blood pressure (inspiratory fall in blood pressure; expiratory rise)	■ Document arrhythmia if present, verify pressure using another method, then follow hospital procedure for care of this type of patient ■ NIBP software usually compensates for normal variation
No NIBP readings or questionable values in the presence of shock	■ Peripheral vascular changes experienced during shock may reduce the reliability of blood pressure readings obtained with any indirect method. Peripheral pulses may be diminished or absent.	■ Consider invasive pressure measurements in patients with symptoms of shock or in any patient who rapidly becomes unstable for unknown reasons