



November 8, 2023

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

Subject: FCC Class II Permissive Change Application for LIFEPAK 35 Defibrillator/Monitor External Modem containing the WCDMA/LTE LE910Cx cellular module.

Radio Description: LE910C4-NF

FCC ID Number: CZ023L4N1, grant date 18-October-2023

To whom it may concern:

Physio-Control, Inc., now a part of Stryker, is herein submitting a Class II permissive change filing for the GSM Module as listed in supporting exhibits. Pursuant to FCC Part 22H, 24E, 27, and 90, Physio-Control Inc. is submitting a request to change PCS Licensed Transmitter Grant. Installers and end users must be provided with antenna installation instructions and transmitter operating conditions and instructions for satisfying RF exposure compliance. The final product operating with this transmitter must include operating instructions and antenna installation instructions, for end-users and installers to satisfy RF exposure compliance requirements."

This Class II permissive change seeks enable portable use (<20 cm distance from persons) in the 'LIFEPAK 35' host device. The LIFEPAK 35 also contains the FCC Part 15C and 15E radio described on Physio-Control grant under FCC ID: CZ023LP35. Physio-Control, Inc. here seeks to provide evidence pertinent to the listed concerns to obtain the new grant due to changes in the device classification.

Sincerely,

Timothy Rand
Sr Compliance Standards Engineer
Stryker Emergency Care
11811 Willows Rd NE
Redmond WA 98052
+1 425 867 4477
tim.rand@stryker.com
www.physio-control.com

Emergency Care

11811 Willows Road NE, Redmond, WA 98052 USA | P +1 425 867 4000 | Toll-free +1 800 442 1142 | stryker.com