



6600 Washington Avenue S.
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

www.starkey.com

September 23, 2019

Federal Communications Commission
445 12th Street SW
Washington, DC 20554
Attention: Application Examiner

Reference:	Certification:	Parts 15.209 and 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz)
		Low Power Device (10.281 MHz)
	Model #:	Livio RIC R Hearing Aid
		FCC ID: EOA-24LIVIORCHG

Dear Sir/Madam:

Enclosed please find an FCC EOA-24LIVIORCHG Class 2 permissive change application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Livio hearing aids.

No changes have been made to the device hardware; the change is a software change enabling a new transmission mode with the 2 MB/s Bluetooth Low Energy protocol as defined in version 5.0 of the Bluetooth Core Specification that has a 1 MHz occupied bandwidth. Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink, appearing to read "William J. Mitchell", is written over a horizontal line.

William J. Mitchell, PE
Principal Regulatory Affairs Engineer
952-918-5683
bill_mitchell@starkey.com