



Hear better. Live better.

6700 Washington Avenue S
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

Starkey.com

1 April 2026

Federal Communications Commission
45 L Street, NE
Washington, DC 20554
Attention: Application Examiner

Reference:	Certification:	Part 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz)
	Model #:	Starlink Remote Control 2.0
		FCC ID: EOA-24WR2

Dear Sir/Madam:

Enclosed please find an FCC EOA-24WR2 Class 2 permissive change application. This device is a remote control for our 2.4 GHz wireless hearing aids that uses the Bluetooth Low Energy protocol.

No changes have been made to the device hardware; the change is a software change enabling a new transmission mode with the 2 MB/s Bluetooth Low Energy protocol as defined in version 5.0 of the Bluetooth Core Specification that has a 1 MHz occupied bandwidth.

Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

William J. Mitchell, PE
Principal Wireless and Safety Regulatory Engineer
952-918-5683
bill_mitchell@starkey.com