

BSM 170/170B

User's Manual



Please note the important information below before reading this manual.



Warning

Failure to comply with safety warnings and regulations can cause serious injury or death.



Caution

Failure to comply with safety cautions and regulations can cause injury or property damage.

InBody

For any problems with BSM170/170B or any clinical questions, please contact us at the following:

InBody Co., Ltd. [HEAD OFFICE]

InBody Bldg., 54, Nonhyeon-ro 2-gil, Gangnam-gu, Seoul, Korea

TEL: +82-2-501-3939 FAX: +82-2-578-2716

Website: <http://www.inbody.com>

E-mail: info@inbody.com

Reproduction, adaptation, or translation of this manual is prohibited without prior written consent from InBody Co., Ltd under the copyright laws. This manual might have typographical errors, and its content can be changed without a prior notice. InBody Co., Ltd. shall not be liable for any errors, incidental, or consequential damages that occurred by not complying with the content of the User's Manual.

Visit our website <http://www.inbody.com> to view and download further information about the functions of the BSM170/170B, the explanation of results output, and more. InBody Co., Ltd reserves the right to modify the appearance, specifications, and etc. of the BSM170/170B to improve the quality of the product, without prior notice for reasons of performance improvement.

Customer Service Information

InBody USA [USA]

13850 Cerritos Corporate Dr., Unit C, Cerritos, CA 90703, USA

TEL: +1-323-932-6503 FAX: +1-323-952-5009 Website: <http://www.inbodyusa.com> E-mail: USA@biospaceamerica.com

InBody Japan Inc. [JAPAN]

Second Floor Ayabe Bldg., 2-17-3 Sotokanda, Chiyoda-ku, Tokyo JAPAN

TEL: +81-03-5298-7667 FAX: +81-03-5298-7668 Website: <http://www.inbody.co.jp> E-mail: inbody@inbody.co.jp

EU Representative. [EUROPE]

DongBang Acuprime, 1 Forrest Units, Hennock Road East, Marsh Barton, Exeter EX2 8RU, U.K

TEL: +44-1392-829500 FAX: +44-1392-823232 Website: <http://www.inbody.com> E-mail: info@acuprime.com

Biospace China. [CHINA]

904, Xing Di Plaza, No. 1698 Yishan Road, Shanghai, 201103, CHINA

TEL: +86-21-64439738, 9739, 9705 FAX: +86-21-64439706 Website: <http://www.biospacechina.com> E-mail: info@biospacechina.com

Australian Sponsor. [AUSTRALIA]

Emergo AUSTRALIA, Level 20, Tower II, Darling Park, 201 Sussex Street, Sydney, NSW 2000, AUSTRALIA

TEL: +61-2-9006-1662 FAX: +61-2-9006-1010 Website: <http://www.emergogroup.com> E-mail: Sponsor@emergogroup.com

BSM170/170B

User's Manual Contents

I. Stadiometer Installation

A. Product Components	4
B. Operating Environment	5
C. Installation Instructions	5
D. Stadiometer Settings	12
E. Maintenance	18

II. Height Measurement

A. Precautionary Measurements	19
B. Measurement Instructions	20
C. Measurement Posture	21

III. Transportation and Storage

A. Repacking Instructions	22
B. Transportation and Storage Environment	22

IV. Frequently Asked Questions 23

V. Others

A. Exterior and Functions	24
B. Symbols	26
C. Specifications	27

I. Stadiometer Installation






A. Product Components

The BSM170/170B contains of the following components.

Make sure all of the following components are present.

* Please inspect each component of the BSM170/170B for damage prior to installation.

- ❶ Footplate 1 EA
- ❷ Stand (upper) 1 EA
- ❸ Stand (lower) 1 EA
- ❹ Measurement unit 1 EA
- ❺ Measurement tape 1 EA
- ❻ Footplate mounting screws 4 EA
- ❼ Stand mounting screws 8 EA
- ❽ User's manual 1 EA
- ❾ AA batteries 4 EA
- ❿ Test Jig 1 EA



B. Operating Environment

Please make sure that the environment is adequate for the BSM170/170B installation. This equipment is designed for indoor use. If installing outdoors, the following requirements must be fulfilled.

Temperatures range	10 ~ 40°C (50 ~ 104°F)
Relative humidity	30 ~ 75% RH
Atmospheric pressure	70 ~ 106kPa

Warning

- Before installing BSM170/170B, check that the ceiling is at least 2.3 m high. The maximum height of this equipment is 2.19m. Keep a space of at least 3cm between the ceiling and the top of the BSM170/170B.

C. Installation Instructions

1. After opening the BSM170/170B box, remove the packaging materials and take all product components out.


Warning

- Keep the packaging materials provided for repacking the equipment in the future. Other wastes should be disposed of according to relevant laws and regulations.
- Install the BSM170/170B on a leveled, non-vibrating surface. Installing the equipment on an uneven surface may cause the examinee to fall down. Measurement results may also be inaccurate.
- Never clean the BSM170/170B with liquid spray or detergent directly. The equipment may corrode and/or malfunction if the liquid or detergent leaks inside. Additionally, BSM170/170B measurement tape may be damaged. When cleaning the BSM170/170B, gently wipe it with the cloth.


2. Insert the “Stand (lower)” into the footplate, ensuring that the arrow (▼) located on the bottom right of the “Stand (lower)” and the arrow (▲) located on the footplate face one another.

* When inserting into the “Stand (lower)”, part without screw holes must face downward.

* The “Stand (lower)” and the footplate are designed to be assembled together in only one particular direction. Forcing the stand to be inserted in other directions could damage the product.




3. Hold the “Stand (lower)” in place to prevent it from sliding out of the insertion hole of the footplate, tilt the equipment sideways towards the floor.




4. Screw each of the 4 footplate leveling screws into the holes located on the opposite side of the stand insertion hole of the footplate.

5. Tighten the Stand mounting screws by turning them clockwise with the Phillips screwdriver.




6. Once all four screws are fastened tightly, return the “Stand (lower)” and footplate to an upright position.
7. Insert the black end of “Stand (upper)” into the “Stand (lower)” as illustrated below.



 Caution

- Hold the “Stand (upper)” as vertically as possible so that it is not inserted in a slanting position.


8. After inserting the “Stand (upper)” into “Stand (lower)”, fasten the two stands together using the mounting screws and Phillips screwdriver.



 Caution

- Avoid using excess force to fasten the mounting screws.


9. Insert the measuring tape into the “Stand (upper)” with the ‘↓’ facing down. Refer to the illustration below. After the measurement tape is completed inserted, remove the sticker attached to the end of the measurement tape to affix the tape to the stand.



 Caution

- The measurement tape must be inserted with the arrow marked ‘↓’ facing down.
- Correct measurements cannot be taken if the measurement tape is inserted reversed.

10. Tilt the Stand to a comfortable position and insert the measuring unit onto the stand as illustrated below.



 Caution


- The LCD on the measurement unit and the measurement tape must be visible on the same side.

11. The BSM170/170B uses batteries as its power source.

* When the measurement standby time expires, the equipment automatically turns off.


- ❶ BSM170 standby time: 10 seconds.
- ❷ BSM170B standby time: 2 minutes if it is not connected to the InBody, 10 seconds if it is connected to the InBody.

- 1) Open the battery cover located on the right side of the LCD screen.
- 2) Insert 4 AA batteries.
- 3) Cover the battery socket with the battery socket lid.



12. After placing the BSM170/170B on a flat surface, the footplate is not level, use the footplate adjustment screws to adjust until level.


* Leveling the equipment is necessary for accurate height measurements.



13. When taking measurements with the BSM170/170B, the operator must move the measurement unit up and down manually. By default, the adjustment screw is tight to keep the measurement unit from sliding down too freely. If the screw is too tight, adjust the screw to the preferred tightness.


* By default, the adjustment screw is tight to keep the measurement unit from sliding down too freely. If the screw is too tight, adjust the screw to the preferred tightness.

* If the screw is tightened too tight, the measurement unit may not move at all.



14. Turns on the BSM170/170B.

15. Move the measurement unit up or down until the 'Move' text disappears. The equipment will enter the measurement standby mode and the display will change as shown below.



D. Stadiometer Settings

1. Required Settings

1. After turning on the BSM170/170B, press and hold the  button and the  button simultaneously for at least 3 seconds to enter the Setup menu.




2. In the Setup menu, set the measurement value hold time, beep, screen brightness, units, and offset value.

1) Measurement value hold time: When the  button is pressed after a height measurement is recorded, the value displayed will stay locked on the LCD screen for this amount of time, even if the measurement unit is moved.

- a. Use the  button to move to 'Hold' and then press the  button to select.
- b. Use the  button to set a hold time for the locked height value.

* The hold time can be set to off or on for anywhere between 1 to 10 seconds.




Measurement value hold time setup

- c. Save changes by pressing the  button.

2) Beep: Enable or disable the beep alert.

- a. Use the  button to move to 'Beep' and then press the  button.
- b. Use the  button to turn on or off the beep alert.

* Beep can be set to on or off.




Beep setup

- c. Save changes by pressing the  button.

3) Screen brightness: Brightness of the LCD screen.

- a. Use the  button to move to 'Light' and then press the  button.
- b. Use the  button to set the screen brightness that is appropriate for the test environment.

* Screen brightness can be set to 100%, 50%, or off.



Screen brightness setup

- c. Save changes by pressing the  button.