

TEST FACILITY AND EQUIPMENT LIST

FACILITIES:

Radiated	ANSI C63.4-2003 (FCC OET/55) open field 3 metre test range. This test range is protected from the cold and moisture by a non-conductive enclosure.
Conducted	2.5m Anechoic Chamber

EQUIPMENT:

Anritsu 2601A Spectrum Analyzer
Advantest R3261A Spectrum Analyzer
Hewlett-Packard RF generator # 8640 B with an 002 doubler
A.H. Systems biconical antenna; 20 MHz to 330 MHz
A.H. Systems log periodic antenna; 300 MHz to 1.8 GHz
Compliance Design P950 Preamp (16 dB) ... 25 MHz to 1.0 GHz

NOTE:

The Anritsu 2601A Spectrum Analyzer and the Advantest R3261A Spectrum Analyzer are calibrated annually, and that calibration is directly traceable to the National Research Council of Canada. (NRC)
This equipment is only used by qualified technicians and only for the purpose of EMI measurements. The three metre test range has been carefully evaluated to the ANSI document C63.4-2003 and will be remeasured for reflections and losses every three years.

ADDITIONAL TEST EQUIPMENT LIST

1. Spectrum Analyzer: HP 8591EM, S/N 3639A00995, (9KHz - 1.8GHz), Calibration Due June 2006
2. Spectrum Analyzer: ANRITSU 2601A, S/N MT64544, (10KHz - 2.2GHz), Calibration Due June 2006
3. Spectrum Analyzer: IFR AN940, S/N 635001039, (9KHz - 26.5GHz), Calibration Due April 2006
4. Preamp: HP 8449B, S/N 3008A00378, (1 - 26.5GHz), Calibration Due August 2006
5. Horn Antenna: Q-PAR 6878/24, S/N 1721, (1.5-18GHz)
6. Horn Antenna: A. H. Systems SAS 572, S/N 164 (18 - 26.5GHz)
7. Line Impedance Stabilization Network.: Marstech, Calibration Due July 2006
8. Horn Antenna: Radar System (Flange 3/4" Square) MIL F 3922/68 (26.5 - 40GHz)
9. OML Mixer: M28HWD, S/N Ka31114-1 (26.5 - 40GHz), Calibration Due November 2005
10. OML Diplexer: DPL.313A (Unit plugs into M28HWD)
11. Semflex Cable: Used with M28HWD and DPL.313A