

APPENDIX B: SAR DIPOLE VERIFICATION PLOTS

FCC ID: IPH-A5126	RF EXPOSURE TEST REPORT	Approved by: Technical Manager
DUT Type: Portable Digital Transceiver	This report is Blockchain Verified. To verify the report authenticity, please visit HTTPS://WWW.BLOCKCHAIN-VERIFIED.COM or by clicking on the Blockchain logo. 	APPENDIX B

ELEMENT

DUT: Dipole 2450.000 MHz; Type: D2450V2 - SN797

Communication System: UID: 0, CW; Frequency: 2450.000 MHz

Medium: Head Simulating Liquid; Medium parameters used:

$f = 2450.000$ MHz; $\text{cond} = 1.80$ S/m; $\text{perm} = 38.5$; $\text{density} = 1000$ kg/m³

Phantom Section: Flat; Space: 10 mm

Test Date: 06/29/2026; Ambient Temp: 19.3°C; Tissue Temp: 20.9°C

Probe: EX3DV4 - SN7539; ConvF:(7.12,6.76,7.63); Calibrated: 2026-03-11

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn1450; Calibrated: 2026-03-10

Phantom: Twin-SAM V8.0; Serial: 1966

Measurement SW: DASY Module SAR V16.4.0.5005

2450.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: $dx=5.0$ mm, $dy=5.0$ mm, $dz=1.5$ mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 11.2 W/kg

SAR(1 g) = 5.47 W/kg; SAR(10 g) = 2.57 W/kg

Deviation (1 g) = 3.99%; Deviation (10 g) = 3.63%



ELEMENT

DUT: Dipole 2450.000 MHz; Type: D2450V2 - SN797

Communication System: UID: 0, CW; Frequency: 2450.000 MHz

Medium: Head Simulating Liquid; Medium parameters used:

f = 2450.000 MHz; cond = 1.75 S/m; perm = 38.0; density = 1000 kg/m³

Phantom Section: Flat; Space: 10 mm

Test Date: 07/23/2026; Ambient Temp: 22.2°C; Tissue Temp: 20.4°C

Probe: EX3DV4 - SN7539; ConvF:(7.12,6.76,7.63); Calibrated: 2026-03-11

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn1450; Calibrated: 2026-03-10

Phantom: Twin-SAM V8.0; Serial: 1966

Measurement SW: DASY Module SAR V16.4.0.5005

2450.0 MHz System Verification at 20.0 dBm (100 mW)


Area Scan (40.0 x 80.0): Measurement grid: dx=10.0 mm, dy=10.0 mm

Zoom Scan (30.0 x 30.0 x 31.2): Measurement grid: dx=5.0 mm, dy=5.0 mm, dz=1.5 mm; Graded Ratio: 1.5

Peak SAR (extrapolated) = 11.3 W/kg

SAR(1 g) = 5.64 W/kg; SAR(10 g) = 2.66 W/kg

Deviation (1 g) = 7.22%; Deviation (10 g) = 7.26%



ELEMENT

DUT: Dipole 5250.000 MHz; Type: D5GHzV2 - SN1191

Communication System: UID: 0, CW; Frequency: 5250.000 MHz
Medium: Head Simulating Liquid; Medium parameters used:
f = 5250.000 MHz; cond = 4.50 S/m; perm = 35.7; density = 1000 kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 07/13/2026; Ambient Temp: 22.0°C; Tissue Temp: 21.1°C

Probe: EX3DV4 - SN3914; ConvF:(6.02,5.45,5.48); Calibrated: 2026-05-08

Sensor-Surface: 1.4mm (VMS + 6p)

Electronics: DAE4 Sn728; Calibrated: 2026-05-07

Phantom: Twin-SAM V8.0; Serial: 2065

Measurement SW: DASY Module SAR V16.4.0.5005

5250.0 MHz System Verification at 17.0 dBm (50 mW)


Area Scan (40.0 x 80.0): Measurement grid: dx=10.0 mm, dy=10.0 mm

Zoom Scan (24.0 x 24.0 x 22.9): Measurement grid: dx=4.0 mm, dy=4.0 mm, dz=1.4 mm; Graded Ratio: 1.4

Peak SAR (extrapolated) = 15.4 W/kg

SAR(1 g) = 3.79 W/kg; SAR(10 g) = 1.09 W/kg

Deviation (1 g) = -3.32%; Deviation (10 g) = -3.11%



ELEMENT

DUT: Dipole 5600.000 MHz; Type: D5GHzV2 - SN1191

Communication System: UID: 0, CW; Frequency: 5600.000 MHz
Medium: Head Simulating Liquid; Medium parameters used:
 $f = 5600.000$ MHz; $\text{cond} = 4.86$ S/m; $\text{perm} = 34.9$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 07/13/2026; Ambient Temp: 22.0°C; Tissue Temp: 21.1°C

Probe: EX3DV4 - SN3914; ConvF:(5.21,4.72,4.74); Calibrated: 2026-05-08
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn728; Calibrated: 2026-05-07
Phantom: Twin-SAM V8.0; Serial: 2065
Measurement SW: DASY Module SAR V16.4.0.5005

5600.0 MHz System Verification at 17.0 dBm (50 mW)


Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (24.0 x 24.0 x 22.9): Measurement grid: $dx=4.0$ mm, $dy=4.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.4

Peak SAR (extrapolated) = 17.1 W/kg

SAR(1 g) = 3.89 W/kg; SAR(10 g) = 1.11 W/kg

Deviation (1 g) = 0.39%; Deviation (10 g) = -0.45%



ELEMENT

DUT: Dipole 5750.000 MHz; Type: D5GHzV2 - SN1191

Communication System: UID: 0, CW; Frequency: 5750.000 MHz
Medium: Head Simulating Liquid; Medium parameters used:
f = 5750.000 MHz; cond = 5.01 S/m; perm = 34.6; density = 1000 kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 07/13/2026; Ambient Temp: 22.0°C; Tissue Temp: 21.1°C

Probe: EX3DV4 - SN3914; ConvF:(5.37,4.86,4.88); Calibrated: 2026-05-08
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn728; Calibrated: 2026-05-07
Phantom: Twin-SAM V8.0; Serial: 2065
Measurement SW: DASY Module SAR V16.4.0.5005

5750.0 MHz System Verification at 17.0 dBm (50 mW)


Area Scan (40.0 x 80.0): Measurement grid: dx=10.0 mm, dy=10.0 mm

Zoom Scan (24.0 x 24.0 x 22.9): Measurement grid: dx=4.0 mm, dy=4.0 mm, dz=1.4 mm; Graded Ratio: 1.4

Peak SAR (extrapolated) = 17.3 W/kg

SAR(1 g) = 3.83 W/kg; SAR(10 g) = 1.10 W/kg

Deviation (1 g) = 1.59%; Deviation (10 g) = 2.33%



ELEMENT

DUT: Dipole 5850.000 MHz; Type: D5GHzV2 - SN1191

Communication System: UID: 0, CW; Frequency: 5850.000 MHz
Medium: Head Simulating Liquid; Medium parameters used:
 $f = 5850.000$ MHz; $\text{cond} = 5.11$ S/m; $\text{perm} = 34.5$; $\text{density} = 1000$ kg/m³
Phantom Section: Flat; Space: 10 mm

Test Date: 07/13/2026; Ambient Temp: 22.0°C; Tissue Temp: 21.1°C

Probe: EX3DV4 - SN3914; ConvF:(5.34,4.84,4.86); Calibrated: 2026-05-08
Sensor-Surface: 1.4mm (VMS + 6p)
Electronics: DAE4 Sn728; Calibrated: 2026-05-07
Phantom: Twin-SAM V8.0; Serial: 2065
Measurement SW: DASY Module SAR V16.4.0.5005

5850.0 MHz System Verification at 17.0 dBm (50 mW)

Area Scan (40.0 x 80.0): Measurement grid: $dx=10.0$ mm, $dy=10.0$ mm

Zoom Scan (24.0 x 24.0 x 22.9): Measurement grid: $dx=4.0$ mm, $dy=4.0$ mm, $dz=1.4$ mm; Graded Ratio: 1.4

Peak SAR (extrapolated) = 18.5 W/kg

SAR(1 g) = 3.98 W/kg; SAR(10 g) = 1.12 W/kg

Deviation (1 g) = 3.65%; Deviation (10 g) = 2.28%

