

Disclaimer of Liabilities

The ScottCare Corporation, its parent, affiliates, agents, officers and employees, shall not be liable for special, incidental, or consequential loss or damages of any kind resulting from or caused by any defect, failure or malfunctioning of the equipment described herein, whether a claim for such loss or damages is based upon warranty, contract or otherwise. The ECG/SpO2/Voice Transmitter/Receiver system, manufactured by the ScottCare Corporation is designed to amplify and transmit Rhythm Strips and SpO2 values along with voice transtelephonically, and the SCOTTCARE CORPORATION MAKES NO CLAIMS OR WARRANTIES, EXPRESS OR IMPLIED (INCLUDING WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE) AS TO ITS EFFECTIVENESS AS A DIAGNOSTIC, LIFE SAVING OR LIFE SUPPORT TOOL.

FCC Part 15 Information

The Telemetry Transmitter provided complies with the limits for a biomedical telemetry device pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation.

Operation of the Telemetry equipment is subject to the following two conditions:

1. The Telemetry Transmitter may not cause harmful interference.
2. The Telemetry Transmitter must accept any interference received, including interference that may cause undesired operation.

This Biomedical Telemetry device transmitter must be operated in accordance to the provisions of CFR47 Part 15.242 (a)(d)(e)(f)(g) and (h).

Any changes or modification not expressly approved by ScottCare could void the user's authority to operate the Telemetry equipment.