



**RETLIF
TESTING
LABORATORIES**

Put Us To The Testsm

101 New Boston Road
Goffstown, NH 03045 USA

Tel: (603) 497-4600 • Fax: (603) 497-5281 • www.retlif.com

FCC Part 15, Subpart C Test Report

On

13.56 MHz Patient Transmitter
FCC ID: KNK917

Customer Name: Secure Care Products

Customer P.O.: NIPO00229

Date of Report Revision: March 17, 2026

Test Report No.: R-6966H-1, Rev. A

Test Start Date: March 4, 2025

Test Finish Date: March 5, 2025

Test Technician: M. Seamans

Approved By: T. Hannemann

Report Prepared By: M. Johnson



Our letters, procedures and reports are for the exclusive use of the customer to whom they are addressed and their communication or the use of the name of Retlif Testing Laboratories must receive our prior written approval. Our letters, procedures and reports apply only to the sample tested and are not necessarily indicative of the qualities of apparently identical or similar products. The letters, procedures, reports and the name of Retlif Testing Laboratories or insignia are not to be used under any circumstances in advertising to the public. This report shall not be reproduced, except in full, without prior written approval of Retlif Testing Laboratories.



45 YEARS OF TESTING EXCELLENCE

Corporate Headquarters:
795 Marconi Avenue
Ronkonkoma, NY 11779 USA
Tel: (631) 737-1500
Fax: (631) 737-1497

3131 Detwiler Road
Harleysville, PA 19438 USA
Tel: (215) 256-4133
Fax: (215) 256-4130

Certification and Signatures

We certify that this report is a true representation of the results obtained from the tests of the equipment stated. We further certify that the measurements shown in this report were made in accordance with the procedures indicated and vouch for the qualifications of all Retlif Testing Laboratories personnel taking them.



Matt Seamans
EMC Test Technician
iNARTE Certified Technician EMC-48491-T



Mike Hippert
Laboratory Supervisor
iNARTE Certified Technician EMC-49190-T

Non-Warranty Provision

The testing services have been performed, findings obtained and reports prepared in accordance with generally accepted laboratory principles and practices. This warranty is in lieu of all others, either expressed or implied.

Non-Endorsement

This test report contains only findings and results arrived at after employing the specific test procedures and standards listed herein. It is not intended to constitute a recommendation, endorsement or certification of the product or material tested. This test report may not be used by the client to claim product endorsement by ANSI National Accreditation Board (ANAB).



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Revision History

Revisions to this document are listed below; the latest revised document supersedes all previous issues of this document.

Revision	Date	Pages Affected
-	February 19, 2026	Original Release
Rev. A	March 17, 2026	Page 7 • Update Results statement to include reason why a quasi-peak measurement was not recorded.



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Technical Information

Report Number: R-6966H-1, Rev. A

Applicant: Secure Care Products
39 Chenell Drive
Concord, NH 03301

Manufacturer: Valstone Systems Corporation USA, Inc.

Manufacturer Address: 7405 Rte. Transcanadienne Suite 100
Montreal, Quebec, Canada H4T 1Z2

Test Sample: 13.56 MHz Patient Transmitter

Model Number: A21350917

Brand Name: Secure Care Products

Power Requirements: 3 VDC

Frequency of Operation: 13.56 MHz

Antenna Type: Ferrite Loop 13.56 MHz Tuned Tank Antenna

Equipment Use: Portable transmitter



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Specification:

FCC Rules and Regulations Part 15, Subpart C, Section 15.209

Test Procedure:

ANSI C63.10:2020 + Corrigendum 1

ANSI C63.10a:2024 + Errata 1

Test Facility:

Retlif Testing Laboratories

101 New Boston Road

Goffstown, NH 03045

FCC Designation Number: US5327

EUT Description:

The Patient Transmitter provides positional awareness of personnel to security systems.

Tests Performed

The test methods performed on the EUT are shown below:

FCC Part 15, Subpart C	Test Method
15.209(a)	Field Strength of Fundamental
15.209 (a)	Field Strength of Spurious

General Test Requirements

1. The measurement procedures of ANSI C63.10 were utilized as specified in FCC Part 15, Subpart C, Section 15.31(a)(3).
2. All measurements were performed at a 3 meter test distance.
3. The EUT was rotated throughout 360 degrees for all radiated emissions measurements as specified in FCC Section 15.31(f)(5).
4. All readily accessible EUT controls were adjusted in such a manner as to maximize the level of emissions in accordance with FCC Section 15.31(g).
5. Appropriate accessories were attached to all EUT ports during the performance of radiated emissions measurements as required by FCC Section 15.31(i).



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Requirements and Test Results

Requirement:

FCC Section 15.209 (a)

Field Strength of Fundamental

FCC Section 15.209(a) – The field strength of any emission within the band 1.705 MHz – 30 MHz shall not exceed 30 uV/M at 30 meters.

Field Strength Measurement & Calculation:

The following spectrum analyzer settings were used:

RBW = 1 MHz for $f \geq 1$ GHz, 100 kHz for $f \leq 1$ GHz

VBW $\geq$ RBW

Detector Function = Peak or Average as applicable

Trace = Max Hold

Sweep = Auto

The maximized field strength of the emission was calculated as follows:

$$F_C = M_R + C_F$$

Where:

F_C = Corrected Field Strength Reading in dB μ V/m

M_R = Uncorrected Meter Reading in dB μ V

C_F = Correction Factor in dB (Pre-Amp + Antenna Factor + Cable Loss + Distance Factor)

For frequencies below 30 MHz a distance factor of -40dB/decade was utilized



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Requirements and Test Results (con't)

15.209 (a) Field Strength of Fundamental

Radiated Emissions Measurement Procedure:

The field strength of the fundamental emission was measured with a spectrum analyzer or EMI Receiver. The EUT was placed on an 80cm high test stand located 3 meters from the test antenna on a FCC listed open area test site. Emissions from the EUT were maximized by re-orientating the test sample, rotating the test sample 360 degrees, changing the orientation of the receive antenna and raising and lowering the test antenna from 1 – 4 meters. The maximized field strength of each observed emission was measured, recorded and compared to the specified limits of 15.209(a) as appropriate.

- **Results:** The maximized measured field strength of the fundamental emission was below the specified test limit of 15.209(a). Because the Peak field strength at the fundamental frequency was below the specified Quasi-Peak limit, a quasi-peak measurement was not recorded. See test data.

Requirement:

15.209 Radiated Emission Limits; General Requirements

Fundamental Frequency (MHz)	Field Strength of Fundamental microvolts/meter	Measurement Distance
0.009 to 0.490	2400/F(kHz)	300
0.490 to 1.705	24000/F(kHz)	30
1.705 to 30.0	30	30
30.0 to 88.0	100	3
88.0 to 216.0	150	3
216.0 to 960.0	200	3
Above 960.0	500	3



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Equipment Lists

FCC Section 15.209(a) – Field Strength of Fundamental

EN	Manufacturer	Model No.	Description	Serial No.	Due Date
012	ETS / EMCO	6502	ANTENNA, ACTIVE LOOP, 10 kHz - 30 MHz	8811-2255	10/31/2025
4029B	RETLIF	RNH	OPEN AREA TEST SITE, ATTENUATION, 3 / 10	001	10/31/2025
5242	TELEDYNE MICROWAVE	PR90-195-1275, 106'	CABLE, COAXIAL, 10 kHz - 6 GHz	N/A	3/31/2025
5274	ROHDE & SCHWARZ	ESW44	RECEIVER, EMI, 1 Hz - 44 GHz	103158	12/31/2025
5287	GOVEE	H5075	HYGROMETER, 0 - 50 deg. C, 0 - 99 %RH	SY51D31405244	12/31/2025

FCC Section 15.209(a) – Field Strength of Spurious

EN	Manufacturer	Model No.	Description	Serial No.	Due Date
012	ETS / EMCO	6502	ANTENNA, ACTIVE LOOP, 10 kHz - 30 MHz	8811-2255	10/31/2025
3427B	ETS / EMCO	3104	ANTENNA, BICONICAL, 20 - 200 MHz	2315	5/31/2025
4029B	RETLIF	RNH	OPEN AREA TEST SITE, ATTENUATION, 3 / 10 Meters	001	10/31/2025
443	ELECTRO-METRICS	LPA-25	ANTENNA, LOG PERIODIC, 200 MHz - 1000 MHz	1014	2/28/2026
5242	TELEDYNE MICROWAVE	PR90-195-1275, 106'	CABLE, COAXIAL, 10 kHz - 6 GHz	N/A	3/31/2025
5274	ROHDE & SCHWARZ	ESW44	RECEIVER, EMI, 1 Hz - 44 GHz	103158	12/31/2025
5287	GOVEE	H5075	HYGROMETER, 0 - 50 deg. C, 0 - 99 %RH	SY51D31405244	12/31/2025



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

**Test Photographs
Field Strength of Fundamental
FCC Part 15, Subpart C, Section 15.209(a)**



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs
Field Strength of Fundamental



Test Configuration



X Axis Configuration



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs
Field Strength of Fundamental



Y Axis Configuration



Z Axis Configuration



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

**FCC Section 15.209(a) – Field Strength of Fundamental
Test Data**



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

EMISSIONS TEST DATA SHEET

Method:	Field Strength of Emissions - Fundamental Field Strength
Test Specification:	FCC Part 15, Subpart C Paragraph: 15.209(a)
Job Number:	R-6966H-1
Customer:	Secure Care Products
Test Sample:	13.56 MHz Patient Transmitter
Model Number:	A21350917
Operating Mode:	Transmitting modulated signal at 13.56 MHz
Technician:	M. Seamans
Date(s):	March 4 th , 2025
Notes:	Test Distance: 3 meters Detector: Peak Resolution BW: 9 kHz

TEST PARAMETERS

Frequency		Measured Peak Level	Correction Factor	Corrected Peak Reading		Peak Reading	Converted Peak Reading	Quasi-Peak Limit at 3m
MHz		dBuV	dB	dBuV/m		dBuV/m	uV/m	uV/m
13.56		47.14	11.51	58.65		58.65	856.05	3000.00



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

**Test Photographs
Field Strength of Spurious
FCC Part 15, Subpart C, Section 15.209(a)**



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs
Field Strength of Spurious



Test Setup, X Axis Configuration



Test Setup, Y Axis Configuration



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs
Field Strength of Spurious



Test Setup, Z Axis Configuration



Test Setup, 9 kHz to 30 MHz



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs Field Strength of Spurious



Test Setup, 30 MHz to 200 MHz, Horizontal Antenna Polarization



Test Setup, 30 MHz to 200 MHz, Vertical Antenna Polarization



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

Test Photographs Field Strength of Spurious



Test Setup, 200 MHz to 1 GHz, Horizontal Antenna Polarization



Test Setup, 200 MHz to 1 GHz, Vertical Antenna Polarization



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

**FCC Part 15, Subpart C, Section 15.209(a) –
Field Strength of Spurious
Test Data**



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

EMISSIONS TEST DATA SHEET

Method:	Field Strength of Spurious Emissions
Test Specification:	FCC Part 15, Subpart C Paragraph: 15.209(a)
Job Number:	R-6966H-1
Customer:	Secure Care Products
Test Sample:	13.56 MHz Patient Transmitter
Model Number:	A21350917
Operating Mode:	Transmitting modulated signal at 13.56 MHz
Technician:	M. Seamans
Date(s):	March 4 th , 2025
Notes:	Test Distance: 3 meters Detector: Quasi-Peak (Average for 480kHz) Limits: Quasi-Peak (Average for the frequency bands 9-90kHz and 110-490kHz)

TEST PARAMETERS

Test Frequency	Antenna Position		Peak Reading	Correction Factor	Corrected Reading		Converted Peak Reading	Limit at 3M
MHz	(H/V)		dBuV	dB	dBuV/m		uV/m	uV/m
0.009	-	-	-	-	-	-	-	1769.912
	-	-	-	-	-	-	-	
0.480*	N/A	-	40.92	11.78	52.70	-	431.52	
	-	-	-	-	-	-	-	
0.490	-	-	-	-	-	-	-	1769.912
0.490	-	-	-	-	-	-	-	176.991
	-	-	-	-	-	-	-	
1.300*	N/A	-	23.25	12.05	35.30	-	58.21	
	-	-	-	-	-	-	-	
1.705	-	-	-	-	-	-	-	176.991
1.705	-	-	-	-	-	-	-	3000.000
	-	-	-	-	-	-	-	
1.720*	N/A	-	22.68	12.02	34.70	-	54.33	
2.150*	N/A	-	20.89	12.01	32.90	-	44.16	
5.000*	N/A	-	19.84	12.16	32.00	-	39.81	
25.000*	N/A	-	24.02	7.48	31.50	-	37.58	
	-	-	-	-	-	-	-	
	-	-	-	-	-	-	-	
30.000	-	-	-	-	-	-	-	3000.00

No EUT emissions were observed throughout the given frequency spectrum. *This emission is not from the EUT. It is a measurement of minimum measurement system sensitivity (Noise Floor).



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A

EMISSIONS TEST DATA SHEET

Method:	Field Strength of Spurious Emissions
Test Specification:	FCC Part 15, Subpart C Paragraph: 15.209(a)
Job Number:	R-6966H-1
Customer:	Secure Care Products
Test Sample:	13.56 MHz Patient Transmitter
Model Number:	A21350917
Operating Mode:	Transmitting modulated signal at 13.56 MHz
Technician:	M. Seamans
Date(s):	March 4 th , 2025
Notes:	Test Distance: 3 meters Detector: Quasi-peak Limits: Quasi-Peak

TEST PARAMETERS

Test Frequency	Antenna Position		Peak Reading	Correction Factor	Corrected Reading		Converted Peak Reading	Limit at 3M
MHz	(H/V)		dBuV	dB	dBuV/m		uV/m	uV/m
30.00	-	-	-	-	-	-	-	100.00
	-	-	-	-	-	-	-	
35.00*	H	-	7.16	12.34	19.50	-	9.44	
	-	-	-	-	-	-	-	
88.00	-	-	-	-	-	-	-	100.00
88.00	-	-	-	-	-	-	-	150.00
	-	-	-	-	-	-	-	
110.00*	H	-	11.72	14.38	26.10	-	20.18	
195.00*	H	-	6.07	19.43	25.50	-	18.84	
205.00*	H	-	6.38	18.12	24.50	-	16.79	
	-	-	-	-	-	-	-	
216.00	-	-	-	-	-	-	-	150.00
216.00	-	-	-	-	-	-	-	200.00
	-	-	-	-	-	-	-	
600.00*	H	-	3.38	23.72	27.10	-	22.65	
	-	-	-	-	-	-	-	
960.00	-	-	-	-	-	-	-	200.00
960.00	-	-	-	-	-	-	-	500.00
	-	-	-	-	-	-	-	
995.00*	H	-	3.12	30.88	34.00	-	50.12	
	-	-	-	-	-	-	-	
1000.00	-	-	-	-	-	-	-	500.00

No EUT emissions were observed throughout the given frequency spectrum. *This emission is not from the EUT. It is a measurement of minimum measurement system sensitivity (Noise Floor).



Retlif Testing Laboratories

Report No. R-6966H-1, Rev. A