

To whom it may concern

Stäfa (Switzerland), March 2022/ G.Borrett, Senior Manager Regulatory Affairs

Product Equality Declaration

We, Sonova AG, hereby declare under our own responsibility that the products listed below as “Hardware Equivalent Products - Phonak Brand” are in all relevant parts and hardware construction identical to the corresponding product identified as “Products with basis Hardware”. The following standards and/or technical regulations and corresponding test reports fully apply accordingly:

Standards Phonak brands
Hearing Aid standards: ACOUSTIC: IEC 60118-0-1-2-6 and NSH 7.0 (including Annex A); EMC: IEC 6011 8-13; EMC immunity: ANSI C63.19 Degrees of protection provided by enclosures (IP code): IEC 60529:1989-11+A1:1999; EN 60529:1991-10 (incl. Corrigendum: 1993-05)+A1:2000-02 Europe: HEALTH & SAFETY: IEC/EN 60601-1, IEC/EN 60601-2-66, IEC/EN 62368-1; IEC/EN 62133; IEC/EN 62479; EMC: IEC/EN 60601-1-2; EN 301489-1,-3,-17; SPECTRUM: EN 300328; EN ETSI 303417; REACH; RoHS-III USA: 47 CFR Part 15(B), Part 15(C): 15.249, Part 2: 2.1091, 2.1093; Canada: RSS-Gen, ICES-003, RSS-210, RSS-102; RSS-247 ; RSS-216 Japan: ARIB 166, Ordinance regulating Radio Equipment (2005-08) Ar12 item 19

Sonova AG
Laubisrütistrasse 28
8712 Stäfa
Switzerland

Phone +41 58 928 01 01
Fax +41 58 928 20 11
info@sonova.com
www.sonova.com

The only difference between the basis product and the equivalent ones is a set of different audiological features per performance level and marketing product name. The difference between devices indicated as “left” and “right” are minor changes to plastic parts only. There will be no impact on the wireless characteristics.

Products with basis Hardware (tested representative model)	Hardware Equivalent Products – Phonak Brand
Phonak Slim P90-R left HW version: 050-0850-xx (left) SW version: Target 7.2 FW version: 067-1462	Phonak Slim P90-R right (with S/M/P/MAV receivers) HW version: 050-0851-xx (right) SW version: Target 7.2 FW version: 067-1462
	Phonak Slim P70-R left Phonak Slim P70-R right (with S/M/P/MAV receivers) HW version: 050-0848-xx (left),050-0849-xx (right) SW version: Target 7.2 FW version: 067-1463
	Phonak Slim P50-R left Phonak Slim P50-R right (with S/M/P/MAV receivers) HW version:050-1009-xx (left),050-1010-xx (right) SW version: Target 7.2 FW version: 067-1470
	Phonak Slim P30-R left Phonak Slim P30-R right (with S/M/P/MAV receivers) HW version:050-1007-xx (left),050-1008-xx (right) SW version: Target 7.2 FW version: 067-1471
	Phonak Slim P-R Trial left Phonak Slim P-R Trial right (with S/M/P/MAV receivers) HW version:050-0853-2A (left),050-0854-2A (right) SW version: Target 7.2 FW version: 067-1462

Furthermore, this document declares that the following devices, which are non-Phonak branded, differ from the model under test "Phonak Slim P90-R left" mainly with respect to color, naming and some minor changes to top and bottom housing (plastic parts only), that are not considered to be significant enough to cause any measurable effect for electrical safety or RF performance according to the list of standards indicated below.

Standards non-Phonak brands

Hearing Aid standards:

Degrees of protection provided by enclosures (IP code): IEC 60529:1989-11+A1:1999; EN 60529:1991-10 (incl. Corrigendum: 1993-05)+A1:2000-02

Europe: HEALTH & SAFETY: IEC/EN 60601-1, IEC/EN 60601-2-66, IEC/EN 62368-1; IEC/EN 62133; IEC/EN 62479; EMC: IEC/EN 60601-1-2; EN 301489-1,-3,-17; SPECTRUM: EN 300328; EN ETSI 303417; REACH; RoHS-III

USA: 47 CFR Part 15(B), Part 15(C): 15.249, Part 2: 2.1091, 2.1093;

Canada: RSS-Gen, ICES-003, RSS-210, RSS-102; RSS-247 ; RSS-216

Japan: ARIB 166, Ordinance regulating Radio Equipment (2005-08) Ar12 item 19

Unitron	Vista	HANSATON	Specsavers
Moxi B9-RS Moxi B7-RS Moxi B5-RS Moxi B3-RS Moxi B1-RS Moxi B-RS FLEX:TRIAL	Vista B9-RS Vista B7-RS Vista B5-RS Vista B3-RS Vista B1-RS Vista B-RS trial Vista B-RS Select	AQ sound ST 9-S AQ sound ST 7-S AQ sound ST 5-S AQ sound ST 3-S AQ sound ST 1-S AQ sound ST S myChoice	Advance 871-RS Advance 771-RS Advance 671-RS Advance 471-RS Advance 271-RS

where xx – means color identification

Staefa, 31 March 2022



Blathnaid Feldman

Senior Director Global Quality Management

Staefa, 31 March 2022



Glenn Borrett

Senior Regulatory Affairs Manager

Type:	PDL	Status:	CURRENT
Document No.:	PDL-7789	Effective Date:	04-Apr-2022
Rev.:	2		
Title:	DoE Phonak Slim P-R		
Process:	Innovation Management		
Owner:	11TKIEFER Thomas Kiefer		

Attributes

Attribute Type	Value	Description
Affected Site	1100	Stafa (CH)
Project	SlimDesign	
Project Phase	Design Output	

Approvals

Level	Actor	Job Title	Sign-off Date	Sign-off By
1	Nataya Jaiman	TQM Regulatory Affairs Assistant	31-Mar-2022	11NJAIMAN
2	Glenn Borrett	Senior Manager Regulatory Affairs	04-Apr-2022	11GBORRETT
2	Blathnaid Feldman	Director Quality Management & Regulatory Affairs	01-Apr-2022	11BFELDMAN

Revision Notes

Access Activity	Note	Accessed By	Accessed Date
Remark	Added a phrase to the Phonak Slim DoE to justify the equality among the different performance levels	11NJAIMAN	31-Mar-2022