

Appendix B – System Check Plots

Test Date : 2026-03-24 | Ambient Temp : 21.7 °C | Tissue Temp : 21.4 °C

System Performance Check

System Performance Check at 2450 MHz

Verification Source Properties

Manufacturer	Model No.	Serial No.	Input Power [dBm]
SPEAG	D2450	1087	17.0

Exposure Conditions

Phantom Section	Group	Frequency [MHz]	Conversion Factor	TSL Permittivity	TSL Conductivity [S/m]
Flat	CW	2450.000	6.6	39.0	1.76

Hardware Setup


Phantom	Tissue Simulating Liquid	Probe Calibration Date	DAE Calibration Date
ELI V8.0 (20deg probe tilt) - 2180	HBBL-600-10000	EX3DV4 - SN7757 / 2025-09-25	DAE4 Sn1743 / 2025-09-11
Measurement Software Version		16.4.0.5005	

Scan Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 30.0
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
Grading Ratio	N/A	1.5

Measurement Results

	Area Scan	Zoom Scan
psSAR-1g [W/kg]	2.71	2.61
psSAR-10g [W/kg]	1.28	1.25
Power Drift [dB]		0.01
TSL Correction	Positive only	Positive only



Test Date : 2026-03-24 | Ambient Temp : 21.7 °C | Tissue Temp : 21.4 °C

System Performance Check**System Performance Check at 6500 MHz****Verification Source Properties**

Manufacturer	Model No.	Serial No.	Input Power [dBm]
SPEAG	D6500	1081	20.0

Exposure Conditions

Phantom Section	Group	Frequency [MHz]	Conversion Factor	TSL Permittivity	TSL Conductivity [S/m]
Flat	CW	6500.000	4.56	32.1	5.79

Hardware Setup


Phantom	Tissue Simulating Liquid	Probe / Calibration Date	DAE / Calibration Date
ELI V8.0 (20deg probe tilt) - 2180	HBBL-600-10000	EX3DV4 - SN7757 / 2025-09-25	DAE4 Sn1743 / 2025-09-11
Measurement Software Version		16.4.0.5005	

Scan Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	51.0 x 85.0	22.0 x 22.0 x 22.0
Grid Steps [mm]	8.5 x 8.5	3.4 x 3.4 x 1.4
Sensor Surface [mm]	3.0	1.4
Graded Grid	N/A	Yes
Grading Ratio	N/A	1.4

Measurement Results

	Area Scan	Zoom Scan
psSAR-1g [W/kg]	24.0	29.5
psSAR-8g [W/kg]	6.02	6.61
psSAR-10g [W/kg]	4.99	5.47
psAPD (1.0 cm ² , sq) [W/m ²]		295
psAPD (4.0 cm ² , sq) [W/m ²]		131
Power Drift [dB]		-0.02
TSL Correction	Positive only	Positive only



Test Date : 2026-03-24 | Ambient Temp : 22.4 °C

System Performance Check**System Performance Check at 10GHz****Verification Source Properties**

Manufacturer	Model No.	Serial No.
SPEAG	H10GVx	1060

Exposure Conditions

Phantom Section	Group	Frequency [MHz]	Conversion Factor
5G	CW	10000.0	1.0

Hardware Setup

Phantom	Medium	Probe / Calibration Date	DAE / Calibration Date
mmWave - 1102	Air	EUmmWV4 - SN9639_F1- 55GHz / 2025-09-11	DAE4 Sn1741 / 2025-07-11
Measurement Software Version		v3.2.2.2358	

Scan Setup

	5G Scan
Grid Extents [mm]	60.0 x 60.0
Grid Steps [mm]	0.125 x 0.125
Sensor Surface [mm]	10.0

Measurement Results

	5G Scan
Avg. Area [cm ²]	4.00
psPD n+ [W/m ²]	59.9
psPD tot+ [W/m ²]	60.2
psPD mod+ [W/m ²]	56.6
E max [V/m]	61.5
Power Drift [dB]	0.01

