

EMERSON | AMS | MODEL A9530 | MADE IN THAILAND

WARNING: Replace Battery only with Tadiran P/N TL-4920M or Emerson A0702PPU. Reference D25902 for installation instructions.



GSA-CUR
Class I Div
Class I Zone
-40°C to 55°C
Type 4x D



ATEX GROUP
II 1 G Ex
-40°C to 55°C

IECEx GROUP
Ex II IC
Ta -40°C to 55°C





CSA-C/US CSA20CA80028624

Class I Div.1 Grps. A,B,C & D

Class I Zone 0 AEx / Ex ia IIC T4 Ga

-40°C ≤ Ta ≤ 85°C

Type 4x Enclosure



ATEX CSANo. 20ATEX 2042X

II 1 G Ex ia IIC T4 Ga (135°C)

-40°C ≤ Ta ≤ 85°C IP66

IECEX CSA 20.0002X

Ex ia IIC T4 Ga

Ta -40°C ≤ Ta ≤ 85°C



FCC ID: NL5-A9530M1

IC ID: 3434A-A9530M1

Knoxville USA 37932

