



APPENDIX 6

TEST FACILITY



TEST FACILITY

Location: No. 15, 14 Line, Chin Twu Chi, Lu Chu Hsiang, Taoyuan, Taiwan, R.O.C.

Description: There are two 3/10m open area test sites and two line conducted labs for final test, and one 3/10m open area test site for engineering lab. The Open Area Test Sites and the Line Conducted labs are constructed and calibrated to meet the FCC requirements in documents ANSI C63.4: 1992 and CISPR 22/EN 55022 requirements.

Site Filing: A site description is on file with the Federal Communications Commission, 7435 Oakland Mills Road, Columbia, MD 21046.

Registration also was made with Voluntary Control Council for Interference (VCCI).

Site Accreditation: Accredited by NEMKO (Authorization #: ELA 124) for EMC & A2LA (Certificate #: 824.01) for Emission

Instrument Tolerance: All measuring equipment is in accord with ANSI C63.4 and CISPR22 requirement that meet industry regulatory agency and accreditation agency requirement.

Ground Plane: Two conductive reference ground planes were used during the Line Conducted Emission, one in vertical and the other in horizontal. The dimensions of these ground planes are as below. The vertical ground plane was placed distancing 40 cm to the rear of the wooden test table on where the EUT and the support equipment were placed during test. The horizontal ground plane projected 50 cm beyond the footprint of the EUT system and distanced 80 cm to the wooden test table. For Radiated Emission Test, one horizontal conductive ground plane extended at least 1m beyond the periphery of the EUT and the largest measuring antenna, and covered the entire area between the EUT and the antenna. It has no holes or gaps having longitudinal dimensions larger than one-tenth of a wavelength at the highest frequency of measurement up to 1GHz.

Site #1 and # 3 Line Conducted Test Site: Vertical ground plane (2.2m x 2.2m)
Horizontal ground plane (2.5m x 2.5m)

Site #4 Line Conducted Test Site: At Shielding Room



APPENDIX 7

TEST EQUIPMENT

TEST EQUIPMENT LIST (EN 55022)

Instrumentation: The following list contains equipment used at C & C Laboratory, Co., Ltd. for testing. The equipment conforms to the CISPR 16-1 / ANSI C63.2-1988 Specifications for Electromagnetic Interference and Field Strength Instrumentation from 10kHz to 1.0 / 5.0 GHz.

Equipment used during the tests:

Open Area Test Site: ; 1/4# 1; ; 1/4# 3; ; 1/2# 4

Open Area Test Site # 1					
EQUIPMENT TYPE	MFR	MODEL NUMBER	SERIAL NUMBER	LAST CAL.	CAL. DUE
Q.P Adaptor	HP	85650A	2811A01399	05/05/2000	05/04/2001
RF Pre-selector	HP	85685A	2947A01064	05/05/2000	05/04/2001
Spectrum Analyzer	HP	8568B	3001A05004	05/05/2000	05/04/2001
S.P.A Display	HP	8568B	3014A18846	05/05/2000	05/04/2001
Precision Dipole	R&S	HZ-12	846932/0004	07/14/2000	07/13/2001
Precision Dipole	R&S	HZ-13	846556/0008	07/14/2000	07/13/2001
Bilog Antenna	CHASE	CBL6112A	2309	02/13/2000	02/12/2001
Turn Table	EMCO	2081-1.21	N/A	N.C.R	N.C.R
Antenna Tower	EMCO	2075-2	9707-2604	N.C.R	N.C.R
Controller	EMCO	2090	N/A	N.C.R	N.C.R
RF Switch	ANRITSU	MP59B	M54367	N.C.R	N.C.R
Site NSA	C&C	N/A	N/A	11/10/1999	11/09/2000

Open Area Test Site # 3					
EQUIPMENT TYPE	MFR	MODEL NUMBER	SERIAL NUMBER	LAST CAL.	CAL. DUE
Spectrum Analyzer	ADVANTEST	R3261C	71720533	10/25/1999	10/24/2000
Pre-Amplifier	HP	8447D	2944A09173	02/01/2000	01/31/2001
EMI Test Receiver	R&S	ESVS20	838804/004	12/24/1999	12/23/2000
Precision Dipole	R&S	HZ-12	846932/0004	07/14/2000	07/13/2001
Precision Dipole	R&S	HZ-13	846556/0008	07/14/2000	07/13/2001
Bilog Antenna	CHASE	CBL6112A	2179	11/27/1999	11/26/2000
Turn Table	EMCO	2081-1.21	9709-1885	N.C.R	N.C.R
Antenna Tower	EMCO	2075-2	9707-2060	N.C.R	N.C.R
Controller	EMCO	2090	9709-1256	N.C.R	N.C.R
RF Switch	ANRITSU	MP59B	M53867	N.C.R	N.C.R
Site NSA	C&C	N/A	N/A	01/30/2000	01/29/2001



Open Area Test Site # 4					
EQUIPMENT TYPE	MFR	MODEL NUMBER	SERIAL NUMBER	LAST CAL.	CAL. DUE
Spectrum Analyzer	ADVANTEST	R3261C	81720301	09/02/2000	09/01/2001
Pre-Amplifier	HP	8447F	2944A03748	10/22/1999	10/21/2000
EMI Test Receiver	R&S	ESCS30	845552/030	12/04/1999	12/03/2000
Precision Dipole	R&S	HZ-12	846932/0004	07/14/2000	07/13/2001
Precision Dipole	R&S	HZ-13	846556/0008	07/14/2000	07/13/2001
Bilog Antenna	CHASE	CBL 6112B	2462	01/13/2000	01/12/2001
Turn Table	Chance most	N/A	N/A	N.C.R	N.C.R
Antenna Tower	Chance most	N/A	N/A	N.C.R	N.C.R
Controller	Chance most	N/A	N/A	N.C.R	N.C.R
RF Switch	ANRITSU	MP59B	M51067	N.C.R	N.C.R
Site NSA	C&C Lab.	N/A	N/A	12/26/1999	12/25/2000



Conducted Emission Test Site: i 1/4# 3 ; i 1/4# 4

Conducted Emission Test Site # 3					
EQUIPMENT TYPE	MFR	MODEL NUMBER	SERIAL NUMBER	LAST CAL.	CAL. DUE
EMI Test Receiver	R&S	ESCS30	847793/012	11/06/1999	11/05/2000
LISN	EMCO	3825/2	9003-1628	07/12/2000	07/11/2001
LISN	R&S	ESH3-Z5	848773/014	10/22/1999	10/21/2000

Conducted Emission Test Site # 4					
EQUIPMENT TYPE	MFR	MODEL NUMBER	SERIAL NUMBER	LAST CAL.	CAL. DUE
EMI Test Receiver	R&S	ESHS10	843743/015	12/10/1999	12/09/2000
LISN	EMCO	3825/2	9003/1382	01/10/2000	01/09/2001
LISN	R&S	ESH2-Z5	843250/010	12/06/1999	12/05/2000

The calibrations of the measuring instruments, including any accessories that may effect such calibration, are checked frequently to assure their accuracy. Adjustments are made and correction factors applied in accordance with instructions contained in the manual for the measuring instrument.



APPENDIX 8

BLOCK DIAGRAM OF TEST SETUP

BLOCK DIAGRAM OF TEST SETUP

System Diagram of Connections between EUT and Simulators

EUT: POCKET PC

Trade Name: Intermec, HTC

Model Number: PB10

DC Power Cable: Unshielded, 1.8m

