

SmartPill - Identification of equipment according to FCC section 2.925(f)


The FCC identifier and label will not be placed on the device itself for two reasons:
(i) the Capsule is so small that placing the label on it impracticable; and
(ii) once the SmartPill capsule package is opened it should be activated and swallowed by the patient, therefore having the label on the Capsule would be of little use.

Consistent with FCC 47 CFR § 2.925(f) and the accompanying note (“a device intended to be implanted within the body of a . . . person would probably require an alternate method of identification”), the Applicant proposes, as an alternative method of affixing the label, to affix it to the Capsule’s packaging as shown below and in the User’s Manual, which is provided as a separate attachment accompanying the application.

Given Imaging is submitting in Application for certification the **EMC Emission Report**, which demonstrates the Capsule’s compliance with the FCC standard requirements.

The SmartPill Capsule is packaged in a 5-pack box as presented below:





One pack are not sold individually, still each 1 –pack will also be labeled as follows:

