



Dedicated to improving lives.

AXIS TRANSMITTER
Operator's Manual
Life Sensing Instruments, Inc.
December 16, 2025

Axis Transmitter Overview

The Axis Transmitter is a component of a wireless medical patient-monitoring system designed to acquire and transmit electrocardiogram (ECG) physiological data to a telemetry receiver. The receiver forwards the data to a central monitoring station for display, trending, and analysis by healthcare professionals.

The Axis uses four patient lead wires connected to disposable ECG electrodes to derive three ECG channels. The analog ECG signals are digitized, processed, and transmitted wirelessly using the Wireless Medical Telemetry Service (WMTS) frequency bands.

Axis Keypad/Buttons

The keypad provides a protective cover for the LCD, contains the assigned TrensCenter Channel Number, and has (4) buttons:

- **POWER BUTTON:**
 - This button has a dual functionality – the button is used for powering the transmitter and displays high-level status information.
 - **Power Functionality:** Press and hold the button for (2) seconds to power on or off the transmitter. The transmitter will make an audible beep to indicate the transmitter is powering on or off.
 - **Status Functionality:** When the PMT is powered on, the Power Button will be illuminated with different colors to indicate the transmitter's status:
 - **Orange:** Transmitter is powered on without an active communication link with the Nexus Telemetry Receiver. To resolve, ensure that the transmitter is in the pre-defined wireless range and confirm that the Nexus Telemetry Receiver is powered on.
 - **Purple:** Transmitter is powered on with an active communication link with the Nexus Telemetry Receiver.
 - **Blue:** Transmitter is powered on with an active communication link with the Nexus Telemetry Receiver, and the transmitter has been admitted to TrensCenter software.
 - **Red:** Transmitter has reached a state of low battery. Replace batteries to resolve.
 - **Main Screen Functionality:** Press this button to navigate to the Main Screen. **Note:** The Main Screen only displays when the Transmitter is admitted to TrensCenter.
- **VIEW BUTTON:**
 - **Lead Off Functionality:** Press this button to navigate to the View Screen which displays the connection status of each ECG electrode. A **Green Checkmark** indicates that the electrode is connected. A **Red X** indicates that the electrode is disconnected. **Note:** This button only has functionality when the transmitter is admitted to TrensCenter.
- **STATUS BUTTON:**
 - This button is used to access the Status Screen for support purposes.
- **REMOTE RECORD BUTTON:**
 - Press and hold this button for (1) second to store a "Remote" Event in TrensCenter. **Note:** This button only has functionality when the transmitter is admitted to TrensCenter.

Axis Display Overview

The display has three modes of operation that are designed to maximize battery life.

- **Active Mode:** The LCD is turned on and at full brightness. This mode is initiated when the keypad buttons are pressed, or the transmitter is powered on. Active Mode uses the most current and impacts battery life.
- **Inactive Mode:** The LCD is turned on but has dimmed the brightness level due to inactivity.
- **Sleep Mode:** The LCD is turned off due to inactivity. Sleep Mode uses the least amount of current and increases battery life.

The display is configured to display **Status Indicators** at the top of all screens. Examples of **Status Indicators** include Channel Number, Filter/Monitor, Pacer, Remote Record, Primary/Auxiliary, Battery Strength, and Signal Strength.

The **Main Screen** displays the patient's **Heart Rate** (as determined and reported by TrensCenter) and the patient's configured **Target Heart Rate Range**.

Instructions for Use

Insert the Batteries

Unscrew each battery cap by grasping the battery cap and turning it counterclockwise until the cap has been full removed. Insert (1) AAA battery into each battery well following the battery direction indicator on the back of the transmitter. Repeat the above instructions for the battery caps but rotate the battery cap clockwise to tighten.

CAUTION: MAKE SURE THE BATTERIES ARE INSTALLED CORRECTLY. IF THE BATTERIES ARE INSTALLED INCORRECTLY, THE TRANSMITTER WILL NOT FUNCTION.

CAUTION: MAKE SURE THAT THE BATTERY CAP IS TIGHT. IF THE CAP IS NOT TIGHTENED PROPERLY, THE TRANSMITTER SIGNAL MAY BE INTERMITTENT, AND THE WATERPROOFING SEAL MAY NOT SEAL WHICH MAY DAMAGE TO THE TRANSMITTER.

Power on the Transmitter

Press and hold the button for (2) seconds to power on or off the transmitter. The transmitter will make an audible beep to indicate the transmitter is powering on or off. The Power Button's LED will flash **Purple** to indicate the transmitter has a communication link with the Nexus Telemetry Receiver.

Connect the Lead Wires & Electrodes

The patient lead wires have a **color-coded connector** that plugs into a **corresponding color-coded transmitter connector**. When inserting the lead wires, insert the connector until the cables fit firmly into the transmitter.




CAUTION: ONLY USE THE LEAD WIRES SUPPLIED BY LSI FOR OPTIMAL PERFORMANCE.

Gently abrade the electrode sites with an alcohol prep pad until slight redness is observed. Dry the site before applying the electrodes. Attach the electrodes to the patient on the prepared sites. Apply pressure to the outer edges of the electrode, not the center. Applying pressure to the center may cause the electrode conductive jelly to ooze out, resulting in a loose electrode. Attach the electrodes to the lead wires.

For optimum monitoring, the electrodes should be placed in the following locations:

- **WHITE (Right Arm):** Right Clavicle and Right Sternal Border in the interclavicular notch
- **RED (Left Leg):** Lower Left Rib Cage, at the V6 chest lead position
- **BLACK (Left Arm):** Left Clavicle and Left Sternal Border in the interclavicular notch
- **GREEN (Right Leg):** Grounding electrode, Lower Right Rib Cage



Verify Electrode Placement

Admit the patient in TrensCenter to the designated Channel. Confirm that the waveform is accurate. Adjust electrodes as necessary to optimize waveform display.

Remote Record Button

Press and hold the Remote Record Button for (1) second to store a "Remote" Event in TrensCenter.

Battery Life Indicator

The Battery Life Indicator on the LCD displays the approximate remaining battery capacity. The icon uses a segmented battery graphic – two white cells indicate maximum battery life, while one yellow cell indicates approximately 50% remaining. When the battery level reaches a critical threshold, the battery icon turns red and the power button begins flashing red. This low-battery alert provides at least 15 minutes of warning before the transmitter powers down automatically.

Signal Strength Indicator

The Signal Strength Indicator located on the LCD displays the quality of the wireless connection between the Axis transmitter and the Nexus Telemetry Receiver. The icon uses four bars to represent signal strength – four bars indicate an ideal connection, while a single red dot indicates the weakest signal. If the indicator shows low strength, move the transmitter closer to the ceiling-mount antenna or ensure there are no obstructions between the transmitter and the ceiling-mount antenna.

Support Services

The authorized support provider for the Axis Transmitter is Life Systems International (LSI), located in Charlotte, NC. Support services include hardware support available through the Repair Depot. To learn more about the support services offered by LSI, call 800-846-1279 x1.


Repair Depot

For repair services of the Axis Transmitter, contact the authorized support provider:

LSI
Attn: Repair Depot
2102 Cambridge Beltway Dr., Ste. A
Charlotte, NC 28273
800-846-1279 x1
support@lsi-medical.com

Axis Transmitter Specifications

MECHANICAL	
PHYSICAL DIMENSIONS (WxHxD)	2.6" W x 3.4" H x .9" D
WEIGHT (WITHOUT BATTERIES)	3.7 oz
WEIGHT (WITH BATTERIES, CABLES)	6.0 oz
BATTERY COMPARTMENT	2 COMPARTMENTS, POLARITY PROTECTED
DATA	
ECG SAMPLE RATE	256/SECOND
ELECTRICAL	
BATTERY TYPE	(2) AAA BATTERIES
INPUT VOLTAGE	1.5 – 3.0 VDC (2 x AAA BATTERIES)
RUN TIME	40 HOURS
LOW BATTERY ALERT	AT LEAST 15 MINUTES PRIOR TO SHUTDOWN
WIRELESS	
MODULATION	4GFSK (WMTS), FSK 150 KHz (ISM)
FREQUENCY BAND (TRANSMIT)	608.00-614.00, 1395.00-1400.00, 1427.00-1432.00 MHz
FREQUENCY BAND (RECEIVING)	910.00 – 920.00 MHz
ANTENNA	
TYPE	DIPOLE FLEX FILM (INTERNAL)



Tested to Comply with FCC Standards
Product Name: Axis Transmitter
Model Number: AXIS
FCC ID: OF8-AX1

Responsible Party:

Life Sensing Instruments, Inc.
2102-A1 Cambridge Beltway Dr.
Charlotte, NC 28273
800-624-2732

Life Sensing Instruments, Inc. declares that the above device complies with the requirements of the Federal Communications Commission.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- 1) this device may not cause harmful interference, and
- 2) this device must accept any interference received, including interference that may cause undesired operation.

Changes or modifications not expressly approved by Life Sensing Instruments, Inc. could void the user's authority to operate the equipment.

This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

This equipment complies with FCC Part 95 Subpart H for Wireless Medical Telemetry Service (WMTS). Operation is subject to the condition that this device does not cause harmful interference. Use of this equipment is restricted to healthcare facilities by authorized health care providers or their designated agents. Manufacturers, installers, and users of WMTS equipment are cautioned that the operation of this equipment could result in harmful interference to other nearby medical devices, in accordance with FCC 95.2395. In accordance with FCC 95.2309(h), this WMTS transmitter must be registered with an FCC-designated frequency coordinator prior to use.

Use of the equipment requires prior coordination with the ASHE frequency coordination designator. For further guidance, visit the ASHE website at: www.wmtssearch.com. Operation in the 608 – 614 MHz band must be coordinated with an FCC-designated WMTS frequency coordinator. Operation in the 1395 – 1400 MHz and 1427 – 1432 MHz bands may also require frequency coordination to avoid interference with coexisting systems.