

4/15/2019

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

To whom it may concern:

Medtronic Minimed, Inc, hereby requests a Permissive Change Class II of the radio equipment certification for the product name certification number, FCC ID: OH27731. This product differs from the originally approved product in the following manor:

(a) Associated digital circuitry:

- C48: Filter cap at the positive output of the balun
- R42 & R43: Bias resistors for the RF chip
- D2: ESD diode on the antenna
- R40: Pull up resistor for the chip select on the SPI bus
- EEPROM

(b) Functional capabilities: No change.

(c) Antenna Characteristics: No change

(d) Cosmetic differences – None

I do attest that the Field strength and/or RF Output readings remained the same or are lower than the originally approved product and that they did not increase therefore qualifying this product for this application type.

Sincerely,

A handwritten signature in black ink, appearing to read "Bob Vitti", is located below the "Sincerely," text.

Bob Vitti

Sr Program Manager| RF-EMC-Safety Certification
Medtronic, Inc. | Diabetes