

NOTES:

SPURIOUS EMISSIONS AT ANTENNA TERMINAL, FCC 2.1051
 E EXEMPT FROM FCC 90.210 EMISSION MASK REQUIREMENT
 DIRECT CONNECTION

MEAS TYPE:

- ☐ Radiated Prescan
☐ Radiated Final
☐ Conducted
☐ Disturbance Power
☒ Other **1 ANTENNA TERMINAL**

ANTENNA/COUPLER:

- ☐ 9124 Bicon
☐ 3146 Log Per
☐ 3106 Horn
☐ 3109 Bicon
☐ 3115 Horn
☐ CBL6112B Bilog
☐ CBL6140 X-Wing
☐ MDS-21 Clamp
☐ NSLK 8126 LISN
☐ NNB-4/63TL LISN
☐ NNB-4/200X LISN
☒ NA

POLARIZATION:

- ☐ Horizontal
☐ Vertical
☒ NA

DISTANCE:

- ☐ 3 Meter
☐ 10 Meter
☐ Meter
☒ NA

LOCATION:

- ☐ OATS
☐ Semi-Anechoic
☐ Shielded Room
☐ Factory Floor
☐ Other



20:38:24 OCT 12 2000

GE MEDICAL MODEL 340 EMISSION MASK

MARKER

13.71 MHz
 -50.31 dBm


ACTV DET: PEAK

MEAS DET: PEAK QP AVG
 MKR 13.71 MHz
 -50.31 dBm

EXTAMP -11.5 dB

11.5 dB TOTAL ADDED ATTENUATION

LOG REF -15.0 dBm



20:33:30 OCT 12 2000

GE MEDICAL MODEL 340 EMISSION MASK

MARKER


237.4 MHz
 -46.58 dBm

ACTV DET: PEAK

MEAS DET: PEAK QP AVG
 MKR 237.4 MHz
 -46.58 dBm

EXTAMP -11.5 dB

LOG REF 10.0 dBm



NOTES:

SPURIOUS EMISSIONS AT ANTENNA TERMINAL, FCC 2.1051
 DIRECT CONNECTION

MEAS TYPE:

☐ Radiated Prescan
☐ Radiated Final
☐ Conducted
☐ Disturbance Power
☒ Other **ANTENNA TERMINAL**

ANTENNA/COUPLER:

☐ 9124 Bicon
☐ 3146 Log Per
☐ 3106 Horn
☐ 3109 Bicon
☐ 3115 Horn
☐ CBL6112B Bilog
☐ CBL6140 X-Wing
☐ MDS-21 Clamp
☐ NSLK 8126 LISN
☐ NNB-4/63TL LISN
☐ NNB-4/200X LISN
☒ **N/A**

POLARIZATION:

☐ Horizontal
☐ Vertical
☒ **NA**

DISTANCE:

☐ 3 Meter
☐ 10 Meter
☐ Meter
☒ **NA**

LOCATION:

☐ OATS
☐ Semi-Anechoic
☐ Shielded Room
☐ Factory Floor
☒ Other




20:45:53 OCT 12, 2000
 GE MEDICAL MODEL 340 EMISSION MASK

MARKER
 460.6628 MHz
 11.19 dBm

ACTV DET: PEAK
 MEAS DET: PEAK QP AVG
 MKR 460.6628 MHz
 11.19 dBm
 EXTAMP -11.5 dB

LOG REF 15.0 dBm



20:24:47 OCT 12, 2000
 GE MEDICAL MODEL 340 EMISSION MASK

MARKER
 4.141 GHz
 -49.15 dBm

ACTV DET: PEAK
 MEAS DET: PEAK QP AVG
 MKR 4.141 GHz
 -49.15 dBm
 EXTAMP -11.5 dB
 PREAMP ON

LOG REF -15.0 dBm

