




Drawings for Proposed Labels



Gasman II Portable Gas Detector

Crowcon Detection Instruments Limited
2 Blacklands Way
Abingdon Business Park
Abingdon
Oxfordshire
OX14 1DY
United Kingdom

DRG. No.	GASMAN-2677-A3	ISS.	1
USED ON	STOCK No.		



CLIP LABELS







CASE LABELS

1. MATERIAL: 0.076 THK. TAMPER EVIDENT POLYESTER, SELF ADHESIVE.
2. TEXT: THERMALLY PRINTED WITH UV STABLE, SOLVENT RESISTIVE INK. BLACK CHARACTERS ON LIGHT BACKGROUND.
3. MINIMUM TEXT HEIGHT - 1.5mm, FUTURA FONT.
4. LOCATION: FIXED INTO 0.25mm DEEP RECESS.
5. FOR INFORMATION ON BATTERY TYPES SEE DRAWINGS GASMAN-2637-A4 & GASMAN-2638-A4.
6. OTHER INFORMATION MAY APPEAR IN THIS AREA THAT IS NOT RATIFIED BY DEMKO.
7. THE "CE" IS NOT RATIFIED BY DEMKO.



ISSUE	DCL. No.	DATE
1	—	3/00

This drawing is the property of CROWCON Detection Inst. Ltd and must not be loaned, copied in whole or in part or used for manufacturing or marketing purposes without their consent. Copyright Reserved.		MATERIAL		SEE DRG		TOLERANCES DTE= ±1 ANG= ±1°		CROWCON 2 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 3DT	
		DO NOT SCALE DRAWING. IF IN DOUBT ASK DRAWN TO B.S. 308/3939		FINISH		CLEAN		TITLE GASMAN II TX CENELEC CERTIFICATION LABELS	
PROJECT No		DRN		DWG		APPR		DRG. No. GASMAN-2677-A3	
ISS.		SCALE		SHT SIZE		A3		ISS. 1	

DRG. No.	GASMAN-2678-A3	ISS.	1
USED ON	STOCK No.		



CLIP LABELS



CASE LABELS

1. MATERIAL: 0.076 THK. TAMPER EVIDENT POLYESTER, SELF ADHESIVE.
2. TEXT: THERMALLY PRINTED WITH UV STABLE, SOLVENT RESISTIVE INK. BLACK CHARACTERS ON LIGHT BACKGROUND.
3. MINIMUM TEXT HEIGHT - 1.5mm, FUTURA FONT.
4. LOCATION: FIXED INTO 0.25mm DEEP RECESS.
5. FOR INFORMATION ON BATTERY TYPES SEE DRAWINGS GASMAN-2637-A4 & GASMAN-2638-A4.
6. OTHER INFORMATION MAY APPEAR IN THIS AREA THAT IS NOT RATIFIED BY UL.

ISSUE	DATE
1	3/00

This drawing is the property of CROWCON Detection Inst. Ltd and must not be copied, copied in whole or in part or used for manufacturing or marketing purposes without their consent. Copyright Reserved.		MATERIAL SEE DRG		TOLERANCES OD= ±1 ANG= ±1°		2 Blacklands Way Abingdon Business Park, Abingdon, Oxfordshire O14 4BY	
		FINISH CLEAN		DIMS mm SHT SIZE A3		TITLE GASMAN II TX - UL CERTIFICATION LABELS	
DO NOT SCALE DRAWING. IF IN DOUBT ASK. DRAWN TO B.S. 308/3939 PROJECT No.		DFN DWG APPR		ORIGINAL SCALE 2:1		DWS. No. GASMAN-2678-A3 BS. 1	