

CLIP LABELS

**GASMAN II**
FLAMMABLE
GAS DETECTOR
PART OF
DEMKO 97D.121939



WARNING:-SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION

**GASMAN II**
OXYGEN
GAS DETECTOR
PART OF
DEMKO 97D.121939



WARNING:-SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION

**GASMAN II**
TOXIC
GAS DETECTOR
PART OF
DEMKO 97D.121939



WARNING:-SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION

31

35

CASE LABELS

**GASMAN II** Flam Gas Detector
DEMKO 97D.121939

**EEx ia IIC T3**
(Tamb=40°C)



WARNING:- BATTERIES MUST
ONLY BE CHARGED IN AREAS
KNOWN TO BE NON HAZARDOUS
GROUP IIC GASES REQUIRE
USE OF BOOT M04437
FCC ID: OXOGASMANII

**GASMAN II** Oxygen Gas Detector
DEMKO 97D.121939

**EEx ia IIC T4**
(Tamb=50°C)



WARNING:- BATTERIES MUST
ONLY BE CHARGED IN AREAS
KNOWN TO BE NON HAZARDOUS
GROUP IIC GASES REQUIRE
USE OF BOOT M04437
FCC ID: OXOGASMANII

**GASMAN II** Toxic Gas Detector
DEMKO 97D.121939

**EEx ia IIC T4**
(Tamb=50°C)



WARNING:- BATTERIES MUST
ONLY BE CHARGED IN AREAS
KNOWN TO BE NON HAZARDOUS
GROUP IIC GASES REQUIRE
USE OF BOOT M04437
FCC ID: OXOGASMANII

1. MATERIAL: 0.076 THK. TAMPER EVIDENT POLYESTER, SELF ADHESIVE.


2. TEXT: THERMALLY PRINTED WITH UV STABLE, SOLVENT RESISTIVE INK. BLACK CHARACTERS ON LIGHT BACKGROUND.

3. MINIMUM TEXT HEIGHT – 1.5mm, FUTURA FONT.
4. LOCATION: FIXED INTO 0.25mm DEEP RECESS.

5. FOR INFORMATION ON BATTERY TYPES SEE DRAWINGS GASMAN-2637-A4 & GASMAN-2638-A4.

6. OTHER INFORMATION MAY APPEAR IN THIS AREA THAT IS NOT RATIFIED BY DEMKO.

7. THE “CE” IS NOT RATIFIED BY DEMKO.

3	1408	6/00		DO NOT SCALE DRAWING. IF IN DOUBT ASK. DRAWN TO B.S. 308/3939	PROJECT No	FINISH CLEAN	DRN DWV	CHKD	APPR	TOLERANCES GEN= ±1 ANG= ±1°	 2 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY
2	1389	3/00									
1	—	3/00									
ISSUE	D.C.I. No.	DATE									

DIMS
mm

SHT SIZE
A3

ORIGINAL
SCALE
2:1

 TITLE GASMAN II TX CENELEC CERTIFICATION LABELS | DRG. No. GASMAN-2677-A3 | ISS. 3 |