

D

C

B

A

D

C

B

A

REVISION RECORD				
REV	ECO	DATE	DESCRIPTION	DRFT
--	----	-----	SEE SHEET 1	



警告 WARNING 警告

打开时有不可见激光辐射。

避免直接暴露于激光束下。暴露于激光下可能导致永久眼损伤。绕过互锁装置（维修互锁解除）将使系统带激光运行！

开启此装置可由柯达医疗影像授权的维修人员进行。请不要擅自进行用户手册中未提及的操作和调整。

開放すると不可視レーザー光が放出されます。

本装置を直接照らさないでください。レーザー光の目に直接的な暴露をすることが可能性があります。ブypassインターロック装置（メンテナンス用インターロック）を解除すると、システムはレーザーが起動した状態で稼働します。

カバーの取り外しはユダックヘルスイメージングの認定サービス技術者のみが行わなければなりません。ユーザーマニュアルに記載されていない操作や調整は絶対に行わないでください。

INVISIBLE LASER RADIATION PRESENT WHEN OPEN.

Avoid direct exposure to beam. Exposure to laser light may result in permanent eye damage. Bypassing interlocks other than the service interlock will allow system to run with the laser energized! Removal of the cover should only be done by service personnel authorized by Kodak Health Imaging. Please Do Not perform operations or adjustments which are not specified in the user manual.

警告 WARNING 警告

危険电压

能造成严重伤害或死亡。维修机前请先断开电源。

危険な電圧

傷害事故、または死亡事故が起きる可能性があります。機器をサービスする前に必ず電源を外してください。

HAZARDOUS VOLTAGE

Can cause severe injury or death. Disconnect power before servicing machine.

小心 CAUTION 注意

静电敏感设备

维修设备之前，请佩戴静电腕带以正确接地。

静电敏感组件

この機器をサービスする前にパーソナルアースストラップで接地してください。

STATIC SENSITIVE EQUIPMENT

Connect a personal grounding strap to appropriate ground before servicing this machine.

一类激光产品

クラス 1 レーザー

Class 1 Laser Product

Produit Laser de catégorie 1

Klasse 1 Laser Produkt

Prodotto Laser di Classe 1

Klass 1 Laser Produkt

小心 CAUTION 注意

此设备产生、使用并能发射射频能量。如未按指示安装和使用，可能对无线电造成有害干扰。

この機器は無線周波エネルギーを生成、使用、および放射する可能性があります。使用法に明記されている方法で設置および使用されていない場合、無線伝導障害を起こす可能性があります。

THIS EQUIPMENT GENERATES, USES, AND CAN RADIATE RADIO FREQUENCY ENERGY. IF NOT INSTALLED AND USED IN ACCORDANCE WITH THE INSTRUCTIONS, IT MAY CAUSE HARMFUL INTERFERENCE TO RADIO COMMUNICATIONS.

FCC: THIS DEVICE COMPLIES WITH PART 15 OF THE FCC RULES, OPERATION IS SUBJECT TO THE FOLLOWING TWO CONDITIONS:
(1) THIS DEVICE MAY NOT CAUSE HARMFUL INTERFERENCE AND
(2) THIS DEVICE MUST ACCEPT ANY INTERFERENCE RECEIVED, INCLUDING INTERFERENCE THAT MAY CAUSE UNDESIRABLE OPERATION.

FCC ID: PA481907E2527

* THIS CLASS A DIGITAL APPARATUS COMPLIES WITH CANADIAN ICES-003

* CET APPAREIL NUMERIQUE DE LA CLASSE A EST CONFORME A LA NORME NMB-003 DU CANADA

IC: 1016B-0180

THIS PRODUCT COMPLIES WITH CDRH REGULATION 21 CFR SUBCHAPTER J.

販売名：コダック画像診断用イメージャ

輸入販売業者名：コダック株式会社

住所：東京都中央区新川2-27-1

許可番号：13BY6496

Japanese Import License

ART WORK FOR 7F1192.E

USE DIE LINE FOR 7F1192.E

LABEL NOTES:

COLORS USED:

	White Substrate
	PMS Black
	PMS #485-CVC Red
	PMS #123-CVC Yellow
	DIE LINE

FSCM NO.	SIZE	DRAWING NO.	REV
	D	7F1192	E
DO NOT SCALE DRAWING		SHT 2 OF 2	

Label Placement

