



November 27, 2017

Federal Communications Commission
445 12th Street SW
Washington, DC 20554
Attention: Application Examiner

Reference: Certification: Part 15.249 Intentional Radiator
Applicant: IntriCon Corporation
Equipment: Low Power Device (2400-2483.5 MHz)
Model #: Lumen 155
FCC ID: **POMIINHBC1B0**


Dear Sir/Madam:

Enclosed please find an FCC certification application for **POMIINHBC1B0**. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid, to receive streamed audio signals from an appropriate accessory using IntriCon proprietary protocol, and used during normal usage to communicate control information to and from any properly configured Bluetooth Low Energy compliant accessory.

Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,



27 Nov 2017

Mitch Johnson, RAC
Director, Regulatory Affairs
651-604-9572
mjohnson@intricon.com