

PHILIPS PROPRIETARY:
THIS DOCUMENT AND/OR ELECTRONIC FILE IS THE PROPERTY OF PHILIPS MEDICAL SYSTEMS AND CONTAINS CONFIDENTIAL AND PROPRIETARY INFORMATION OF PHILIPS. NEITHER IT NOR THE INFORMATION CONTAINED THERIN SHALL BE DISCLOSED TO OTHERS OR DUPLICATED OR USED BY OTHERS EXCEPT AS AUTHORIZED BY PHILIPS MEDICAL SYSTEMS IN WRITING.

NOTE: BLUE LINES REPRESENT DIE LINES & ARE NOT PART OF THE ARTWORK.

PHILIPS Telemetry Monitor 5500
1.4 GHz Radio

Philips Medizin Systeme
Böblingen GmbH
Hewlett-Packard Strasse 2
71034 Böblingen, Germany

Made in USA from foreign
& domestic components. (01) XXXXXXXXXX
YYYY-MM-DD (21) CCXXXXNNNN

REF 867232

SN CCXXXXNNNN

Service # 4535XXXXXX1

MAC AAAAAAAAAA

FCC ID: PQC-5500SH1A4

IPX8

Operation of this equipment requires the prior coordination
with a frequency coordinator designated by the FCC for the
Wireless Medical Telemetry Service.

NOTE 1:

ALL TEXT/GRAPHICS TO BE THERMAL PRINTED (EXCLUDING THE "REFER TO IFU", RFID AND MR UNSAFE SYMBOL WHICH ARE PRE-PRINTED ON THE RAW LABEL STOCK.)

NOTE 2:

BOTH 2D BARCODES SHOWN ON ARTWORK ARE FOR SIZE/ LOCATION ONLY. THEY ARE NOT VALID.

E	See EMDR	Added warning for WMTS operation.	F. ENSLIN	23-Jul-2025
D	See EMDR	Removed GMDN, WEE symbol and Non-ionizing radiation symbol. Corrected Note 1.	F. ENSLIN	18-May-2025
C	See EMDR	Added MR unsafe symbol, removed "CONTAINS" in front of FCC ID	F. ENSLIN	04-Jun-2024
B	SEE EDMR	CHANGED FONT, GS1 TO UDI, MOVED GMDN	K.CHAPMAN	19JUL2021
REV	REF	REVISION DESCRIPTION	APPR	DATE
NAME:	DATE:	ALLEGRO 1.4 GHz (ECG ONLY) PRIMARY PRODUCT ID LABEL ARTWORK	PHILIPS	
DESIGNER: K.CHAPMAN	23MAR2021			
A.L'Bahy	23MAR2021		LABEL STOCK PART NUMBER: 453564635681	
RELEASE TO PROD:				
453564635681-M-ADWG	REFERENCE MECH. DRG	MASTER ARTWORK	SCALE: 1 : 1	453564962651-M-MART