



Federal Communication Commission
Equipment Authorization Division, Application
Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

C2PCPX Letter according to Notification 202109-001

FCC ID: PQC-OBRSBV1, granted on 11/18/2013

TO WHOM IT MAY CONCERN

Pursuant to KDB 388624 D02 (Pre-Approval Guidance List) v18r03, Notification 202109-001, respectively, we, Philips Medizin Systeme Böblingen GmbH hereby declare that our product can meet the following items:

Notification 202109-001 item	Comment
1) The requirements of § 2.1043 are fulfilled, i.e., the device's block functions for the fundamental frequency, primary modulator circuit, maximum power, or field strength ratings shall remain unchanged.	confirmed
2) Transmitter PCB layout and parts changes are only permitted if there is no change in identifying a device's form, functional specification, as initially granted or previously approved under a Class II permissive change.	
3) PCB changes are limited to non-substantive modifications layout changes to the same size physical circuit board previously granted.	
4) C2PCPX is not permitted to add, remove, augment, or change capabilities, such as transmitters, increased bandwidth, additional rule parts, bands, etc.	
5) In the PAG submission for item C2PCPX, the applicant shall provide complete information on testing demonstrating that the proposed changes for fundamental	



Philips Medizin Systeme Böblingen GmbH

Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202
Geschäftsführer: Dr. Dieter Haase (Sprecher), Günter Gegner, Sang Do Kim, Florian Kuhn
Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, www.philips.com/healthcare

emissions are unchanged within the normal, acceptable tolerances and out-of-band; emissions do not exceed the appropriate limits.	
6) The modified device shall not be marketed under the existing grant of certification before confirmation that the C2PCPX PAG is approved and granted.	
7) Software Defined Radio (SDR) grants that use the C2PCPX procedure are not permitted to make subsequent Class III permissive changes.	
8) The C2PCPX PAG procedure has no impact on the provisions of V) of this publication for non-SDR software-only changes; thus, adding an equipment class when related to rule changes is still permitted.	
9) Class I permissive changes are not permitted under this C2PCPX procedure.	

If you have any questions feel free to contact us at the address shown below.

Sincerely,

Anna Brouhard

(date, signature)

Anna Brouhard, Principle Regulatory Affairs Specialist Radio Frequency, HPM
Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Strasse 2
71034 Böblingen, Germany
004970314630