

January 22, 2002

To: Certification Manager

Re: Class II Permissive change for PQCM3813A2

Enclosed, please find an application for a class II permissive change of the Philips Medical Scale Transmitter Model M3813A (FCC ID: PQCM3813A2). The purpose of this application is to add two additional configurations to the grant of authorization. The two configurations include the the pod mounted to a floor standing weight sensor and the pod mounted to a floor standing weight sensor via a metal pedestal. Please note that the electronics have not changed in any way. Only the mounting location and length of the external cable are different. Upon testing, it was discovered that there was a degradation of EMI performance at a few spurious emission frequencies.

Michael Buchholz
Senior EMC Manager
Curtis-Straus LLC