

SenseWear®

BioTransceiver for Blood Glucose Monitoring

Supplemental Instructions

Please consult the AccuChek® Advantage® First Time Guide and AccuChek® Advantage® Owner's Booklet for instructions on setting up the meter, taking a reading, and performing routine control solution verification.

Taking a reading

1. Make sure the SenseWear® BioTransceiver is **unplugged** from the glucose meter.
2. Insert a test strip and perform glucose meter reading as described in the AccuChek® literature.
3. When your reading is displayed, remove the strip.

Sending your readings

1. Locate the glucose meter and BioTransceiver within 15 feet of the SenseWear® Wireless Gateway.
2. Plug the BioTransceiver into the glucose meter as shown below.
3. The Wireless Gateway will “beep” and the “RECEIVE” light will flash yellow when data has been sent.



Certifications

Disclaimer

This product is not intended to be a medical device or a substitute for proper medical diagnosis, care, or treatment related to your condition. Seek professional consultation if you have any questions related to the readings provided by this product.

FCC statement

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit separate from the receiver.
- Consult the dealer or an experienced radio/TV technician for help.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

CAUTION: Changes or modifications to this equipment not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

FCC 47CFR 15C TCB - 47 CFR Part 15 Subpart C Intentional Radiator Certification Test

FCC 47CFR 15B cIB - 47 CFR Part 15 Subpart B Unintentional Radiator Class B Verification

IEC 60601-1-1 (2000-12) - Medical Electrical Equipment - Part 1-1: General Requirements for Safety - Collateral Standard: Safety Requirements for Medical Electrical Systems Second Edition