



Abbott Neuromodulation
6901 Preston Rd.
Plano, TX 75024
T: +1 818 493 2085

October 1, 2025

Element Materials Technology
6775 NE Evergreen Pkwy
#400
Hillsboro, OR 97124

FCC ID: PX2-GMIPG1

To whom it may concern,

We, Abbott Laboratories (Neuromodulation Division 6901 Preston Road, Plano, TX 75024), hereby submit a Class II Permissive Change application for equipment authorization pertaining to the following model, which is part of a neurostimulation therapy system.

Eterna™ SCS System, Model 32400
Liberta RC™ DBS System, Model 62400

The IPG model operates within the torso and utilizes Bluetooth Low Energy (BLE) technology for wireless communication. The device complies with FCC Rule Part 15.247, as outlined in KDB 558074 Section 8, and the maximum operational output power is 0 dBm or less, considering its low duty cycle transmission characteristics.

This submission reflects design modifications that do not alter the RF characteristics in a manner that would invalidate the original certification. As such, the changes qualify under the scope of a Class II Permissive Change, in accordance with 47 CFR § 2.1043.

Further details regarding the design modifications, along with the radio section of the impacted design documents and supporting test reports, are provided in the accompanying submission package.

Manasi Khare

Digitally signed by
KHAREMX
Date: 2025.10.15
09:33:43 -05'00'

Director, Regulatory Affairs