

Label

Product Name : Medical Tablet PC
Model No : M101MDT
Rating : +19V $\pm$ 3.15A
For use only Power Supply :
GlobTek (GTM96600-6019-T3)
Input : 100-240Vac, 50-60Hz, 1.5A,
Output : 19VDC 3.15A Max.
Tablet PC Input Rating : 19VDC 3.15A

This device complies with part 15 of the FCC Rules.
Operation is subject to the following two conditions:
(1) This device may not cause harmful interference, and
(2) This device must accept any interference received,
including interference that may cause undesired
operation. This Class B digital apparatus meets all
requirements of the Canadian Interference-Causing
Equipment Regulations. Cet appareil numérique de la
classe B respecte toutes les exigences du Règlement sur
le matériel brouilleur du Canada.



ISO 7000-1641
Operating instructions



E489236

Complies with 21CFR 1040.10 and
1040.11 except for deviations pursuant
to Laser Notice No.50,dated June 24 2007

MEDICAL - GENERAL MEDICAL
EQUIPMENT AS TO ELECTRICAL
SHOCK, FIRE AND MECHANICAL
HAZARDS ONLY IN ACCORDANCE
WITH ANSI/AAMI ES 60601-1 (2005)
+ AMD (2012) and CAN/CSA-C22.2
No.6060-1 (2014)



Contains
FCC ID : PX9-AC9260NGW
IC : 9911A-AC9260NGW

Winmate Inc.
9F, No.111-6, Shing-De Rd., San-Chung Dist.,
New Taipei City 24158, Taiwan

Manufactured Date :

Part number of this label 91D00000344H

MADE IN TAIWAN

Label location

