



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 115-17 Date when calibrated 10/17/2017 Page 1 of 2

Item calibrated Signal Generator Extension Module SGX 050
E8257DV15 # US54250106

Description of measurement standard / measuring instrument / identification

Customer Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

Method of calibration GOST 20271.1, MK KL 05.3-2014

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising
signature



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration is performed by using

- 1 Wattmeter M 568
- 2 Wattmeter M 546
- 3 Voltmeter V7-34
- 4 Frequency meter RCH3-72
- 5 Signal generator MG3694C
- 6 Attenuator D3-37
- 7 Attenuator D3-38

Calibration conditions

Temperature: 21.9 °C.
Humidity: 42.0 %.
Pressure: 100.1 kPa.

Calibration results are given in the Measuring report # 115-17

#	Parameter	Specifications required	Specifications tested and measured
1	RF Frequency Band	50 – 75 GHz	Corresponds
2	Multiplication Factor (Low / High)	4 / 2	Corresponds
3	Low Frequency RF Input	12.5 – 18.75 GHz	Corresponds
4	Low Freq. RF Input Power (Typical / Damage)	10 dBm ± 3dB / 16 dBm	Corresponds
5	High Frequency RF Input	25 – 37.5 GHz	Corresponds
6	High Freq. RF Input Power (Typical / Damage)	0 dBm ± 3dB / 6 dBm	Corresponds
7	Output Power (Typical / Minimum)	20 dBm / 17 dBm	Corresponds (Table 1)

**Signature of the person who has
performed calibration**



M. Kasperovich/ Engineer

Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk
220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager



M. Svirid

October 17, 2017

MEASURING REPORT # 115-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Signal Generator Extension Module SGX 050 E8257DV15 # US54250106
Method of calibration:	GOST 20271.1, MK KL 05.3-2014
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 21.9 °C	Humidity: 42.0 %	Pressure: 100.1 kPa
----------------------	------------------	---------------------

MEASURING EQUIPMENT

#	Measuring equipment	Serial number
1	Wattmeter M 568	164
2	Wattmeter M 546	163
3	Voltmeter V7-34	0067787
4	Frequency meter RCH3-72	931200
5	Signal generator MG3694C	133805
6	Attenuator D3-37	2
7	Attenuator D3-38	8

MEASURING RESULTS

Table 1

RF output frequency, GHz	50	55	65	75
RF input frequency, GHz	12.50	13.75	16.25	18.75
RF input power, dBm	10.0	10.0	10.0	10.0
RF output power, dBm	20.33	19.41	22.10	20.28
Expanded uncertainty, dB	0.68	0.68	0.68	0.68

Engineer  M. Kasperovich

Quality Manager  A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1,
Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 116-17 Date when calibrated 10/17/2017 Page 1 of 2

Item calibrated Signal Generator Extension Module SGX 051
E8257DV10 # US53250009

Description of measurement standard / measuring instrument / identification

Customer Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

Method of calibration GOST 20271.1, MK KL 05.3-2014

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising
signature



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration is performed by using

- 1 Wattmeter M 534
- 2 Wattmeter M 546
- 3 Voltmeter V7-34
- 4 Frequency meter RCH3-72
- 5 Signal generator MG3694C
- 6 Attenuator D3-38
- 7 Attenuator AP-20

Calibration conditions

Temperature: 21.9 °C.
Humidity: 42.0 %.
Pressure: 100.1 kPa.

Calibration results are given in the Measuring report # 116-17

#	Parameter	Specifications required	Specifications tested and measured
1	RF Frequency Band	75 – 110 GHz	Corresponds
2	Multiplication Factor (Low / High)	6 / 3	Corresponds
3	Low Frequency RF Input	12.5 – 18.33 GHz	Corresponds
4	Low Freq. RF Input Power (Typical / Damage)	10 dBm ± 3dB / 16 dBm	Corresponds
5	High Frequency RF Input	25 – 36.67 GHz	Corresponds
6	High Freq. RF Input Power (Typical / Damage)	0 dBm ± 3dB / 6 dBm	Corresponds
7	Output Power (Typical / Minimum)	14 dBm / 10 dBm	Corresponds (Table 1)

**Signature of the person who has
performed calibration**


M. Kasperovich/ Engineer
Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk
220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager



M. Svirid

October 17, 2017

MEASURING REPORT # 116-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Signal Generator Extension Module SGX 051 E8257DV10 # US53250009
Method of calibration:	GOST 20271.1, MK KL 05.3-2014
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 21.9 °C	Humidity: 42.0 %	Pressure: 100.1 kPa
----------------------	------------------	---------------------

#	Measuring equipment	Serial number
1	Wattmeter M 534	161
2	Wattmeter M 546	163
3	Voltmeter V7-34	0067787
4	Frequency meter RCH3-72	931200
5	Signal generator MG3694C	133805
6	Attenuator D3-38	8
7	Attenuator AP-20	4

MEASURING RESULTS

Table 1

RF output frequency, GHz	75	92.5	110
RF input frequency, GHz	12.50	15.42	18.33
RF input power, dBm	10.0	10.0	10.0
RF output power, dBm	13.34	15.03	11.89
Expanded uncertainty, dB	0.68	0.82	0.82

Engineer  M. Kasperovich

Quality Manager  A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1,
Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.



Virginia Diodes, Inc

979 2nd St. SE
Suite 309
Charlottesville, VA 22902
Phone: 434-297-3257
Fax: 434-297-3258

Certificate of Conformance

To: Keysight Technologies, Inc.
1400 FOUNTAINGROVE PARKWAY
DOCK 2LS
Santa Rosa, CA 95403-1799
United States

From: Virginia Diodes, Inc
979 2nd St. SE
Suite 309
Charlottesville, VA 22902

Packing List No: 173364

Shipping Date: 12/21/17

Today's Date: 12/22/17


PO Number: 9000796586

Attn: Mike Skaggs
Phone: 1-707-577-3492
S.O.#: 3757528.TWS-571

Quantity				<u>Order-Job</u>
<u>Shipped</u>	<u>Unit</u>	<u>Description</u>		<u>Number</u>
1	EA	VDIWR6.5SGX E8251-80078 Rev - 001; SN: SGX 223.		17497-01
1	EA	VA6R E8257-80039 Rev - 001; SN: 19022.		17497-02
1	EA	EXTENDED WARRANTY-WR6.5SGX E8257-80097 Rev - 001; SN: SGX 223.		17497-03

The VDI product(s) in this shipment meet(s) the guidelines for performance specifications established in accordance with the corresponding Purchase Order. Data presented in the User Guide, where applicable, has been obtained in accordance with VDI's Quality Management System. All instruments, used to obtain data, which require calibration have been calibrated with equipment traceable to the National Institute of Standards and Technology (NIST) and through NIST to the International System of Units (SI).

Authorized Signature
Virginia Diodes, Inc



Virginia Diodes, Inc

Certificate of Conformance

To: Agilent
1400 Fountaingrove Parkway
Dock 4LS
Santa Rosa, CA 95403

From: Virginia Diodes, Inc

Packing List No: 131150
Shipping Date: 07/02/13

Today's Date: 07/02/13
PO Number: 9000612686

Attn: Mike Skaggs

Phone: 1-707-577-4607

Quantity				<u>Order-Job</u>
<u>Shipped</u>	<u>Unit</u>	<u>Description</u>		<u>Number</u>
1	EA	VDIWR5.1SAX WR5.1PXAX - Frequency Extender PXA 140-220GHz; SN: VDI PXAX 007.		13056B-01
1	EA	VDIWR5.1SGX WR5.1PSGX - Frequency Extender PSG 140-220GHz; SN: VDI PSGX 007.		13056B-02

Handwritten initials 'MS' in black ink.

Inspector's Initials
Virginia Diodes, Inc

Handwritten signature in black ink.

Authorized Signature
Virginia Diodes, Inc

☒ 校正實驗室
33383 桃園市龜山區
文明路29巷8號
TEL:+886-3-3280026

財團法人台灣電子檢驗中心

校正報告

CALIBRATION REPORT



☐ 新竹校正實驗室
30075 新竹市科學園區
園區二路47號205室
TEL:+886-3-5798806

工服 NO. 18-06-BCC-337-01L

ELECTRONICS TESTING CENTER, TAIWAN

Page 1 of 5

申請者(Applicant): 香港商立德國際商品試驗有限公司新竹分公司

地址(Address): 新竹市科學園區力行一路1號E-2

供校儀器 ITEM CALIBRATED

儀器名稱: PSG Analog Signal Generator

Nomenclature

型別: E8257D

Model No.

校正依據: 詳如說明4所示

Cal. Procedure Used

校正資料: ☒ 僅量測

☐ 調整

Cal. Info. Cal.Only

Adjusted

實際環境: 溫度: 23 °C

相對濕度: 46 %

Real Condition Temperature

Relative Humidity

製造商:

KEYSIGHT

Mfg.

識別號碼:

MY53401987

ID. No.

收件日期:

Jun.22,2018

Receipt Date

校正日期:

Jun.26,2018

Cal. Date

建議再校日期: Jun.25,2019

Recommended Recal. Date

使用標準器及附配件 STANDARD AND ACCESSORIES

儀器名稱
Nomenclature

廠牌/型號
Mfg. / Model No.

識別號碼
ID. No.

校正日期
Date Cal.

有效日期
Due Date

EPM Series Power Meter

HP E4419B

13050609-001

2018/03/21

2019/09/20

Power Sensor

AGILENT 8487D

13053508-004

2017/09/15

2021/03/14

Measuring Receiver

R&S FSMR

13054413-001

2017/11/16

2021/05/15

Universal Counter

HP 53132A

13060804-001

2018/01/12

2018/07/11

Frequency Counter

AGILENT 53152A

13060808-001

2015/11/30

2019/05/29

追溯源 CALIBRATION SOURCE

儀器名稱
Nomenclature

校正單位(認可編號)
Cal. Source(ACRED Code)

報告號碼
Cal. Report No.

校正日期
Date Cal.

有效日期
Due Date

Power Sensor

AGILENT(ANAB AC-1498)

1-9230656670-1

2017/09/15

2021/03/14

Measuring Receiver

R&S(DAKK S D-K-15195-01-00)

438369/D-K-15195-01-01

2017/11/16

2021/05/15

Rubidium Atomic Frequency Standard

CHT(TAF N0815)

FTC-2018-04-16

2018/04/30

2019/10/29

Power Meter

NML(TAF N0688)

U180022A

2018/03/30

2019/09/29

ETC hereby certifies that the equipment noted herein has been compared with the above listed standards. The Standards used to perform this calibration are traceable to NML/ROC,NIST/USA or other countries. The calibration services from ETC are capable of performing services in compliance with the requirements of ISO/IEC 17025.

財團法人台灣電子檢驗中心特此證明報告內記載之受校儀器已與上列標準做過比較校正,用以校正之標準器可追溯至中華民國國家度量衡標準實驗室,美國標準及技術研究院,或其它國家之度量衡國家標準。本中心的校正服務均符合ISO/IEC 17025之規定。

校正地點: 財團法人台灣電子檢驗中心校正實驗室

財團法人台灣電子檢驗中心
ELECTRONICS TESTING CENTER,
TAIWAN



實驗室主管
Laboratory Head



報告簽署人
Signature



校正報告

財團法人台灣電子檢驗中心

工 服NO.18-06-BCC-337-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 2 of 5

1. Frequency Accuracy Check:

Reading	Standard	Expanded Uncertainty
250.0000 kHz	249.99999054 kHz	1.2×10^{-7}
1.000000 MHz	0.9999999620 MHz	1.2×10^{-7}
10.000000 MHz	9.9999996197 MHz	1.2×10^{-7}
100.000000 MHz	99.999996197 MHz	1.2×10^{-7}
200.000000 MHz	199.99999240 MHz	1.2×10^{-7}
500.000000 MHz	499.99998101 MHz	1.2×10^{-7}
1.000000000 GHz	0.999999962 GHz	1.2×10^{-7}
2.000000000 GHz	1.999999924 GHz	1.2×10^{-7}
5.000000000 GHz	4.999999811 GHz	1.2×10^{-7}
10.000000000 GHz	9.999999622 GHz	1.2×10^{-7}
20.000000000 GHz	19.999999242 GHz	1.2×10^{-7}
30.000000000 GHz	29.999998861 GHz	1.2×10^{-7}
40.000000000 GHz	39.999998732 GHz	1.2×10^{-7}
46.000000000 GHz	45.999998545 GHz	1.2×10^{-7}

2. Output Level And Frequency Response Accuracy Check:

Test Freq.	Setting(dBm)	Measured(dBm)	Expanded Uncertainty(dB)
250.0000 kHz	0.00	0.06	0.22
1.000000 MHz	0.00	-0.03	0.22
10.000000 MHz	0.00	-0.05	0.22
100.000000 MHz	0.00	-0.04	0.22
200.000000 MHz	0.00	-0.04	0.22
500.000000 MHz	0.00	-0.02	0.22
1.000000000 GHz	0.00	-0.06	0.22
2.000000000 GHz	0.00	0.00	0.27
3.000000000 GHz	0.00	-0.06	0.27
4.000000000 GHz	0.00	-0.04	0.27
5.000000000 GHz	0.00	-0.05	0.27
6.000000000 GHz	0.00	-0.07	0.27
7.000000000 GHz	0.00	-0.07	0.27
8.000000000 GHz	0.00	-0.11	0.27
9.000000000 GHz	0.00	-0.12	0.27
10.000000000 GHz	0.00	-0.10	0.27
11.000000000 GHz	0.00	-0.13	0.27
12.000000000 GHz	0.00	-0.33	0.27
13.000000000 GHz	0.00	-0.48	0.32
14.000000000 GHz	0.00	-0.17	0.32
15.000000000 GHz	0.00	-0.14	0.32
16.000000000 GHz	0.00	-0.12	0.32
17.000000000 GHz	0.00	-0.05	0.32
18.000000000 GHz	0.00	-0.12	0.38

校正報告

財團法人台灣電子檢驗中心

工 服NO. 18-06-BCC-337-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 3 of 5

2. Output Level And Frequency Response Accuracy Check: (Continued)

Test Freq.	Setting(dBm)	Measured(dBm)	Expanded Uncertainty(dB)
19.000000000 GHz	0.00	-0.09	0.38
20.000000000 GHz	0.00	-0.14	0.38
21.000000000 GHz	0.00	-0.08	0.38
22.000000000 GHz	0.00	-0.08	0.38
23.000000000 GHz	0.00	-0.09	0.38
24.000000000 GHz	0.00	-0.10	0.38
25.000000000 GHz	0.00	-0.20	0.38
26.000000000 GHz	0.00	-0.12	0.38
27.000000000 GHz	0.00	-0.20	0.61
28.000000000 GHz	0.00	-0.16	0.61
29.000000000 GHz	0.00	-0.27	0.61
30.000000000 GHz	0.00	-0.10	0.61
31.000000000 GHz	0.00	-0.09	0.61
32.000000000 GHz	0.00	-0.06	0.61
33.000000000 GHz	0.00	-0.22	0.61
34.000000000 GHz	0.00	-0.02	0.61
35.000000000 GHz	0.00	-0.13	0.61
36.000000000 GHz	0.00	-0.21	0.61
37.000000000 GHz	0.00	-0.06	0.61
38.000000000 GHz	0.00	-0.11	0.61
39.000000000 GHz	0.00	-0.15	0.61
40.000000000 GHz	0.00	0.02	0.91
41.000000000 GHz	0.00	-0.05	0.91
42.000000000 GHz	0.00	-0.04	0.91
43.000000000 GHz	0.00	-0.42	0.91
44.000000000 GHz	0.00	-0.46	0.91
45.000000000 GHz	0.00	-0.02	0.91
46.000000000 GHz	0.00	-0.56	0.91
47.000000000 GHz	0.00	-0.63	0.91
48.000000000 GHz	0.00	-0.45	0.91
49.000000000 GHz	0.00	-0.31	0.91
50.000000000 GHz	0.00	-0.54	0.91
100.000000 MHz	10.00	9.98	0.68
100.000000 MHz	-10.00	-10.07	0.68
100.000000 MHz	-20.00	-20.07	0.68
100.000000 MHz	-30.00	-30.08	0.68
100.000000 MHz	-40.00	-40.08	0.68
100.000000 MHz	-50.00	-50.11	0.68
100.000000 MHz	-60.00	-60.10	0.68
100.000000 MHz	-70.00	-70.12	0.68
100.000000 MHz	-80.00	-80.12	0.68
100.000000 MHz	-90.00	-90.15	0.68

校正報告

財團法人台灣電子檢驗中心

工 服NO. 18-06-BCC-337-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 4 of 5

2. Output Level And Frequency Response Accuracy Check: (Continued)

Test Freq.	Setting(dBm)	Measured(dBm)	Expanded Uncertainty(dB)
100.000000 MHz	-100.00	-100.14	0.68
1.000000000 GHz	10.00	10.00	0.68
1.000000000 GHz	-10.00	-10.08	0.68
1.000000000 GHz	-20.00	-20.10	0.68
1.000000000 GHz	-30.00	-30.12	0.68
1.000000000 GHz	-40.00	-40.14	0.68
1.000000000 GHz	-50.00	-50.12	0.68
1.000000000 GHz	-60.00	-60.12	0.68
1.000000000 GHz	-70.00	-70.13	0.68
1.000000000 GHz	-80.00	-80.12	0.68
1.000000000 GHz	-90.00	-90.13	0.68
1.000000000 GHz	-100.00	-100.11	0.68
10.000000000 GHz	10.00	9.93	1.1
10.000000000 GHz	-10.00	-10.07	1.1
10.000000000 GHz	-20.00	-20.06	1.1
10.000000000 GHz	-30.00	-30.10	1.1
10.000000000 GHz	-40.00	-40.11	1.1
10.000000000 GHz	-50.00	-50.09	1.1
10.000000000 GHz	-60.00	-60.12	1.1
10.000000000 GHz	-70.00	-70.12	1.1
10.000000000 GHz	-80.00	-80.13	1.1
10.000000000 GHz	-90.00	-90.14	1.1
10.000000000 GHz	-100.00	-100.14	1.1
20.000000000 GHz	10.00	9.86	1.4
20.000000000 GHz	-10.00	-10.13	1.4
20.000000000 GHz	-20.00	-20.14	1.4
20.000000000 GHz	-30.00	-30.14	1.4
20.000000000 GHz	-40.00	-40.15	1.4
20.000000000 GHz	-50.00	-50.19	1.4
20.000000000 GHz	-60.00	-60.16	1.4
20.000000000 GHz	-70.00	-70.15	1.4
20.000000000 GHz	-80.00	-80.14	1.4
20.000000000 GHz	-90.00	-90.17	1.4
20.000000000 GHz	-100.00	-100.17	1.4

校正報告

財團法人台灣電子檢驗中心

工 服NO. 18-06-BCC-337-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 5 of 5

3. Harmonic Distortion Measurement Check:

Test Freq.	Harmonic(dBc)	Expanded Uncertainty(dB)
250.0000 kHz	-32.1	0.74
1.00000 MHz	-34.5	0.74
10.0000 MHz	-40.0	0.74
100.0000 MHz	-61.5	0.74
500.0000 MHz	-37.5	0.74
1.000000000 GHz	-40.1	0.74
5.000000000 GHz	-62.4	1.4
10.000000000 GHz	-76.7	1.9
20.000000000 GHz	-69.7	2.0

4. 10 MHz Ref. Out. Frequency Accuracy Check:

Reading(MHz)	Standard(MHz)	Expanded Uncertainty
10	9.9999996223	1.2×10^{-7}

說明:

- 1.本校正報告內的擴充不確定度評估與表示是依據「ISO Guide 98-3量測不確定度表示方式指引」，擴充不確定度 $U = ku_c$ ，其中 u_c 為組合標準不確定度， $k = 2.0$ ，為信賴水準約95%之涵蓋因子。
- 2.環境管制條件: 溫度: $(23 \pm 2) ^\circ\text{C}$; 相對濕度: $(50 \pm 10) \%$ 。
- 3.報告內之建議再校日期為應申請者要求列入。
- 4.「信號產生器之校正程序書」，B00-CD-369，3rd Edition。