

18 December 2006

AmericanTCB
6731 Whittier Avenue
McLean VA 22101
Attn: Director of Certification

RE: Declaration of Compliance to RF Exposure Limits for Humans

Cardionet Sensor

RF Exposure Statement

1. Standard Applicable


According to 1.1307 (b)(1), systems operating under the provisions of this section shall be operated in a manner that ensure that the public is not exposed to radio frequency energy level in excess of the Commission's guideline.

This is a portable device operating in 902-928 MHz range.

2. Measurement Result:

This is a portable device and the max peak output power is 13.64 dBm (.023 W). Lower than low threshold 60/f GHz mW (65 mW), d<2.5 cm, general population category.

The SAR measurement is not necessary.



Joel T. Schneider
Senior Engineer
Minnesota Wild River Lab
TÜV America
Product Service
Tel: 651 604 4582
Fax: 651 638 0298
jschneider@tuvam.com