



Attention

Risk of Electromagnetic Interference

The use of the external stimulator adjacent to or stacked with other devices should be avoided, as this may lead to the external stimulator operating incorrectly.

- Where usage in such a manner is unavoidable, you should monitor the external stimulator and the other device(s) being used with it in order to ensure that they are all working correctly.



Attention

Risk of Electromagnetic Interference through the Use of Unauthorized Accessories

The use of accessories, transducers or cables not listed by BIOTRONIK or of accessories other than those specified by BIOTRONIK, can produce elevated electromagnetic emissions or cause degradation in the device's resistance to electromagnetic interference. Such effects can lead to the faulty operation of the device.

- Only use accessories authorized by BIOTRONIK.

Country-Related Information

International Radio Certification

Telemetry Information for Australia



This device is in compliance with the Australian "Radiocommunications Act 1992" and, therefore, it is labeled according to the "Radiocommunications (Compliance labeling – Devices) Notice".

Telemetry Information for the USA

The stimulator will be registered with the Federal Communications Commission under the following number:

- External stimulator:
FCC ID: QRI-SCSTS
- Implantable stimulator:
FCC ID: QRI-SCSIPG

This device complies with part 15 of the FCC Rules.

Operation is subject to the following two conditions:

1. This device may not cause harmful interference, and
2. this device must accept any interference received, including interference that may cause undesired operation.

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

Symbols on the Implantable Stimulator

Symbol	Meaning
	Identification of port A and port B placement