



Indexsar Limited
Oakfield House
Cudworth Lane
Newdigate
Surrey RH5 5BG

Tel: +44 (0) 1306 632 870
Fax: +44 (0) 1306 631 834
e-mail: enquiries@indexsar.com

Calibration / Conformance statement
Balanced Validation dipole

Type:	IXD -090 900MHz
-------	------------------------

Manufacturer:	IndexSAR, UK
---------------	---------------------

Serial Number:	0050
----------------	-------------

Place of Calibration:	IndexSAR, UK
-----------------------	---------------------

IndexSAR Limited hereby declares that the IXD series dipole named above has been checked for conformity to the specifications given in the draft IEEE 1528 and CENELEC En 50361 standards on the date shown below.


Date of Calibration/Check:	1st February 2002
----------------------------	-------------------------------------

The dipole named above should be periodically re-checked using the procedures set out in the dipole calibration document. It is important that the cautions regarding handling of the dipoles (given in the calibration document) are adhered to. Provided the dipole is undamaged and the periodic checks show that it continues to meet the return loss figures quoted in the calibration document, the recommended period before for re-calibration is due is 5 years.

Next Calibration Date:	February 2007
------------------------	----------------------

The calibration measurements were carried out using the methods described in the calibration document.
Where applicable, the standards used in the calibration process are traceable to the UK's National Physical Laboratory.


Calibrated By:


Approved By: