



St. Jude Medical Cardiac Rhythm
Management Division (CRMD)
15900 Valley View Court
Sylmar, CA 91342 USA
Main 818 362 6822
Fax 818 364 5814

June 7, 2022

Federal Communications Commission
445 12th Street SW
Washington, DC 20554

Subject: FCC Class II Permissive Changes Application for FCC ID: RIA-MER3700BLE

Reference: FCC ID: RIA-MER3700BLE
Original Grant Date: 06/06/2022
Applicant: St. Jude Medical
Product Description: BLE Radio

To Whom It May Concern:

Enclosed please find a Class II Permissive Changes certification application.

Original FCC ID: RIA-MER3700BLE; certification date: 06/06/2022

The BLE Radio itself has not changed; the change filed under this application are:

1. The RF Module operation is limited with BLE spec shown in the FCC Form 731.
2. The new Trogon antenna p/n: SR42W009; peak antenna gain of 5.03dBi in 2.4MHz band
3. The RF Module is used (and was tested) being installed in the Host Device p/n: MER3700.

Sincerely,

A handwritten signature in blue ink, appearing to read "Perry Li".

Perry Li
Principal Engineer, R&D