



2024-07-25

TUV SUD America Inc. 10 Centennial Drive  
Peabody, MA 01960, USA

Dear Sir or Madam,

We, St. Jude Medical 15900 Valley View Court Sylmar, CA 91342, hereby certify the differences described below for all model families listed are valid and true.

Yours sincerely,

A handwritten signature in black ink, appearing to read 'Perry Li', with a small dot below it.

Name: Perry Li

Title: Associate Research Fellow



| Product Name     | Bipolar DF4                                                                                                                                                                                       | Bipolar DF-1                                                                                                                                                                                      |
|------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Avant™ HF        | CDHFA700D<br>Tested Model (Baseline)                                                                                                                                                              | CDHFA700B<br>• Tested Model (Baseline)                                                                                                                                                            |
| Neutrino™ NxT HF | CDHFA800D<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    | CDHFA800B<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    |
| Gallant™ HF      | CDHFA500D<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    | CDHFA500B<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    |
| Neutrino™ NxT HF | CDHFA600D<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    | CDHFA600B<br>• No hardware or software difference from baseline.<br>• Different product name and model number.                                                                                    |
| Entrant™ HF      | CDHFA300D<br>• Software change to accommodate different therapy settings.<br>• No change to RF software.<br>• No hardware difference from baseline.<br>• Different product name and model number. | CDHFA300B<br>• Software change to accommodate different therapy settings.<br>• No change to RF software.<br>• No hardware difference from baseline.<br>• Different product name and model number. |