



St. Jude Medical
Implantable
Electronic Systems
15900 Valley View Court
Sylmar, CA 91342 USA
Main 818 362 6822
Fax 818 364 5814

May 25th, 2016

TUV SUD BABT FCB
Octagon House,
Segensworth Road,
Fareham,
Hampshire,
PO15 5RL

Model Names and Differences

We, St. Jude Medical IESD, hereby request Family certification for the following RF electrically identical implantable pulse generators. The differences consist of battery capacities (either 5.3 or 7.5 Ah), type of therapy (Tonic or non-Tonic SCS/DBS or Dorsal Root Ganglion therapy (DRG), controlled by software) and method of connecting patient electrodes to the device (SJM, 1.27 mm or DRG header).

Model 3660 Orion IPG, 5.3Ah, SCS, SJM header
Model 3661 Orion IPG, 5.3Ah, SCS, 1.27mm header
Model 3662 Orion IPG, 7.5Ah, SCS, SJM header
Model 3663 Orion IPG, 7.5Ah, SCS, 1.27mm header
Model 6660 Orion IPG, 5.3Ah, DBS, Tonic, SJM header
Model 6661 Orion IPG, 5.3Ah, DBS, Tonic, 1.27mm header
Model 6662 Orion IPG, 7.5Ah, DBS, Tonic, SJM header
Model 6663 Orion IPG, 7.5Ah, DBS, Tonic, 1.27mm header
Model 3665 Orion IPG, 5.3Ah, SCS, Tonic, SJM header
Model 3667 Orion IPG, 7.5Ah, SCS, Tonic, SJM header
Model 3664 Proclaim™ 5 IPG, 5.3Ah, DRG, DRG header

Sincerely,

Name: Jorge Amely-Velez
Title: Senior Principal Engineer
St. Jude Medical IESD
15900 Valley View Court
Sylmar, CA 91342