



LABELLING REQUIREMENTS

Labelling Requirements

The sample label shown shall be *permanently affixed* at a conspicuous location on the device and be readily visible to the user at the time of purchase.

<EXPD 3643N1>



< Rear side >

LABELLING REQUIREMENTS

<EXPD 4343N1>

DRTECH

<http://www.drtech.com>

DRTECH Corporation
Suite No. 1, 2 Floor / Suite No.2, 3 Floor,
29, Dunchon-daero 541beon-gil,
Jungwon-gu, Seongnam-si, Gyeonggi-do, 13216, Republic of Korea

Product Name : Flat Panel Digital X-ray Detector Model : EXPD 4343N1 Contains FCC ID : RNH-TOP8812BU Canada IC No. : 29808-TOP8812BU Weight : 3.5kg (Incl. battery) Dimension(WxLxH) : 460.0 x 460.0 x 15.5mm Input : (Tether Cable) 24V =, 1A (USB-C Adaptor) 20V =, 5A (Wireless Charger) 20V =, 0.75A (Battery) 3750mAh, 11.1V = or 3750mAh, 7.4V =	MA : SN : UDI :
---	--------------------------------------

EC REP

DRTECH Europe GmbH
Otto-Volger-Straße 3a, 65843 Sulzbach (Taunus), Germany

No field serviceable component inside. – Do not attempt to open the case.
Aucun composant réparable sur site ne se trouve à l'intérieur. - N'essayez pas d'ouvrir le boîtier.







MD



06269-24-16839



< Rear side >