

Current Date(日期): 2017-06-25

Subject: Application for New Certification, IC ID: 10913A-FDIRV3

Dear Application Examiner,

The purpose of this filing is to apply for a new certification of a BT 4.0 single mode device. The BT function of the device is used to transfer information to a Smart Phone with a proprietary app installed. The device utilizes 11 channels in the frequency range 2402 to 2480MHz with a conducted output power of ~0.71mW. The antenna is a PCB Antenna

Your efforts in reviewing this application are greatly appreciated. Please contact me if there are questions or additional information needed concerning this request.

Sincerely,

Signature: 

Eddy Zong
SGS-CSTC Standards Technical Services (Shanghai) Co., L

Authorized agent of iHealth Las, Inc.

Tel: +86 (0)21 6191 5664

Address: No. 588 West Jindu Road, Songjiang District, Shanghai, China. 201612