
Resting Metabolic Rate Test

--

Metabolism

Metabolism can be understood as the conversion by the human body between food and accumulated fat into energy. The energy is used by the body to maintain constant temperature, to move and to make all the organ function. Measure of metabolism is: calories (cal).

Total Metabolic Rate

The total metabolic rate are the total calories that the human body needs in order to actuate the daily functional activities.

Resting Metabolic Rate (RMR)

Resting Metabolic Rate represents the calories that the vital organs need to properly operate at rest (heart, brain, lungs, liver, kidneys etc.). RMR represents between 60 % and 75 % of the human's total metabolism.

Importance to measure RMR

A knowledge of the RMR is very helpful in order to understand the nutritional needs and to properly manage it.

Measure of the rest metabolic rate with indirect calorimetry

Energy expenditure can be measured directly by putting a person in a calorimeter and measuring the amount of heat produced by the body mass.

This is expensive and very impractical in the clinical setting. Energy expenditure can be measured indirectly with a metabolic cart by analysis of respired gases (usually expired) to derive volume of air passing through the lungs, the amount of oxygen extracted from it (i.e., oxygen uptake VO_2) and the amount of carbon dioxide, as a by-product of metabolism, expelled to atmosphere (CO_2 output – VCO_2). With these measurements the resting energy expenditure (RMR) and respiratory quotient (RQ) can be calculated.

The RQ represents the ratio of carbon dioxide exhaled to the amount of oxygen consumed by the individual. RQ is useful in interpreting the results of the RMR. The abbreviated Weir equation is probably the most common calculation of RMR.

Abbreviated Weir equation:

$$\text{RMR} = [3.9 (\text{VO}_2) + 1.1 (\text{VCO}_2)] 1.44$$

How to perform a RMR test

For best results, when having a REE done, there are certain conditions that need to be controlled and others that just require documenting at the time of the test. During the test the individual is interfaced with a metabolic measurement system by means of a facemask.

A mouthpiece with a nose clip is also sometimes used, but it may create overly stressful conditions to a subject (patient).

Important considerations or conditions to improve the RMR measurement:

- No food for at least 12 hours and no smoke for at least 2 hours before the test.
- Maintain quiet surroundings when the test is in progress and normal temperature. The individual should not move arms or legs during the test.
- Medications taken should be noted, such as stimulants or depressants.
- The first 5 minutes of acquisition should be discarded by the computation of RMR
- Steady state should be achieved, which would be identified clinically by the following criteria: 5 minute period when average minute VO_2 and VCO_2 changes by less than 10%, average RQ changes by less than 5%
- Stable interpretable measurements should be obtained in a 15 to 20 minute test.
- Renal failure patients requiring hemodialysis should not be tested during dialysis therapy.

Recommendations

Resting metabolic rate test using the face mask

1. Since the ventilation is very low (normally <10 litres/min), the turbine calibration has to be performed with very slow manoeuvres (each complete manoeuvre in about 10-15 seconds), to obtain the best accuracy.
2. Use the following correction for the dead space (VD):
 - 50 ml for the small mask
 - 60 ml for the medium mask
 - 70 ml for the large mask

Resting metabolic rate test using the canopy option

1. Verify, before and during the test, that the FeCO_2 falls into the range 0.5%-0.8% and adjust the flow rate of the pump. If the FeCO_2 is too low, increase the flow, if it is too high decrease the flow. In fact, if the FeCO_2 is too low the measurement could be not reliable, while an high FeCO_2 could be dangerous for the patient.
2. In order to perform a correct ERGO calibration and to obtain more reliable data from the test, it is recommended to use a calibration cylinder with the following concentrations: 1% CO_2 , 20% O_2 , balance N_2 . If you use this cylinder, please remember to modify the reference values, as explained in the chapter *Calibration*.

Performing a test using the face mask

Calibrations

Before the test, it is necessary to perform an ergo calibration (see Calibration chapter) and it is advisable to perform also a turbine calibration (see Recommendations in this chapter).

How to prepare a patient

The patient interfaces with the equipment by means of a face mask, like in the stress exercise. The mask has to be tight to the face, in order to avoid any air leakage.

Start the test

1. Enter in the ergometry program
2. Select a patient or add a new one (**File/Patients...**)
3. Select **Start test** from **Test** menu



The 'Execute Test' dialog box contains the following fields and options:

- Height (cm): 178.0
- Weight (Kg): 76.0
- Mode: Radio buttons for Gas, ECG, Gas + ECG, Best ECG, **RMR** (selected), and Simulated test.
- Ergometer: [no one] (dropdown)
- Protocol: RMR (dropdown)
- Workspace: RMR (dropdown)
- Buttons: OK, Other data..., Details..., Cancel

The 'Modify test information' dialog box contains the following fields and options:

- ID code: 1000
- Last name: BOND
- First name: JAMES
- Birth Date: 06/03/1957
- Sex: Radio buttons for Male (selected) and Female
- Ethnic Cor. (%): 100
- Height (cm): 178.0
- Weight (Kg): 76.0
- HR max (bpm): 181
- EEV1 (l): 0.00
- UN (g/day): 0.0
- VD (ml): 0
- Temperature (°C): 25
- Humidity (%): 50
- Press. (mmHg): 760
- Temp. flowm. (°C): 34
- Hum. flowm. (%): 100
- Notes: incremental test - cycloergometer
- Distance: 0.00
- Load 1: Load, Unit of meas.: Watt
- Load 2: Real Load, Unit of meas.: Watt
- Load 3: Revolution, Unit of meas.: RPM
- Buttons: OK, Cancel, ?


4. Enter the patient's data and select the **RMR** mode (1st picture).
5. Press **Other Data...** and enter the dead space value (50ml Small mask, 60ml Medium mask and 70ml Large mask). It is possible to enter the Ureic Nitrogen value NU (2nd picture).
6. Confirm and start the test by pressing **OK**.

Selecting **RMR** the system set automatically the following options:

- Data acquisition with a 30 seconds average
- RMR protocol, which is:
 - 5 minutes discarded;
 - 10 minutes with data acquisition, of which the software will make an average at the end of the test;
 - automatic end of the test after the 16th minute.
- Selection of the RMR workspace (windows placement);


The test is fully automatic, the software will stop it and save the data at the end of the 16th minute.

The real time view is as shown in the following picture:



Viewing the test

At the end of the test, it will be opened automatically a window with the test results.



At the end of the test, or if it is selected **View/RMR**, the main results are shown:

- The average time interval (default: 10 minutes)
- Average values of VO_2 , VCO_2 , R, RMR, RF, VE, HR, FAT% and CHO% and predicted values if available.
- Body Mass Index (BMI) and interpretation
- Graph of the energetic expenditure for all the data acquisition interval, highlighting the selected average interval.

In order to verify the goodness of the test, check that the ventilation and respiratory frequency are similar to the predicted ones (12 breaths/min for the respiratory frequency and 6 litres/min for the ventilation), and the heart rate is the rest heart rate of the patient.




Nota: The percentage of used Proteins (PRO%) is calculated assuming 12 grams of Ureic Nitrogen in 24 hours. You can modify this value selecting **View/Information...** -> **Modify...**

How to modify the average interval

If the average interval (automatically identified by the software) is not satisfying, for example because the patient was speaking in the first minutes, it is possible to modify the interval of the average.

Right-click and select **Edit RMR....** It is possible to move the start and the end lines.



To move the start line, left-click on the exact time in which you want to start the calculations, for the end line, right-click.

Print

The print of the current window generates a report similar to the one in the following page.



COSMED s.r.l.
P.O. BOX 3, 00040 Rome, Italy
tel: +39-069315492; fax: +39-069314580
<http://www.cosmed.it>; E-mail: info@cosmed.it

Last name: BOND		First name: JAMES	
ID code: 1000	Test number: 3	Barometric press. (mmHg): 737	
Sex: M	Test date: 13/03/1997	Temperature (degrees C): 27	
Age: 40	Test time: 00:00	STPD: 0.799	
Height (cm): 178.0	N. of steps: 32	BTPS insp: 1.087	
Weight (Kg): 76.0	Duration (hh:mm:ss): 00:16:00	BTPS exp: 1.020	
HR max (bpm): 180	BSA (m ²): 1.9	BMI (Kg/m ²): 23.9	


Notes:

Constant Load Exercise - cycloergometer

Parameter	Rest	Pred.	% Pred.
Time (mm:ss)	09:30	10:00	95.0%
RMR (Kcal/day)	1981.1	1731.9	114.3%
R (---)	0.82	0.85	96.8%
VO2 (ml/min)	286	266	107.7%
VCO2 (ml/min)	235	226	104.2%
Rf (b/min)	13.7	12.0	114.2%
VE (l/min)	9.0	6.0	150.4%
HR (bpm)	72		
FAT% (%)	59.5		
CHO% (%)	40.9		
BMI (Kg/m ²)	23.9	Normal	

BMI Statements

Underweight	<18.5
Normal	18.5-24.9
Overweight	25.0-29.9
Obesity class I	30.0-34.9
Obesity class II	35.0-39.9
Obesity class III	>=40



Performing a test using the canopy option

The principle of a ventilated bubblehood system is that a stream of air is forced to pass across the face of a subject and mixes with the air which is collected by a transparent hood, placed over the subject's head. A measurement system, knowing the flow rate, calculates the oxygen consumption and the CO₂ production and, starting from these values, the energy expenditure.

Calibrations

Before the test, it is necessary to perform an ergo calibration (see *Calibration* chapter) and it is advisable to perform also a turbine calibration (see *Recommendations* in this chapter).


How to prepare the canopy and the patient

Replacement of the power plug

If the power plug does not fit into the mains socket, replace it with the one in the packaging.

In order to replace the plug:

1. Extract the plug from the battery charger
2. Insert the proper plug in the battery charger.



Connecting the Canopy

1. Connect the Canopy unit to the mains by means of the medical grade AC/DC adapter provided.
2. Insert the bubblehood adapter into the bubblehood from the outside and fix it screwing the ring from the inside, being careful to insert it in the proper hole, as shown in the following picture.



3. Connect the bubblehood to the wrinkled tube, interposing a bacterial filter.
4. Connect the wrinkled tube to the unit through the *Flow in* connector.
5. Connect the optoelectronic reader of the Quark to the *Flow out* connector of the Canopy unit through the spirometry adapter.



6. Fix the vail to the bubblehood through the velcro strips.

How to prepare the patient

1. Switch on the Canopy unit. If there are no problems, the red led on the front panel of the unit flashes for few seconds and the alarm beeps. If the led does not flash and/or the alarm does not beep, the test cannot be performed, because the backup battery is exhausted or there is no backup battery.
2. When the green led turns on, the test can start. If the green led does not turn on, the red led flashes and the alarm beeps, the test cannot be performed because the pump does not work or the mains does not power the system.
3. After these checks, put the patient in a supine position.
4. Place the bubblehood with the veil on the patient's head. The tube has to be placed near the patient's mouth.



Performing the test



1. Enter in the ergometry program
2. Select a patient or add a new one (**File/Patients...**)
3. Select **Start test** from **Test** menu.

A screenshot of a software window titled "Execute Test". It contains several input fields and a list of modes. The "Height (cm)" field is set to 180, and the "Weight (Kg)" field is set to 80. The "Mode" section has radio buttons for Gas, ECG, Gas + ECG, Rest ECG, RMR, Canopy (which is selected), and Simulated test. Below this, there are dropdown menus for "Ergometer:" (set to [no one]), "Protocol:" (set to RMR), and "Workspace:" (set to RMR). At the bottom right are buttons for "OK", "Other data...", "Details...", and "Cancel".

4. Enter the patient's data and select the **Canopy** mode.
5. Confirm and start the test by pressing **OK**.
6. In the first part of the test the flow rate of the pump has to be adjusted by means of the *Flow adjustment* handle on the front panel of the Canopy unit, in order to measure an FeCO_2 between 0.5% and 0.8%. FeCO_2 values can be read on the right side of the PC monitor.

-
7. When the FeCO_2 remains within the acceptability range, press **F2** to start the data acquisition. Verify, also during the test, that the measured FeCO_2 is within the 0.5%-0.8% range. Otherwise, adjust it by means of the *Flow adjustment* handle.

Warning: *If the green led turns off during the test, the red led flashes and the alarm beeps, abort the test, because the pump does not work or the mains does not power the system. In the last case, the pump works only because of the backup battery.*

The test is fully automatic, the software will stop it and save the data at the end.

Viewing the test

At the end of the test, it will be opened automatically a window with the test results.

At the end of the test, or if it is selected **View/RMR**, the main results are shown:

- The average time interval (default: 10 minutes)
- Average values of VO_2 , VCO_2 , R, RMR, VE, HR, FAT% and CHO% and predicted values if available.
- Body Mass Index (BMI) and interpretation
- Graph of the energetic expenditure for all the data acquisition interval, highlighting the selected average interval.

In order to verify the goodness of the test, check that the FeO_2 and FeCO_2 values are within the acceptability ranges (20.2%-20.8% and 0.5%-0.8% respectively), and the heart rate is the rest heart rate of the patient.



Note: The percentage of used Proteins (PRO%) is calculated assuming 12 grams of Ureic Nitrogen in 24 hours. You can modify this value selecting **View/Information... -> Modify...**

How to modify the average interval

If the average interval (automatically identified by the software) is not satisfying, for example because the patient was speaking in the first minutes, it is possible to modify the interval of the average.


Right-click and select **Edit RMR....** It is possible to move the start and the end lines.

To move the start line, left-click on the exact time in which you want to start the calculations, for the end line, right-click.

Print

The print of the current window generates a report similar to the one of the RMR test using the face mask.

Sub-maximal Exercise Testing



Introduction

Several physiological responses to exercise are used to evaluate cardiorespiratory fitness, including oxygen consumption, heart rate, and blood pressure. Measuring these variables during exercise, particularly maximum exercise, increase the chance of detecting any coronary artery disease or pulmonary disease.

Unfortunately, maximum exercise tests are impractical because they are expensive, require extensive clinical supervision, and subject individuals to levels of physical stress that may be unnecessary depending on the objectives of the test. Consequently, maximal testing is reserved for clinical assessments, athletic evaluation, and research.

A sub-maximal exercise test costs less and carries a lower risk for the individual. Although less sensitive and specific for detecting disease or estimating maximal oxygen consumption, correctly performed sub-maximal tests can provide a valid estimate of cardiorespiratory fitness.

Pre-test screening

Pre-test health screening is essential for risk stratification and for determining the type of test that should be performed and the need for an exercise test prior to exercise training. A thorough pretest health screening includes the following:

- Complete medical history
- Medical contraindications to exercise
- Symptoms suggesting cardiac or pulmonary disease
- Angina or other forms of discomfort at rest or during exercise
- Unusual shortness of breath at rest or during exercise
- Dizziness or light-headedness
- Orthopaedic complications that may prevent adequate effort or compromise the validity of test results
- Other unusual signs or symptoms that may preclude testing
- Risk factors for coronary heart disease
- History of major cardiorespiratory events
- Current medications
- Activity patterns
- Nutritional habits
- Reading and signing an informed consent form

Sub-maximal exercise testing

Heart rate varies linearly with VO_2 to the point of maximum exertion; thus, $\text{VO}_{2\text{max}}$ may be estimated using the relation between heart rate and VO_2 without subjecting the individual to maximum levels of physical stress. During sub-maximal exercise testing, predetermined workloads are used to elicit a steady state of exertion (plateau of heart rate and VO_2). The steady-state heart rate at each work level is displayed graphically and extrapolated to the VO_2 at the age-predicted maximal heart rate ($\text{HR} = 220 - \text{age}$). A variety of protocols for different exercise modalities (i.e., treadmill, stationary cycle, and step increments) can be used as long as the VO_2 requirements of each selected workload can be estimated with accuracy.

The objectives of cardiorespiratory fitness assessments in the apparently healthy population are as follows:

- Determine the level of cardiorespiratory fitness and establish fitness program goals and objectives.
- Develop a safe, effective exercise prescription for the improvement of cardiorespiratory fitness.
- Document improvements in cardiorespiratory fitness as a result of exercise training or other interventions.
- Motivate individuals to initiate an exercise program or comply with an established program.
- Provide information concerning health status.

A few assumptions regarding testing are necessary to ensure the highest degree of accuracy when using sub-maximal exercise testing to estimate $\text{VO}_{2\text{max}}$:

- Selected workloads are reproducible. A steady-state heart rate is obtained during each stage of the test. Usually, workload durations of 3 minutes or more are used to ensure steady state.
- The maximal heart rate for a given age is uniform ($\text{HR} = 220 - \text{age}$).
- Heart rate and VO_2 have a linear relation over a wide range of values; thus, the slope of HR/VO_2 regression can be extrapolated to an assumed maximum heart rate.
- Mechanical efficiency (i.e., VO_2 at a given work rate) is consistent.

Although if done correctly, sub-maximal exercise tests provide valuable information concerning cardiorespiratory fitness, they have extremely limited diagnostic capabilities and should not be used as a replacement for clinical exercise tests or other clinical treatment or management modalities. Health care professionals should avoid detailed interpretation beyond the scope of the information obtained.

Considerations with sub-maximal exercise testing

Considerations for selection of protocol and equipment include any physical or clinical limitations that may preclude certain types of exercise (i.e., age, weight, arthritis, orthopaedic complications, individual comfort, level of fitness, type of exercise training that will be performed, and individual preference).

For example, some individuals may perform better on a non-weight-bearing modality (cycle versus treadmill), while others may not have the required range of motion in the hip or knee to pedal and may perform better walking. Deconditioned, weak, or elderly persons may have to start the test at a low work level and increase the workload in small increments. Also, field tests may not be appropriate for those who require strict supervision during testing, who do not understand the concept of pacing, or who cannot be expected to put forth a good effort. More consistent results may be obtained by testing in a controlled environment such as a laboratory setting. Creativity when selecting protocols may allow adaptations of commonly used protocols to accommodate athletes competing in specific sports. Regardless of the type of exercise and protocol selected, the same type of exercise and protocol should be used for repeat testing if between-test comparisons are important.

Staffing

Staff members should be able to do the following:

1. Establish rapport with the subject and make him or her feel comfortable.
2. Recognize normal acute and chronic responses to exercise.
3. Recognize abnormal signs and symptoms during exercise.
4. Provide basic life support measures competently.
5. Adhere to established procedures and protocols.
6. Clearly explain test results to the individual.

Test termination

Sub-maximal tests should be terminated according to ACSM or other accepted guidelines (see table in the following). In the event of an abnormal response, the test should be terminated, the medical director of the facility and the individual's primary care physician notified, and all specified follow-up procedures performed. In the event of mechanical or electrical failure that may compromise the accuracy of the test results or monitoring capabilities, the test should be terminated until the problem is corrected.

General Indications for Stopping an Exercise Test in Apparently Healthy Adults

Onset of angina or angina-like symptoms

Significant drop (20 mmHg) in systolic blood pressure or a failure of the systolic blood pressure to rise with an increase in exercise intensity

Excessive rise in blood pressure: systolic pressure >260 mmHg or diastolic pressure >115 mmHg

Signs of poor perfusion: tight-headedness, confusion, ataxia, pallor, cyanosis, nausea, or cold and clammy skin

Failure of heart rate to increase with increased exercise intensity

Noticeable change in heart rhythm

Subject requests to stop

Physical or verbal manifestations of severe fatigue

Failure of the testing equipment

Assuming that testing is non-diagnostic and is being performed without direct physician involvement or electrocardiographic monitoring.

Considerations for accuracy

The ability to obtain valid and reproducible results is essential to ensure that any differences between pre-treatment and post-treatment test results are due to exercise training rather than variations in testing procedures. Some inconsistencies that are inherent may increase variability:

- Sub-maximal heart rate is influenced by time of day, eating, smoking, and familiarization with test procedures.
- Prediction equations for estimating $\text{VO}_{2\text{max}}$ may overestimate trained individuals and underestimate untrained individuals.
- The efficiency of motion during walking, running, and cycling varies.
- Cardiac output and VO_2 have a test-retest variability of 3-4%.

Psychological factors, such as pre-test anxiety, may influence the heart rate, especially at rates below 120 beats per minute and at low workloads. It is not unusual for the heart rate and/or blood pressure to be higher at rest than during the initial stages of exercise in these cases. Having the subject repeat the first test may improve reliability, particularly if the subject has never previously performed such a test.

Factors that can cause variation in the heart rate response to testing:

- Dehydration
- Prolonged heavy exercise prior to testing
- Environmental conditions (e.g., heat, humidity, ventilation)
- Fever
- Use of alcohol, tobacco, or caffeine 2 to 3 hours prior to testing

Because of these inherent inconsistencies, standard procedures for each test must be strictly followed to ensure the greatest accuracy and reproducibility possible:

- Standard testing protocol
- The same testing modality and protocol for repeat testing
- A constant pedal speed throughout cycle ergometry testing
- Cycle seat height properly adjusted, recorded, and standard for each test
- The time of day for repeat testing consistent
- All data collection procedures standardized and consistent
- Test conditions standard
- Subjects free of infection and in normal sinus rhythm
- Prior to the test, no intense or prolonged exercise for 24 hours, smoking for 2-3 hours, caffeine for 3 hours, or heavy meal for 3 hours
- Room temperature 18-20°C (64-68°F) with air movement provided

Performing the test


In this chapter it is supposed that the user is able to:

- perform an exercise test
- create exercise protocols
- view, edit and print tests

If this is not the case, please read the *Exercise testing* chapter.

To perform a sub-maximal test, follow these instructions:

1. Create a proper protocol (procedural guidelines for several sub-maximal testing protocols are provided in [ACSM's *Guidelines for Exercise Testing and Prescription*, 6th Edition Philadelphia: Williams&Wilkins, 2000:22-29]).
2. Start an exercise test.
3. Perform the test as it were a maximal exercise test, ending it when the heart rate reaches the 85% of the H_{max}, or it happens an event listed in the section *Test termination*.
4. Display a VO₂/Kg vs. HR plot
5. Right-click on the graph and select **VO₂ submax** from the pop-up menu.



If the predicted HR max (calculated as 220-age) is not suitable for the patient tested, it is possible to edit the HR max value from the **View/Information...** page.

An example of testing protocol

An example of protocol is reported here. The YMCA cycle ergometry protocol is defined as follows.

1st step: workload 150 kgm/min

2nd step: if the HR at the end of the 1st step is:

	<80	80-89	90-100	>100
set the workload at (kgm/min)	750	600	450	300


3rd step: if the HR at the end of the 2nd step is:

	<80	80-89	90-100	>100
set the workload at (kgm/min)	900	750	600	450

4th step: if the HR at the end of the 3rd step is:

	<80	80-89	90-100	>100
set the workload at (kgm/min)	1050	900	750	600


Spirometry



Setting spirometry options

The software allows to configure some options selecting **Configure** from the **Option** menu.

Spirometry



Automatic Interpretation

K4 b² has the function of interpreting each test performed by a patient visualising an automatic diagnosis. The algorithm has been calculated basing on “Lung Function Testing: selection of reference values and interpretative strategies, A.R.R.D. 144/1991:1202-1218”.

The automatic diagnosis is calculated at the end of the FVC Test if:

- the automatic diagnosis option is enabled.
- the patient’s anthropometric data allow the calculation of the LLN (Lower Limit of Normal range).
- at least one FVC test has been performed.

To enable/disable the automatic diagnosis:


1. Click on **Enable Automatic Interpretation** checkbox to enable or disable the calculation and the visualisation of the automatic interpretation.
2. Select the LLN (Lower Limit of Normal Range) criteria among the ATS (LLN=Pred-0.674*SD), ERS (LLN=Pred-1.647*SD) or 80%Pred (LLN=Pred*0.8) specifications.

Quality control

K4 b² allows a quality test control. The calculation has been carried out referring to “Spirometry in the Lung Health Study: Methods and Quality Control, A.R.R.D. 1991; 143:1215-1223”. The messages concerning the quality control are shown at the end of the test.

To enable/disable the quality control, click on **Enable Quality Control** checkbox.

Parameters manager



The program allows to calculate a huge number of parameters; it is advisable, in order to simplify the analysis of the results, to view, to print and to sort the desired parameters only. Select the menu item **Options/Parameters...**

View

Move the parameters to view into the *Selected parameters* list.

Print

Move the parameters to print into the *Selected parameters* list.

Sort

Drag the parameter up or down with the mouse.


Customise

Add, modify and delete custom parameters.



If it is necessary to restore the default parameters press the button in the left corner of the window to initialise the parameters database.

Predicted values manager



The program contains a preset of predicted equations, but the user is allowed to customise its own predicted sets. Select **Predicteds...** from **Options** menu.

The window is divided into two forms: **Predicteds set** and **Formula definition**.

Predicteds set

This form allows the user to manage the set of predicted. The following information define a set:

Name: identifies the set and cannot be duplicated;

Description: free field;

Age: the adult predicted starts since this age.

To enter a new set of predicted click on the **New** button. The field **Name** must be filled and must be unique. To stop without saving click on the **Cancel** button. To save the set, click on the **Save** button.

To delete a set of predicted click on the **Delete** button. If a set is deleted, also the associated formulae are deleted.

It is possible to generate a new set of predicted with the same attributes and the same formulae of the selected one. To do this click on the **Copy...** button and specify a new Name.

To import a set of predicted click on the **Import...** button and select a file of Predicted files type.

To export a set of predicted click on the **Export...** button.

In the list **Set current predicted** choose the current predicted for printing and viewing.

Set the current predicted

K4 b² allows to calculate the predicted values according to the following configurable sets:

Adult	Paediatric
ERS 93	Zapletal
Knudson83	Knudson83
ITS white	ITS white
ITS black	ITS black
LAM	LAM
MC Barcellona	MC Barcellona
Nhanes III	Nhanes III

Select the desired choice in the group **Predicted**.

Formula definition

The screenshot shows the 'Predicteds' dialog box with the 'Formula definition' tab selected. On the left is a list of predicted sets. The main area contains a 'Description' field, a radio button to 'Use the predicted formulae' with a dropdown menu, and another radio button to 'or the customized formulae'. Under the customized formulae section, there are radio buttons for 'Male' and 'Female', and a table with columns 'Formula' and 'Standard Deviation' for 'Young' and 'Adult' groups. At the bottom are buttons for 'Copy', 'Paste', 'Parameter...', 'Save', 'Delete', 'OK', 'Cancel', 'Apply', and 'Help'.

This form allows the user to manage the formulae associated to a set of predicted.

Select the set of predicted from the list **Predicted** set.

To insert a new parameter click on the **New...** button.

The parameter formulae can be:

- calculated according to the predicted values in the list **Use the predicted formulae**;
- customised by the user with the option **...or the customised formulae**.

The **Delete** button deletes the selected parameter.

The **Copy** button stores the selected parameter in memory.

The **Paste** button inserts a new parameter from the one copied. If the name is not unique, the user is asked whether to specify a new name or to replace the existing parameter.

Page set-up

Select **Page Setup...** from the **File** menu.

Header

All the printouts carried out by the program are preceded by 3 rows of customisable header (usually they contain the name and the address of the Hospital using the spirometer).

Data

Patient and visit information are printed below the header. These data are reported on 3 columns and 5 rows. the user may configure the disposition, change and eventually cancel the fields, as he prefers.

Margins

Configures the print margins from the borders of the paper. The unit of measure is decided in **Units of measurements**.

Footer

Configures information at the bottom of the page.

Printed file name

Defines the automatic name to be assigned to the pdf file, if the report will be printed in this format.

In the example it has been set to create a filename composed by <date of the test> followed by <last name> and <first name>.

Spirometry tests



Note: Read carefully the contraindications in Chapter 1.

Once completed the phases of the introduction of the patient's data and the visit data, it is possible to carry out the spirometric tests.

K4 b² allows to perform the following tests:

Key	Test
FVC pre	Forced Vital Capacity
FVC post	Forced Vital Capacity after bronchial stimulation
SVC	Slow Vital Capacity
MVV	Maximum Voluntary Ventilation

Before performing any test make sure that:

1. K4 b² is properly connected to your PC and the selected serial port (COM1, COM2) corresponds to the one effectively use.
2. The name shown on the status bar corresponds to the patient who is to carrying out the tests.
3. The today's visit card exists.

Forced Vital Capacity (pre)

FVC is a reference test to verify obstructive (airflow limitations) and restrictive disorders (lung volume limitations). To achieve good test results it is fundamental a good manoeuvre (quality control messages, real time plots ...)

The main parameters measured during FVC tests are:

FVC	Forced Vital Capacity
FEV1	Forced Expiratory Volume in 1 second
FEV1/FVC%	FEV1 as a percentage of FVC
PEF	Peak Expiratory Flow
FEF25-75%	Forced mid-Expiratory Flow

The two representative plots are the Flow/Volume and Volume/Time loops.

By comparing FVC, FEV1 and FEV1/FVC% values the software allows an automatic interpretation concerning the levels of obstructive and/or restrictive disorders.

Recommendations

- The flowmeter has to be disconnected from the breathing valve
- The patient should wear the nose clips
- The turbine has been recently calibrated (ATS recommends a daily calibration)
- The paper mouthpiece or the antibacterial filter is properly connected to the flowmeter through the corresponding adapter

For hygienic reasons, we strongly recommend the use of a bacterial filter.

If a kid must perform the test it is recommended to enable the encouragement function which shows exactly the manoeuvre of the FVC test.

Perform a FVC (pre) test



1. Select **Forced Vital Capacity pre** from the **Test** menu and wait for the green led is prompted on the right side of the screen.
2. Explain the manoeuvre to the patient and press the **F2** key.
3. Wait some seconds and perform the test.
4. After having performed the test, press **F3** or wait for the automatic end (5 seconds without flow), so that the software displays the F/V and V/t graphs, the main parameters, and the predicted values.
5. In order to visualise the F/V and V/t graph and the main parameters press the following buttons:



view Flow Volume graph



view Volume Time graph



view data of the test




6. Press **Alt+F3** to stop the acquisition discarding the results.
7. Repeat the test until it is correctly performed (ATS recommends 3 times).
8. Press **Exit** to visualise the test list carried out during current session together with the results of the main parameters.
9. Select the test that you want to save (the Best Test according to the ATS criteria is highlighted as default) and press **OK**.

Test encouragement

During FVC manoeuvre you might experience some lack of collaboration with kids or with other patients. In this case you may find a good help in using the encouragement software tool.

Perform the FVC test with the encouragement

1. Select **Encouragement** from **View** menu.
2. Perform the test as explained in the previous paragraph.



Slow Vital Capacity

Important test for assessing COPD (chronic obstructive pulmonary disease) patients affected by this disease might present a the Slow Vital Capacity could be higher than the Forced one (FVC).

The main parameters measured during SVC tests are:

EVC	Expiratory Slow Vital Capacity
IVC	Inspiratory Slow Vital Capacity
ERV	Expiratory Reserve Volume
IRV	Inspiratory Reserve Volume

If the inspiratory/expiratory maximal manoeuvre is preceded by a some breaths at tidal volume the software allows to measure the Respiratory Pattern, represented by the following parameters:

VE	Ventilation per minute
Vt	Tidal volume
Rf	Respiratory frequency
Ttot	Breath time
Ti/Ttot	Inspiratory time/Ttot
Vt/Ti	Vt/Ti

Perform a SVC test



1. Select **Slow Vital Capacity** from the **Test** menu and wait for the green led is prompted on the right side of the screen.
2. Press **F2** and instruct the Patient to breath normally until the message “carry out...” is prompted; then ask to perform a Slow Vital Capacity (deep inhalation, maximal slow expiration and deep inhalation again).
3. Press **F3** or wait for automatic interruption (5 seconds without flow) in order to visualise the V/t graph together with the main parameters compared to the predicted values

4. To visualise the V/t graph and the main parameters press the follow buttons:



view Volume Time graph



view data of the test



5. Press **Alt+F3** to stop the acquisition discarding the results.
6. Repeat the test until it is correctly performed (ATS recommends 3 times).
7. Press **Exit** to visualise the test list carried out during current session together with the results of the main parameters.
8. Select the test that you want to save (the Best Test according to the ATS criteria is highlighted by default) and press **OK**.

The reference for the ERV calculation is displayed on the V/T graph.

Maximum Voluntary Ventilation

Test for assessing the maximum ventilatory capacity. In the past, it was commonly performed during routine PF tests, however its clinical use declined over the years. Today MVV test is most commonly performed as part of the exercise tolerance tests, where it is used as an index of maximum ventilatory capacity. Test consists in breathing in and out deeply and rapidly for 12, 15 seconds. The expired volume during this short period is then extrapolated

The most important measured parameter is the following:

MVV Maximum Voluntary Ventilation

Perform a MVV test



1. Select **Maximum Voluntary Ventilation** from the **test** menu and wait for the green led is prompted on the right side of the screen.
2. Press **F2** and make the Patient breath as much deeply and rapidly as possible for at least 12 seconds.
3. Press **F3** or wait for automatic interruption (5 seconds without flow) in order to visualise the V/t graph together with the main parameters compared to the predicted values
4. To visualise the V/t graph and the main parameters press the follow buttons:



view Volume Time graph



view data of the test



5. Press **Alt+F3** to stop the acquisition discarding the results.
6. Repeat the test until it is correctly performed (ATS recommends 3 times).
7. Press **Exit** to visualise the test list carried out during current session together with the results of the main parameters.
8. Select the test that you want to save (the Best Test according to the ATS criteria is highlighted as default) and press **OK**.

Bronchial Provocation Test

Bronchodilator test



Note: Read carefully the contraindications in Chapter 1.

Bronchodilators are administered routinely in the b^2 laboratory to determine whether airflow obstruction is reversible. Bronchodilators increase airway calibre by relaxing airway smooth muscle.

The test consists of comparing results between the reference FVC (FVC PRE) and the FVC POST performed after the administration of the drug. Increasing value of 13-15% of FEV₁, respect to the basal value (FVC Pre) is considered as a reversible condition.

Main parameters are the following:

D_{FEV1}%pre Change of FEV₁ as a percentage of test PRE

D_{FVC}%pre Change of FVC as a percentage of test PRE

D_{PEF}%pre Change of PEF as a percentage of test PRE

Some authors states that the above mentioned parameters are too dependent from the FVC Pre, hence latest reference (ERS93, [A comparison of six different ways of expressing the bronchodilating response in asthma and COPD; reproducibility and dependence of pre bronchodilator FEV₁: E. Dompeling, C.P. van Schayck et Al; ERJ 1992, 5, 975-981]) recommend the following parameters:

D_{FEV1}%pred Change of FVC as a percentage of predicted value

D_{FEV1}%poss Change of FEV₁ as a percentage of "possible value"

Methacholine and Histamine Bronchial provocation Tests

The most common indication for performing methacholine and histamine bronchial challenges is to diagnose hyperresponsive airways. Some patients demonstrate normal baseline pulmonary function despite complaints of "tightness" wheezing, cough, and a little or not response to bronchoconstrictor. Other patients demonstrate spirometric improvement after use of bronchoconstrictor have diurnal variation in peak flows. In this groups aerosolised bronchial challenges are used to confirm a diagnosis of Asthma.

We can summarise the use of the test as follows:

1. Diagnose asthma
2. Confirm a diagnosis of asthma
3. Document the severity of hyperresponsivness
4. Follow changes in hyperresponsivness

When patients with hyperresponsive airways inhale certain pharmacologic agents (i.e. Methacholine or histamine) the airways respond by constricting.

Test consists of executing repeated FVC following the pharmacologic agents inhalation according to an established protocol. The fall of the FEV₁ parameter is used to calculate the bronchial hyperresponsivness. The most important parameter is the PD₂₀ that is amount of drug (mg/ml) that causes a reduction of 20% of the FEV₁ respect the basal value (without drug).

Main parameters are:

P₁₀ Dose that causes a 10% fall of FEV₁.

P₁₅ Dose that causes a 15% fall of FEV₁.

P₂₀ Dose that causes a 20% fall of FEV₁.

The representative plot is the *Dose/response curve*, showing the percentage variation of FEV₁ versus the Drug dose in logarithmic scale.

The program assumes as the **baseline test** the best **FVC pre** carried out during the today's visit. You can change the reference pre test editing the **Post** test.

The name of the drug, its quantity and its unit of measurement, can be typed immediately before any **FVC post** manoeuvre (manual protocol) or can be stored in a database of bronchoprovocation (**File/Bronchial Provocation protocols Database...**).

Perform the test



(During 1st step only) select **Protocol...** from the **Test** menu and choose the name of the bronchoprovocation protocol that you are going to use (**manual protocol** if you want to type the information about the agent before any manoeuvre)

1. Select **FVC post** from the **Test** menu.
2. Select an existing protocol or click on “manual protocol”, and wait the green leds turned on.
3. Press **F2**, or the button by side, to start the test.
4. Press **F3**, or the button by side, to achieve the test.
5. In order to visualise the V/t graph and the main parameters press the follow buttons:



view Flow Volume graph



view data of the test



view bronchial provocation response

6. Press **Alt+F3** to stop the acquisition discarding the results.
7. Repeat the test until it is correctly performed (ATS recommends 3 times).
8. Press **Exit** to visualise the test list carried out during current session together with the results of the main parameters.
9. Select the test that you want to save (the Best Test according to the ATS criteria is highlighted as default) and press **OK**.

Bronchial Provocation protocols Database

The response to a bronchoprovocator is usually assessed in terms of change in the FEV1, vital capacity or airways resistance on the basis of serial measurements (FVC manoeuvres) in which the results of the initial test constitute the reference values. The international literature proposes several standardised protocols in order to address the methodological issues of the various available techniques.

The possibility to store a bronchoprovocation protocol in a database is useful to simplify and automate the sequence of operations that the Physician need to execute during the bronchoprovocation tests.

The typical sequence of activities to carry out a bronchoprovocation test are:

1. Typing and storing a bronchoprovocation protocol in the database (usually only once).
2. Selection of protocol among the list of the ones already present in the database before carrying out the FVC post tests (the selection of “manual protocol” allows to execute the test fully manually).
3. Performing the Post tests.

Enter a new Bronchial provocation protocol in the archive

1. Select **Bronchoprov. protocols database** from the **File** menu.
2. Type the Protocol name, the Bronchoprovocator name and the unit of measurement in the proper input fields.
3. If the bronchoprovocator has a cumulative effect select the cumulative check button.
4. Enter the quantities for each step and press the button .

Viewing results



All the visualisation functions refer to the test carried out by the Current Patient, whose name is indicated on the left-side of the status bar.

To view tests results:

1. Select the **Patients** from the **File** menu
2. Select the patient corresponding to the test you want to view.
3. Select in the list box of the tests up to 5 tests of the kind (FVC, VC/IVC, or MVV) and press **OK**.

To switch between graph and or data use the following buttons on the toolbar:



view Flow Volume graph (F5)



view Volume Time graph (F6)



view data of the test (F7)



view bronchial provocation response.

If you need more than one visualisation meantime use the **New Window** function from the **Window** menu.

If you need to display a list of visits:

- Select **Visits list...** from the **File** menu.
- Type the name of the Company and/or the time interval desired or simply confirm for the complete list.

Tests of the current patient

If a **current patient** has been selected you can quickly view his tests selecting **Test current patient...** from the **View** menu.



Delete a test

1. Select **Patients** from the **File** menu or press the button by side.
2. Select the test that you want to eliminate from the list of the tests referred to the Current Patient and press **Delete**.

Printing results

You can print out in three different ways:

- printing the Report
- printing the Active Window
- printing a series of reports

Printing Reports



To print a report of the current visit, select **Print report...** from **File** menu. The software will choose automatically the best performed test.

The standard Report is composed by 1, 2 or 3 pages depending if you wish to printout the FVC data and the graphs together on the first page or if you wish to printout the bronchoprovocation response.

- Selecting the option **One page (no ATS)** the report will contain, on one page, the F/V and V/t graphs of the best test, overlapped on the **FVC Post**, the patient data, the notes, the diagnosis and the test results.
- Otherwise the report will contain two pages, the first with the patient data, the graphs and the diagnosis, and the second one with the measured parameters, according to the ATS recommendations.
- The 3rd page will contain the bronchoprovocation response.

Select the desired options:

FVC graph	Prints the F/V and V/t curves for the best FVC test.
One page (no ATS)	Prints data and graphs on the first page.
Response	Prints the bronchoprovocator response.
Preview	Views a report preview on the screen.

Printing the active window



This printout function is only enabled when the active window (title bar highlighted) is one of the following objects:

- Any kind of Graph.
- Numeric data
- List of visit

To print the active window

1. Ensure that the active window is one of the preceding objects.
2. Select **Print Active window** from **File** menu.

Printing a series of reports

Sometimes it is useful to printout automatically a series of reports (all tests carried out with the employees, all tests carried out in the today's session).

To print out proceed as follows:

1. Select **Visit List** from the **File** menu
2. Set the criteria of the visits to be added in the list (from, to,...)
3. Select **Print Report** from the **File** menu.

Electronic reports (*.pdf)

If an Adobe PDF writer "Printer Driver" is installed and set as the default printer, it is possible to store the printout report automatically in any location of the HD or eventually LAN paths according to a customizable filename format.

It is possible to define the created filename format selecting **File/Page Set up...** (see Page set-up).

Export data

With this function you can export the test data in 4 different formats:

- *.txt (ASCII)
- *.xls (Microsoft Excel)
- *.wk1 (Lotus 123)
- *.xpo (Cosmed)


Export a test

1. Select **Export tests** from the **File** menu.
2. Select the test to export from the list box and press **OK**.
3. Type the name and the format of the file in the dialog **Save as**. If the ASCII format is selected, the Text button in the dialog box Save as allows you to configure the separators for character based files.

With the *.xpo Cosmed file format it is possible to import data from another K4 b² archive. Press **OK** to confirm.

4. Select the folder for the export and type the file name. Press **OK** to confirm. A status bar will show the file creation.

External devices



GPS initialisation

The GPS operates on information gathered from satellites. To gather this information, take your GPS on outside and find large, open area that has a clear view of the sky (a nearby park would work fine). The GPS needs to receive at least three strong satellite signals to find your location.

At the first power on the GPS needs to be initialized; the initialisation is a fundamental procedure for obtaining accurate and reliable data and should be performed on a large area where the sky is fully "visible".

After the initial self test is complete, the GPS will begin the process of satellite acquisition and tracking. The acquisition process is fully automatic and, under normal circumstances, will take approximately 45 seconds to achieve a position fix (15 seconds if ephemeris data is known).

Like all GPS receivers, COSMED GPS utilizes initial data such as last stored position, date and time as well as satellite orbital data to achieve maximum acquisition performance. If significant inaccuracy exists in the initial data, or if the orbital data is obsolete, it may take 5.0 minutes to achieve a navigation solution. The GPS Autolocate™ feature is capable of automatically determining a navigation solution without intervention from the user. This procedure may be required if one of the following situations occurs:

- 1) Transportation over distances further than 1500 kilometers.
- 2) Failure of the internal memory battery without system standby power.
- 3) Stored date/time off by more than 30 minutes.

The GPS will automatically update satellite orbital data as it operates. The intelligence of the GPS combined with its hardware capability allows these data to be collected and stored without intervention from the host system.

Initialize the GPS

- 1) If the receiver is not operated for a period of six (6) months or more, the unit will "search the sky" in order to collect satellite orbital information. This process is fully automatic and, under normal circumstances, will take 3-4 minutes to achieve a navigation solution.
- 2) If the memory backup battery of the GPS fails, the receiver will search the sky as described above. Should the memory battery discharge, the unit needs to be powered on for several hours to insure a sufficient recharge to maintain several months of clock operation and memory storage.
- 3) If the initial data is significantly inaccurate, the receiver perform an operation known as AutoLocate™. This procedure is fully automatic and, under normal circumstances, will require 1.5 minutes to calculate a navigation solution.

During the acquisition process a message "acquiring satellites...." is prompted on the display of the Portable unit.

The AutoLocate™ function can be manually forced selecting **GPS AutoLocate** from the **Calibration** menu, in order to obtain the best accuracy.

Fixing the antenna to the subject

The GPS Receiver has to be positioned onto the harness of the K4b² according the following pictures, paying attention to keep the receiver in a position so that the sky will be always "visible" during the test.

Some applications such as cycle racing and rowing may require different positioning of the antenna.



Operating sequence

Test with GPS module can be carried out with K4 b² system in Holter Data Record or Telemetry Data Transmission mode only. In addition to the Operating sequence of this mode you must carry out the following operation.

Run a test with GPS

1. Connect the receiver antenna to the Portable Unit plugging phone jack into the RS232 port at the bottom of the PU.



2. Select **Settings** then **External device** and press **Enter**.
3. Enable the GPS option by moving the “*” sign on **GPS** and press **Enter** to confirm settings.
4. Check the GPS module functionality choosing **Calibration** then **GPS Control** and press **Enter**. Display will show latitude and longitude.

```
LAT: 41°43.130
LONG: 12°36.701
```

5. With the use of **Up** and **Down** key verify that the displayed altitude value (Alt) is different from zero. In case displayed altitude value is fixed on zero, please be sure that the antenna receiver is well plugged in, the “sky” is visible and wait until the Altitude value is shown.

```
VEL: 15.4
ALT: 670
```

Monitoring GPS parameters in real time

To monitor in real time GPS parameters during Telemetry Mode Transmission or as soon as test has been stored or downloaded, go to the PC software and select **Parameters to view/Test execution...** (real time) or **Parameters to view/Test visualization...** (after download) from the **Options** menu.

Select the following parameters:

Velocity	GPS Vel (m/sec)
Distance	GPS Dist (meters, incremented during exercise phases only)
Latitude	Lat (DD°MM.MMM' N/S)
Longitude	Long (DD° MM.MMM' E/W)
Altitude	Alt (meters)

Only when test has been stored or downloaded you can verify the **Graphical path** (automatically drawn on a scaled X/Y plane oriented to North) selecting on the PC software **Visualization** and **GPS track**.

***Note:** Distance is automatically calculated only during the "exercise" phases.*

Pulse Oximeter (option)

The oximeter option is useful to monitor SpO₂ value during the test. Test with this option can be carried out with K4b² system in Holter Data Record or Telemetry Data Transmission mode only.


In addition to the Operating sequence of this mode you must carry out the following operation:

Operating Sequence

1. Connect the Oximeter module to the Portable Unit plugging phone jack into the RS232 port at the bottom of the PU
2. Go to the K4b² control panel and select Oximeter like External device connected by choosing **Settings** than **External device** and press **Enter**
3. Enable the Oximeter option by moving the “*” sign on **Oximeter** and press **Enter** to confirm settings.
4. Positioning Finger or Ear Clip on the patient and fix well the cable with Velcro stripes on the harness to minimize motion artifact

To monitoring in real time SpO₂ value during Telemetry Mode Transmission or as soon as test has been stored or downloaded, go to the PC software, select **Parameters to view/Test execution...** (real time) or **Parameters to view/Test visualization...** (after download) from the **Options** menu and select SpO₂ parameter.

System maintenance



System maintenance

All service operations which are not specified in this user manual should be performed by qualified personnel in accordance with the service handbook (to be required to the manufacturer).

Rubber mouthpieces, face masks, breathing valve and the other parts are not shipped sterile. They should be disinfected before using according to the following instructions.

All materials used in the construction of the K4 b² are non toxic and pose no safety risk to the patient or operator.

Prior to the device cleaning, disinfection and inspection it is necessary to switch off the device and to disconnect adapters from the supply mains.

In order to guarantee the highest accuracy of measurements we recommend you to disinfect the turbine periodically.

Use disposable anti-bacterial filters or disinfect each part in contact with the patient before each test.

Cleaning and disinfection

Cleaning and disinfecting instructions are of fundamental importance to control infections and assure patient safety. In fact aspiration of residue, particles and contaminated agents are life – threatening.

In this handbook we strongly recommend you to follow the rules worked out by ATS and ERS (see: "Lung Volume Equipment and Infection Control" – ERS/ATS WORKSHOP REPORT SERIES, European Respiratory Journal 1997; 10: 1928 – 1932), which are summarised as follows:

- Accessible internal as well as external surfaces of equipment exposed to expired gas should be washed and disinfected prior to testing of subsequent patients.
- Liquid disinfection can be used if the equipment is well cleaned first (no droplets of saliva/sputum remain).
- Disposable gloves should be worn when handling mouthpieces, when cleaning equipment exposed to saliva or sputum and especially when drawing blood.
- Laboratory staff should wash hands prior to testing of each patient.
- Adopt particular precautions when testing patients with recognised high – risk communicable diseases (e.g. tuberculosis, multidrug – resistant staphylococcus). In these cases, the clinical need for such testing should justify the risks.

During the disinfection:

- do not use alcohol or other liquids containing gluteraldehyde on the exterior surfaces of the equipment. Actually they can damage polycarbonates plastics and may produce unhealthy substances.
- do not use abrasive powders or glass cleaners containing alcohol or ammonia on the plexiglas components of the equipment
- do not steam autoclave any parts of the equipment unless it is clearly specified.
- do not immerse the optoelectronic reader.

Preparing the disinfecting solution

The following recommendations are retrieved from:

APIC (*Association for Professionals in Infection Control and Epidemiology, Inc.*): *APIC Guidelines for Selection and Use of Disinfectants*; William A. Rutala, PhD, MPH, CIC. *American Journal of Infection Control*, vol.24, N.4, pp. 313-342, August 1996 - <http://www.apic.org/pdf/gddisinf.pdf>

As disinfecting solution it is suggested:

- Sodium hypochlorite 0.5% (5000 ppm) prepared fresh for use within 24 hours.
- Sodium hypochlorite 1% (10000 ppm) prepared fresh for use within 30 days.

The first solution can be easily prepared by adding 1 part household bleach (sodium hypochlorite 5.25%) to 9 parts water, the second one by adding 1 part household bleach to 4 parts water.




Warning: Do not use alcoholic solutions for the turbine, otherwise there can be damages to the plastic material.


Cleaning the turbine flowmeter

It is necessary to disinfect periodically the turbine for sanitary measures or/and for the correct device function.

The disinfecting procedure is easy and may be effected every time the user needs, keeping attention to some precautions:



1. Take out the turbine.
2. Dip it in a disinfectant solution (non alcoholic based) for about 20 minutes.
3. Rinse the turbine in a vessel, filled of clean water, shaking gently to remove the disinfectant (do not clean the turbine by putting it under running water!).
4. Let it dry to air.
5. After cleaning the turbine, check if the turbine propeller rotates freely even with a low speed air flow.
6. Connect the turbine to the reader.



Precautions during the cleaning of the turbine

1. Do not expose the turbine to high heat and do not put it under running water.
2. Do not ever dip the optoelectronic reader in any kind of solution, the liquid infiltration would damage the internal circuit.
3. Do not use alcoholic solutions to clean the turbine.

Suggested disinfection solutions

Helipur H Plus	Braun Melsungen AG
Gigasept FF	Schulke & Mayr GmbH
Dismozon pur	Bode Chemie GmbH
TETA-S	Fresenius AG
CIDEX	Johnson & Johnson

Masks cleaning and disinfection


The face masks should be cleaned and sterilised after each test.

Disassembling the different parts of the mask

1. Remove the valves from their place.
2. Remove the adapter for the optoelectronic reader.

Cleaning the mask

1. Clean the mask with hot water and a soap solution to remove the impurities.
2. Rinse the mask with energy in running hot water.



Warning: Do not use synthetic or petroleum-based products for the masks cleaning.

Disinfecting the mask

It's possible disinfecting the mask following these procedures:

- Standard autoclaving method
Rapid cycles of autoclave lasting 10 minutes at 132°C (270°F)
Heavy cycles of autoclave lasting 30 minutes at 121°C (250°F)
Pre vacuum cycles of autoclave lasting 30 minutes at 121°C (250°F)
- Ethylene oxide method (ETO)
The ethylene oxide doesn't deteriorate the face masks. Sterilisation by this method is not advised unless sufficient data is available regarding the time required for complete out-gassing of residual ETO. If you use this method, follow carefully the instruction provided by the maker of the sterilising product.
- Pasteurisation
The disinfecting with hot water is a sterilising method that may be used with the silicone masks.

Permapure maintenance

- Do not bend, squash or deform it.
- Do not keep it in open air, if not used, especially in crowded or smoky places.
- If saliva is entered in the tube, replace it immediately, because it lost its functions.
- Periodically grease the o-ring on the connector in order to simplify the flowmeter connection.
- Replace it every 100 test / 6 month.

Inspections

The equipment requires easy inspections to be carried out in order to assure a proper electrical and mechanical safety level in the years.

These inspections are highly recommended after a rough use of the equipment or after a period of storage in unfavourable environmental conditions.

Referring to the electrical safety, it is important to check the conditions of insulation materials of cables, plugs and any other visible part by means of simple inspection, when the equipment is switched off and adapters (or electrical feeders) are disconnected from the supply mains.

Mechanical parts to check are: the turbine and breathing circuits.

Follow these instructions:


- extract the turbine from the optoelectronic reader;
- verify, by inspection, that the turbine axis fits correctly its seats and the blade is strongly fastened on the axis itself (it can be useful to shake slightly the turbine in order to note any anomalous movement).

Check if there are any torn or broken components in the breathing circuits: remember that they can create safety risk to patients during tests.


Replace the fuses

The fuses can be replaced easily in the following way:


1. Open the power supply cover using a screwdriver as shown in the picture.



2. Extract the fuse holder as shown in the picture




3. Replace the damaged fuse(s).



***Note:** Be careful to use proper fuses:
A 680 023 500 (Time lag fuses 5x20 250V T500mA)*

Appendix



Warranty and limitation of liability

COSMED provides a one (1) year limited warranty from the date of the original sale of COSMED products. All COSMED products are guaranteed to be free from defect upon shipment. COSMED's liability for products covered by this warranty is limited exclusively to replacement, repair, or issuance of a credit for the cost of a defective product, at the sole discretion of COSMED. COSMED shall not be liable under the foregoing warranty unless (i) COSMED is promptly notified in writing by Buyer upon discovery of defect; (ii) the defective product is returned to COSMED, transportation charges prepaid by Buyer, (iii) the defective product is received by COSMED no later than four weeks after the last day of the one (1) year limited warranty period; and (iv) COSMED's examination of the defective product establishes, to COSMED's exclusive satisfaction, that such defect was not caused by misuse, neglect, improper installation, unauthorised repair or alteration, or accident. If the product is manufactured by a third-party, COSMED shall make available for the Buyer's benefit only those warranties which COSMED has received from the third-party manufacturer(s). COSMED hereby specifically disclaims any and all warranties and/or liabilities arising from defect(s) and/or damage(s) to and/or caused by products manufactured by third-party manufacturers. Buyer must obtain written authorisation from COSMED prior to the repair or alteration of COSMED product(s). Failure of Buyer to obtain such written authorisation shall void this warranty.

COSMED hereby specifically disclaims any and all other warranties of any kind, whether express or implied, in fact or by law, including, but without limitation, any and all warranties of merchantability and/or fitness for a particular purpose.

COSMED shall not be liable for special, indirect and/or consequential damages, nor for damages of any kind arising from the use of any COSMED's products, whether said products are used alone or in combination with other products or substances.

Determination of the suitability of any of COSMED's product(s) furnished hereunder for the use contemplated by Buyer is the sole risk and responsibility of Buyer, and COSMED has no responsibility in connection therewith. Buyer assumes all risks and liabilities for loss, damage or injury to persons or property of Buyer or others arising out of the use or possession of COSMED's products.

The limited warranty as herein above set forth shall not be enlarged, diminished, modified or affected by, and no obligation or liability shall arise or grow out of, the renderings of technical advice or service by COSMED, its agents or employees in connection with Buyer's order or use of the product(s) furnished hereunder.

Return goods policy for warranty or non warranty repair

Goods shipped to COSMED for repair are subject to the following conditions:

1. Goods may only be returned after your receipt of a **Service Return Number (SRN)** from COSMED S.r.l.
2. Place your SRN report and Packing List outside the package.
3. Goods returned must be shipped with freight and insurance charges prepaid. **Collect shipments will not be accepted.**
4. The following list of goods are not eligible for return unless proven defective.
 - Special order items
 - Expendable products
 - Goods held over 30 days from COSMED's invoice date.
 - Used goods not in original shipping containers.
 - Goods which have been altered or abused in any way.
5. The following parts are not covered by warranty:
 - consumables
 - fragile glass or plastic parts
 - rechargeable batteries
 - damages at the
 - damages due to use of the device not conforming to the indication reported in this manual

Repair Service Policy

Goods returned to seller for Non-Warranty repair will be subject to conditions 1, 2, 3, 4. The returned goods need to re-enter COSMED together with the customs documents (Pro-forma Invoice and Customs Paper) as requested by the Italian law.

- The shipment has to be qualified as a Temporary Export.
- All the goods returned to COSMED without the customs papers will not be accepted.

For European Community members:

Pro-Forma invoice complete with:

- Number
- Description of the goods
- Quantity
- Serial Number
- Value in €
- Number of parcel
- Gross weight
- Net weight
- Reason for resent (i.e. Resent for repair)

In case you should send the system for repair please contact the nearest service centre or contact COSMED at the following address:

COSMED S.r.l.

Via dei Piani di Monte Savello 37
P.O. Box 3
00040 Pavona di Albano - Rome, Italy
tel. +39 (06) 9315492
fax +39 (06) 9314580
E-mail: customersupport@cosmed.it

For USA customers only please contact:

COSMED USA Inc

2758 North Paulina
Chicago IL 60614 USA
Phone: +1 (773) 528-8113
Fax: +1 (773) 528-8116
email: usa.sales@cosmed.it

To ensure that you receive efficient technical assistance, please specify as precisely as possible the nature of the problem as it is specified on the assistance information form.

We advise you to save the original packaging. You may need it in case to ship the unit to a technical assistance centre.

Privacy Information

Dear Customer,

we inform you that your personal data are gathered and will be used by Cosmed Srl in conformity with the requirements of the Italian privacy law (Decreto Legislativo 196/2003). We believe it is important for you to know how we treat your personal data.

Personal data treatment and purposes

We request and process your personal data:

- a. to place an order, register a product, request a service, answer a survey, enter a contest, correspond with us (all of the above, in the following: “service”) and, if necessary, to supply the Competent Authorities with the required information;
- b. in order to define your commercial profile;
- c. in order to use your commercial profile for own marketing and advertising purposes;
- d. for accounting purposes, including e-mailing of commercial invoices;
- e. for providing your information to selected business partners (also abroad), in order to supply the service;

How your personal data are treated

Your personal data will be stored in electronic format, and protected at the best from destruction, loss (even accidental), not authorized accesses, not allowed treatment or use not in conformity with the purposes above listed.

The consent is optional, but...

If you deny the consent, we regret we cannot supply the service.

Holder of the treatment

The holder of the treatment is Cosmed Srl, Via dei Piani di Monte Savello 37, Pavona di Albano Laziale (RM). The responsible of the personal data treatment is indicated in the documentation stored by Cosmed Srl itself.

Customer rights

In accordance with art.7 of the Law, you can:

- a. obtain confirmation of the existence of your personal data and their communication in intelligible form;
- b. obtain:
 - updating, correction or integration of your data;
 - deletion or transformation in anonymous form of your personal data;
- c. deny your consent to the treatment of your personal data;

These rights can be exercised directly requesting in writing to the holder of the treatment.

Converting factors configuration



You can edit the parameters shown in Control Panel by selecting **Control Panel** from the **Calibration** menu in the calibration program, then pressing the button by side.

Configuration parameters

0x0000 [O2]
0x0001 [CO2]
0x0002 [Ambient temp.]
0x0003 [Internal temp.]
0x0005 [Barometric press.]
0x0006 [Analyzers press.]
0x0007 [Battery voltage]
0x0009 [Heart rate]
0x000A [Turbine Flow]
0x000B [Turbine Volume]

Raw data
Name: O2
Unit of meas.: %
Factor: .01
Precision: 2

$Y = (mV - BL) * Gain / 1000$

Base line (mV): -24
Gain ins: 1004
Gain exp: 1000

OK
Cancel
Help

You might configure the following options:

Name: identify the parameter

Unit of meas.: unit of measurement

Base line and Gain: factors used to convert the acquired raw data (mV) into the final format according to $Y=(mV-BL)*Gain$. The value entered for gain must be multiplied by 1000 (for Gain=1, enter 1000).

Precision: the number of decimals shown as 0

Calculations references

VO₂ and VCO₂

"Energy Expenditure and Fuel Selection in Biological Systems: The Theory and Practice of Calculations Based on Indirect Calorimetry and Tracer Methods": M. Elia, G. Livesey, World Rev. Nutr. Diet. Basel, Karger, 1992, vol 70, pp 68-131.

"Nutritional Assessment in Critical Care, A Training Handbook": Donald C. Zavala

Anaerobic threshold (modified V-Slope)

The break-point or intercept of the two slopes can be selected by a computer program that defines the VO₂ above which VCO₂ increases faster than VO₂, without hyperventilation.

During an incremental exercise above the Lactate Threshold, the net increase in lactic acid production results in an acceleration of the rate of increase in VCO₂ relative to VO₂. When these variables are plotted against each other (squared graph without recovery points), the relationship is composed of two apparently linear components, the lower of which has a slope of slightly less than 1.0, whereas the upper component has a slope steeper than 1.0. The intercept of these two slopes is the LT or AT point measured by gas exchange.

The increase in VCO₂ in excess of that derived from aerobic metabolism must be generated from the buffering of lactic acid. This is an obligatory gas exchange phenomenon seen in all subjects who exercise to work levels above their LT. This technique is referred to as the V-Slope method.

OVS, Original V-Slope method: "A new method for detecting anaerobic threshold by gas exchange", Beaver, Wasserman, Whipp, JAP 1986, 60:2020-2027.


MVS, Modified V-Slope method: "Metabolic acidosis during exercise in patients with chronic obstructive pulmonary disease", Sue, Wasserman, CHEST 1988, 94:931-938.

O₂ kinetics

"Delayed Kinetics of VO₂ in the Transition from prior Exercise. Evidence for O₂ Transport Limitation of VO₂ Kinetics: A Review"; R.L. Hughson and M.A. Morrissey, Int. J. Sports Med. 4 (1983) 31-39

ISO 8996: Ergonomics – Determination of metabolic heat production, 1990

In the following picture it is shown how the O₂ debit and deficit values are computed.



ATS 94 recommendations

Reference: "Standardization of Spirometry: 1994 Update" "American J. Respiratory Critical Care Medicine", Vol. 152, 1107-1136; 1995.

ATS recommendations

Volume range: 8l (BTPS)
Flow range: ± 14 l/sec
Volume accuracy: $\pm 3\%$ or < 50 ml
Flow accuracy: $\pm 5\%$ or < 200 ml/sec
Flowmeter resistance: < 1.5 cmH₂O da 0 a 14 l/sec

Reproducibility: the 2 largest of 3 acceptable FEV₁ and FVC values should be within 5% or 150 ml.

The end of test: no change in volume for 1 second with at least 6 seconds of collected volume.

Accumulation time: the maximum time allowed for volume accumulation during the VC manoeuvre should be at least 30 seconds and at least 15 seconds during the FVC.

The spirometer should be store at least 8 FVC manoeuvres.

FEV₁ should be calculated by using the "back extrapolation" method to detect the start of the test, extrapolated volume must not be higher then 5% FVC or 150ml.

The graphic resolution of the printed report must be as in the following:

Volume: 10 mm/l
Flow: 5 mm/l/sec
Time: 20 mm/sec
F/V ratio: 2:1

The total number of error (FVC e FEV₁ $> \pm 3.5\%$, FEF_{25-75%} $> 5.5\%$) during the measurement of the 24 standard waveforms must be lower than 4.

Predicted values

ERS93

Standardized Lung Function Testing: Official Statement of the European Respiratory Society, The European Respiratory Journal Volume 6, Supplement 16, March 1993.

Compilation of reference values for lung function measurements in children: Ph. H. Quanjer, J. Stocks, G. Polgar, M. Wise, J. Karlberg, G. Borsboom; ERJ 1989, 2, Supp.4, 184s-261s.

KNUDSON 83

Changes in the Normal Maximal Expiratory Flow-Volume Curve with Growth and Aging: J. Knudson, D. Lebowitz, J. Holdberg, B. Burrows; ARRD 1983; 127:725-734

ITS

Intermountain Thoracic Society: Clinical Pulmonary Function Testing, second edition (1984) pp 101, 144

LAM

A survey of ventilatory capacity in Chinese subjects in Hong Kong: Lam Kwok-Kwong, Pang Shing et Al. Annals of Human Biology, 1982, vol. 9, No. 5, 459-472.

Multicéntrico de Barcelona

Spirometric reference values from a Mediterranean population: J. Roca, J. Sanchis, A. Agusti-Vidal, F. Segarra, D. Navajas. R. Rodriguez-Roisin, P. Casan, S. Sans. Bull. Eur. Physiopathol. Respir. 1986, 22, 217-224.

Nhanes III

Spirometric reference values from a sample of the general US population: John L. Hankinson, John. R. Odencrantz and Kathleen B. Fedan. Am J Respir Critr Care Med 1999, 159, 1798-187.

Pneumobil (Brazil)

Valores extraídos do *Programa Pneumobil/Brasil* para a Tese de Doutorado do Dr. Carlos Alberto de Castro Pereira. (Boehringer).

Gutierrez (Chile)

Gutierrez et Al. Reference values for Chile population

Knudson, Morris and Bass

The maximal Expiratory Flow-Volume curve: Knudson et al. ARRD Vol. 123, p. 659-664, 1981

Spirometric Standard for healthy non-smoking adults: ARRD Vol. 10-3, p. 57-67, 1971

Pereira (Brazil)

Pereira CAC; Barreto SP; Simões JG; Pereira FWL; Gerstler JG; Nakatani J. Valores de Referência para Espirometria em uma amostra da população brasileira adulta. Jornal de Pneumologia 1992; 18: 10-22.

Mallozi MC. Valores de referência para espirometria em crianças e adolescentes, calculados a partir de uma amostra da cidade de São Paulo. Valores finais publicados em : Pereira CAC; Lemle A; Algranti E; Jansen JM; Valença LM; Nery LE; Mallozi M; Gerbasi M; Dias RM; Zim W. I Consenso Brasileiro sobre Espirometria. Jornal de Pneumologia 1996; 22:105-164.

Scalambrini Costa F, Scueiri CEB, Silva Jr WC, Pereira CAC, Nakatani J. Valores de referência para espirometria em uma amostra da população brasileira adulta da raça negra. J Pneumologia 1996;22: 165-170.

Neder JA; Andreoni S; Castelo-Filho A; Nery LE. Reference values for lung function tests. I. Static Volumes. Brazilian Journal Medical and Biological Research 1999; 32:703-17.

Neder JA, Andreoni S, Lerario MC, Nery LE. Reference values for lung function tests. II. Maximal respiratory pressures and voluntary ventilation. Braz J Med Biol Res 1999 ;32:719-27

DLCO

Standardized Lung Function Testing: Official Statement of the European Respiratory Society, The European Respiratory Journal Volume 6, Supplement 16, March 1993.

Compilation of reference values for lung function measurements in children: Ph. H. Quanjer, J. Stocks, G. Polgar, M. Wise, J. Karlberg, G. Borsboom; ERJ 1989, 2, Supp.4, 184s-261s.

Reference Values for Residual Volume, Functional Residual Capacity and Total Lung Capacity - ATS workshop on Lung Volume measurements, official statement of the European Respiratory Society; J. Stocks, Ph. H. Quanjer: ERJ, 1995, 8, 492-506

Single Breath Oxygen Test

Buist SA, Ross BB: Quantitative Analysis of the Alveolar Plateau in the Diagnosis of Early Airway Obstruction. ARRD 108: 1081, 1973

Mansell A, Bryan C, Levison H: Airway Closure in Children. JAP 33: 711-714, 1972

Buist SA, Ross BB: Predicted Values for Closing Volumes Using a Modified Single Breath Test. ARRD 107: 744-751, 1973.

Rint

Lombardi E, Sly PD, Concutelli G, et al. Reference values of interrupter respiratory resistance in healthy preschool white children. Thorax 2001; 56: 691-695.

Mip/Mep

Leo F. Black, Robert E. Hyatt: Maximal Respiratory Pressures: Normal Values and Relationship to Age and Sex, American Review of Respiratory Disease, Volume 99, 1969

Vincken W, Ghezzi H & Cosio MG (1987). Maximal static respiratory pressures in adults: normal values and their relationship to determinants of respiratory function. Bull Eur Physiopathol Resp 23: 435-439.

Automatic diagnosis (algorithm)

Reference: "Lung Function Testing: selection of reference values and interpretative strategies", A.R.R.D., 144/ 1991:1202-1218.

$LLN = Pred - 0.674 * SD$ (ATS, 50° percentile)

$LLN = Pred - 1.647 * SD$ (ERS, 95° percentile)

$LLN = Pred * 0.8$ (80%Pred)

Message interpretation	Criterion
Normal Spirometry	FVC and FEV1/FVC > LLN
Obstructive abnormality (it may be physiological)	% Pred FEV1 >= 100
Obstructive abnormality: mild	% Pred FEV1 < 100 and >= 70
Obstructive abnormality: moderate	% Pred FEV1 < 70 and >= 60
Obstructive abnormality: moderately severe	% Pred FEV1 < 60 and >= 50
Obstructive abnormality: severe	% Pred FEV1 < 50 and >= 34
Obstructive abnormality: very severe	% Pred FEV1 < 34
Restrictive abnormality: mild	FVC < LLN and % Pred FVC >= 70
Restrictive abnormality: moderate	% Pred FVC < 70 and >= 60
Restrictive abnormality: moderately severe	% Pred FVC < 60 and >= 50
Restrictive abnormality: severe	% Pred FVC < 50 and >= 34
Restrictive abnormality: very severe	% Pred FVC < 34

Quality Control Messages

Reference: Spirometry in the Lung Health Study: Methods and Quality Control, ARRD 1991; 143:1215-1223.

Message	Criterion
Start faster	VEXT >5% of the FVC and >150ml
Blast out harder	PEFT >120 msec
Avoid coughing	50% drop in the flow in first second
Blow out longer	FET100% <6 sec.

Blow out more air	flow >0.2l/s within 20 ml of FVC
Blow out harder	dPEF<10%
Take a deeper breath	dFVC<200ml and 5% best FVC
Blow out faster	dFEV1<200ml and 5% FEV1
That was a good test	No errors
FVC reproducible	diff. 2 max FVC within 0.2 l
FEV1 reproducible	diff. 2 max FEV1 within 0.2 l
PEF reproducible	diff. 2 max PEF within 10 %
MVV time too short	MVV time less than 12 sec

References

Gas Exchange References

["On line computer analysis and breath by breath graphical display of exercise function tests."; Beaver, Wasserman, Whipp, JAP , 34(1):128-132, 1973]

["Measurement and analysis of gas exchange during exercise using a programmable calculator"; Sue, Hansen, Blais, Wasserman, JAP, 49(3), 1980:456-461]

["Principles of exercise testing and interpretation, 2^o edition"; Wasserman et Al, 1994]

["Clinical Exercise Testing, 3rd edition", Jones 1988]

ERS task force on standardization of clinical exercise testing. "Clinical exercise testing with reference to lung disease: indications, standardization and interpretation strategies." J. Roca, B. Whipp, S. Anderson, R. Casaburi, J.E. Cotes, P. Palange...., ERJ 1997; 10: 2662-2689.

Indirect calorimetry

["Energy Expenditure and Fuel Selection in Biological Systems: The Theory and Practice of Calculations Based on Indirect Calorimetry and Tracer Methods": M. Elia, G. Livesey, World Rev. Nutr. Diet. Basel, Karger, 1992, vol 70, pp 68-131.]

["Nutritional Assessment in Critical Care, A Training Handbook": Donald C. Zavala]

Spirometry

ATS '94: "Standardization of Spirometry: 1994 Update", American J. Respiratory Critical Care Medicine, Vol. 152, 1107-1136; 1995

ERS '93: "Standardised Lung Function Testing: Official Statement of the European Respiratory Society", The European Respiratory Journal Volume 6, Supplement 16, March "

Lung function", J.E. Cotes, Blackwell scientific publications

"Guidelines for Clinical Exercises Testing Laboratories", I.L. Pina, G.J. Balady, P. Hanson, A.J. Labovitz, D.W. Madonna, J. Myers. American Heart Association. 1995; 91, 912.

Sub-maximal testing

["Cardiorespiratory Assessment of Apparently Healthy Populations", Timothy R. McConnell, in ACSM's Resource Manual for Guidelines for Exercise Testing and Prescription, 4th Edition, pp. 361-366]

[Franklin BA, ed. ACSM's Guidelines for Exercise Testing and Prescription, 6th Edition Philadelphia: Williams&Wilkins, 2000:22-29]
